Microwave Phase Detection at the Level of 10^{-11} rad

Eugene N. Ivanov^{*} and Michael E. Tobar[†]

University of Western Australia, School of Physics M013, 35 Stirling Hwy., Crawley 6009 WA, Australia

(Dated: October 30, 2018)

Abstract

We report on a noise measurement system with the highest spectral resolution ever achieved in the microwave domain. It is capable of detecting the phase fluctuations of rms amplitude of $2 \times 10^{-11} rad/\sqrt{Hz}$ at Fourier frequencies above few kHz. Such precision allows the study of intrinsic fluctuations in various microwave components and materials, as well as precise tests of fundamental physics. Employing this system we discovered a previously unknown phenomenon of down-conversion of pump oscillator phase noise into the low-frequency voltage fluctuations.

PACS numbers: 07.60.Ly, 06.20.-f, 07.50.Hp, 84.40.-x

^{*}Electronic address: eugene@physics.uwa.edu.au

[†]Electronic address: mike@physics.uwa.edu.au



FIG. 1: Microwave noise measurement system based on an interferometer with a distributed resonator

At frequencies, ν , much less than a few THz at room temperature ($\nu \ll k_B T_0/h$), Nyquist thermal noise[1, 2] restricts the resolution of phase (amplitude) measurements. For the conventional interferometric measurement system [3] containing a test sample, this limit is set by the Standard Thermal Noise Limit (STNL), given by $\sqrt{2k_B T_0/(P_{inc}\alpha)} rad/\sqrt{Hz}$ where, k_B is the Boltzmanns constant, T_0 is the ambient temperature, α is the loss in the test sample, P_{inc} is the power incident on the test sample, and h is Planks constant. This work describes a new microwave noise measurement system with a phase noise floor of $2 \times 10^{-11} rad/\sqrt{Hz}$, which is 10dB below the STNL and more sensitive than any laser interferometer operating at the same level of signal power[4, 5].

A schematic diagram of the noise measurement system is shown in Fig. 1. It consists of a standing wave interferometer based on a Magic Tee waveguide coupler and a microwave readout which includes a low-noise microwave amplifier (LNA), reference phase-shifter and a double-balanced mixer (DBM). The distinctive feature of the measurement system in Fig. 1 is a distributed resonator formed by an inductive diaphragm and a short-circuited piece of waveguide with a variable phase-shifter. A test sample can be either inserted directly into the waveguide or (if it is a coaxial device) coupled to it with an adapter. The smaller the loss in the test sample, the higher the sensitivity due to the extended lifetime of the distributed resonator. The signal reflected from the distributed resonator interferes destructively with a fraction of the incident signal at the interferometers "dark port" (port 4 in Fig. 1). This cancels the carrier of the difference signal while preserving the noise modulation sidebands caused by non-thermal fluctuations in the test sample. The noise sidebands are amplified and converted into the voltage noise at the output of the LNA and DBM assembly. Depending on the setting of the reference phase shifter (θ in Fig. 1) the output voltage noise varies synchronously with either phase or amplitude fluctuations of the test sample. The use of the LNA overcomes the relatively high technical fluctuations in the DBM making the effective noise temperature of the readout system T_{RS} close to its physical temperature $T_{RS} \approx T_0 + T_{amp}$, where T_{amp} is the effective noise temperature of the microwave LNA.

The main reasons for choosing the waveguide components instead of micro-strip (used in [3, 6]) were (i) to increase the efficiency of power recycling (by reducing the distributed loss in the interferometer arms) and (ii) to minimize the effect of technical noise sources inside the interferometer as some micro-strip components tend to exhibit an excess noise when exposed to relatively high levels of microwave power (of the order of 1W).

The phase sensitivity of the measurement system $(du/d\phi)$ was optimized by adjusting the aperture of the inductive diaphragm and its distance from the symmetry plane of the Magic Tee. A piece of hollow waveguide approximately 10cm long was initially used as a test sample. At $P_{inc} = 1W$ the highest value of phase sensitivity was measured to be 1.4kV/rad, corresponding to a factor of 4 enhancement relative to the conventional interferometer operating at the same level of input power. This sensitivity enhancement was accompanied by more effective use of signal power, as only a few percent of the incident power was reflected from the input of the Magic Tee back to the pump source.

The Single Side Band (SSB) phase noise floor of the above measurement system is shown in Fig.2 (curve 1). The rough peaks in the noise spectrum are of vibration and acoustic origin. At Fourier frequencies, f > 5kHz, the SSB phase noise floor approaches -213dBc/Hz. This is almost 13dB lower than the noise floor of a power recycled laser interferometer operating at the same level of signal power[4, 5]. Also, it should be noted that the above resolution was achieved in real time, as opposed to slow measurements associated with the cross-correlation noise measurement technique first suggested in [7] and later applied to interferometric measurements [8]. The horizontal line in Fig.2 at the level -203dBc/Hzcorresponds to the SSB STNL of the conventional interferometer calculated for a set of parameters typical for our experiments: $P_{inc} = 1W$, T = 300K and $\alpha = 0.5dB$. As follows from Fig. 2, the improvement I in the resolution of noise measurement relative to the STNL is approximately 10dB. This is slightly less than what one would expect from the



FIG. 2: SSB phase noise floors of a 9GHz measurement system; noise floor of the entire system (1); noise floor due to fluctuations in the microwave electronics of the readout system (2)

enhancement in the phase sensitivity E mentioned earlier (factor of 4). This is because: $I = E\sqrt{T_0/T_{RS}}$, where, in our case, $T_{RS} \approx 390K$.

The above measurements were repeated with the input of the LNA terminated to evaluate the contribution of the microwave electronics to the overall uncertainty of the phase measurements. The resulting phase noise floor is given by curve 2 in Fig.2. The increasing divergence between two noise spectra at low Fourier frequencies could be attributed to intrinsic fluctuations inside the interferometer likely to be associated with the ambient temperature upsetting the interferometers balance

At low Fourier frequencies the spectral density of phase fluctuations seems to be independent on the type of components used to form the interferometer. Such conclusion could be derived from examining the phase noise floors of different interferometric measurement systems tested (Fig. 3): (i) conventional coaxial interferometer; (ii) coaxial interferometer with recycling [6] and (iii) current waveguide based interferometer. At Fourier frequencies f < 1Hz all spectra in Fig. 3 converge to the same power law: $\frac{1}{2}S_{\phi} \approx 3 \times 10^{-16}/f^2$ rad^2/Hz) which, we believe, is a signature of ambient temperature fluctuations. The above statement should be taken with some caution, as we did not take special measures to isolate our experimental setup from the vibration and acoustical influences.

A different situation was observed at f > 10Hz where the results of noise measurements with coaxial and micro-strip components were often irreproducible showing a large scatter



FIG. 3: Noise floors of different 9GHz interferometric measurement systems: conventional coaxial interferometer (1); coaxial interferometer with power recycling (2); waveguide interferometer with signal recycling (3).

from one experimental trial to another: in some cases the noise floor of the coaxial system was close to that of the waveguide system, while in others it was an order of magnitude higher. The noise performance of the waveguide based interferometers has always remained highly consistent. In the above experiments we used conventional continuously variable attenuators and phase-shifters, except for the waveguide interferometer, which was based on the precision rotary vane attenuator and phase shifter.

It should be noted that not only phase fluctuations in a test sample can be measured with the ultra-high precision: the measurements of amplitude fluctuations can equally benefit from the introduction of a distributed resonator. The point is that neither phase nor amplitude noise spectra are measured directly: they are both inferred from voltage noise at output of the measurement system when the latter is tuned either phase or amplitude sensitive. In case of the measurement system in Fig.1, the spectral density of its output voltage noise was practically independent on the type of tuning. This meant equal contribution from both amplitude and phase fluctuations to the output voltage noise and, therefore, the possibility of their measurements with equal precision.

We also tested a microwave counterpart of the optical Michelson interferometer with power recycling. In such a case, an inductive diaphragm was removed from the side port of the Magic Tee and placed in the path of the incident signal. The aperture of the diaphragm



FIG. 4: Reflection and transmission coefficients of the microwave interferometer with power recycling

and its distance from the symmetry plane of the Magic Tee were adjusted to minimise power of the reflected signal. When viewed with a microwave vector network analyser the input reflection coefficient of the interferometer resembled transfer function of a notch filter with the depth of 24 dB (Fig.4). At $P_{inc} = 1W$ the maximum phase sensitivity of the power recycled interferometer was measured to be 0.9kV/rad.

The high resolution of the phase measurements observed in the above experiments was achieved by paying serious attention to technical noise sources. First, a careful balancing of the interferometer was required to avoid the excess 1/f voltage noise associated with the saturation of the LNA. Secondly, the signal of the microwave pump source was bandpass filtered to remove the higher order harmonics from its spectrum. Such harmonics are not attenuated when the interferometer is balanced and can saturate the LNA. Also, we have to deal with the enhanced sensitivity of the measurement system to pump oscillator frequency noise due to dispersion of the distributed resonator. The phase noise floor of the measurement system due to the pump oscillator phase fluctuations can be evaluated from:

$$S_{\phi}^{n/f} = S_{\phi}^{osc}(f) (\frac{du/d\nu}{du/d\phi})^2 f^2.$$
 (1)

Here $du/d\nu$ is the frequency sensitivity of the measurement system, which was measured to be 20mV/kHz when the phase sensitivity was maximized (i.e. $du/d\phi = 1.4kV/rad$).



FIG. 5: RF spectra at the output of the microwave LNA: 1- conventional interferometer: 2interferometer with signal recycling

Substituting the above values of sensitivities into (1) and assuming the phase noise of a high performance commercial frequency synthesizer, say E8257C from Agilent Technologies, one could conclude that the spectral resolution beyond the STNL would be only feasible at Fourier frequencies f < 3kHz. For this reason, most of our measurements (including the noise spectra in Fig. 2) were performed with a composite source consisting of a low-phase noise 9GHz oscillator based on Sapphire Loaded Cavity (SLC) resonator [9] and a high power amplifier needed to generate a few Watts of useful power. Within the range of Fourier frequencies 100Hz to 200kHz the phase noise of the composite source was dominated by the phase fluctuations of the high power amplifier. This, however, did not affect the resolution of noise measurements at f < 700 kHz. At higher Fourier frequencies the rise of oscillators phase noise (associated with the loss of gain of oscillator frequency control loop, as seen from Fig. 5) was the main factor responsible for the loss of spectral resolution. It was also responsible for a tall peak around 10MHz in the spectrum of mixer voltage noise. This high intensity voltage noise was saturating the pre-amplifier in front of the FFT spectrum analyzer (Fig. ??) rising the white noise floor of the entire measurement system [9]. Introduction of a passive low-pass filter in front of the pre-amplifier solved this problem.

We also discovered an additional type of noise inherent to narrow-band interferometers. It was a white noise with spectral density weakly influenced by interferometers amplitude mismatch, but showing strong dependence on interferometers phase mismatch. Power at the interferometer "dark port" must be kept below approximately 50dBm to prevent the loss of phase resolution due to this phenomenon. The origin of this noise was found to be related to the down-conversion of high frequency phase fluctuations of the pump oscillator signal. The point is that the narrow-band interferometer does not filter out fast phase fluctuations of the pump oscillator. This is illustrated by Fig. 5, which shows the RF spectra at the interferometer dark port. The high intensity phase noise at Fourier frequencies of order 10MHzis synchronously demodulated to DC inside the DBM when interacting with the replica of itself arriving at the second port of the DBM. We simulated these noise-noise interactions by considering the propagation of a phase-modulated signal through the dispersive network. We also managed to induce this effect artificially by modulating frequency of the pump oscillator with a band limited random voltage noise centered around 10MHz. The problem of the excess noise in the recycled interferometer was solved by the narrow-band filtering the pump signal. In our case, this was achieved by taking the output signal of the low-phase noise oscillator from the filtered port of the SLC resonator.

Concluding, we demonstrated the possibility of real time noise measurements at microwave frequencies with the phase noise floor 10dB below the STNL at Fourier frequencies above few kHz. This was achieved by combining the principles of microwave circuit interferometry with more efficient use of signal power and minimizing the influence of technical noise sources on the measurement process. We also discovered that under certain conditions an excess white noise appears at the output of the narrow-band interferometer due to the self-demodulation of high frequency phase fluctuations in the spectrum of the incident signal.

A narrow-band interferometer is ideally suited for studying the noise phenomena in lowloss components and materials at microwave frequencies. We used it to characterize the intrinsic phase fluctuations in ferrite circulators. So far, we could only claim that there is no general rule describing the phase noise in such devices: in some cases the circulator phase fluctuations were easily observable, while in others no noise was detected.

Recently, it was shown that a sensitive measurement of the isotropic Lorentz violation parameter, κ_{tr} , within the Standard Model Extension (SME)[10, 11], could be attained by placing a magnetically permeable material inside the microwave interferometer [12]. Assuming a 10*cm* long piece of a ferrite filled waveguide inserted in the recycled interferometer and the rotation rate of about 0.2Hz [13], one can constrain κ_{tr} within the interval of $\pm 10^{-8}$ only after a single day of observations provided that no technical fluctuations are present in the ferrite [12]. This is about an order of magnitude better than current best limits [14, 15]. Because our measurements did not reveal any such fluctuations in magnetically shielded ferrite devices (including a ferrite loaded waveguide), the system looks promising for some future precise tests of the SME.

Acknowledgments

This work was funded by the Australian Research Council.

- [1] H. Nyquist, *Phys. Rev.*, **32**, 110, (1928)
- [2] B.M. Oliver, *Proc. IEEE*, 53(5), 436 (1965)
- [3] E. N. Ivanov, M. E. Tobar, and R. A. Woode, IEEE Trans. on Ultrason. Ferroelect. Freq. Contr. 45 1526 (1998).
- [4] B. Lantz, P. Fritschel, H. Rong, E. Daw, J. Opt. Soc. Am. A, 19(1), pp. 91-100, (2002)
- [5] P. Fritschel, G. Gonzalez, B. Lantz, P. Saha, M. Zucker, Phys. Rev. Lett, 80(15), pp. 3181-4, (1998)
- [6] E. N. Ivanov and M. E. Tobar, IEEE Trans. on UFFC, vol. 49, no. 8, 2002, pp. 1160-1161
- [7] F. L. Walls, S. R. Stein et al., in Proc. of the 30th Symp. on Frequency Control, pp. 269-274, 1976
- [8] E. Rubiola, V. Giordano and J. Groslambert, IEEE Trans. on Instr. and Meas., 48(2), pp. 642-646, 1999
- [9] E. N. Ivanov and M. Tobar, IEEE Trans. on UFFC, 56(2), pp.263-269, 2009.
- [10] V.A. Kostelecky, M. Mewes, Phys.Rev.Lett. 87, 251304 (2001)
- [11] Q. G. Bailey, V.A. Kostelecky, Phys.Rev.D 70, 076006 (2004)
- [12] M.E. Tobar, P. Wolf, A. Fowler, J.G. Hartnett, Phys. Rev. D 71, 025004 (2005).
- [13] P.L. Stanwix, M.E. Tobar, P. Wolf, et. al., PRL 95, 040404 (2005)
- [14] M. Hohensee, A. Glenday, C-H Li, et. al., Phys. Rev. D 75, 049902 (2007)
- [15] S. Reinhardt et. al., Nature Phys. 3, 861 (2007)