

Emerging Ideas in Nanocantilever based Biological Sensors

Ankit Jain* and Muhammad Ashraful Alam[#]

School of ECE, Purdue University, West Lafayette, IN, USA, 47907

e-mail: *jankiit@gmail.com, [#]alam@purdue.edu

Abstract

In this review article, we focus on emerging nanocantilever based biological sensors and discuss the response of nanocantilevers towards bio-molecules capture. The article guides the reader through various modes of operation (e.g., static or dynamic) to detect the change in cantilever's characteristics (e.g., mass, stiffness, and/or surface stress) due to adsorption of bio-molecules on cantilever surface. First, we explain the classical linear resonant mode mass sensors and static stress based sensors. The effect of operating the cantilever in nonlinear regime is then illustrated through examples of bifurcation based mass sensors and electromechanical coupling based Flexure-FET biosensors. We believe that a new class of nonlinear sensors, with their extraordinary sensitivity towards bio-molecules capture, could be the potential candidate for low cost point-of-care applications.

1. INTRODUCTION

Detection of biological molecules e.g., viruses, proteins, DNA, etc., is essential for food safety, early warning of biological attack, early stage diagnosis of cancer, and genome sequencing. Nanoscale devices are widely regarded as a potential candidate for ultra-sensitive, low-cost, label-free detection of bio-molecules and are considered as a technology alternative to the existing chemical or optical detection schemes. Label-free schemes detect bio-molecules using their intrinsic properties, e.g., size, mass or charge of a molecule, instead of using extrinsic optical or magnetic labels attached to the target molecule. Among the various label-free technologies, significant research has focused on developing ultra-sensitive biological sensors based on nanocantilevers [1], [2].

The use of a cantilever as a sensor dates back to 1943 when Norton proposed a hydrogen gas sensor based on a cantilever [3]. The opportunity to develop the cantilever as a highly sensitive biosensor, however, had to wait the invention and wide-spread adoption of atomic force microscope (AFM) [4]. An AFM measures the forces between the tip of a cantilever and the sample surface using the tip deflection (contact mode AFM) or changes in the resonance frequency of a vibrating cantilever (dynamic mode AFM). As we will see in Sec. 2, nanocantilever based biosensors operate in a closely related principle, where interaction with biological molecules changes the bending (static mode) or resonance frequency (resonant mode) of the cantilever [5]. Note that these mechanical sensors offer an advantage of detecting both charged as well as neutral bio-molecules; in contrast,




Fig.1: (a) Schematic of a generic biological sensor operating in fluidic environment. The receptor, target and parasitic molecules are also shown. (b) Drawing of a nanocantilever in which capture of target molecules changes the cantilever's static or dynamic response through change in the mass, stiffness and/or surface stress. (c) A chart showing various linear and nonlinear cantilever biosensors to be discussed in this article. S is the sensitivity of respective biosensors and N_s is the areal density of captured bio-molecules on the cantilever surface.

their electronic counterpart, e.g., ISFETs (Ion Sensitive Field Effect Transistors [6]) can only detect charged molecules. Today, cantilever-based devices find applications in broad range of fields such as communication, computation, optics, scanning probe microscopy, and sensing. Specific examples include RF-MEMS capacitive/ohmic switches [7], varactors, tunable oscillators [8], NEMS relays [9], NEMFET [10], deformable mirrors [11], displays [12], accelerometers [13], and chemical/gas sensors. In this article, however, we only focus on the use of cantilever as a biological sensor, and discuss both classical and emerging modes of biosensor operation.

Figures 1(a)-(b) show the schematic of a sensor surface operating in a fluidic environment. The sensor surface is functionalized with receptor molecules so that it can subsequently conjugate to the target molecules (without any optical or magnetic labels) contained in the fluid. For example, if the target is an antibody, then receptor is corresponding antigen, or if the target is a DNA base (e.g., A), then receptor is its

conjugate base (e.g., C). The specific binding of the target and receptor molecules changes the sensor characteristics and the change is measured as a signature of detection. For example, the binding may induce an additional surface stress or may change the mass or stiffness of the cantilever, resulting in bending or change in the resonance frequency [5].

There are three key metrics of any sensing scheme (electronic or mechanical), namely, response time [14], selectivity [15], and sensitivity [16]. *Response time* is the time required to capture a certain number of target molecules to get a detectable output signal. Interestingly, response time depends on the geometry of the sensor surface and cylindrical geometry allows the smallest response time [14]. On the other hand, *selectivity* is associated with the problem of non-specific binding, i.e., binding of parasitic molecules with receptor molecules, producing a “false-positive” signal. In a highly selective sensor, receptor molecules should only bind to the target molecules and not to any other molecules in the solution. Finally, *sensitivity* can be defined in number of ways; in general, it is measured as the change in the sensor characteristics (e.g., resonance frequency of a cantilever [5] or drain current of ISFET [17]) in response to the capture of a given number of target molecules. Note that, response time and selectivity of a sensor do not depend on the sensing scheme, whereas sensitivity depends on the sensing scheme. Therefore, in this article, we only discuss sensitivity related issues of nanocantilever based biological sensors. Note that, sensitivity of the sensor depends whether it is operated in linear or nonlinear regime. Figure 1(c) summarizes all linear and nonlinear cantilever biosensors to be discussed in the following sections.

The rest of the article is organized as follows. In section 2, we discuss the spring-mass model of cantilever and illustrate classical mass and stress based sensors in section 3. We then present the emerging nonlinear biosensors like bifurcation based mass sensors and Flexure-FET in section 4. We finally conclude in section 5.

2. CANTILEVER BASED SENSORS AS SPRING MASS SYSTEM

The static as well as dynamic response of cantilever based sensors is best described by Euler-Bernoulli beam equation [18]. In this article, we however use a lumped parameter, spring-mass system (Fig. 2(b)) of a cantilever to illustrate its key features. The equation of motion of lumped parameter spring-mass system is given by-

$$m \frac{d^2 y}{dt^2} + \frac{m\omega_0}{Q} \frac{dy}{dt} - k(y_0 - y) - k'(y_0 - y)^3 = F_{ext}, \quad (1)$$

where m is the effective mass of the cantilever, y is the position of vibrating cantilever, t is time, ω_0 is the natural frequency of cantilever, Q is the quality factor, k is the effective spring constant of the cantilever such that $\omega_0 = \sqrt{k/m}$, y_0

is the position of the cantilever in its rest position, and k' is the constant associated with cubic nonlinearity of spring. Note that, $k = \frac{\alpha E W H^3}{(1-\nu)L^3}$ is the spring constant of the cantilever where α is a geometrical factor, E is the Young's modulus, ν is the Poisson's ratio, W is the width, H is the thickness, and L is the length. F_{ext} is the external force acting on the cantilever, e.g., surface forces, electrostatic forces, etc. Historically, the cantilever based sensors have been operated without applying any external force (i.e., $F_{ext} = 0$) and in linear response regime i.e., $k' \approx 0$ (section 3). We will explore the nonlinear $F_{ext} \neq 0$ & $k' \neq 0$ operation in section 4.

3. CLASSICAL LINEAR BIOSENSORS

3.1 Resonant mode mass sensors

In resonant mode sensing, vibrating nanocantilever can be used as a microbalance and bio-molecules can be detected by observing the change in dynamic response of the cantilever [1], [5]. Fundamentally, dynamic response of a cantilever is governed by its resonance frequency f_0 that is given by (using Eq. 1 with $k' = 0$ and $F_{ext} = 0$)-

$$f_0 = \frac{\omega_0}{2\pi} = \frac{1}{2\pi} \sqrt{\frac{k}{m}}. \quad (2)$$

Once the target molecules are captured, change in m (and/or k) shifts the resonance frequency (Δf) to indicate the capture of bio-molecules. Experimentally, resonance frequency of the cantilever can be obtained by measuring the amplitude-frequency spectrum of vibrating cantilever using optical techniques. Figure 2(c) shows amplitude-frequency spectrum of a vibrating cantilever for three different conditions: before functionalizing with receptor molecules (blue circles), after functionalizing (red squares), and after capture of target molecules (black diamond) [19]. By definition, the peak in the amplitude-frequency spectrum corresponds to the resonance frequency f_0 . As expected, resonance frequency decreases following the attachment of receptor molecules due to the added mass on the cantilever. It decreases *further* after capture of target molecules by receptor molecules due to further increase in the mass. Change in the resonance frequency due to the adsorption of the molecules can be obtained using Eq. (2), and is given by-

$$\frac{\Delta f}{f_0} \approx -\frac{\Delta m}{2m} + \frac{\Delta k}{2k}, \quad (3)$$

where Δm is the mass of added molecules and Δk is the change in stiffness. Figure 2(d) shows Δf as a function of Δm for two different cantilevers [20] and linear dependence of Δf on Δm confirms Eq. 3 (assuming $\Delta k = 0$). Equation (3) suggests that the sensitivity $S \equiv \Delta f/f_0$ of nanocantilever biosensors can only vary




Fig. 2: Dynamic resonant mode sensing using nanocantilever. **(a)** Schematic of a vibrating cantilever whose resonance frequency changes due to capture of target bio-molecules. **(b)** Equivalent spring-mass model of the vibrating cantilever. **(c)** Amplitude vs. frequency spectrum of a free standing cantilever (\circ), after receptor molecules attachment (\square), and after capture of target molecules (\diamond) [19]. **(d)** Change in resonance frequency as a function of added mass of the bio-molecules for two different cantilevers [20].

linearly with Δm (assuming $\Delta k = 0$) and therefore $S \propto N_s$ with N_s being the areal density of captured bio-molecules on cantilever surface. We emphasize that these biosensors – with careful design and appropriate instrumentation – can be extraordinarily sensitive; indeed, zeptogram mass detection has been reported [21]. It is also important to realize that the linear sensitivity with Δm is achieved only if the change in stiffness (Δk) due to capture of bio-molecules is negligible (Eq. (3)). In general, the capture of target molecules increases k [19]. If Δk compensates Δm , Eq. (3) suggests that there may be no change in resonance frequency at all (i.e., $\Delta f \sim 0$) and the sensitivity could be vanishingly small. Therefore, one must independently measure the change in k to decouple the ‘mass effect’ from ‘stiffness effect’, so that the mass of the adsorbed molecule can be correctly estimated [22], [23].

The sharpness of the peak (or the width of the amplitude-frequency spectrum) vibrating cantilever is characterized by its quality factor (Q) (Eq. (1)) and depends on the damping due to the surrounding medium. As Q increases, resonance peak becomes sharper and width of the spectrum is reduced. Unfortunately, value of minimum detectable Δf increases as Q is reduced [24]. Therefore, measurements in

vacuum or air can be more sensitive (capable of resolving small Δf and therefore smaller Δm) as compared that in fluidic environment [24]. This dependence of Δf on Q has inspired design of suspended microchannel resonators [25–27] that do not suffer from Q degradation due to the surrounding fluid. Such resonators are operated either in vacuum or in air and the fluid containing the target bio-molecules flows through the microchannel, embedded within the cantilever itself.

To summarize, the response of resonant mode nanocantilever based biosensors is linear with respect to the added mass of bio-molecules. Minimum detectable mass depends on the quality factor of the vibrating cantilever. And, suspended microchannel resonators can detect lower masses due to their high quality factors.

3. 2 Stress based static mode sensors

Another class of nanocantilever sensor involves operation in the static mode, in which capture of target molecules introduces a surface stress [5]. Changes in the surface stress can be the result of an adsorption process or electrostatic interactions between charged molecules on the surface or conformational changes of the adsorbed molecules. This change in the surface stress bends the cantilever as shown in Fig. 3(a). The deflection of the tip of the cantilever Δy is then measured as a signature of bio-molecules capture. Stoney's equation [28] relates the deflection Δy with the change in the surface stress $\Delta\sigma$ as follows-

$$\Delta y = \frac{3L^2(1-\nu)}{EH^2}\Delta\sigma, \quad (4)$$

where L is the length, ν is the Poisson's ratio, E is the Young's modulus, and H is the thickness of the cantilever. Note that, Eq. (4) can be obtained from Eq. (1) (with time derivatives and $k' = 0$) with appropriately chosen $F_{ext} = -\frac{3aWH}{L}\Delta\sigma$.

Deflection Δy of the cantilever tip can be measured optically (e.g., using a laser and photodiode) or electrically (e.g. using an integrated piezo-resistor). Figure 3(b) shows deflection Δy as a function of the target bio-molecules concentration in the solution for two different cantilevers having different geometrical dimensions [29]. The response Δy is *sub-linear* with respect to the concentration and it depends on the geometrical dimensions of the cantilever. Using Eq. (4) and data shown in Fig. 3(b), $\Delta\sigma$ can be calculated and is shown in Fig. 3(c). Interestingly, $\Delta\sigma$ for two different cantilevers follow a single curve (Fig. 3(c)), suggesting that $\Delta\sigma$ only depends on the concentration of the bio-molecules and not on the cantilever properties.

Instead of optical measurement of Δy or $\Delta\sigma$ through laser-photodiode system, one can measure $\Delta\sigma$ by measuring the change in the resistance of a piezoresistor




Fig. 3: Stress based sensing using the static response of a nanocantilever. (a) Schematic of a bent cantilever due to capture of bio-molecules on its surface. (b) Displacement of the tip of the cantilever as a function of bio-molecules concentration for two different cantilevers [29]. (c) Corresponding change in the surface stress of the cantilever. Symbols denote experimental data and solid line is just guide to the eye. (d) Ratio of the resistance of piezoresistive material attached to the cantilever after (R) and before (R_0) capture of bio-molecules [30]. Symbols denote experimental data and dotted line is just guide to the eye.

attached to the cantilever [30]. For these piezoresistive based cantilever biosensors, the sensitivity is defined as the ratio of resistance after (R) and before (R_0) the capture of bio-molecules i.e., $S_{Piezo} \equiv R/R_0$. Figure 4(d) shows S_{Piezo} for two different cantilevers, suggesting that response S_{Piezo} only varies logarithmically with bio-molecules concentration. We, therefore, conclude that these static mode nanocantilever sensors respond only sub-linearly i.e., $S \sim \ln(C/C_s)$ (Figs. 3(b)-(d)) to target analyte concentration (Figs. 3(b)-(d)).

4. EMERGING NONLINEAR BIOSENSORS

In the previous section, we have discussed classical linear biosensors that can either be operated in static or dynamic mode. Now, we discuss a new class of emerging nonlinear biosensors that utilize inherent instability of nanocantilever static/dynamic response to achieve better sensitivity towards bio-molecules capture.

4. 1 Bifurcation based mass sensors

As discussed above, classical resonant mode biosensors rely on the change in resonance frequency due to capture of bio-molecules. Note that, when operated in the linear regime (under small amplitude limit), the amplitude-frequency spectrum is symmetric and bell-shaped, as shown in Fig. 4(a). In this case, detection of bio-molecules is achieved by observing the shift in the peak (i.e., Δf , see Fig. 4(a)), as discussed in Sec 3.1.

In the large amplitude nonlinear response regime, however, higher order spring nonlinearities (i.e., $k' \neq 0$ in Eq. (1)) distorts the response, and amplitude-frequency spectrum is no longer symmetric [31]. Figure 4(b) shows one such amplitude-frequency spectrum with softening nonlinearity ($k' < 0$ and $F_{ext} = F_0 \sin(2\pi f_{ex} t)$ in Eq. (1) with F_0 being the excitation amplitude and f_{ex} is the excitation frequency) for a Duffing like resonator. Interestingly, spectrum exhibits sudden jumps at points P and S representing saddle-node bifurcations. The hysteretic behavior shown in Fig. 4(b) is achieved, when $F_0 > F_c$ with F_c being a critical threshold. Kumar et al., has proposed a bifurcation based mass sensor that utilizes these sudden jumps and rely on the shift in the amplitude and not on the shift in the frequency to signal bio-molecules capture [32]. In bifurcation based sensing, the resonator is operated near one of the critical point (say P). Capture of the bio-molecules reduces the fundamental frequency f_0 and increases f_{ex}/f_0 resulting in the sudden change in the amplitude of oscillation ΔA , as shown in Figs. 4(b)-(c). Measurement of ΔA (using laser Doppler vibrometer (LDV)) is then used as the signature of capture of bio-molecules. It should be appreciated that this sensing scheme is very sensitive to small quantities of added molecules due to the amplification offered by inherent instability of mechanical system.

4. 2 Electromechanical coupling based Flexure-FET biosensors

We have discussed both linear and nonlinear cantilever based biosensors that rely on optical readout of y or f_0 . Now, we discuss a new class of nonlinear biosensors called Flexure-FET [16] that utilize the electromechanical coupling between a suspended beam and a field effect transistor to achieve much higher sensitivity compared to traditional biosensors. Flexure-FET consists of a channel biased through a thin-film suspended gate (Fig. 5(a)). While the structure is similar to that of a suspended-gate FET [33], NEMFET [10] or resonant gate transistor [34], we call the sensor Flexure-FET to emphasize its distinctive nonlinear operation specifically optimized for ultrasensitive detection of bio-molecules. In a Flexure-FET, any change in the mechanical property of the suspended gate is directly reflected in the change of drain current of underneath field effect transistor and




Fig. 4: Comparison of classical linear mass sensors with nonlinear bifurcation based mass sensors. Amplitude-frequency spectrum of (a) classical linear and (b) bifurcation based mass sensors proposed in ref. [32]. (c) Amplitude as a function of time in bifurcation based mass sensors. Classical sensors rely on the change in frequency Δf whereas bifurcation based sensors rely on change in the amplitude ΔA due to capture of bio-molecules.

thereby enables electrical readout. As shown in Fig. 5(b), ultra high sensitivity arises from the coupling of two nonlinear responses, namely (i) spring-softening [35] in which stiffness decreases nonlinearly with the applied gate bias V_G and vanishes at the pull-in point (for detailed discussions on pull-in instability, see Ref. [36], [37]), and (ii) sub-threshold electrical conduction [38] in which current depends exponentially on the surface potential. Such nonlinear electro-mechanical coupling enables exponentially high sensitivity for Flexure-FET sensors, which is fundamentally unachievable by exclusive use of existing nanoscale electronic or mechanical biosensors.

It should be noted that from a mechanical perspective, Flexure-FET operates close to pull-in instability, a critical point. Similar critical point sensing has also been reported for vapor sensors that operates close to bucking-instability [39], [40] and for mass sensor that operates close to saddle-node bifurcation [32] (discussed in Sec. 4.1) and their higher sensitivity have been confirmed experimentally. However, beyond the critical point sensing, the integrated transistor-action in the sub-threshold regime provides the Flexure-FET an *additional* exponential sensitivity (and simpler DC read-out) that could not be achieved by the classical nonlinear sensor schemes.

The operating principle of Flexure-FET can be understood based on a spring-mass system coupled to electrostatic actuation, see Fig. 6 [10], [34]. With the application of gate bias V_G , the gate moves downward towards the dielectric (y vs. V_G curve in Fig. 5(b)) and the corresponding increase in gate capacitance is reflected in the increased drain current I_{DS} , as shown in Fig. 5(b). The static behavior of the device is dictated by the balance of spring and electrostatic forces (Eq. (1) with time derivatives zero and $k' = 0$), i.e.,




Fig. 5: Static stiffness based nonlinear electromechanical sensing with electrical readout. (a) Schematic of Flexure-FET bio-sensor and (b) y and I_{DS} vs. gate voltage characteristic of Flexure-FET. V_T is the threshold voltage and V_{PI} is the pull-in voltage of Flexure-FET.

$$k(y_0 - y) = F_e \equiv \frac{1}{2} \epsilon_0 E_{air}^2 A, \quad (5)$$

where E_{air} is the electric field in the air, and $A = WL$ is the area of the gate electrode. The electric field below the membrane E_{air} is equal to $\epsilon_s E_s(\psi_s)$, where, ϵ_s is the dielectric constant of the substrate, and

$$E_s(\psi_s) = \sqrt{\frac{2qN_A}{\epsilon_0 \epsilon_s} \left[\psi_s + \left(e^{\frac{q\psi_s}{k_B T}} - 1 \right) \frac{k_B T}{q} - \left(\frac{n_i}{N_A} \right)^2 \left(\psi_s - \left(e^{\frac{q\psi_s}{k_B T}} - 1 \right) \frac{k_B T}{q} \right)^2 \right]^{\frac{1}{2}}}, \quad (6a)$$

where, $E_s(\psi_s)$ is the electric field at the substrate-dielectric interface (see Ref. [38] page 64 for a detailed derivation of Eq. 6(a)), ψ_s is the surface potential, q is the charge of an electron, N_A is the substrate doping, k_B is the Boltzmann constant, T is the absolute temperature, and n_i is the intrinsic carrier concentration in the substrate. The voltage drop in air ($y\epsilon_s E_s(\psi_s)$), dielectric ($\frac{y_d}{\epsilon_d} \epsilon_s E_s(\psi_s)$), and substrate (ψ_s) can be related to the applied gate bias V_G as follows-




Fig. 6 (a)-(b) Equivalent spring-mass model of Flexure-FET. Stiffness changes from k to $k + \Delta k$ after the capture of biomolecules, and therefore, position of gate changes from y to $y + \Delta y$ which results in the modulation of drain current from I_{DS1} to I_{DS2} .

$$V_G = \left(y + \frac{y_d}{\epsilon_d} \right) \epsilon_s E_s(\psi_s) + \psi_s, \quad (6b)$$

where, y_d is the dielectric thickness. Equations (5)-(6) are solved self-consistently for y and ψ_s at each V_G . The corresponding inversion charge density (Q_i) in the channel and drain current (I_{DS}) are given by,

$$Q_i = \frac{qn_i^2}{N_A} \int_0^{\psi_s} \frac{e^{\frac{q\psi}{k_B T}} - 1}{E_s(\psi)} d\psi, \quad (7)$$

$$I_{DS} = \mu_n L Q_i \frac{V_{DS}}{W}, \quad (8)$$

where, μ_n is the channel mobility for electrons, V_{DS} is the applied drain to source voltage. Figure 5(b) shows the steady-state response of Flexure-FET as a function of biasing voltage V_G , obtained from the numerical simulations of Eqs. (5)-(8).

For transduction, Flexure-FET biosensor utilizes the change in suspended-gate stiffness from k to $k + \Delta k$, due to the capture of bio-molecules. The change in stiffness due to the capture of bio-molecules has been demonstrated by several recent experiments of mass sensing using nanocantilever based resonators [19], [41–43]. This well-known observation of stiffness change has been attributed to the change in the membrane thickness, Young's modulus, and/or surface stress of the beam [19], [41–43]. Indeed, Craighead in Ref. [44] suggests its use as a basis of a new class of mechanical biosensor.




Fig. 7: Change in the sensor characteristics due to capture of target molecules on the surface of the gate, (a) y vs. V_G before and after capture, and (b) corresponding change in the position of gate electrode Δy vs. V_G . Δy increases rapidly near pull-in due to spring-softening effect. The capture of target molecules is directly mirrored in the change in I_{DS} . (c) I_{DS} vs. V_G for before and after capture, and (d) corresponding ratio of the two currents I_{DS1} (before capture) and I_{DS2} (after capture) as a function of Δy . Orders of magnitude change in I_{DS} can be easily achieved for typical surface density of $N_s = 5 * 10^{12} cm^{-2}$, projected area of the bio-molecule, $A_t = \pi R_t^2$ with $R_t = 1nm$, and $H_t = 5.1nm$. These parameters translate to just an equivalent $\Delta k \sim 6\%$. The device considered has the following typical parameters: $L = 4\mu m$, $W = 1\mu m$, $H = 40nm$, $E = 200GPa$, $y_0 = 100nm$, $y_d = 5nm$, $\epsilon_s = 11.7$, $\epsilon_d = 3.9$, $N_A = 6 * 10^{16} cm^{-3}$.

In the following analysis, we model change in k by the change in the effective thickness H of the gate (ΔH), *although it should be stressed that the conclusions do not depend on the particular hypothesis regarding Δk* . For now, we ignore the details of the spatial distribution of molecules associated with random sequential adsorption [15], and assume a uniform distribution of adsorbed molecules on the sensor surface. Therefore, the conservation of volume suggests $\Delta H = N_s A_t H_t$, where N_s is the areal density, A_t is the effective cross-sectional area, and H_t is the effective thickness of the target molecule. Using the fact that $k = \frac{\alpha E W H^3}{(1-\nu)L^3}$, change in

stiffness Δk due to ΔH ($\ll H$) can be related to adsorbed molecule density N_s as follows,

$$\frac{\Delta k}{k} \approx \frac{3N_s A_t H_t}{H}. \quad (9)$$

It can be shown that if Flexure-FET is operated close to pull-in and in sub-threshold regime, sensitivity S (using Eqs. (5)-(9), see Ref. [16] for details of derivation) is given by-

$$S_{Flexure} \equiv \frac{I_{DS1}}{I_{DS2}} \approx \exp(\gamma_1 \sqrt{N_s} - \gamma_2 N_s), \quad (10)$$

where γ_1/γ_2 are two sensor geometry dependent constant. Equation 10 confirms the exponential sensitivity of Flexure-FET towards bio-molecules capture.

The results for the change in sensor characteristics due to the capture of bio-molecules are summarized in Fig. 7. For example, Fig. 7(a) shows y vs. V_G before and after capture of target molecules. After the capture, gate moves up (for a fixed V_G) due to increased restoring spring force (because of increase in k , see Fig. 7(a)). Interestingly, change in gate position Δy is maximum close to pull-in due to spring-softening effect, as shown in Fig. 7(b). The change in gate position Δy is directly reflected in change in I_{DS} . Figure 7(c) shows I_{DS} vs. V_G before and after capture of bio-molecules. Interestingly, I_{DS} *decreases* after capture due to increased separation between the gate and the dielectric (hence decreased capacitance). The corresponding ratio of the currents I_{DS1} (before capture) and I_{DS2} (after capture) increases exponentially with Δy (Fig. 7(d)), and becomes maximum near pull-in.

5. CONCLUSIONS

In this review article, we have discussed various ways of detecting bio-molecules using nanocantilevers. Classical resonant mode biosensors detect the change in the resonance frequency of vibrating cantilever and require complex optical instrumentation for detection, especially when very high sensitivity is desired. Stress based static mode sensors detect the deflection of the tip of the cantilever and responds linearly to change in the stress. To achieve better sensitivity than achieved by classical linear biosensors, critical-point nonlinear bio-sensors have started to appear in the literature. For example, we discussed bifurcation based mass sensors that operate close to a saddle-node bifurcation. Finally, we have discussed Flexure-FET biosensor that integrates a transistor for direct electrical readout and utilizes nonlinear electromechanical coupling for its exponential sensitivity. We believe that these critical point nonlinear biosensors with electrical readout will offer opportunity to integrate highly sensitive sensors in low cost point-of-care applications.

References

- [1] J. L. Arlett, E. B. Myers, and M. L. Roukes, "Comparative advantages of mechanical biosensors," *Nature Nanotechnology*, vol. 6, pp. 203-215, 2011.
- [2] J. Tamayo, P. M. Kosaka, J. J. Ruz, A. San Paulo, and M. Calleja, "Biosensors based on nanomechanical systems," *Chemical Society reviews*, vol. 42, no. 3, pp. 1287-311, Jan. 2013.
- [3] F. J. Norton, "Gas Analyzer," U.S. Patent 2,307,8001943.
- [4] G. Binnig and C. F. Quate, "Atomic Force Microscope," *Physical Review Letters*, vol. 56, no. 9, pp. 930-933, Mar. 1986.
- [5] N. V. Lavrik, M. J. Sepaniak, and P. G. Datskos, "Cantilever transducers as a platform for chemical and biological sensors," *Review of Scientific Instruments*, vol. 75, pp. 2229-2253, 2004.
- [6] P. Bergveld, "Thirty years of ISFETOLOGY - What happened in the past 30 years and what may happen in the next 30 years," *Sensors and Actuators B-Chemical*, vol. 88, pp. 1-20, 2003.
- [7] G. M. Rebeiz and J. B. Muldavin, "RF MEMS switches and switch circuits," *IEEE Microwave Magazine*, vol. 2, pp. 59-71, 2001.
- [8] J. J. Yao, "RF MEMS from a device perspective," *Journal of Micromechanics and Microengineering*, vol. 10, p. R9-R38, 2000.
- [9] V. Pott, H. Kam, R. Nathanael, J. Jeon, E. Alon, and T.-J. King Liu, "Mechanical Computing Redux: Relays for Integrated Circuit Applications," *Proceedings of the IEEE*, vol. 98, no. 12, pp. 2076-2094, Dec. 2010.
- [10] H. Kam, D. T. Lee, R. T. Howe, and T.-J. King, "A new nano-electro-mechanical field effect transistor (NEMFET) design for low-power electronics," in *IEDM*, 2005, pp. 463-466.
- [11] T. Bifano, "Adaptive imaging: MEMS deformable mirrors," *Nature Photonics*, vol. 5, no. 1, pp. 21-23, Jan. 2011.
- [12] Q. M. Technologies, "Value Proposition of mirasol Displays," 2011.

- [13] L. M. Roylance and J. B. L. A.-E. Angell, "A batch-fabricated silicon accelerometer," *IEEE Transactions on Electron Devices*, vol. 26, pp. 1911-1917, 1979.
- [14] P. R. Nair and M. A. Alam, "Performance limits of nanobiosensors," *Applied Physics Letters*, vol. 88, no. 23, p. 233120, Jun. 2006.
- [15] P. R. Nair and M. A. Alam, "Theory of 'Selectivity' of label-free nanobiosensors: A geometro-physical perspective.," *Journal of applied physics*, vol. 107, no. 6, p. 64701, Mar. 2010.
- [16] A. Jain, P. R. Nair, and M. A. Alam, "Flexure-FET biosensor to break the fundamental sensitivity limits of nanobiosensors using nonlinear electromechanical coupling.," *Proceedings of the National Academy of Sciences of the United States of America*, vol. 109, no. 24, pp. 9304-8, Jun. 2012.
- [17] P. R. Nair and M. A. Alam, "Screening-limited response of nanobiosensors," *Nano Letters*, vol. 8, pp. 1281-1285, 2008.
- [18] C. Ziegler, "Cantilever-based biosensors," *Analytical and Bioanalytical Chemistry*, vol. 379, no. 7-8, pp. 946-959, 2004.
- [19] A. K. Gupta et al., "Anomalous resonance in a nanomechanical biosensor," *Proceedings of the National Academy of Sciences of the United States of America*, vol. 103, pp. 13362-13367, 2006.
- [20] K. L. Ekinci and M. L. Roukes, "Nanoelectromechanical systems," *Review of Scientific Instruments*, vol. 76, 2005.
- [21] Y. T. Yang, C. Callegari, X. L. Feng, K. L. Ekinci, and M. L. Roukes, "Zeptogram-scale nanomechanical mass sensing," *Nano Letters*, vol. 6, pp. 583-586, 2006.
- [22] F. Sadeghian, H. Goosen, A. Bossche, and F. van Keulen, "Application of electrostatic pull-in instability on sensing adsorbate stiffness in nanomechanical resonators," *Thin Solid Films*, vol. 518, pp. 5018-5021, 2010.
- [23] E. Gil-Santos et al., "Nanomechanical mass sensing and stiffness spectrometry based on two-dimensional vibrations of resonant nanowires," *Nature Nanotechnology*, vol. 5, pp. 641-645, 2010.

- [24] K. L. Ekinici, "Ultimate limits to inertial mass sensing based upon nanoelectromechanical systems," *Journal of Applied Physics*, vol. 95, no. 5, p. 2682, Mar. 2004.
- [25] T. P. Burg et al., "Weighing of biomolecules, single cells and single nanoparticles in fluid.," *Nature*, vol. 446, no. 7139, pp. 1066-9, May 2007.
- [26] J. Lee, W. Shen, K. Payer, T. P. Burg, and S. R. Manalis, "Toward attogram mass measurements in solution with suspended nanochannel resonators.," *Nano letters*, vol. 10, no. 7, pp. 2537-42, Jul. 2010.
- [27] J. L. Arlett and M. L. Roukes, "Ultimate and practical limits of fluid-based mass detection with suspended microchannel resonators," *Journal of Applied Physics*, vol. 108, no. 8, p. 084701, Oct. 2010.
- [28] G. G. Stoney, "The Tension of Metallic Films Deposited by Electrolysis," *Proc. R. Soc. Lond. A*, vol. 82, p. 172, 1909.
- [29] G. H. Wu, R. H. Datar, K. M. Hansen, T. Thundat, R. J. Cote, and A. Majumdar, "Bioassay of prostate-specific antigen (PSA) using microcantilevers," *Nature Biotechnology*, vol. 19, pp. 856-860, 2001.
- [30] K. W. Wee et al., "Novel electrical detection of label-free disease marker proteins using piezoresistive self-sensing micro-cantilevers," *Biosensors and Bioelectronics*, vol. 20, pp. 1932-1938, 2005.
- [31] H. A. C. Tilmans, M. Elwenspoek, and J. H. J. Fluitman, "Micro resonant force gauges," *Sensors and Actuators A: Physical*, vol. 30, no. 1-2, pp. 35-53, Jan. 1992.
- [32] V. Kumar et al., "Bifurcation-based mass sensing using piezoelectrically-actuated microcantilevers," *Applied Physics Letters*, vol. 98, 2011.
- [33] N. Abele, R. Fritschi, K. Boucart, F. Casset, P. Ancey, and A. M. Ionescu, "Suspended-Gate MOSFET: bringing new MEMS functionality into solid-state MOS transistor," in *IEDM*, 2005, vol. 0, no. c.
- [34] H. C. Nathanson, W. E. Newell, R. A. Wickstrom, and J. R. Davis, "The Resonant Gate Transistor," *Transactions on Electron Devices*, vol. 14, no. 3, pp. 117-133, 1967.
- [35] H. Torun, K. K. Sarangapani, and F. L. Degertekin, "Spring constant tuning of active atomic force microscope probes using electrostatic spring softening effect," *Applied Physics Letters*, vol. 91, 2007.

- [36] S. Krylov and R. Maimon, "Pull-in dynamics of an elastic beam actuated by continuously distributed electrostatic force," *Journal of Vibration and Acoustics-Transactions of the Asme*, vol. 126, pp. 332-342, 2004.
- [37] S. Krylov, "Lyapunov exponents as a criterion for the dynamic pull-in instability of electrostatically actuated microstructures," *International Journal of Non-Linear Mechanics*, vol. 42, pp. 626-642, 2007.
- [38] T. Yuan and N. T. H., *Fundamentals of Modern VLSI Devices*. Cambridge University Press.
- [39] D. R. Southworth, L. M. Bellan, Y. Linzon, H. G. Craighead, and J. M. C.-A. 163503 Parpia, "Stress-based vapor sensing using resonant microbridges," *Applied Physics Letters*, vol. 96, 2010.
- [40] D. J. Joe et al., "Stress-based resonant volatile gas microsensor operated near the critically buckled state," *Journal of Applied Physics*, vol. 111, no. 10, p. 104517, May 2012.
- [41] J. Tamayo, D. Ramos, J. Mertens, and M. Calleja, "Effect of the adsorbate stiffness on the resonance response of microcantilever sensors," *Applied Physics Letters*, vol. 89, 2006.
- [42] P. S. Waggoner and H. G. Craighead, "The relationship between material properties, device design, and the sensitivity of resonant mechanical sensors," *Journal of Applied Physics*, vol. 105, 2009.
- [43] D. Ramos, J. Tamayo, J. Mertens, M. Calleja, and A. Zaballos, "Origin of the response of nanomechanical resonators to bacteria adsorption," *Journal of Applied Physics*, vol. 100, 2006.
- [44] H. Craighead, "Nanomechanical systems - Measuring more than mass," *Nature Nanotechnology*, vol. 2, pp. 18-19, 2007.