

AIRBORNE ULTRASONIC TACTILE DISPLAY BRAIN–COMPUTER INTERFACE PARADIGM

KATSUHIKO HAMADA

The University of Tokyo, Tokyo, Japan

HIROMU MORI

Life Science Center of TARA, University of Tsukuba, Japan

HIROYUKI SHINODA

The University of Tokyo, Tokyo, Japan

TOMASZ M. RUTKOWSKI*

*Life Science Center of TARA, University of Tsukuba, Japan
RIKEN Brain Science Institute, Japan*

ABSTRACT. We study the extent to which contact-less and airborne ultrasonic tactile display (AUTD) stimuli delivered to the palms of a user can serve as a platform for a brain computer interface (BCI) paradigm. Six palm positions are used to evoke combined somatosensory brain responses, in order to define a novel contact-less tactile BCI. A comparison is made with classical attached vibrotactile transducers. Experiment results of subjects performing online experiments validate the novel BCI paradigm.

CONTENTS

1. Introduction	2
2. Materials and Methods	2
3. Results and Conclusions	3
Author contributions	5
Acknowledgements	5
References	5

E-mail addresses: tomek@bci-lab.info.

*Corresponding author.

1. INTRODUCTION

State-of-the-art brain computer interfaces (BCIs) are typically based on mental visual or auditory paradigms, as well as motor imagery paradigms, which require extensive user training and good eyesight or hearing. In recent years, alternative solutions have been proposed to make use of a tactile modality [MPSNP06, BVE10, MMM⁺12] to enhance brain-computer interfacing efficiency. The concept reported in this paper further extends the brain's somatosensory channel by the application of a contact-less stimulus generated with an airborne ultrasonic tactile display (AUTD) [ITS08]. The rationale behind the use of the AUTD is that, due to its contact-less nature, it allows for a more hygienic application, avoiding the occurrence of skin ulcers (bedsores) in patients in a locked-in state (LIS). This paper reports very encouraging results with AUTD-based BCI (autdBCI) in comparison with the classical paradigm of vibrotactile transducer-based somatosensory stimulus (vtBCI) attached to the user's palms [MMM⁺12]. The rest of the paper is organized as follows. The next section introduces the materials and methods used in the study. The results obtained in online experiments with 13 healthy BCI users are then discussed. Finally, conclusions are formulated and directions for future research are outlined.

2. MATERIALS AND METHODS

Thirteen male volunteer BCI users participated in the experiments. The users' mean age was 28.54, with a standard deviation of 7.96 years. The experiments were performed at the Life Science Center of TARA, University of Tsukuba, at the University of Tokyo and at RIKEN Brain Science Institute, Japan. The online (real-time) EEG autdBCI and vtBCI paradigm experiments were conducted in accordance with the *WMA Declaration of Helsinki - Ethical Principles for Medical Research Involving Human Subjects* and the procedures were approved and designed in agreement with the ethical committee guidelines of the Faculty of Engineering, Information and Systems at University of Tsukuba, Japan. The AUTD stimulus generator produced vibrotactile contact-less stimulation of the human skin via the air using focused ultrasound [ITS08, Ham14]. The effect was achieved by generating an ultrasonic radiation static force produced by intense sound pressure amplitude (a nonlinear acoustic phenomenon). The radiation pressure deformed the surface of the skin on the palms, creating a tactile sensation. An array of ultrasonic transducers mounted on the AUTD created the focussed radiation pressure at an arbitrary focal point by choosing a phase shift of each transducer appropriately (the so-called phased array technique). Modulated radiation pressure created a sensation of tactile vibration similar to the one delivered by classical vibrotactile transducers attached to the user's palms. The AUTD device developed by the authors [ITS08, Ham14] adhered to ultrasonic medical standards and did not exceed the permitted skin absorption levels (approximately 40 times below the limits). The effective vibrotactile sensation was set to 50 Hz [Ham14]. As a reference, in the second vtBCI experiment, contact vibrotactile stimuli were also applied to locations on the users' palms via the transducers H1HX09C005-8. Each transducer in the experiments was set to emit a square acoustic frequency wave at 50 Hz, which was delivered from the ARDUINO micro-controller board with a custom battery-driven and isolated power amplifier and software developed in-house and managed from a MAX 6 visual programming environment. The two experiment set-ups above are presented in Figure 1. Two types of experiments were performed with the volunteer users. Psychophysical experiments with button-press responses were conducted in order to test uniform stimulus difficulty levels from response accuracy and time measurements. The subsequent online BCI EEG experiments evaluated the autdBCI paradigm efficiency and allowed for a comparison with the classical skin contact-based vtBCI reference. In both the above experiment protocols, the users were instructed to spell sequences of six digits representing the stimulated positions on




FIGURE 1. Tactile stimulus set-ups for autdBCI and vtBCI experiments.

their palms. The training instructions were presented visually by means of the *BCI2000* program [SM10] with the numbers 1 – 6 representing the palm locations as depicted in the left panel of Figure 1. The EEG signals were captured with an EEG amplifier system g.USBamp by g.tec Medical Engineering GmbH, Austria, using 16 active electrodes. The electrodes were attached to the head locations: *Cz*, *Pz*, *P3*, *P4*, *C3*, *C4*, *CP5*, *CP6*, *P1*, *P2*, *POz*, *C1*, *C2*, *FC1*, *FC2*, and *FCz*, as in the 10/10 extended international system. The ground electrode was attached to the *FPz* position, and the reference was attached to the left earlobe. No electromagnetic interference was observed from the AUTD or vibrotactile transducers operating with frequencies notch-filtered together with power line interference from the EEG. The EEG signals captured were processed online with a *BCI2000*-based application [SM10], using a stepwise linear discriminant analysis (SWLDA) classifier [KSC⁺06] with features drawn from the 0 – 800 ms ERP intervals. The stimulus length and inter-stimulus-interval were set to 400 ms, and the number of averages in BCI trials was set to 15 in order to collect enough data for the classifier training. The EEG recording sampling rate was set at 512 Hz, and the high and low pass filters were set at 0.1 Hz and 60 Hz, respectively. The notch filter to remove power line interference was set for a rejection band of 48 ~ 52 Hz. Each user performed three experiment sessions (randomized 90 targets and 450 non-targets each), which were later averaged for the online SWLDA classifier.

3. RESULTS AND CONCLUSIONS

The averaged evoked responses to targets and non-targets are depicted together with standard error bars in Figure 2. The BCI six digit sequences spelling accuracy analyses for both the experiments for the various averaging options are summarized in Figure 3. The chance level was of 16.6%. The mean accuracies for 15-trial averaged ERPs were 63.8% and 69.4% for autdBCI and vtBCI, respectively. The maximum accuracies were 78.3% and 84.6% respectively. The differences were not significant, supporting the concept of AUTD-based tactile stimulus usability for BCI. However, a single trial classification offline analysis of the collected responses resulted in mean accuracies of 83.0% for autdBCI and 53.8% for vtBCI, leading to a possible 19.2 bit/min and 7.9 bit/min, respectively. In the case of the autdBCI, only a single user's results were bordering on the level of chance, and four subjects attained 100% (10 trials averaging). On average, lower accuracies were obtained with the classical vtBCI, with which three users bordered on the level of chance, and only one user scored 100% accuracy level in SWLDA-classified averaged responses. This case study demonstrates results obtained with a novel six-command-based aut-




FIGURE 2. The autdBCI (blue - targets; red - non-targets) and vtBCI (green - targets; black - non-targets) grand mean averaged ERP responses, together with standard error bars. Due to limited space, of the 16 measured, only electrodes *Cz* and *Pz* are presented.

dBCI paradigm. We compared the results with classical vibrotactile transducer stimuli already generated. The experiment results obtained in this study confirm the validity of the contact-less autdBCI for interactive applications and the possibility to further improve the results with the utilization of single trial-based linear classification. The EEG experiment with the paradigm confirms that contact-less (airborne) tactile stimuli can be used to create six command-based interfaces. The results presented offer a step forward in the development of novel neurotechnology applications. Due to the still not very high interfacing rate achieved by users in the case of online BCI, the current paradigm obviously requires improvement and modification. These requirements determine the major lines of study for future research.

However, even in its current form, the proposed autdBCI can be regarded as a practical solution for LIS patients (locked into their own bodies despite often intact cognitive functioning), who cannot use vision or auditory-based interfaces due to sensory or other disabilities.




FIGURE 3. Averaged autdBCI and vtBCI six digits spelling accuracy results colour coded separately for each user, with standard error bars depicted.

AUTHOR CONTRIBUTIONS

Designed and performed the EEG experiments: KH, HM, TMR. Analyzed the data: KH, HM, TMR. Conceived the concept of the AUTD-based BCI paradigm: TMR, HS. Supported the project: HS, TMR. Wrote the paper: KH, TMR.

ACKNOWLEDGEMENTS

Hiromu Mori and Tomasz M. Rutkowski were supported in part by the Strategic Information and Communications R&D Promotion Program (SCOPE) no. 121803027 of The Ministry of Internal Affairs and Communications in Japan.

REFERENCES

- [BVE10] Anne-Marie Brouwer and Jan B F Van Erp. A tactile P300 brain-computer interface. *Frontiers in Neuroscience*, 4(19), 2010.
- [Ham14] Katsuhiko Hamada. Brain-computer interface using airborne ultrasound tactile display. Master thesis, The University of Tokyo, Tokyo, Japan, March 2014. In Japanese.
- [ITS08] Takayuki Iwamoto, Mari Tatezono, and Hiroyuki Shinoda. Non-contact method for producing tactile sensation using airborne ultrasound. In Manuel Ferre, editor, *Haptics: Perception, Devices and Scenarios*, volume 5024 of *Lecture Notes in Computer Science*, pages 504–513. Springer Berlin Heidelberg, 2008.
- [KSC⁺06] Dean J Krusienski, Eric W Sellers, François Cabestaing, Sabri Bayouduh, Dennis J McFarland, Theresa M Vaughan, and Jonathan R Wolpaw. A comparison of classification techniques for the P300 speller. *Journal of Neural Engineering*, 3(4):299, 2006.
- [MMM⁺12] Hiromu Mori, Yoshihiro Matsumoto, Shoji Makino, Victor Kryssanov, and Tomasz M. Rutkowski. Vibrotactile stimulus frequency optimization for the haptic BCI prototype. In *Proceedings of The 6th International Conference on Soft Computing and Intelligent Systems, and The 13th International Symposium on Advanced Intelligent Systems*, pages 2150–2153, Kobe, Japan, November 20–24, 2012.
- [MPSNP06] G.R. Muller-Putz, R. Scherer, C. Neuper, and G. Pfurtscheller. Steady-state somatosensory evoked potentials: suitable brain signals for brain-computer interfaces? *Neural Systems and Rehabilitation Engineering, IEEE Transactions on*, 14(1):30–37, March 2006.
- [SM10] G. Schalk and J. Mellinger. *A Practical Guide to Brain-Computer Interfacing with BCI2000*. Springer-Verlag London Limited, 2010.