



23rd Conference on Application of Accelerators in Research and Industry, CAARI 2014

A New Proton CT Scanner

G. Coutrakon^a, G. Blazey^a, S. Boi^a, A. Dyshkant^a, B. Erdelyi^a, D. Hedin^a,
E. Johnson^a, J. Krider^a, V. Rykalin^a, S. A. Uzunyan^{a,*}, V. Zutshi^a, R. Fordt^b,
G. Sellberg^b, J. E. Rauch^b, M. Roman^b, P. Rubinov^b, P. Wilson^b, M. Naimuddin^c

^aDepartment of Physics, Northern Illinois University, DeKalb, IL 60115, USA

^bFermi National Accelerator Laboratory, Batavia, IL 60510, USA

^cDelhi University, 110007, India

Abstract

The design, construction, and preliminary testing of a second generation proton CT scanner is presented. All current treatment planning systems at proton therapy centers use X-ray CT as the primary imaging modality for treatment planning to calculate doses to tumor and healthy tissues. One of the limitations of X-ray CT is in the conversion of X-ray attenuation coefficients to relative (proton) stopping powers, or RSP. This results in more proton range uncertainty, larger target volumes and therefore, more dose to healthy tissues. To help improve this, we present a novel scanner capable of high dose rates, up to 2 MHz, and large area coverage, 20 x 24 cm², for imaging an adult head phantom and reconstructing more accurate RSP values.

© 2014 Published by Elsevier Ltd.

Selection and peer-review under responsibility of the Organizing Committee of CAARI 2014.

Keywords: proton Computed Tomography, calorimeter, fiber tracker

1. Introduction

Northern Illinois University in collaboration with Fermi National Accelerator Laboratory (FNAL) and Delhi University has been designing and building a proton CT scanner for applications in proton treatment planning. In proton therapy, the current treatment planning systems are based on X-ray CT images that have intrinsic limitations in terms of dose accuracy to tumor volumes and nearby critical structures. Proton CT aims to overcome these limitations by determining more accurate relative proton stopping powers directly as a result of imaging with protons. At present, the proton RSPs for various tissues, as derived from X-ray CT, produce range uncertainties [1] of about 3 to 4%. We hope to reduce this to approximately 1% of the total range using proton CT. In addition, three to five times lower doses than X-ray CT are possible and absence of artifacts from high density dental or other implants will add to higher quality images. The proton CT imaging requires reconstruction of the individual proton tracks and their energy losses in the scanned volume. The number of protons to acquire for the head-size volume scan is of order 1 billion. To finish scan

*Corresponding author. E-mail address: serguei@nicadd.niu.edu

in a time acceptable for the patients the track collection rate should be of of order 2 MHz, that requires fast tracker and energy detectors. To date two proton CT scanners are under development in the United States. The system that uses silicon strip technology for the tracker planes and five plastic scintillators for the range measurements was built in the Santa Cruz Institute of Particle Physics (SCIPP) and is undergoing testing at Loma Linda University Medical Center [2]. We describe a proton CT scanner based on fiber tracker and scintillator stack range detector that has been developed at Northern Illinois University in conjunction with FNAL in Batavia, IL. Fig. 1 shows a schematic of a proton CT scanner. It consists of eight planes of tracking




Fig. 1. Four (X,Y) stations measure the proton trajectory before and after the patient. A stack of 3.2 mm thick scintillator tiles is used to measure the residual energy or range after the patient.

detectors; two X and two Y coordinate measurements both before and after the patient. This provides the information for finding the trajectory through the head to correct, as much as possible, for multiple coulomb scattering in the patient. A “most likely path” formalism [3] is used to find which voxels, of order 1 mm^3 , are crossed by every track. In addition, a calorimeter consisting of a stack of thin scintillator tiles is used to determine the water equivalent path length (WEPL) of each track through the head. The X-Y coordinates and WEPL are required input for image reconstruction software to find RSP value of each voxel in the head and generate corresponding 3D image.

2. Design specifications

In addition to a high data rate of 2 MHz, we wish to cover a large enough area to image an adult human head so that table motion is not required or that we do not need to splice data from multiple scans to make an image long enough along the body axis. For head scans, we have chosen a maximum head size of 23 cm diameter and a length along the body axis of 20 cm. This will allow imaging of the head down to the jaw bone in one 360° gantry rotation. A fixed incident proton beam energy of 200 MeV with a range of 26 cm in water can be used for head size imaging. This proton CT detector is compatible with the geometric constraints of most proton treatment nozzles and patient positioners. Beam spreading from an effective source in the nozzle sets the detector sizes required for cone beam geometry. Multiple coulomb scattering in the tracking detectors requires a reduction of the mass of the detectors as much as possible. For this reason, each tracking plane has a water equivalent thickness less than 1 mm.

3. Detector Design and Construction

In order to have low mass detectors, with high proton rates, and continuous area coverage over a large area, the tracker was constructed from 0.5 mm diameter polystyrene scintillating fibers by Kuraray [4]. Fibers were initially cut to 50 cm length, then laid flat, and doubled layered (see Fig. 2) on a low density, 0.03 g/cm^3 , 2 mm thick rohocell substrate with machined grooves and glued to hold the fibers in place with

close spacing to avoid gaps in detecting passing protons. The entire assembly is supported on carbon fiber frames. A photograph of one tracker plane, 20 x 24 cm, is shown in Fig. 3a. Fibers are grouped in triplet, called bundles, according to Fig. 2 which give a pitch between bundles of 0.94 mm. Each bundle is readout into silicon photo multipliers (SiPMs), produced by CPTA [5] which are mounted on Techtron blocks that connect each of them to a fiber triplet. The SiPMs chosen have the best chromatic (or wavelength) match to the Kuraray scintillators. One end of each fiber is polished and mirrored. The other end is polished and mechanically pressed to an SiPM on a block shown in Fig. 3b. The rms spatial resolution of each tracker plane is given by the pitch divided by $\sqrt{12}$, or 0.27 mm. The integrated water equivalent thickness (WET) of each tracker along the beam direction is less than 1 mm. With four planes of 20 x 24 cm² in area and four planes with 24 x 30 cm² in area, there are about 2100 channels of readout for the entire tracker.




Fig. 2. The double layer configuration of fibers which are bundled into triangular triplets for readout through a single SiPM. Spacing between bundles is ~0.94 mm.




Fig. 3. (a) An (X,Y) fiber tracker station assembled on 2 mm rohocell substrate. (b) The fiber ends are attached to SiPMs through an interface board shown on right side.

The calorimeter chosen for this design is a proton range detector which consists of a stack of 96, 3.2 mm thick, polyvinyltoluene (PVT) scintillating tiles, with 0.006 mm aluminized mylar between adjacent tiles. Each tile, 27 x 36 cm² in area, is machine grooved to embed a 1.2 mm diameter wavelength shifting (WLS) fiber that weaves 4 times across the tile for improved light collection efficiency. Both ends of the WLS fiber are read out through SiPMs. This requires 192 channels of readout for the calorimeter. Each SiPM signal is amplified and digitized for later analysis for fitting to the shape of a Bragg peak to determine the proton range in the calorimeter. Water equivalent blocks can be used to calibrate range measured in calorimeter [6].

An intrinsic limitation in any proton calorimeter is the combined range (or energy) straggling due to the mass represented by the patient plus calorimeter. In near water equivalent materials such as brain tissue and PVT scintillator, the sum of energy straggling in the human head and calorimeter is almost constant and

approximately equal to ± 3.6 mm [7]. Therefore, there is little incentive to produce tiles less than 3 mm thickness.

The 96 tile calorimeter has been built and undergone first tests with 200 MeV proton beam at Central DuPage Hospital in Warrenville, IL. Examples of pedestal distribution and a single photo-electron distribution from a calorimeter tile are shown in Fig. 4a. Figure 4b shows the Bragg peak from a sample of 200 MeV protons. To measure signal to noise ratio of the fiber bundles a fiber tracker plane prototype was also exposed to a 200 MeV proton beam. The results are shown in Fig. 4c, with 15 to 20 photo-electrons per proton per channel in the beam spot area.




Fig. 4. (a) Pulse height spectrum in ADCs showing the pedestal (first peak) and the single photo-electron noise signal (second) peak for a single calorimeter tile. (b) The ADC distribution as a function of tile number for 200 MeV protons. The error bars represent ± 1 rms about the average. (c) The average photo-electron counts per fiber tracker bundle for 200 MeV protons measured with a fiber tracker plane prototype.

4. Electronics

The electronics that read out the SiPMs consists of a custom board with preamplifiers, digitizers, and ethernet readout (PAD-E in Fig. 5). This custom board uses COTS components to provide readout for up to 32 channels of SiPM in a 220 mm x 100 mm format that fits into a standard 3U sub-rack. The same board is used for readout of the trackers and the calorimeter. The digitization of the signals from SiPMs, after appropriate amplification and shaping, is 12 bits per channel at 75 MSPS. The PAD-E is completely self contained and generates the bias for the SiPMs (one bulk voltage but with a 3 V adjustment range for each SiPM). It also contains an FPGA for processing all of the data generated by the SiPMs, memory for buffering up to 128 MB of data and a gigabit ethernet interface for pushing data directly to the data acquisition (DAQ) system. Other support circuitry includes temperature sensors for the SiPMs, clock management and a high speed USB port for debugging. Parameters such as the board's ethernet address or the correct bias voltage for the SiPMs are stored in a small flash memory on the board. The PAD-E is powered by a single 5 V power supply and has a power consumption of up to 15 W for 32 SiPM electronics channels. Each board locks to the clock provided by one board in each of the 9 sub-racks. Each of these boards in turn locks to a "master" board which has a free running crystal clock. Run control is accomplished by communicating with this master over ethernet.

The scanner is "self triggered" in the sense that any channel with a signal above threshold will be time stamped and stored in a local buffer for readout. A synchronous signal allows all boards to provide a timestamp that is used by the DAQ system to associate the data from different parts of the detector for a single proton history. Data from signals in the detector is highly compressed (only fiber address and timestamp from the trackers, compressed amplitude and time stamp from the calorimeter) and sent to the DAQ as soon as it is available. A synchronization signal which circulates across all boards approximately once per millisecond initiates a packet or "frame" of data readout from PAD-E memory to DAQ memory via 1 Gbit/s ethernet with only slight dead time penalty. A "footer" with error messages can be sent with each packet as well. Organizing the data into these one millisecond "time frames" allows for a relatively small timestamp (16 bits of 75 MHz clock cycles) and allows the DAQ to monitor the integrity of the data.




Fig. 5. A diagram of the front end electronics (PAD-E) for amplifying, digitizing and storing data before shipping to the DAQ at one millisecond intervals.

5. Data Acquisition System

The structure of the DAQ system [8] for the pCT scanner is shown on the left side of Fig. 6. The front end electronics will send data to the DAQ via 1 Gbit/s ethernet lines using UDP protocol. Each proton event contains data for 8 tracker planes and the 96 tile scintillator stack. We calculate that each event will generate about 25 bytes from the 8 hits on the 8 planes and about 75 bytes from the 96 tiles. For a 10 minute scan with 90 projection angles at a data rate of 2 million protons per second, we expect 200 MB/s written to RAM by 24 data collectors running on six interconnected Linux workstations. At the end of the scan, the back end DAQ will write data to disk and subsequently, through post processing of the data, obtain proton histories in the format for image reconstruction, i.e., 4 X and 4 Y coordinates, WEPL, and beam (or phantom) rotation angle.




Fig. 6. DAQ computer cluster (left) and GPU cluster computer (right) for image reconstruction. Data collectors hold hit locations and ADC amplitudes of the scintillator tiles with a time stamp for coalescing into tracks during post processing. Events are then sent to the image reconstruction computer cluster shown on the right side of the figure.

6. SUMMARY

The NIU Phase II proton CT scanner is fully assembled and installed for tests in a 200 MeV proton beam in Warrenville, IL, USA. Figure. 7 shows the scanner mounted on a cart in a treatment room. After system commissioning, a CIRS head phantom [9] will be inserted between tracker planes to collect data for image reconstruction on a CPU/GPU compute cluster [10]. This compute cluster has been tested with data acquired with an earlier prototype scanner [11, 12] and has demonstrated high quality 3D image reconstruction of a 14 cm spherical Lucy phantom from Standard Imaging, Inc. [13]. The first 3D head scan images are expected to be obtained in fall of 2014. The detailed project documentation can be found at [14].




Fig. 7. Fully assembled proton CT scanner at CDH Proton center. From right to left, beam enters upstream tracker planes followed by downstream tracker planes and finally the calorimeter. The gap in the middle is where the rotation stage for rotating the head phantom in the horizontal plane is placed.

7. Acknowledgements

We wish to thank Central DuPage Hospital and Loma Linda University Medical Center for the generous use of their facilities. This project was supported by funds from the Department of Defense.

References

- [1] U. Schneider, "Proton Radiography as a tool for quality control in proton Therapy", *Med. Phys.* 22, p.353 (1994).
- [2] H. F.-W. Sadrozinski *et al.*, "Development of a head scanner for proton CT", *Nucl. Instrum. Meth. A* 699, 205-210 (2013).
- [3] B. Erdelyi, "A comprehensive study of the most likely path formalism for proton computed tomography", *Physics in Medicine and Biology* 54, p.6095 (2009).
- [4] KurarayCo., Ltd, Japan, <http://kuraraypsf.jp/psf/index.html>
- [5] CPTA Ltd., www.cpta-apd.ru, Moscow, Russia.
- [6] R. F. Hurley *et al.*, "Water equivalent path length calibration of a prototype CT scanner", *Med. Phys.* 39 (5), p. 2438 (2012).
- [7] J. Janni, "Proton range-energy tables, 1 keV- 10 GeV", *Atomic Data and Nuclear Data Tables* 27(2-5), p.147-429 (1982).
- [8] S. Uzunyan *et al.*, "Development of a proton Computed Tomography (pCT) scanner at NIU", *Proceedings of the New Trends in High-Energy Physics*, p. 152-157, Alushta, Crimea, Ukraine, Sep. 2013, arXiv:1312.3977 (2013), ISBN 978-966-02-7015-2.
- [9] Computerized Imaging Reference Systems, Inc., www.cirsinc.com, Norfolk, VA.
- [10] K. Duffin *et al.*, "An analysis of a distributed GPU implementation of proton computed tomographic reconstruction", *Proceedings of the 2012 Super Computing Companion: high performance computing, networking storage and analysis*, pg.166-175, ISBN no. 978-0-7695-4956-9, Nov. 2012, Seattle, WA.
- [11] G. Coutrakon et al., "Design and Construction of the 1st Proton CT Scanner", *AIP Conference Proceedings*, No. 1525, Application of Accelerators in Research and Industry, Ft. Worth, TX, Aug. 2012, p. 327-331.
- [12] H. F.-W. Sadrozinski *et al.*, *IEEE NSS-MIC Conference*, 4457-4461, (2011).
- [13] www.standardimaging.com, see Lucy 3D QA phantom, Middleton, WI 53562.
- [14] <http://www.niu.edu/nicadd/research/medical>.