

# Site-resolved imaging of ytterbium atoms in a two-dimensional optical lattice

Martin Miranda,\* Ryotaro Inoue, Yuki Okuyama, and Mikiyo Kozuma  
*Department of Physics, Tokyo Institute of Technology,  
 2-12-1 O-okayama, Meguro-ku, Tokyo 152-8550, Japan*  
 (Dated: November 27, 2018)

We report a high-resolution microscope system for imaging ytterbium atoms trapped in a two-dimensional optical lattice. By using the ultra-violet strong transition combined with a solid immersion lens and high resolution optics, we achieved a 377(13) nm optical resolution, and resolved individual sites in an optical lattice with 543 nm spacing. The deep potential required for the imaging process was prepared using a combination of a shallow ground state and a deep excited state potentials. This system is suitable for alkaline-earth-metal-like atoms, where other sub-Doppler cooling techniques are not available.

PACS numbers: 37.10.De, 37.10.Gh, 37.10.Jk, 67.85.Hj, 67.85.Jk

Quantum gas microscopes are high resolution fluorescence imaging devices capable of resolving single-site atoms trapped in a two-dimensional optical lattice [1, 2]. The highly controllable optical lattice potential shape combined with a high resolution microscope system provides a fascinating tool for simulating quantum systems, such as observing superfluidity to Mott insulator phase transitions [3], quantum magnetism in the Bose-Hubbard regime [4] and quantum dynamics of super-exchanging interactions [5].

So far, quantum gas microscope experiments have been restricted to rubidium atoms [6]. Here, we report a system to address ytterbium (Yb) atoms trapped in a two-dimensional optical lattice with single-site resolution, which extends the quantum gas microscope technology to new atomic species. The Yb atom, having no electronic spin in the ground state, is robust against decoherence due to magnetic fluctuations, and therefore, is a promising atomic species for quantum information processing [7, 8]. Yb also has a variety of stable isotopes: five bosonic and two fermionic isotopes. One of those isotopes ( $^{171}\text{Yb}$ ) has a 1/2 nuclear spin and is ideal for implementing a quantum bit [9–11]. Polarization dependent fluorescence detection combined with the quantum microscope would also enables us to directly read the spin state of the system [12]. Additionally, Yb has two long-lived metastable state transitions, one of which is useful to tune inter-atomic interactions [13]. The wide variety of isotopes also makes possible the study of rich Mott insulators systems [14].

To observe atoms trapped in an optical lattice, the conventional method is to excite the atoms and observe the resulting fluorescence using a high numerical aperture (NA) microscope. As atoms rapidly heat during this process, a deep lattice potential and effective laser cooling mechanism are required. In the case of rubidium atoms, a polarization gradient cooling technique [1, 2, 15] was used to cool down the atoms while observing the resultant fluorescence. However, the bosonic isotopes of Yb are completely spin-less, giving a very simple energy struc-

ture in which sub-Doppler cooling mechanisms such as polarization gradient cooling and Raman cooling [16–18] are not applicable. Under these circumstances, Doppler cooling is the only possible mechanism available to cool down the atoms. A promising method to cool down Yb atoms trapped in an optical lattice with a dual molasses technique was recently realized by Kyoto group [19].

In this paper we report a method to load bosonic  $^{174}\text{Yb}$  atoms into a two-dimensional optical lattice near the surface of a solid immersion lens (SIL), and image the atoms with single-site resolution using a high NA microscope system. The SIL is a hemispherical lens that increases the resolution of the microscope by a factor of 1.5 (the index of refraction of the fused-silica glass) [20]. Additionally, it allows us to create a very stable optical system by fixing the position of the lattice respect to the SIL flat surface [21].




FIG. 1. (color online). Cooling mechanism for atoms trapped in an optical lattice. The lattice light at 1080 nm is near resonant with the upper  $^1P_1$ - $^1S_0$  transition (natural linewidth 3 MHz), creating a large light shift in the  $^1P_1$  state. The cooling beam is red detuned to the shifted  $^1S_0$ - $^1P_1$  strong dipolar transition (wavelength 399 nm, natural linewidth 29 MHz).

Our lattice cooling mechanism consists in selecting an optical lattice wavelength such that the potential in the excited state is deep and the one in the ground state is

shallow. A cooling light, which is tuned to the strong dipole-allowed transition, couples these states to provide a deep effective potential, as shown in Fig. 1. For our experiment we choose a wavelength of  $\lambda_{\text{acc}} = 1080$  nm. This wavelength is red-detuned by only 3 nm to the upper  $^1P_1$ - $^1S_0$  transition (wavelength 1077 nm, natural linewidth 3 MHz), which creates the desired potential shape. More specifically, for this wavelength, the light shift in the excited  $^1P_1$  state is roughly 400 times larger than that in the ground state. The cooling light is then tuned to the strong transition  $^1S_0$ - $^1P_1$  (wavelength 399 nm, natural linewidth 29 MHz).

To see the advantages of this system, we consider first the case when the cooling light is not present in the system. As the lattice light is far off-resonant to any of the transitions from the ground state, a shallow potential in the ground state results in small scattering rates ( $\tau \approx 180$  s for a 10  $\mu\text{K}$  of potential depth), which is an ideal condition for creating a BEC and a Mott Insulator.

During the imaging process, the near-resonant cooling light illuminates the atoms and couples the lower  $^1S_0$  and  $^1P_1$  states. Atoms then experience an average potential between the deep excited state potential and the shallow ground one [22], which is deep enough to trap atoms during the imaging process [23].

Another important feature of this potential shape is that, cooling light will always remain red-detuned for any position of the lattice. This ensures that the Doppler cooling mechanism works effectively in the entire lattice.




FIG. 2. (color online). Lateral view of the SIL attached to the 3 mm thick glass cell. Atoms are evaporated cooled in crossed-beam dipole trap. A “conveyor belt” generated by reflecting back the vertical ODT beam is used to transport the atoms closer to the surface.

The experimental setup consists of a glass cell containing the SIL connected to a metal vacuum chamber at a pressure of  $2 \times 10^{-9}$  Pa. The glass cell is made from a 3 mm thick optical quality fused-silica glass, and the upper surface has a through hole with a diameter of 12 mm, where the SIL is fitted (see Fig. 2). The SIL is a hemispherical lens made of fused silica glass, and it is divided into three parts. The upper part is a spherical cap

with a radius of 8 mm and a height of 1.32 mm, corresponding to a NA of 0.55. The middle part consists of a 3 mm thick disc with a diameter of 24 mm, and the lower part is a 3.68 mm thick disc with a diameter of 11.5 mm. All three parts were joined together by optical contact bonding, and the lower flat and upper spherical surface were super polished to reduce stray light. The SIL was glued to the glass cell using a vacuum leak sealant (SPI Vacseal) compatible with ultra high vacuum.

Our experiment starts by loading ultracold  $^{174}\text{Yb}$  atoms into a magneto-optical trap and transferring the atoms into the glass cell using an optical tweezer, as explained in our previous experimental work [21]. The horizontal optical dipole trap (ODT) beam transports  $1.0 \times 10^7$  atoms at a temperature of 25  $\mu\text{K}$  to a distance of 2.9 mm to the surface of the SIL. A vertical ODT beam is then introduced, entering from the upper flat surface of the SIL, and exiting at the center of the lower flat facet. The vertical ODT beam intersects the horizontal beam to form a crossed-beam ODT. Evaporated cooling is then performed by ramping down the intensity of the horizontal ODT beam in 3 s, resulting in  $1.0 \times 10^6$  atoms at a temperature of 2  $\mu\text{K}$ . Vertical transport of the atoms to the surface of the SIL is realized by adiabatically converting the vertical ODT into an optical “conveyor belt” and moving the retro-reflection mirror (see Fig. 2). Using this method, we can transfer  $8.0 \times 10^5$  atoms at a temperature of 2  $\mu\text{K}$  to a distance of 20  $\mu\text{m}$  below the surface of the SIL. It is important to note that the incident angle of the vertical ODT beam is set to the Brewster one (internal angle 34.2 degrees), which is slightly bigger than the aperture of the spherical cap (33.4 degrees). This ensures that the beam will enter and exit from flat surfaces, thus avoiding any interference that may complicate the transport process. In our previous experimental work [21], the aperture of the SIL was larger than the critical incident angle, requiring a complicated system of three ODT beams to transport the atoms to the surface.

After transporting the atoms, the optical accordion technique is used to create a Bose-Einstein condensate (BEC) and compress the atoms to create a two-dimensional condensate. This technique consists in reflecting a laser beam in a shallow angle into the SIL, creating a standing wave with a periodicity dependent on the incident angle. The incident angle can be manipulated to compress the atomic cloud. Details of this technique are described in our previous report [21]. For this experiment, we changed the accordion light source from a pulsed light source at 780 nm to a continuous-wave (CW) source at 1080 nm. The reason for using a CW source here is that an optical lattice can be easily created by retro-reflecting the optical accordion beam. Also, as we need to create a two-dimensional optical lattice, we replaced a single optical accordion with two orthogonal optical accordions aligned to the same incident angle, and shifted by a frequency of 440 MHz. Each of




FIG. 3. (color online). Experimental apparatus to trap and image atoms in a two-dimensional optical lattice. The microscope system is formed by an objective lens with a NA of 0.55, that combined with a SIL improves the NA of the system to 0.81. A retro-reflected optical accordion is used to create the two dimensional optical lattice.

the beams has a beam waist of  $35 \mu\text{m}$ .

To load the atoms into the accordion potential, the intensity of the optical accordion is gradually increased while the vertical ODT is transformed back to a propagating wave by reducing the intensity of the retro-reflected beam. As a result, atoms are transferred to a crossed trap between the two optical accordions and the vertical ODT. A condensate is then generated by evaporated cooling, first reducing the power of the optical accordions, and later removing the vertical ODT completely, resembling an evaporation cooling technique using a dimple trap [24]. The number of atoms in the condensate was  $8.0 \times 10^4$  after 800 ms of evaporation. The accordion incident angle is then gradually changed from  $0.7^\circ$  to  $6^\circ$  in 500 ms to compress the condensate into a pancake shape. Due to three-body losses during the compress process, only  $2 \times 10^4$  atoms remain in the trap.

Finally, atoms are loaded into the two-dimensional optical lattice by retro-reflecting each of the optical accordions. A system comprised by two polarized beam splitters and a half-wave plate mounted into a hollow stepping motor, is used to gradually increase the reflected beam during 100 ms (see Fig. 3) and avoid heating of the atoms.

Neglecting the Gaussian envelope of the laser beam




FIG. 4. (color online). a) False colored typical fluorescence image of atoms trapped in a two-dimensional optical lattice. One CCD camera count represents roughly  $1/8$  incident photons. b) Intensity profile along the center for three consecutive occupied lattice sites (inset image). Fitting using three point spreads function resulted in a resolution of  $377(13)$  nm. A single pixel represents  $90$  nm in real space.

and losses in the reflections, the final lattice has the following shape:

$$V_{\text{lat}} = V_0 (\cos^2(k_x x) + \cos^2(k_y y)) \times \sin^2(k_z z) \quad (1)$$

where  $V_0$  is the maximum potential depth, and  $z = 0$  is the surface of the SIL (see Fig. 3 for the coordinate system). For an optical accordion with an incident angle of  $\theta = 6^\circ$ , the wave-numbers for orthogonal directions are  $k_x = k_y = \frac{2\pi}{\lambda_{\text{acc}}} \cos \theta$  and  $k_z = \frac{2\pi}{\lambda_{\text{acc}}} \sin \theta$ , corresponding to a lattice spacing of  $543$  nm and  $5.1 \mu\text{m}$  in the horizontal (x,y) and vertical (z) directions, respectively. Initially, all the atoms in the pancaked shaped cloud are located in the first layer at  $z = 2.6 \mu\text{m}$ . The initial lattice depth is approximately  $3 \mu\text{K}$  or  $64 E_r$ , with trap frequencies of  $(\omega_x, \omega_y, \omega_z) = 2\pi(11, 11, 1.2)$  kHz.

The imaging process starts by increasing the lattice depth to  $220 \mu\text{K}$  or  $4700 E_r$  in 10 ms. For this lattice depth, the light shift in the  $^1P_1$  state was calculated to be  $60$  mK. The cooling light is aligned into the same optical axis as the optical accordion (see Fig. 3). This laser

beam was derived from a frequency doubled Ti:S laser, and it has a power of 30 mW and a beam waist of 100  $\mu\text{m}$ . The lattice fluorescence was found to be optimal for a detuning of  $\Delta/2\pi = -1.4$  GHz, which is consistent with the expected light shift on the excited state.

After 100  $\mu\text{s}$  of fluorescence imaging time, a single shot image of the Yb atoms trapped in a two-dimensional optical lattice was successfully observed (see Fig. 4a). The fluorescence profile of three isolated consecutive lattices sites was used to calculate the resolution of our microscope system. The profile is fitted using a sum of three pointing spread functions (see Fig. 4b). The resultant resolution (distance to the first dark circular ring surrounding the central Airy disc) was 377(13) nm. A similar result was obtained from a single site fitting. For a microscope system with a NA of 0.81, and an imaging wavelength of  $\lambda = 399$  nm, the expected theoretical resolution is  $0.61\lambda/\text{NA} = 300$  nm. The difference between the experimentally obtained resolution and the theoretical one might be related to the fact that the atoms are not completely at rest during the imaging process.




FIG. 5. Total photons per atom (in CCD counts) as a function of the cooling light exposure time. The profile is fitted using an integrated exponential decay, resulting in a life-time of 17(1)  $\mu\text{s}$ .

For a quantum gas microscope system, it is important to ensure that during the imaging process atoms remain in the site without hopping, and also be able to determine the parity of the number of atoms residing in each site [1–5]. To satisfy both of these conditions, the most common method is to calculate the total fluorescence in each site and plot the result into a histogram. If the result shows two equidistant peaks having a Poisson distribution, it is possible to affirm that the atoms remain still during the imaging process. A condition for observing these peaks is that the imaging time should be smaller compared to the life-time of the atoms.

In the experiment, the life time of the atoms was estimated from the imaging-time dependence of the average number of photons emitted from each atom, as shown in Fig. 5. Assuming that the number of atoms decrease

exponentially, the result was fitted using an integrated exponential curve, resulting in a lifetime of  $\tau = 17(1)$   $\mu\text{s}$ . Even when the lifetime was one order of magnitude shorter than what we expected from computer simulations results (and consequently, the total number of emitted photons was also one magnitude smaller), we were able to obtain well resolved images. By incrementing the life-time to  $\tau \sim 50$   $\mu\text{s}$ , we expect that this system would be able to detect Mott insulator shell structures in very short imaging times ( $\sim 20$   $\mu\text{s}$ ). We are currently focusing our efforts to detect the causes of this problem.

In conclusion, we have demonstrated a high resolution microscope system capable of resolving Yb atoms trapped in a two-dimensional optical lattice with single-site resolution. The experimental setup used to load and image atoms in a two-dimensional optical lattice is completely all-optical, and can be in principle adapted to other Yb isotopes or atomic species.

This work is supported by a grant-in-Aid for Scientific Research on Innovative Areas “Fluctuation & Structure” from the MEXT, the Cabinet Office, Government of Japan, through its “Funding Program for Next Generation World-Leading Researchers”, the Matsuo Foundation and the Murata Science Foundation. One of the authors (M.M.) was supported in part by the Japan Society for the Promotion of Science.

---

\* miranda.m.aa@m.titech.ac.jp

- [1] J. F. Sherson, C. Weitenberg, M. Endres, M. Cheneau, I. Bloch, and S. Kuhr, *Nature* **467**, 68 (2010).
- [2] W. S. Bakr, J. I. Gillen, A. Peng, S. Fölling, and M. Greiner, *Nature* **462**, 74 (2009).
- [3] W. S. Bakr, A. Peng, M. E. Tai, R. Ma, J. Simon, J. I. Gillen, S. Fölling, L. Pollet, and M. Greiner, *Science* **329**, 547 (2010).
- [4] J. Simon, W. S. Bakr, R. Ma, M. E. Tai, P. M. Preiss, and M. Greiner, *Nature* **472**, 307 (2011).
- [5] T. Fukuhara, A. Kantian, M. Endres, M. Cheneau, P. Schauß, S. Hild, D. Bellem, U. Schollwöck, T. Giamarchi, C. Gross, *et al.*, *Nature Physics* **9**, 235 (2013).
- [6] Recently, we had knowledge that the Harvard group succeeded in developing a quantum microscope for lithium atoms.
- [7] A. J. Daley, M. M. Boyd, J. Ye, and P. Zoller, *Phys. Rev. Lett.* **101**, 170504 (2008).
- [8] A. V. Gorshkov, A. M. Rey, A. J. Daley, M. M. Boyd, J. Ye, P. Zoller, and M. D. Lukin, *Phys. Rev. Lett.* **102**, 110503 (2009).
- [9] K. Shibata, S. Kato, A. Yamaguchi, S. Uetake, and Y. Takahashi, *Applied Physics B: Lasers and Optics* **97**, 753 (2009).
- [10] A. Noguchi, Y. Eto, M. Ueda, and M. Kozuma, *Phys. Rev. A* **84**, 030301 (2011).
- [11] M. Takeuchi, N. Takei, K. Doi, P. Zhang, M. Ueda, and M. Kozuma, *Phys. Rev. A* **81**, 062308 (2010).
- [12] Y. Eto, A. Noguchi, P. Zhang, M. Ueda, and M. Kozuma, *Phys. Rev. Lett.* **106**, 160501 (2011).

- [13] S. Kato, S. Sugawa, K. Shibata, R. Yamamoto, and Y. Takahashi, *Phys. Rev. Lett.* **110**, 173201 (2013).
- [14] S. Sugawa, K. Inaba, S. Taie, R. Yamazaki, M. Yamashita, and Y. Takahashi, *Nature Physics* **7**, 642 (2011).
- [15] S. L. Winoto, M. T. DePue, N. E. Bramall, and D. S. Weiss, *Phys. Rev. A* **59**, R19 (1999).
- [16] S. E. Hamann, D. L. Haycock, G. Klose, P. H. Pax, I. H. Deutsch, and P. S. Jessen, *Phys. Rev. Lett.* **80**, 4149 (1998).
- [17] V. Vuletić, C. Chin, A. J. Kerman, and S. Chu, *Phys. Rev. Lett.* **81**, 5768 (1998).
- [18] R. Zhang, N. V. Morrow, P. R. Berman, and G. Raithel, *Phys. Rev. A* **72**, 043409 (2005).
- [19] K. Shibata, R. Yamamoto, and Y. Takahashi, *Journal of the Physical Society of Japan* **83**, 014301 (2014), <http://dx.doi.org/10.7566/JPSJ.83.014301>.
- [20] S. M. Mansfield and G. S. Kino, *Applied Physics Letters* **57**, 2615 (1990).
- [21] M. Miranda, A. Nakamoto, Y. Okuyama, A. Noguchi, M. Ueda, and M. Kozuma, *Phys. Rev. A* **86**, 063615 (2012).
- [22] J. Dalibard and C. Cohen-Tannoudji, *JOSA B* **2**, 1707 (1985).
- [23] Note that, for this approximation to be true, the Rabi frequency should be much larger than the trap frequency in the optical lattice. The strong transition with saturated intensity satisfies this condition.
- [24] D. Jacob, E. Mimoun, L. De Sarlo, M. Weitz, J. Dalibard, and F. Gerbier, *New Journal of Physics* **13**, 065022 (2011).