

---

# Super-resolution structured illumination microscopy: past, present and future

Kirti Prakash<sup>1,2,\*</sup>, Benedict Diederich<sup>3,4</sup>, Stefanie Reichelt<sup>5</sup>, Rainer Heintzmann<sup>3,4,6,\*</sup>, Lothar Schermelleh<sup>7,\*</sup>

**1 National Physical Laboratory, TW11 0LW Teddington, UK**

**2 Department of Paediatrics, Wellcome–MRC Cambridge Stem Cell Institute, University of Cambridge, Cambridge, UK**

**3 Leibniz Institute of Photonic Technology, Albert-Einstein-Straße 9, 07745 Jena, Germany**

**4 Institute of Physical Chemistry and Abbe Center of Photonics, Helmholtzweg 4, Friedrich-Schiller-University, Jena, Germany**

**5 CRUK Cambridge Research Institute, Robinson Way, Cambridge CB2 0RE, UK**

**6 Faculty of Physics and Astronomy, Friedrich-Schiller-University, Jena, Germany**

**7 Micron Advanced Bioimaging Unit, Department of Biochemistry, University of Oxford, Oxford OX1 3QU, UK**

**✉ kirtiprakash25, beniroquai, StefanieReiche6, HeintzmannLab, LSchermelleh**

**\* heintzmann@gmail.com, kirtiprakash2.71@gmail.com, lothar.schermelleh@bioch.ox.ac.uk**

## Abstract

Structured illumination microscopy (SIM) has emerged as an essential technique for 3D and live-cell super-resolution imaging. However, to date, there has not been a dedicated workshop or journal issue covering the various aspects of SIM, from bespoke hardware and software development and the use of commercial instruments to biological applications. This special issue aims to recap recent developments as well as outline future trends. In addition to SIM, we cover related topics such as complementary super-resolution microscopy techniques, computational imaging, visualisation and image processing methods.

## Introduction

Fluorescence light microscopy is a core technique in life sciences that has contributed to countless major discoveries. However, the diffraction limit, first described by Ernst Abbe in 1873 [1], has restricted the optical resolution to about half the wavelength of the light used, i.e. 200-300 nm in the lateral directions (x and y), and to about the wavelength of light, i.e. 500-800 nm along the optical axis (z). In the past two decades, several super-resolution microscopy (SRM) approaches have been developed that allow to overcome this barrier and push the spatial resolution to the 10-100 nm range, thus closing the gap to electron microscopy [25, 70, 71]. The SRM techniques can be further

---

divided into single-molecule localisation microscopy (SMLM) [43], that includes techniques like photoactivatable localisation microscopy (PALM) [7, 29] and stochastic optical reconstruction microscopy (STORM) [67], stimulated emission depletion (STED) [3, 27, 58] microscopy and structured illumination microscopy (SIM) [20, 23]. In recognition of these breakthrough developments Eric Betzig, Stefan Hell and William E. Moerner were awarded the Nobel prize for Chemistry in 2014 [6].

#### Frequently asked questions in super-resolution structured illumination microscopy field

The following questions have not always been agreed upon in the super-resolution/SIM field, and with this special issue, we hope to address some of these:

1. What is super-resolution microscopy and should diffraction-limited linear SIM be classified as ‘super-resolution’?
2. Should High-NA TIRF-SIM, which can achieve lateral resolution down to 84 nm, be considered as diffraction limited?
3. Can non-linear SIM become broadly applicable and live-cell compatible?
4. Do you need ‘switching’ of states for non-linear super-resolution?
5. How can information about single-molecule detection be best combined with the knowledge of the illumination structure?
6. Do high quality SIM images require reconstruction in Fourier space?
7. Can SIM be used for deep tissue imaging?
8. How can the fundamental limitation of SIM, i.e. generating sufficient stripe contrast in densely labelled and/or extended biological structures due to out-of-focus light, be addressed?
9. Should image scanning microscopy be considered a form of SIM and what forms of structured illumination could be used other than stripes?
10. Can SIM be used to improve the resolution of (Rayleigh scattering) transmission microscopy?
11. How does sparse illumination compare to dense illumination in linear and non-linear SIM?
12. Can we generate “true” super-resolution images from simple instruments enhanced with machine-learning-based algorithms?
13. Can research-grade super-resolution (SIM) microscopes be built cost-efficiently?

Recent SRM developments have focussed on combining localisation-based microscopy with modulated illumination to push the resolution (precision) towards a few nanometers as in MINFLUX and SIMFLUX [5, 13, 62, 64]. However, an increase in resolution requires an increase in local light dosage, which increases photobleaching/phototoxicity (SIM, STED, SMLM) or requires the need to introduce spatial and temporal sparsity (MINFLUX) making imaging comparably slow. Moreover, most SRM techniques mentioned above rely primarily on chemically fixed (i.e. dead) cells, are restricted in the imaged sample volume (2D, single plane, small field-of-views) or are restricted to imaging sparse single entities such as vesicles or single molecules.

---

Super-resolution linear SIM is a notable exception [24, 59]. By making use of frequency mixing when exciting samples with a patterned illumination followed by computational unmixing and reconstruction, SIM achieves a 2-fold resolution increase over conventional diffraction-limited fluorescence microscopy in 2D or 3D (Figure 1). While the numerical resolution improvement is moderate compared to other SRM techniques, it pushes SIM into an application sweet spot with most macromolecular structures and their dynamics falling in the size range of 100-300 nm [69, 73, 77]. By not asking for the highest spatial resolution, SIM is less demanding in terms of photon budget and therefore more compatible with live-cell imaging [30, 31, 39, 53, 72, 81]. In addition, it offers volumetric imaging with relatively large field-of-views and comparably high temporal resolution which makes it suitable for high-content and live-cell imaging [50, 83, 84]. Taken together, the uniquely balanced combination of properties have made SIM remarkably successful in a wide range of biological applications and promoting new discoveries [10, 16, 17, 19, 32, 34, 41, 47–49, 51, 55–57, 76, 77, 79].

## Recent and future developments in SIM

SIM has been available in commercial instruments for more than 10 years and has been successfully established in many labs and core facilities around the world. However, wider dissemination of SIM has been curtailed by the complexity of instruments and its proneness to reconstruction artefacts when sample properties, system calibration and parameter settings are not carefully matched [15]. Thus, to lower the activation energy for research labs to venture into super-resolution SIM, many recent developments aim at further performance/resolution enhancement and ‘democratising’ SIM by

- using non-linear SIM approaches [21, 26, 42, 63] or by combinations with single-molecule imaging [5, 13, 18, 37, 65]
- exploiting correlative and combinatorial fluorescence imaging approaches with multifocal [2, 22], 2-photon [33, 45, 75] and light-sheet microscopy [11, 38]
- implementing the technique into smaller and more cost-efficient setups [28, 40, 80]
- making the method more robust against artefacts using alternative illumination schemes [4, 52, 68] and/or intelligent data processing [8, 82]
- increasing its application range e.g. by the implementation of adaptive optics [9, 14, 35], cryo-imaging [60, 79]

Important advancements are being made to improve and simplify the instrumentation. For instance, the newly designed optical setups using an open-sources toolbox [80] or waveguide-based photonic chips [82] or open-source software development [54] may reduce the cost of SR imaging enabling SIM on smaller and more affordable devices. Moreover, next-generation sCMOS cameras (for example, the new 500 fps sCMOS by Photometrics) offer increased frame rates and higher sensitivity with reduced noise levels to enable faster high-quality SIM imaging for observing cellular dynamics [12, 46, 74]. For a maximal imaging speed, the rolling shutter schemes can be combined with SIM data acquisition [78].

In parallel, the rapidly evolving field of computational microscopy is paving the way to circumvent the resolution limit without the necessity of specialised hardware. Such computational techniques not only reduce the complexity of the instrumentation but can also potentially reduce aberration artefacts [8, 40, 61, 66, 80, 82]. Adaptive optics (AO) works well to correct sample aberrations. However, it does not solve the problem of reduced illumination pattern contrast in the presence of significant out of focus

---

background. One solution here is offered by combining SIM with light sheet illumination, 2-photon, or photoswitching methods.

Recent advancements in intelligent microscopy and machine learning (ML) assisted denoising are now being used for improving SIM reconstructions. For example, using deep-learning models, the apparent axial resolution of SIM has been improved by a factor of two [8, 36, 44]. By combining physical data acquisition with simple image processing algorithms, one can recover previously inaccessible information. At the same time, this brings up new problems of validation of the results from ML methods. New standards have to be developed, where computational scientists and mathematicians can help to bring together state-of-the-art algorithms, such as GPU-parallelization for faster data processing or machine-learning-based data interpretation.

## The wider-reaching social implications

SRM has evolved into a highly interdisciplinary field requiring experts from physics, engineering, chemistry, biology and computer sciences to drive innovations for increasing spatial and temporal resolution and ultimately biological applicability. The global advances in super-resolution microscopy need to be synchronized to inspire researchers in the field of microscopy to combine their methods with techniques from different areas.

SRM and SIM, in particular, have become important tools of basic science discovery. However, there are still considerable activation barriers for the researcher to use SRM, as instruments are not yet ‘turn-key’, and using SIM (or any other SRM method) to its full potential requires considerable knowledge. Furthermore, the use of SRM to study human pathologies and for medical diagnosis is still in its infancy, but will foreseeably play a considerable role in the future. New SIM modalities, implementation of AO, machine-learning, and the intelligent algorithms will help to democratise SIM and increase its usage in basic research and biomedical application.

Super-resolution microscopy is very often an expensive undertaking, thus limiting its broader application. In this meeting, we plan to kick-off the idea of an open-source ‘openSIM’ system inspired by its derivatives in the light-sheet community. Open discussions where ideas and requirements for a prototype are gathered are a first and significant step to make science not only affordable but available. This special issue shows current progress in SIM development and application.

## Acknowledgements

The special issue is part of Theo Murphy international scientific meeting organised by the Royal Society called SIMposium: recent advancements in structured illumination microscopy. We thank Mike Shaw for comments on the manuscript. A Twitter (🐦) discussion on this topic which can be followed here or with hashtag SIMposium.

## References

1. E. Abbe. Beitrage zur Theorie des Mikroskops und der mikroskopischen Wahrnehmung. *Archiv für Mikroskopische Anatomie*, 9:413–420, 1873.
2. S. Abrahamsson, H. Blom, A. Agostinho, D. C. Jans, A. Jost, M. Müller, L. Nilsson, K. Bernhem, T. J. Lambert, R. Heintzmann, et al. Multifocus structured illumination microscopy for fast volumetric super-resolution imaging. *Biomedical optics express*, 8(9):4135–4140, 2017.

- 
3. S. C. Baer. Method and apparatus for improving resolution in scanned optical system, Feb. 2 1999. US Patent 5,866,911.
  4. G. Ball, J. Demmerle, R. Kaufmann, I. Davis, I. M. Dobbie, and L. Schermelleh. Simcheck: a toolbox for successful super-resolution structured illumination microscopy. *Scientific reports*, 5(1):1–12, 2015.
  5. F. Balzarotti, Y. Eilers, K. C. Gwosch, A. H. Gynnå, V. Westphal, F. D. Stefani, J. Elf, and S. W. Hell. Nanometer resolution imaging and tracking of fluorescent molecules with minimal photon fluxes. *Science*, 355(6325):606–612, 2017.
  6. E. Betzig, S. W. Hell, and W. E. Moerner. The nobel prize in chemistry 2014. *Nobel Media AB*, 2014.
  7. E. Betzig, G. H. Patterson, R. Sougrat, O. W. Lindwasser, S. Olenych, J. S. Bonifacino, M. W. Davidson, J. Lippincott-Schwartz, and H. F. Hess. Imaging intracellular fluorescent proteins at nanometer resolution. *Science*, 313(5793):1642–1645, 2006.
  8. M. Boland, E. A. Cohen, S. Flaxman, and M. A. Neil. Improving axial resolution in sim using deep learning. *arXiv preprint arXiv:2009.02264*, 2020.
  9. M. Booth, D. Andrade, D. Burke, B. Patton, and M. Zurauskas. Aberrations and adaptive optics in super-resolution microscopy. *Microscopy*, 64(4):251–261, 2015.
  10. A. C. Brown, S. Oddos, I. M. Dobbie, J.-M. Alakoskela, R. M. Parton, P. Eissmann, M. A. Neil, C. Dunsby, P. M. French, I. Davis, et al. Remodelling of cortical actin where lytic granules dock at natural killer cell immune synapses revealed by super-resolution microscopy. *PLoS Biol*, 9(9):e1001152, 2011.
  11. B.-J. Chang, V. D. P. Meza, and E. H. Stelzer. csilsm combines light-sheet fluorescence microscopy and coherent structured illumination for a lateral resolution below 100 nm. *Proceedings of the National Academy of Sciences*, 114(19):4869–4874, 2017.
  12. B.-C. Chen, W. R. Legant, K. Wang, L. Shao, D. E. Milkie, M. W. Davidson, C. Janetopoulos, X. S. Wu, J. A. Hammer, Z. Liu, et al. Lattice light-sheet microscopy: imaging molecules to embryos at high spatiotemporal resolution. *Science*, 346(6208), 2014.
  13. J. Cnossen, T. Hinsdale, R. Ø. Thorsen, M. Siemons, F. Schueder, R. Jungmann, C. S. Smith, B. Rieger, and S. Stallinga. Localization microscopy at doubled precision with patterned illumination. *Nature methods*, 17(1):59–63, 2020.
  14. D. Débarre, E. J. Botcherby, M. J. Booth, and T. Wilson. Adaptive optics for structured illumination microscopy. *Optics express*, 16(13):9290–9305, 2008.
  15. J. Demmerle, C. Innocent, A. J. North, G. Ball, M. Müller, E. Miron, A. Matsuda, I. M. Dobbie, Y. Markaki, and L. Schermelleh. Strategic and practical guidelines for successful structured illumination microscopy. *Nature protocols*, 12(5):988–1010, 2017.
  16. P. Eswaramoorthy, P. W. Winter, P. Wawrzusin, A. G. York, H. Shroff, and K. S. Ramamurthi. Asymmetric division and differential gene expression during a bacterial developmental program requires diviva. *PLoS Genet*, 10(8):e1004526, 2014.

- 
17. R. D. Gray, D. Albrecht, C. Beerli, M. Huttunen, G. H. Cohen, I. J. White, J. J. Burden, R. Henriques, and J. Mercer. Nanoscale polarization of the entry fusion complex of vaccinia virus drives efficient fusion. *Nature microbiology*, 4(10):1636–1644, 2019.
  18. L. Gu, Y. Li, S. Zhang, Y. Xue, W. Li, D. Li, T. Xu, and W. Ji. Molecular resolution imaging by repetitive optical selective exposure. *Nature methods*, 16(11):1114–1118, 2019.
  19. Y. Guo, D. Li, S. Zhang, Y. Yang, J.-J. Liu, X. Wang, C. Liu, D. E. Milkie, R. P. Moore, U. S. Tulu, et al. Visualizing intracellular organelle and cytoskeletal interactions at nanoscale resolution on millisecond timescales. *Cell*, 175(5):1430–1442, 2018.
  20. M. G. Gustafsson. Surpassing the lateral resolution limit by a factor of two using structured illumination microscopy. *Journal of microscopy*, 198(2):82–87, 2000.
  21. M. G. Gustafsson. Nonlinear structured-illumination microscopy: wide-field fluorescence imaging with theoretically unlimited resolution. *Proceedings of the National Academy of Sciences*, 102(37):13081–13086, 2005.
  22. B. Hajj, J. Wisniewski, M. El Beheiry, J. Chen, A. Revyakin, C. Wu, and M. Dahan. Whole-cell, multicolor superresolution imaging using volumetric multifocus microscopy. *Proceedings of the National Academy of Sciences*, 111(49):17480–17485, 2014.
  23. R. Heintzmann and C. G. Cremer. Laterally modulated excitation microscopy: improvement of resolution by using a diffraction grating. In *Optical Biopsies and Microscopic Techniques III*, volume 3568, pages 185–196. International Society for Optics and Photonics, 1999.
  24. R. Heintzmann and M. G. Gustafsson. Subdiffraction resolution in continuous samples. *Nature Photonics*, 3(7):362–364, 2009.
  25. R. Heintzmann and T. Huser. Super-resolution structured illumination microscopy. *Chemical reviews*, 117(23):13890–13908, 2017.
  26. R. Heintzmann, T. M. Jovin, and C. Cremer. Saturated patterned excitation microscopy—a concept for optical resolution improvement. *JOSA A*, 19(8):1599–1609, 2002.
  27. S. W. Hell and J. Wichmann. Breaking the diffraction resolution limit by stimulated emission: stimulated-emission-depletion fluorescence microscopy. *Optics letters*, 19(11):780–782, 1994.
  28. Ø. I. Helle, F. T. Dullo, M. Lahrberg, J.-C. Tinguely, O. G. Hellesø, and B. S. Ahluwalia. Structured illumination microscopy using a photonic chip. *Nature Photonics*, 14(7):431–438, 2020.
  29. S. T. Hess, T. P. Girirajan, and M. D. Mason. Ultra-high resolution imaging by fluorescence photoactivation localization microscopy. *Biophysical journal*, 91(11):4258, 2006.
  30. L. M. Hirvonen, K. Wicker, O. Mandula, and R. Heintzmann. Structured illumination microscopy of a living cell. *European Biophysics Journal*, 38(6):807–812, 2009.

- 
31. X. Huang, J. Fan, L. Li, H. Liu, R. Wu, Y. Wu, L. Wei, H. Mao, A. Lal, P. Xi, et al. Fast, long-term, super-resolution imaging with hessian structured illumination microscopy. *Nature biotechnology*, 36(5):451–459, 2018.
  32. B. Hübner, M. Lomiento, F. Mammoli, D. Illner, Y. Markaki, S. Ferrari, M. Cremer, and T. Cremer. Remodeling of nuclear landscapes during human myelopoietic cell differentiation maintains co-aligned active and inactive nuclear compartments. *Epigenetics & chromatin*, 8(1):1–21, 2015.
  33. M. Ingaramo, A. G. York, P. Wawrzusin, O. Milberg, A. Hong, R. Weigert, H. Shroff, and G. H. Patterson. Two-photon excitation improves multifocal structured illumination microscopy in thick scattering tissue. *Proceedings of the National Academy of Sciences*, 111(14):5254–5259, 2014.
  34. G. Jacquemet, A. Stubb, R. Saup, M. Miihkinen, E. Kremneva, H. Hamidi, and J. Ivaska. Filopodome mapping identifies p130cas as a mechanosensitive regulator of filopodia stability. *Current Biology*, 29(2):202–216, 2019.
  35. N. Ji. Adaptive optical fluorescence microscopy. *Nature methods*, 14(4):374–380, 2017.
  36. L. Jin, B. Liu, F. Zhao, S. Hahn, B. Dong, R. Song, T. C. Elston, Y. Xu, and K. M. Hahn. Deep learning enables structured illumination microscopy with low light levels and enhanced speed. *Nature communications*, 11(1):1–7, 2020.
  37. P. Jouchet, C. Cabriel, N. Bourg, M. Bardou, C. Pouis, E. Fort, and S. Lévêque-Fort. Nanometric axial localization of single fluorescent molecules with modulated excitation. *BioRxiv*, page 865865, 2019.
  38. P. J. Keller, A. D. Schmidt, A. Santella, K. Khairy, Z. Bao, J. Wittbrodt, and E. H. Stelzer. Fast, high-contrast imaging of animal development with scanned light sheet-based structured-illumination microscopy. *Nature methods*, 7(8):637–642, 2010.
  39. P. Kner, B. B. Chhun, E. R. Griffis, L. Winoto, and M. G. Gustafsson. Super-resolution video microscopy of live cells by structured illumination. *Nature methods*, 6(5):339–342, 2009.
  40. M. Lachetta, H. Sandmeyer, A. Sandmeyer, J. S. am Esch, T. Huser, and M. Müller. Simulating digital micromirror devices for patterning coherent excitation light in structured illumination microscopy. *bioRxiv*, 2020.
  41. C. Lesterlin, G. Ball, L. Schermelleh, and D. J. Sherratt. RecA bundles mediate homology pairing between distant sisters during dna break repair. *Nature*, 506(7487):249–253, 2014.
  42. D. Li, L. Shao, B.-C. Chen, X. Zhang, M. Zhang, B. Moses, D. E. Milkie, J. R. Beach, J. A. Hammer, M. Pasham, et al. Extended-resolution structured illumination imaging of endocytic and cytoskeletal dynamics. *Science*, 349(6251), 2015.
  43. K. A. Lidke, B. Rieger, T. M. Jovin, and R. Heintzmann. Superresolution by localization of quantum dots using blinking statistics. *Optics express*, 13(18):7052–7062, 2005.
  44. C. Ling, C. Zhang, M. Wang, F. Meng, L. Du, and X. Yuan. Fast structured illumination microscopy via deep learning. *Photonics Research*, 8(8):1350–1359, 2020.

- 
45. F. Liu, Q. Li, S. Jiang, L. Zhou, J. Zhang, and H. Zhang. Two-photon structured illumination microscopy imaging using fourier ptychography scheme. *Optics Communications*, page 126872, 2021.
  46. T.-L. Liu, S. Upadhyayula, D. E. Milkie, V. Singh, K. Wang, I. A. Swinburne, K. R. Mosaliganti, Z. M. Collins, T. W. Hiscock, J. Shea, et al. Observing the cell in its native state: Imaging subcellular dynamics in multicellular organisms. *Science*, 360(6386), 2018.
  47. U. Manor, S. Bartholomew, G. Golani, E. Christenson, M. Kozlov, H. Higgs, J. Spudich, and J. Lippincott-Schwartz. A mitochondria-anchored isoform of the actin-nucleating spire protein regulates mitochondrial division. *Elife*, 4:e08828, 2015.
  48. K. McArthur, L. W. Whitehead, J. M. Heddleston, L. Li, B. S. Padman, V. Oorschot, N. D. Geoghegan, S. Chappaz, S. Davidson, H. San Chin, et al. Bak/bax macropores facilitate mitochondrial herniation and mtDNA efflux during apoptosis. *Science*, 359(6378), 2018.
  49. V. Mennella, B. Keszthelyi, K. McDonald, B. Chhun, F. Kan, G. C. Rogers, B. Huang, and D. Agard. Subdiffraction-resolution fluorescence microscopy reveals a domain of the centrosome critical for pericentriolar material organization. *Nature cell biology*, 14(11):1159–1168, 2012.
  50. J. Mertz. Optical sectioning microscopy with planar or structured illumination. *Nature methods*, 8(10):811, 2011.
  51. E. Miron, R. Oldenkamp, J. M. Brown, D. M. Pinto, C. S. Xu, A. R. Faria, H. A. Shaban, J. D. Rhodes, C. Innocent, S. de Ornellas, et al. Chromatin arranges in chains of mesoscale domains with nanoscale functional topography independent of cohesin. *Science advances*, 6(39):eaba8811, 2020.
  52. Y. Mo, F. Feng, H. Mao, J. Fan, and L. Chen. Structured illumination microscopy artifacts caused by illumination scattering. *bioRxiv*, pages 2021–01, 2021.
  53. E. Mudry, K. Belkebir, J. Girard, J. Savatier, E. Le Moal, C. Nicoletti, M. Allain, and A. Sentenac. Structured illumination microscopy using unknown speckle patterns. *Nature Photonics*, 6(5):312–315, 2012.
  54. M. Müller, V. Mönkemöller, S. Hennig, W. Hübner, and T. Huser. Open-source image reconstruction of super-resolution structured illumination microscopy data in imagej. *Nature communications*, 7(1):1–6, 2016.
  55. S. Murugesan, J. Hong, J. Yi, D. Li, J. R. Beach, L. Shao, J. Meinhardt, G. Madison, X. Wu, E. Betzig, et al. Formin-generated actomyosin arcs propel t cell receptor microcluster movement at the immune synapse. *Journal of Cell Biology*, 215(3):383–399, 2016.
  56. M. Nozumi, F. Nakatsu, K. Katoh, and M. Igarashi. Coordinated movement of vesicles and actin bundles during nerve growth revealed by superresolution microscopy. *Cell reports*, 18(9):2203–2216, 2017.
  57. F. Ochs, G. Karemore, E. Miron, J. Brown, H. Sedlackova, M.-B. Rask, M. Lampe, V. Buckle, L. Schermelleh, J. Lukas, et al. Stabilization of chromatin topology safeguards genome integrity. *Nature*, 574(7779):571–574, 2019.
  58. V. Okhonin. A method of examination of sample microstructure. SU 1374922 A1. Priority 10.04.1986. DOI: 10.13140/2.1.2588.1922. 1991.



- 
59. K. O'Holleran and M. Shaw. Optimized approaches for optical sectioning and resolution enhancement in 2d structured illumination microscopy. *Biomedical optics express*, 5(8):2580–2590, 2014.
  60. M. A. Phillips, M. Harkiolaki, D. M. S. Pinto, R. M. Parton, A. Palanca, M. Garcia-Moreno, I. Kounatidis, J. W. Sedat, D. I. Stuart, A. Castello, et al. Cryosim: super-resolution 3d structured illumination cryogenic fluorescence microscopy for correlated ultrastructural imaging. *Optica*, 7(7):802–812, 2020.
  61. K. Prakash. Laser-free super-resolution microscopy. *bioRxiv*, 2020.
  62. K. Prakash. At the molecular resolution with minflux? 2021.
  63. E. H. Rego, L. Shao, J. J. Macklin, L. Winoto, G. A. Johansson, N. Kamps-Hughes, M. W. Davidson, and M. G. Gustafsson. Nonlinear structured-illumination microscopy with a photoswitchable protein reveals cellular structures at 50-nm resolution. *Proceedings of the National Academy of Sciences*, 109(3):E135–E143, 2012.
  64. L. Reymond, T. Huser, V. Ruprecht, and S. Wieser. Modulation-enhanced localization microscopy. *Journal of Physics: Photonics*, 2(4):041001, 2020.
  65. L. Reymond, J. Ziegler, C. Knapp, F.-C. Wang, T. Huser, V. Ruprecht, and S. Wieser. Simple: Structured illumination based point localization estimator with enhanced precision. *Optics express*, 27(17):24578–24590, 2019.
  66. C. T. Russell and M. Shaw. mmsim: An open toolbox for accessible structured illumination microscopy. *bioRxiv*, 2021.
  67. M. J. Rust, M. Bates, and X. Zhuang. Stochastic optical reconstruction microscopy (storm) provides sub-diffraction-limit image resolution. *Nature methods*, 3(10):793, 2006.
  68. L. Schaefer, D. Schuster, and J. Schaffer. Structured illumination microscopy: artefact analysis and reduction utilizing a parameter optimization approach. *Journal of microscopy*, 216(2):165–174, 2004.
  69. L. Schermelleh, P. M. Carlton, S. Haase, L. Shao, L. Winoto, P. Kner, B. Burke, M. C. Cardoso, D. A. Agard, M. G. Gustafsson, et al. Subdiffraction multicolor imaging of the nuclear periphery with 3d structured illumination microscopy. *Science*, 320(5881):1332–1336, 2008.
  70. L. Schermelleh, A. Ferrand, T. Huser, C. Eggeling, M. Sauer, O. Biehlmaier, and G. P. Drummen. Super-resolution microscopy demystified. *Nature cell biology*, 21(1):72–84, 2019.
  71. L. Schermelleh, R. Heintzmann, and H. Leonhardt. A guide to super-resolution fluorescence microscopy. *Journal of Cell Biology*, 190(2):165–175, 2010.
  72. L. Shao, P. Kner, E. H. Rego, and M. G. Gustafsson. Super-resolution 3d microscopy of live whole cells using structured illumination. *Nature methods*, 8(12):1044–1046, 2011.
  73. M. Shaw, A. Bella, and M. G. Ryadnov. Creim: Coffee ring effect imaging model for monitoring protein self-assembly in situ. *The journal of physical chemistry letters*, 8(19):4846–4851, 2017.

- 
74. M. Shaw, L. Zajiczek, and K. O'Holleran. High speed structured illumination microscopy in optically thick samples. *Methods*, 88:11–19, 2015.
  75. C. J. Sheppard, M. Castello, G. Tortarolo, G. Vicidomini, and A. Diaspro. Image formation in image scanning microscopy, including the case of two-photon excitation. *JOSA A*, 34(8):1339–1350, 2017.
  76. J.-H. Sir, A. R. Barr, A. K. Nicholas, O. P. Carvalho, M. Khurshid, A. Sossick, S. Reichelt, C. D'Santos, C. G. Woods, and F. Gergely. A primary microcephaly protein complex forms a ring around parental centrioles. *Nature genetics*, 43(11):1147–1153, 2011.
  77. D. Smeets, Y. Markaki, V. J. Schmid, F. Kraus, A. Tattermusch, A. Cerase, M. Sterr, S. Fiedler, J. Demmerle, J. Popken, et al. Three-dimensional super-resolution microscopy of the inactive x chromosome territory reveals a collapse of its active nuclear compartment harboring distinct xist rna foci. *Epigenetics & chromatin*, 7(1):1–27, 2014.
  78. L. Song, H.-W. Lu-Walther, R. Förster, A. Jost, M. Kielhorn, J. Zhou, and R. Heintzmann. Fast structured illumination microscopy using rolling shutter cameras. *Measurement Science and Technology*, 27(5):055401, 2016.
  79. K. F. Sonnen, L. Schermelleh, H. Leonhardt, and E. A. Nigg. 3d-structured illumination microscopy provides novel insight into architecture of human centrosomes. *Biology open*, 1(10):965–976, 2012.
  80. H. Wang, R. Lachmann, B. Marsikova, R. Heintzmann, and B. Diederich. Ucsim2: 2d structured illumination microscopy using uc2. *bioRxiv*, pages 2021–01, 2021.
  81. Y. Wu and H. Shroff. Faster, sharper, and deeper: structured illumination microscopy for biological imaging. *Nature methods*, 15(12):1011–1019, 2018.
  82. L.-H. Yeh, S. Chowdhury, and L. Waller. Computational structured illumination for high-content fluorescence and phase microscopy. *Biomedical optics express*, 10(4):1978–1998, 2019.
  83. A. G. York, P. Chandris, D. Dalle Nogare, J. Head, P. Wawrzusin, R. S. Fischer, A. Chitnis, and H. Shroff. Instant super-resolution imaging in live cells and embryos via analog image processing. *Nature methods*, 10(11):1122–1126, 2013.
  84. A. G. York, S. H. Parekh, D. Dalle Nogare, R. S. Fischer, K. Temprine, M. Mione, A. B. Chitnis, C. A. Combs, and H. Shroff. Resolution doubling in live, multicellular organisms via multifocal structured illumination microscopy. *Nature methods*, 9(7):749–754, 2012.




Figure 1. (Caption next page.)

---

**Figure 1. Biological super-resolution imaging with 3D-SIM.** (A) Mouse C127 mammary epithelial cell nucleus replication labelled with 5-ethenyl-2'-deoxyuridine (EdU, red) for 15 min before fixation with formaldehyde. The thymidine analogue EdU is incorporated into newly synthesized DNA of S-phase cells (here mid-to-late S phase) and detected via click-chemistry with Alexa Fluor 594 azide. DNA is labelled with 4',6-diamidino-2-phenylindole (DAPI, cyan). Single z-section of an image stack is shown with conventional wide-field illumination (top, left), structured illumination (1 of 15 raw images acquired per z-plane with laterally shifted and rotated stripes; bottom, left), and after 3D-SIM reconstruction (right). Note that 3D-SIM resolves higher-order domain organisation of chromatin and DNA-free interchromatin regions (inset), as well as the location of nuclear pores in the peripheral chromatin layer visible as DAPI void dots in the central region of the nucleus. Scale bar: 5  $\mu\text{m}$  and 1  $\mu\text{m}$  (inset). (B) Corresponding frequency distribution of the DAPI signal in Fourier (reciprocal) space. Concentric rings indicate the respective spatial resolution in  $\mu\text{m}$ . Spots in the raw SI frequency plot (arrowheads) correspond to first and second order stripes in the image (generated by a three-beam interference; only coarse first order stripes are visible in the image). (C) Orthogonal cross section and corresponding frequency distribution of the same dataset. The arrowheads indicate the position of the z-section shown in (A). Note the 2-fold extended frequency distribution in the reconstructed data in both lateral and axial direction, that includes the filling-in of the missing frequencies along the z-axis of the wide-field frequency plot.