

# MACHINE LEARNING CHARACTERIZATION OF CANCER PATIENTS-DERIVED EXTRACELLULAR VESICLES USING VIBRATIONAL SPECTROSCOPIES.

<sup>1</sup>Abicumaran Uthamacumaran, <sup>2</sup>Samir Elouatik, <sup>3,4</sup>Mohamed Abdouh, <sup>5</sup>Michael Berteau-Rainville, <sup>6</sup>Zhu-Hua Gao, and <sup>7,8,9\*</sup>Goffredo Arena

<sup>1</sup>*Concordia University, Department of Physics, Montreal, QC, Canada*

<sup>2</sup>*Université de Montréal, Département de chimie, Montreal, QC, Canada*

<sup>3</sup> Cancer Research Program, Research Institute of the McGill University Health Centre, 1001 Decarie Boulevard, Montreal, Quebec, Canada, H4A 3J1

<sup>4</sup>The Henry C. Witelson Ocular Pathology Laboratory, McGill University, Montreal, QC, Canada

<sup>5</sup>*Institut national de la recherche scientifique, Centre Énergie Matériaux Télécommunications, Varennes, QC, Canada, J3X 1P7*

<sup>6</sup> Department of Pathology, Research Institute of the McGill University Health Centre, 1001 Decarie Boulevard, Montreal, Quebec, Canada, H4A 3J1

<sup>7</sup>Istituto Mediterraneo di Oncologia, Viagrande, Italy

<sup>8</sup> Department of Surgery, McGill University, St. Mary Hospital, 3830 Lacombe Avenue, Montreal, Quebec, Canada, H3T 1M5

<sup>9</sup>Fondazione Gemelli-Giglio, Contrada Pollastra, Cefalu'

*\*Correspondence: Goffredo Arena\* (Primary Supervisor): [goffredoarena@gmail.com](mailto:goffredoarena@gmail.com)*

Abicumaran Uthamacumaran: [a\\_utham@live.concordia.ca](mailto:a_utham@live.concordia.ca)

*Keywords:* Cancer; Extracellular Vesicles (EVs); Spectroscopy; Machine Learning; Liquid Biopsy; Diagnostics; Complex Systems.

## **ABSTRACT**

The early detection of cancer is a challenging problem in medicine. The blood sera of cancer patients are enriched with heterogeneous secretory lipid bound extracellular vesicles (EVs), which present a complex repertoire of information and biomarkers, representing their cell of origin, that are being currently studied in the field of liquid biopsy and cancer screening. Vibrational spectroscopies provide non-invasive approaches for the assessment of structural and biophysical properties in complex biological samples. In this study, multiple Raman spectroscopy measurements were performed on the EVs extracted from the blood sera of  $n=9$  patients consisting of four different cancer subtypes (colorectal cancer, hepatocellular carcinoma, breast cancer and pancreatic cancer) and five healthy patients (controls). FTIR(Fourier Transform Infrared) spectroscopy measurements were performed as a complementary approach to Raman analysis, on two of the four cancer subtypes.

The AdaBoost Random Forest Classifier, Decision Trees, and Support Vector Machines (SVM) distinguished the baseline corrected Raman spectra of cancer EVs from those of healthy controls ( $N=18$ ) with a classification accuracy of  $> 90\%$  when reduced to a spectral frequency range of  $1800 - 1940\text{ cm}^{-1}$  and subjected to a 50:50 training: testing split. FTIR classification accuracy on  $N=14$  spectra showed an 80% classification accuracy. Our findings demonstrate that basic machine learning algorithms are powerful tools to distinguish the complex vibrational spectra of cancer patient EVs from those of healthy patients. These experimental methods hold promise as valid and efficient liquid biopsy for machine intelligence-assisted early cancer screening.

## 1. INTRODUCTION

Cancers remain globally one of the leading causes of disease-related death, accounting for nearly 10 million deaths in 2020 (World Health Organization, 2021). Early diagnosis, using different strategies, remains the best way to decrease mortality from neoplastic disease. Sera of cancer patients are enriched with extracellular vesicles (EVs) (Théry et al., 2018). EVs are nanoscopic (~30-200 nm), heterogeneous packets of information released by cells forming long-range, intercellular communication networks (Uthamacumaran, 2020; Théry, et al., 2018). They are emergent, reprogramming machineries used by cancer ecosystems to regulate cell fate dynamics (Guo et al., 2019). Furthermore, EVs can reprogram distant tissue microenvironments into pre-metastatic niches and horizontally transfer malignant traits to target cells located in distant organs (Abdouh et al., 2014; 2016; 2017; 2019; 2020; Arena et al., 2017; Hamam et al., 2016, Guo et al., 2019) as well as therapy resistance to promote aggressive cancer phenotypes (Steinbichler et al., 2019; Keklikoglou et al., 2019). On the other hand, EVs derived from an hESC (human embryonic stem cell) microenvironment have been shown to suppress cancer phenotypes and reprogram a subset of malignant phenotypes to a more benign states (Camussi et al., 2011; Zhou et al., 2017).

EVs are currently being studied as potential biomarkers in liquid biopsy for cancer screening (Zhao et al., 2019). However, the ability to distinguish cancer patients from healthy patients according to EVs patterns remains a challenge due to the lack of known specific EVs cancer markers. Raman spectroscopy provides a non-invasive experimental approach for the assessment of structural and biophysical properties in complex systems (Ember et al., 2017). The Raman technique depends on the change in polarizability of a molecule (due to the inelastic scattering of light by molecules) and measures the relative frequencies at which scattering of radiation occurs, while IR spectroscopy depends on a change in the dipole moment (Smith and Dent, 2005; Larkin, 2011).

The general principles of Raman scattering can be summarized as follows. When a photon interacts with

a molecule, it provides energy to induce a short-lived energy transition of the vibrational state of the molecule towards an unstable, higher-energy transition state. We call this Raman scattering. In contrast, the IR spectra would provide information about Raman inactive vibrational modes better suited for asymmetric polar mode vibrations. Conventionally, vibrational spectroscopy techniques have been used to elucidate the molecular structures of biochemical samples via Raman peak assignment (Brusatori et al., 2017). In the present project we applied Raman and FTIR spectroscopy to human sera EVs to evaluate for the presence of robust patterns able to characterize cancer EVs from non-cancer EVs in liquid biopsy for cancer screening.

The EVs in the patient-derived liquid biopsies consist of a complex mixture of heterogeneous EVs with a diverse set of information. To overcome the challenge represented by the potential inadequacy of these methods to capture the spectral complexity of cancer-derived EVs, we implemented the use of artificial intelligence (AI) to help in the detection of patterns in complex systems such as cancer EVs. Machine Learning algorithms are emerging as state-of-the-art approaches for pattern detection in data science (Uthamacumaran, 2020) and given the complexity of these systems, we hypothesized that the use of machine learning algorithms might provide a more efficient way of detecting characteristic patterns and classifying EVs between healthy and cancer group.

## **2. MATERIALS AND METHODS**

### **2.1 EVs EXTRACTION, PURIFICATION, AND CHARACTERIZATION.**

#### **2.1.1. Blood collection and serum preparation:**

Patients for the current study were recruited from the department of General Surgery at the Royal Victoria Hospital and St-Mary's Hospital (Montreal, Canada) and underwent a written and informed consent for blood collection in accordance to a protocol approved by the Ethics Committee of our institution (SDR-10-057). Blood samples were collected from both healthy individuals and patients who presented to our clinic for a follow-up or those that underwent resection of primary cancer (Table 1). Blood samples (10 ml) were collected from a peripheral vein in vacutainer tubes (Becton Dickinson) containing clot-activation additive and a barrier gel to isolate serum. Blood samples were incubated for 60 min at room temperature to allow clotting and were subsequently centrifuged at 1500 g for 15 min. The serum was collected, and a second centrifugation was performed on the serum at 2000 g for 10 min, to clear it from any contaminating cells. Serum samples were aliquoted and stored at -80°C until use.

**TABLE 1.** Patient cases identification.

<b>Patient ID</b>	<b>Disease</b>
Case 200717	Breast cancer
Case 160517	Pancreatic cancer
Case 426	Colorectal cancer
Case 549	Hepatocellular carcinoma
Case 161216	Healthy control
Case 220916	Healthy control
Case270717.1	Healthy control
Case270717.2	Healthy control
Case270717.3	Healthy control

### 2.1.2. EVs isolation on Iodixanol gradient:

Serum samples were diluted in cold phosphate buffered solution (PBS) and were subjected to a series of sequential differential centrifugation steps (Abdouh et al., 2019). Samples were centrifuged at 500 g for 10 min to remove contaminating cells, followed by centrifugation at 2000 g for 20 min to remove cell debris. Supernatants were passed through a 0.2  $\mu\text{m}$  syringe filter (Corning) and centrifuged at 16,500 g for 20 min at 4°C to remove apoptotic bodies and remaining cell debris. Supernatants were ultracentrifuged at 120,000 g (40,000 rpm) for 80 min at 4°C using 70 Ti rotor in an Optima XE ultracentrifuge machine (Beckman Coulter). The pellets were recovered in PBS/120 mM (4-(2-hydroxyethyl)-1-piperazineethanesulfonic acid) (HEPES) buffer solution. EVs samples were purified using iodixanol (OptiPrep density gradient, Sigma). A density gradient was prepared by serial dilutions of iodixanol stock (60% w/v): (i) 5 volumes of 60% iodixanol were mixed with 1 volume of 0.25 M Sucrose, 0.9 M NaCl and 120 mM HEPES solution (SNH solution, pH 7.4) to obtain 50% iodixanol, (ii) 2 volumes of 50% iodixanol were mixed with 3 volumes of SNH solution to obtain 20% iodixanol, and (iii) 1 volume of 50% iodixanol was mixed with 9 volumes of SNH solution to prepare 5% iodixanol. EVs samples (1.92 ml) were mixed with 2.88 ml of 50% iodixanol to obtain a 30% iodixanol solution. The discontinuous iodixanol density gradient was prepared by carefully layering 2.5 ml of 5% iodixanol and 2.5 ml of 20% of iodixanol sequentially in 13.26 ml Ultra-Clear tubes (Beckman Coulter). EVs in 30% iodixanol solution (4.8 ml) were carefully placed at the bottom of the tubes without disturbing the

gradient. Tubes were centrifuged at 120,000 g (40,000 rpm) for 2 hrs at 4°C using SW41 Ti swinging bucket rotor. Nine fractions of one milliliter each were collected, diluted in PBS and centrifuged at 120,000 g (40,000 rpm) for 80 minutes. The crude EVs pellets were washed twice with PBS at 120,000 g (40,000 rpm) for 80 min at 4°C, resuspended in 200 µl PBS, and stored at -80°C until use. For subsequent analyses, fractions 3-5 were pooled as they were enriched for EVs (Figure 1).

### 2.1.3. EVs characterisation:

EVs were characterized physically using Transmission Electron Microscopy (TEM) and NanoSight Nanoparticle Tracking Analysis (NTA), and phenotypically via Western blot. TEM analyses were performed using a JEM-2010 electron microscope (Jeol Ltd.). 20 µl of EVs samples were loaded on a copper grid and stained with 2% phosphotungstic acid. Samples were dried by incubating them for 10 min under an electric incandescent lamp and scanned.

In Parallel, diluted aliquots of the extracted EVs samples were run on a Nanosight NS500 system (Nanosight Ltd.), and the size distribution was analyzed using the NTA 1.3 software. For immunoblotting, pelleted EVs were lysed in (Radioimmunoprecipitation Assay) (RIPA) buffer containing protease inhibitors (Sigma). Equal amounts of proteins were resolved on 12% sodium dodecyl sulphate– polyacrylamide gel electrophoresis (SDS-PAGE) and transferred to a nitrocellulose membrane (BioRad). Membranes were blocked in tris-buffered saline (TBS) containing 5% non-fat dry milk and exposed overnight at 4°C to mouse-anti-TSG101 (ab83, Abcam), and mouse-anti-Alix (2127, Cell Signaling) antibodies. Membranes were washed 3 times in Tris-buffered saline with 0.05% Tween-20 (TBST) and incubated with anti-mouse horseradish peroxidase (HRP)-conjugated secondary antibody for 1 hr at room temperature. After 3 additional washes in TBST, the blots were developed using Immobilon Western HRP Substrate (Millipore, Etobicoke, Canada) (Théry et al., 2018). (Théry et al.,2018)

## **2.2 RAMAN SPECTROSCOPY PROTOCOL**

Thawed EVs samples prior to Raman measurements were lightly vortexed to resuspend the EVs in the buffer. Local Raman shifts in the samples were measured under ambient conditions using a Renishaw inVia micro-Raman confocal microscope system with a 514 nm excitation green laser. The samples were illuminated using a 50 X objective (numerical aperture of 0.75) and Raman spectra were recorded and analyzed using the Renishaw Wire software. Typical measurements began with the calibration of the instrument with a silicon wafer and background noise determination on a clean Raman-grade calcium fluoride (CaF<sub>2</sub>) slide (Crystran Ltd, UK) at 10% laser power, subsequently followed by measurements of the PBS buffer. The target irradiance on the sample is 2.5 mW/μm<sup>2</sup> to avoid sample degradation. Then, multiple measurements of the patient- derived EVs samples were performed under identical conditions after the 20 μL droplets of the EVs suspension has been air-dried on the CaF<sub>2</sub> slide. Measurements were performed for 30 second exposure time and 10 accumulations to optimize the signal-to-noise ratio (S/N).

## **2.3 FTIR SPECTROSCOPY PROTOCOL**

The FTIR measurements were performed in transmission mode with an Agilent FTIR microscope equipped with a (deuterated triglycine sulfate) DTGS detector. The samples were prepared as dried films of EVs/buffer mixture on CaF<sub>2</sub> windows (identical conditions as the Raman measurements). FTIR and Raman samples were prepared from the same EVs samples. The source power and magnification (15x) are standard and are fixed by the system configuration. The FTIR spectrum acquisition were performed with a 4 cm<sup>-1</sup> spectral resolution and 256 co-added scans.

## **3.4. SOFTWARE AND STATISTICAL ANALYSIS**

Data pre-processing consisted mainly of baseline correction. The infrared spectra were baseline-corrected using the asymmetric least squares smoothing method as implemented in OriginPro 2021. For the FTIR spectra, the asymmetric factor was set to 10<sup>-3</sup>, the threshold at 10<sup>-3</sup> and the smoothing factor at



3. For the Raman spectra, the asymmetric factor was set at  $10^{-4}$ , the threshold at  $10^{-3}$  and the smoothing factor at 3. Peak fitting was performed post-baseline correction (See Supplementary Information).

Machine Learning analysis was performed on Google Colab using the machine learning binary classifiers from the Scikit-learn python library (Pedregosa et al., 2011). The random forest algorithm builds and combines multiple decision trees for a greater accuracy in prediction of patterns in complex datasets. In decision trees, each tree predictor depends on the values of a random vector sampled independently (Breiman, 2001; Bishop, 2006). The random forest (RF) ensemble methods minimize the errors and optimize the variance-bias trade off in the datasets to provide a more accurate prediction in the data classification (Pedregosa et al., 2011; Geron, 2019). Herein, the AdaBoost classifier was used as an ensemble learner to enhance the predictive performance of the RF classifier. The hyperparameters of the three classifiers were tuned as follows:

*SVM*: linear kernel,  $C= 1.0$ , tolerance = 0.01, and all other default hyperparameters were kept.*RF*


*AdaBoost Classifier*:  $n\_estimators = 50$ ,  $C= 1.0$  (learning rate), and the SAMME.R algorithm.

*Decision Trees*: Entropy criterion and default hyperparameter.

### 3. RESULTS


#### 3.1. EVs isolated from serum displayed exosomes and microvesicles characteristics.

Based on the minimal information statement for the study of EVs set by the ISEV (International Society for Extracellular Vesicles) (Théry et al., 2018), we characterized the isolated EVs both physically and phenotypically. By using Western blot analysis, we observed that these vesicles expressed selective markers of EVs (i.e., Alix and TSG101) (Figure 1A). The highest expression levels of these markers were observed in fractions 3-5 (at iodixanol density of 1.107-1.13 g/ml). These fractions were subsequently pooled for further analyses. When assessed by NTA, isolated EVs displayed a mean diameter of 109 nm (range 59-145 nm) (Figure 1B). In addition, TEM analyses showed that the isolated EVs were round-shaped vesicles with a mean diameter of 90 nm (Figure 1C).



**FIGURE 1. ISOLATION AND CHARACTERIZATION OF PATIENT-DERIVED EVs.** (A) EVs were isolated as described under Materials and Methods using iodixanol (OptiPrep) gradient and ultracentrifugation. Proteins isolated from the different fractions were analyzed by Western blot for the expression of specific EVs markers. Note that the highest expression levels of EVs markers are located in fractions 3 to 5. These fractions were pooled for subsequent analyses. MW = Protein molecular weight marker. (B) NTA analysis of pooled fractions 3-5. (C) TEM micrograph of purified EVs (red arrowheads). Scale bar 100 nm.


**3.2. Reinshaw Raman Spectroscopy identifies rich spots of EVs on air-dried CaF<sub>2</sub> slides.** Spectra acquisition consisted of scanning at multiple pink spots on the prepared sample slides (Figure 2A). The spontaneous formation of pyramidal crystals on the slides are shown (Figure 2B).



**FIGURE 2. EVs DETECTION BY SPECTROSCOPY.** (A) EVs SPOT ON RAMAN CONFOCAL MICROSCOPE (50X Objective). The axes represent the microscope field of view sizes in X and Y directions corresponding to the 50x lens that was used for measurement. The Inset bar gives the field of view scale corresponding to the 50x objective lens. The image shows the EVs spot focused on the air-dried CaF<sub>2</sub> slide through the confocal microscope of the inVia Reinshaw Raman system, with the scale bar of 20 μm as indicated by the scale bar on the bottom. These pink spots are regions enriched with EVs and provide a method to infer where the optimal Raman spectra are acquired. The yellow patchy regions correspond to the slide with higher concentrations of the PBS buffer. (B). CRYSTAL FORMATION. Pyramidal crystals spontaneously self-organized in the EVs rich regions (pink spots).

**3.3. Vibrational Spectra highlight characteristic peaks in patient-derived EVs.** The protein bands correspond to the peaks at 1440 cm<sup>-1</sup> (CH<sub>2</sub>, CH<sub>3</sub> deformation), 1580 cm<sup>-1</sup> (amide II bond) and amide I protein band (1600–1690 cm<sup>-1</sup>) visible in Figure 3. They were present in both healthy and cancer samples with no distinct observable patterns distinguishing the two. The lipid- bands (2750–3040 cm<sup>-1</sup>) are clearly distinct in the baseline corrected Raman spectra and have a stronger intensity in some of the


cancer EVs samples (except for Case200717) in comparison to the healthy samples (the EVs membrane is composed of lipids) (Figure 3A-D). The spectra show that the differences in vibrational modes are difficult to generalize amidst the different cancer groups using traditional peak fit assignment.



**FIGURE 3. BASELINE-CORRECTED EVs SPECTRA.** Raman (A-D) and FTIR (E-F) spectra of EVs isolated from health and cancer patient sera. **(A) BASELINE-CORRECTED RAMAN SPECTRA OF Case 426 (COLORECTAL CANCER).** The Raman spectra of cancer sample 426 (in pink) and healthy controls (in green) are shown as measured at a 514 nm excitation laser with 10% laser power, 10 accumulations for 30 seconds. Identical conditions were used for the following Raman spectra in the Graph 1 panel. EVs drops were air dried on the calcium fluoride slides prior to measurement acquisitions. The single peak at 321 cm<sup>-1</sup> corresponds to the CaF<sub>2</sub> substrate. **(B) BASELINE-CORRECTED RAMAN SPECTRA OF 549 (HEPATOCELLULAR CARCINOMA).** **(C) BASELINE - CORRECTED RAMAN SPECTRA OF CANCER Case 200717 (BREAST CANCER).** The single high intensity peak in the 2800-3010 cm<sup>-1</sup> corresponds to the lipid band. **(D) BASELINE-CORRECTED RAMAN SPECTRA OF CANCER Case 160517 (PANCREATIC CANCER).** **(E) BASELINE-CORRECTED FTIR SPECTRA OF 426.** **(F) BASELINE-CORRECTED FTIR SPECTRA OF 549.**

**3.4. PCA (Principal Component Analysis) Dimensionality reduction accurately separates clusters of healthy patients' (control) EVs from cancer patients' EVs.**

The PCA 2D scatter plot shows a pattern with data dispersion and lower cluster resolution along with a significant overlapping between the domains of normal population and cancer population data points (Figure 4). On the PCA plot, we can see a separation of different data domains. In analysis of the Raman spectra shown above, it was deduced that the spectral range from 1800-1940  $\text{cm}^{-1}$  would provide a better PCA clustering separation amidst the two groups (corresponding to Figure 4B and 4D). PCA clustering seems to well separate healthy FTIR spectra from cancer FTIR spectra, as well (figure 4E and 4F).




**FIGURE 4. PCA CLUSTERING ON EVs SPECTRA. (A). PCA Clustering on Raman Spectra without Baseline Correction (N=19).** The raw Raman spectra of 19 spectra (12 cancers and 7 controls) from n= 9 patients have been clustered by their first two principal components. Some of the spectral PCA points are overlapping between the cancer and healthy groups (in pink and green, respectively). **(B) PCA on Frequency Reduced Raman Spectra without Baseline Correction (N=18).** **(C) PCA Clustering on Baseline-**

**Corrected Raman Spectra (N=18).** The complete range Raman spectra when subjected to the baseline correction, shows better PCA separation amidst the binary classes. (D) **PCA Clustering on Baseline-Corrected Raman Spectra with Reduced Frequency (N=18) (1800-1940  $\text{cm}^{-1}$ ).** (E) **PCA Clustering on FTIR Spectra (N=14) without Baseline Correction.** The PCA space of raw FTIR spectra of 8 cancer samples (four of patient sample 426 and four of patient sample 549) and 6 healthy controls shows well-separated clusters even without baseline correction. (F) **PCA Analysis on Baseline-Corrected FTIR Spectra (N=14).**

**3.5. Machine Learning algorithms exhibit poorer performance in the classification of EVs on Raman spectra without baseline correction.** Various binary classification algorithms were trained on raw Raman spectra without baseline correction to assess their performance accuracy in distinguishing cancersamples from healthy controls (Figures 5A-J). All results collectively confirm that without baseline correction, ML algorithms exhibit poor predictive performance.






**FIGURE 5. ML PERFORMANCE ON RAW RAMAN SPECTRA. (A). ADABOOST RANDOM FOREST (RF) ON WHOLE RANGE RAW RAMAN SPECTRA. (B) ADABOOST CROSS-VALIDATION CURVE ON RAW RAMAN SPECTRA FOR WHOLE-RANGE. (C) ADABOOST RF CLASSIFICATION ON RAW RAMAN SPECTRA WITH REDUCED FREQUENCY. (D) CV CURVE FOR ADABOOST ON REDUCED RAW RAMAN SPECTRA. CV**

accuracy score:  $20.00 \pm 24.49\%$ . **(E) ROC CURVE FOR DECISION TREES ON RAW RAMAN SPECTRA NO BASELINE.** **(F) CV CURVE FOR DECISION TREES PERFORMANCE ON RAW RAMAN SPECTRA (at  $1808.25 \text{ cm}^{-1}$ ).** The CV accuracy score was determined as  $60.00 \pm 20.00 \%$ . With an increased sample size of patients and hence, increased training datasets the performance of these algorithms can be better optimized for the intended impact of the presented findings. **(G) ROC CURVE FOR SVM CLASSIFICATION ON RAW RAMAN SPECTRA (COMPLETE).** **(H) CLASSIFICATION REPORT FOR SVM ON RAW RAMAN SPECTRA (FULL RANGE).** CV accuracy score:  $80.00 \pm 24.49 \%$ . **(I) ROC CURVE FOR SVM PERFORMANCE ON RAW RAMAN SPECTRA WITH REDUCED FREQUENCY RANGE.** **(J) CV CURVE FOR SVM PREDICTIONS ON REDUCED SPECTRA.** CV accuracy score:  $90.00 \pm 20.00 \%$ .

**3.6. ML algorithms show near 90% classification accuracy in distinguishing healthy and cancer EVs on Baseline-corrected Raman spectra in the reduced frequency range (1800-1940 cm<sup>-1</sup>).** All tested

binary classifiers well distinguished cancer EVs from healthy EVs with a near 90% classification accuracy on the reduced frequency Raman spectra (Figures 6A-H). Although the SVM algorithm showed the highest classification accuracy on the baseline-corrected Raman spectra within the reduced frequency range, the cross-validation score showed significant uncertainty (standard deviation) (Figures 6 G-H). As such, *the AdaBoost Random Forest classifier was determined to be the most efficient and robust algorithm in distinguishing the spectra between the two patient groups with the highest sensitivity and specificity* (Figures 6A-D). These findings confirm that ML algorithms and hence, AI, can accurately distinguish between cancer and healthy patients- derived EVs on baseline corrected Raman spectra with high sensitivity and specificity.




**FIGURE 6. (A). ML PERFORMANCE ON BASELINE-CORRECTED RAMAN SPECTRA. AdaBoost Classifier’s ROC on Baseline Corrected Raman Spectra (Full Range). (B) ADABOOST CLASSIFIER’S CV LEARNING CURVE ON BASELINE CORRECTED RAMAN SPECTRA. CV accuracy score:  $90.00 \pm 20.00\%$ . (C) AdaBoost Classifier’s ROC on Baseline Corrected Raman Spectra (Reduced Frequency Range). (D) ADABOOST CLASSIFIER’S CV**

**LEARNING CURVE ON BASELINE CORRECTED RAMAN SPECTRA (REDUCED FREQUENCY).** CV accuracy score:  $83.33 \pm 21.08$  %. (E) **CV Curve for Decision Tree on Baseline Corrected Raman Spectra.** CV accuracy score:  $30.00 \pm 24.45$  %. (F) **ROC CURVE FOR SVM ON BASELINE CORRECTED RAMAN SPECTRA (FULL RANGE).** Classification accuracy of 66.66% and MSE: 0.333. The f1 scores of the cancer and healthy group classification were 0.77 (i.e., precision of 1.00 and recall of 0.62) and 0.40 (i.e., precision of 0.25 and recall of 1.00), respectively corresponding to an AUC of 0.81. The CV accuracy score was found to be  $80.00 \pm 40.00$  %. (G) **ROC CURVE FOR SVM PERFORMANCE ON BASELINE CORRECTED RAMAN SPECTRA WITH REDUCED FREQUENCY RANGE.** Classification accuracy of 100% and MSE: 0.0. The turquoise line is not visible and farthest away from the red-dashed line of the ROC curve indicating perfect classification accuracy. The f1 scores were 1.00 for both groups indicating a 100% sensitivity and specificity. (H) **CV CURVE FOR SVM PERFORMANCE ON REDUCED FREQUENCY RAMAN SPECTRA.** CV accuracy score:  $90.00 \pm 20.00$  %.

### **3.7. Tree-based classification algorithms well-distinguish cancer FTIR spectra from healthy FTIR**

**spectra with 90% classification accuracy.** Both tree-based machine learning algorithms, the AdaBoost RF and Decision Tree classifiers exhibited a decent classification accuracy in distinguishing cancer FTIR spectra from healthy patient FTIR spectra with an area under the curve of 0.83 and 1.00, respectively (Figures 7A-D). However, the SVM was found to be a poor classifier of the two patient groups' FTIR spectra (Figures 7E-F). These preliminary results demonstrate that Raman spectra may provide a more robust approach to characterizing and distinguishing cancer EVs from healthy EVs in comparison to FTIR.



**FIGURE 7. ML PERFORMANCE ON BASELINE-CORRECTED FTIR SPECTRA. (A) ROC CURVE OF ADABOOST CLASSIFIER ON BASELINE CORRECTED FTIR SPECTRA. (B) CV CURVE FOR ADABOOST CLASSIFICATION ON BASELINE CORRECTED FTIR. CV score:  $100.00 \pm 0.00\%$ . (C) ROC CURVE FOR DECISION TREES ON RAW FTIR SPECTRA. (D) CV CURVE FOR DECISION TREES ON RAW FTIR SPECTRA. CV score of  $90.00 \pm 20.00\%$ . (E) ROC CURVE FOR SVM ON BASELINE CORRECTED FTIR SPECTRA. 57.14 % classification accuracy and MSE of 0.4. The results confirm again that the baseline correction has poorer classification performance than the raw FTIR spectra. The f1 scores of the cancer and healthy groups' classification**

were found to be 0.73 and 0.00, respectively. **(F) CV CURVE OF SVM PERFORMANCE ON BASELINE**

**CORRECTED FTIR SPECTRA.** CV accuracy:  $75.00 \pm 25.00\%$ .



#### 4. DISCUSSION

ML is emerging as a powerful tool in data science to help identify patterns in complex systems. For instance, ML algorithms applied on Intraoperative Raman spectroscopy can distinguish brain cancer cells from normal brain tissue (including both invasive and dense cancers) with an accuracy, sensitivity, and specificity > 90% (Jermyn et al., 2015). Background subtraction algorithms, feature extraction techniques and autofluorescence removal algorithms can be employed to distinguish the cancer cells' Raman bands from those of healthy tissues (Brusatori et al., 2017; Zhao et al., 2007). ML algorithms provide a robust computational platform to detect patterns in the complex spectra of cancer EVs, reshaping our understanding of these nanoscale communication networks (Uthamacumaran, 2020). Our study is a proof-of-concept demonstration of the promises machine intelligence holds in computational medicine.

Recently, EVs from stem cells were analyzed using Raman spectroscopy (Gualerzi et al., 2019). Following baseline correction and normalization of the obtained Raman spectra, cluster analysis and interpolation found that adult stem cell-derived EVs grouped differently when subjected to simple dimensionality reduction techniques like principal component analysis-linear discriminant analysis (PCA-LDA) (Gualerzi et al., 2019). Importantly, it was demonstrated that the PCA clustering of EVs from different stem cell types depended also on the purity of the EVs. Therefore, purification technique used to extract and isolate the EVs from the liquid biopsies, namely size-exclusion column chromatography or differential ultracentrifugation, plays a critical role in the cluster separation of EVs Raman spectra.

Similarly, surface-enhanced Raman scattering (SERS) signals of exosomes from normal and NSCLC (Non-small-cell lung carcinoma) cancer cells on gold-nanoparticle substrates were performed by Shin et al. (2018). The exosomes/EVs were isolated using size-exclusion column chromatography techniques. SERS provides an enhancement of the Raman scattering signals by use of metallic, nanostructured substrate. As such, even low concentrations of biomolecules can be detected using the SERS technique. The SERS

Raman spectra of cancerous exosomes showed unique vibrational peaks distinguishing NSCLC exosomes from healthy clusters when subjected to spectral decomposition by Principal Component Analysis (PCA) (Park et al., 2017; Shin et al., 2018; Rojalin et al., 2019).

Shin et al. (2018) analyzed 25 spectra for each of the HPAEC (normal), H1299, and PC9 (NSCLC lung cancer cell lines) cell-derived exosomes, as well as the PBS control (pure buffer). PCA clustering well separated the cancer exosomes from the controls. Additionally, the cancerous exosomes tended to be located on the positive side of the PCA loading, which can be screened for biomarkers. Furthermore, correlation analysis on the Raman bands of exosome protein markers identified EGFR as the most unique Raman band for NSCLC exosomes (Shin et al., 2018). Similar findings with peak fitting algorithms and the multivariate curve resolution- alternating least squares (MCR-ALS) algorithm on Raman spectra were used to cluster-classify pancreatic cancer EVs on gold-nanoparticle plated SERS substrates (Banaei et al., 2017).

Romana et al. (2020) used FTIR spectroscopy in the mid-Infrared (mid-IR) range to study exosomes released from human colorectal adenocarcinoma HT-29 cancer cells cultured in different media can be classified using PCA-LDA. Another recent study showed that neural networks optimized with ML algorithms such as principal component analysis– linear discriminant analysis (PCA-LDA) and binary classifiers like SVMs (Support Vector Machines) can distinguish oral cancer patients from healthy individuals by characteristic signatures in the Fourier transform infrared (FTIR) spectroscopy of their salivary exosomes (Zlotogorski-Hurvitz et al., 2019). Thus far, however, FTIR characterization of the complex spectra of patient-derived heterogenous exosomes (EVs) remain uninvestigated. Taken together, these findings collectively suggest simple ML algorithms coupled to spectroscopy techniques can pave the accurate detection and classification of cancers from patient-derived liquid biopsies. Our efforts and results as presented hereby may pave minimally invasive cancer screening with liquid-biopsy

derived EVs characterization.

We showed the emergence of crystals on the prepared EVs sample slides. Caution must be taken to ensure the Raman laser is focused on the EVs solution and not these crystals whose origin is unknown. They may have emerged through the PBS buffer crystallization upon drying effect, and potentially catalyzed by the interaction with the incident laser beam (higher amounts of pyramids were present post-irradiation). Monitoring by time-lapse imaging under controlled heat and conditions is suggested for prospective studies.

***ML algorithms well-classified Baseline-corrected Raman spectra with a near 90% classification accuracy.***

We demonstrated that the Raman spectra at a selected frequency region between 1800-1940  $\text{cm}^{-1}$  wave numbers show several peaks distinguishing healthy from cancer EVs spectra. This selected frequency range corresponds to C=C double bonds and C=O double bonds (strong signals for Raman in this region) and were shown to enhance the classification accuracy of ML classifiers on the Raman spectra. However, their exact structural identity remains unelucidated. There are no major differences distinguishable at the level of the Raman peak assignment. This further supports that our traditional approach of peak assignments and molecular bond characterization are vastly insufficient for the study of complex systems such as cancer EVs. Although no visible peak differences are observed, the ML classifiers were able to distinguish the cancer EVs from healthy EVs via pattern recognition.

The spectra of all samples and their Raman/FTIR peak assignment to molecular bonds is provided in the Supplementary Information. The alignment of the multiple spectral measurements for each sample indicates there are no differences in the spectra due to sample preparation, impurities, or differences due to optical effects present (figures S3-S8). We ruled out differences seen due to spatial heterogeneity instead of differences between samples with repeated measurements. The FTIR data of

both cancer samples 426 (colorectal cancer) and 549 (hepatocellular cancer) revealed lipid bands in the 2850-3010  $\text{cm}^{-1}$  region (corresponding to C-H stretches) (Figure 3A- D). The 3200-3400  $\text{cm}^{-1}$  stretch corresponds to the OH stretch. Some peaks which are tentatively present only within the cancer FTIR spectra were present at 1120  $\text{cm}^{-1}$ , and at 3070  $\text{cm}^{-1}$  (Figure 3E and 3F).

When the Raman spectra was reduced to only the 1800-1940  $\text{cm}^{-1}$  wave number region, a better PCA separation was observed in 18 samples with 27 intensity data points in each (one control was removed as an outlier due to low intensity counts) (Figure 4). The cancer spectra, although diverse and heterogeneous (there are four different cancer subtypes present) seem to cluster in the left quadrant of the PCA space. There is still one green data point (healthy spectrum) clustered with the cancer spectra. In the following ML classification results, *reduced (or selected) frequency* denotes Raman spectra being processed to only the 1800-1940  $\text{cm}^{-1}$  region of interest (unless otherwise specified, as in the case for the Decision Trees). With further detailed spectral analysis, additional spectral ranges can be included in downstream analysis. A correct artifact filtering and peak fitting of the Raman spectra is required to primarily focus on the relevant spectral information only and achieve a more efficient PCA classification. Even simple dimensionality reduction techniques provide a quick tool to characterize and distinguish serum derived EVs from cancer patients with those of healthy individuals.

The AdaBoost RF classifier is a meta-estimator and an iterative ensemble learner available on Sci-kit/sklearn (*Python machine learning library*). The AdaBoost RF classifier was assessed on 1021 data points from 19 raw Raman spectra without baseline correction from the two classes (Figure 5). To assess the sensitivity/specificity of the ML predictions, receiver operating characteristic (ROC) curves are generated to show the diagnostic ability of the binary classifier with varied discrimination thresholds. As shown in Figure 5, the ROC curve for the AdaBoost classifier's performance on the raw Raman spectra (whole range) with a testing size of 0.5 (i.e., the ML is trained on 50% of the data and tested on 50%) is shown (Figure 5A). A 0.5 test size ensures stringent training conditions for the

classification assessment. A smaller test size of 0.2 and 0.3 always showed greater classification accuracy. 50 tree estimators and a learning rate of 1.0 were kept as the default hyperparameters. The classification accuracy was 77.78% with a mean-square error (MSE) of 0.222. As seen, the area under the curve (AUC) was 0.45 indicating a poor classification accuracy. The turquoise curve shows the relationship between the true positive rate (TPR) and false positive rate (FPR). The closer the turquoise curve comes closer to the red dashed curve at 45 degrees of the graph plane, the less accurate the classifier's predictions, and lower AUC of the ROC curve. The ROC curve visually informs us the trade-off between the sensitivity (TPR) and the specificity (1-FPR) (Bishop, 2006). The f1 scores were 0.62 (i.e., a precision of 0.67 and recall of 0.57) for the cancer group and 0.29 (i.e., a precision of 0.25 and recall of 0.33) for the healthy groups. The F1 score of 1.00 indicates a perfect recall and precision. The F1-score is often used as a measure of statistical accuracy for binary classifiers in ML (Breiman, 2001; Geron, 2019).

The poor performance of the RF classifier in Figure 5 indicates the data must be filtered to a narrower spectral range or alternately, undergo a baseline correction, as indicated by the PCA plot of the reduced frequency space above. In binary classification, recall of the positive class is defined as sensitivity while the recall of the negative class is specificity. The precision is defined as the ratio  $tp / (tp + fp)$  where  $tp$  is the number of true positives and  $fp$  the number of false positives. The recall is the ratio  $tp / (tp + fn)$  where  $fn$  is the number of false negatives. The recall denotes the ability of the classifier to find all the positive samples. The f-1 score defines the weighted harmonic mean of the precision and recall, where an f-1 score reaches its best value at 1 and worst score at 0. Here, the f1 score was found to be of 1.00 (i.e., precision and recall were 1.00) implying both, a 100% sensitivity and specificity.

Figure 5B displays the cross-validation learning curve corresponding to graph 5A. It shows that the AdaBoost classifier is not optimally tuned to predict the classes of the newly presented test datasets as

indicated by the vast grey shaded region (indicates training uncertainty). The grey fill space on the plot denotes the standard deviation of the training performance by the classifier as the training size increases. The broad range of grey fill indicates a heavier computational training is required for the classification accuracy to be optimized. The cross-validation score, or also known as out-of-sample testing, indicates the likelihood of the RF classifier's performance when new results are presented with the current amount of training it has undergone. It is a validation technique to generalize the performance of the RF classifier to an independent dataset. The CV accuracy score was found to be  $70.00 \pm 24.49$  %. The curve in turquoise corresponds to the CV score curve and the optimal training score of 1.00 is indicated by the dashed violet curve.

The RF classifier was assessed on the frequency reduced spectra ( $1800\text{-}1940\text{ cm}^{-1}$  wavenumber region) of 18 samples (12 cancers and 6 healthy controls). With a 0.5 test size the classification performance was 50.00 % with a MSE of 0.5 indicating poor performance of the classifier. The AUC is 0.67 indicating a poor sensitivity and specificity (Breiman, 2001; Geron, 2019). An f1 score of 0.86 was observed for the cancer group and of 0.50 for the healthy group (Figure 5C).

Decision trees are a supervised learning technique which use multiple algorithms to decide to split a node into two or more sub-nodes with tree-like diagrams to classify some target variable/data. The performance of the decision trees classifier is shown with a 0.5 test size on a randomly selected single frequency (at  $1808.25\text{ cm}^{-1}$ ). The classification accuracy was found to be 88.89% with a MSE of 0.111. The AUC is 0.94 indicating a high classification accuracy. An f1 score of 0.93 (i.e., precision of 1.00 and recall of 0.88) for cancer and 0.67 for healthy (i.e., precision of 0.50 and recall of 1.00) was observed. The entropy criterion (one of the learning parameters) was used for the tree classification. The results remained unchanged with baseline correction for the Decision Tree performance (Figure 5E).

The classification predictions by the SVM algorithm with a linear kernel is shown on the full range raw Raman spectra. SVM is a supervised ML algorithm which finds the optimal hyperplane that maximizes the margin between the data classes using gradient descent learning. Classification accuracy of 80.00% and MSE: 0.2 were observed. An f1 score of 0.75 for the cancer class (i.e., precision and recall of 0.75) and of 0.83 for the healthy class (i.e., precision and recall of 0.83) were obtained corresponding to an AUC of 0.79 (Figure 5G). The SVM classification accuracy on the raw spectra within the selected frequency range reported above and a 0.5 test size was found to be 88.89 % with a MSE of 0.111. The f1 scores for the cancer and healthy group classification were 0.92 (precision of 0.86 and recall of 1.00) and 0.80 (precision of 1.00 and recall of 0.67), respectively (Figure 5I).

Figure 6 displays the ML results on the baseline corrected Raman spectra. A classification accuracy of 88.89% and MSE of 0.111 was observed for the AdaBoost RF classifier on baseline corrected Raman spectra. The f1 scores of the cancer and healthy group classification were 0.91 (i.e., precision of 1.00 and recall of 0.83) and 0.86 (i.e., precision of 0.75 and recall of 1.00), respectively corresponding to an AUC of 0.92 (Figure 6A). A classification accuracy of 83.33% and MSE of 0.167 was found for the AdaBoost classifier on baseline corrected Raman spectra within the reduced frequency range. The f1 scores of the cancer and healthy group classification were 0.89 (i.e., precision of 1.00 and recall of 0.80) and 0.67 (i.e., precision of 0.50 and recall of 1.00), respectively corresponding to an AUC of 0.90 (Figure 6C). As shown, in comparison to Figure 5, the ML performance enhanced by baseline correction of the spectra. Hence, some of the essential features of the raw Raman spectra can be removed by the baseline correction or perhaps the quality of the baseline correction. Although, the classification accuracy remains roughly the same for both baseline-corrected spectra in the complete range and reduced frequency range, their precision has changed especially for the healthy controls (Figure 6D). Similar classification accuracies were obtained for the Decision Trees and SVM algorithms as well (Figure 6E-H). Their classification

accuracy was enhanced to 100.00% on the reduced frequency spectra (1800-1940  $\text{cm}^{-1}$ ). However, the improved classification accuracy was accompanied by a lowered cross-validation score (i.e., increased standard deviation). These results collectively confirm that the baseline correction improves the ML classification accuracy in detecting patterns distinguishing cancer spectra from healthy spectra.

***Preliminary findings show ML algorithms better predict cancer EVs from Raman spectra than from FTIR spectra.***

As demonstrated in Figure 7, the FTIR was used as a complementary verification of the Raman data.

AdaBoost Classifier was tested on N = 14 baseline corrected FTIR spectra, 857 points each spanning from 698 to 4000  $\text{cm}^{-1}$ . An 80.00% classification accuracy and MSE of 0.2 was observed. The f1 scores of the cancer and healthy groups classification were 0.80 for both leading to an AUC of 0.83. Our findings show the raw data (no baseline correction) performed better than the baseline corrected FTIR spectra.

The opposite trend was observed for the Raman spectra where a better classification accuracy was seen overall with the baseline corrected spectra (Figure 7A). The classification accuracy for a randomly selected frequency (at 698.22975  $\text{cm}^{-1}$ ) was found to be 100.00 % with a MSE of 0.0. The specificity and sensitivity were 1.00. Our findings repeatedly confirm Decision Trees at selected frequencies may be robust predictors of distinguishing healthy from cancer spectra in both cases, the Raman, and FTIR methods (Figure 7C).

Our results demonstrate vibrational spectroscopies coupled with basic machine learning algorithms maybe a robust tool for the detection of cancer in patient-derived liquid biopsies by the characterization of EVs. One limitation of our study is the sample size. Hence, the inclusion of a larger sample size of patients with cancers of different origins and grades/stages is needed to enhance its impact in clinical medicine. Another limitation is the presence of lipoproteins in our patient- derived EVs samples. Lipoproteins will co-isolate with the EVs, even after flotation. Size-exclusion chromatography following the ultracentrifugation could have partially resolved this issue. However, given the



lipoproteins were present in all patients, and our algorithms distinguished between cancer and healthy patients with > 90% accuracy, patterns are consistently observed between the two patient groups which could have not been affected by the lipoproteins.

## 5. CONCLUSION

The presented findings strongly demonstrate that the ML classifiers can accurately distinguish cancer EVs from those of healthy patients using both vibrational spectroscopy techniques. Traditional peak assignment approaches are insufficient in assessing the complex spectra of cancer EVs and distinguishing them from those of healthy controls. Our study demonstrates basic ML classifiers can find patterns which traditional approaches fail to reconcile. It must be reminded that these are heterogeneous complex systems since the EVs were acquired from the blood sera of patients. A wide repertoire of EVs from different tissues within the patient are expected in each sample (i.e., deriving from inflammatory and other organismal cells). Previous studies only investigated cancer EVs obtained from cancer cell lines which reduce the degree of complexity in comparison to our samples. However, regardless, the ML algorithms identified Raman signatures distinguishing the two groups with high degrees of statistical accuracy, sensitivity, and specificity. The ML performance on Raman spectra had a better classification accuracy than those on FTIR spectra. Our findings can be further improved by the use of SERS (Surface Enhanced Raman Spectroscopy) on patient samples with a uniform gold substrate to obtain a better resolution of peaks. These preliminary findings with N= 18 analysis strongly suggest that Raman spectroscopy in combination with machine learning classifiers can be used to develop sensitive liquid biopsies for early detection of cancer in screened patients. Our study displays the impact and scope of interdisciplinary applications of EVs in improving healthcare and medicine.

**DISCLOSURE STATEMENT:** The authors report no conflict of interest.

**FUNDING STATEMENT:** Giuseppe Monticciolo financially supported the research and the experiments described in this paper. The funder had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript.

**DATA AVAILABILITY STATEMENT:** All data generated and analyzed during this study are included in this manuscript and in its supplementary information files.

**ETHICS APPROVAL STATEMENT:** Ethics approval and consent to participate Patients recruited for this study underwent an informed and written consent for blood collection in accordance to a protocol approved by the Ethics Committee of the McGill University Health Centre (Reference. MP-37-2018- 3916 and 10-057-SDR). Mice involved in this study were used with the approval and in compliance with the McGill University Health Centre Animal Compliance Office (Protocols 2012-7280 and 2015-7731).

**PATIENT CONSENT STATEMENT:** The authors declare patient consent was granted for the study.

**AUTHOR CONTRIBUTIONS:**

AU performed the machine learning algorithms, co-wrote, and edited the manuscript.

SE carried out the spectroscopy measurements.

MA extracted and purified the patient-EVs, co-wrote, and edited the manuscript.

MBR performed the baseline corrections and spectral peak fit analysis.

ZHG co-supervised the project.

GA co-supervised the project, co-wrote and edited the manuscript.

## REFERENCES

- 1) Cancer. World Health Organization. <https://www.who.int/news-room/fact-sheets/detail/cancer>  
(visited on June 2021)
- 2) C.; Witwer, K.W.; Aikawa, E.; Alcaraz, M.J.; Anderson, J.D.; Andriantsitohaina, R.; Antoniou, A.; Arab, T.; Archer, F.; Atkin-Smith, G.K.; et al. Minimal information for studies of extracellular vesicles 2018 (MISEV2018): a position statement of the International Society for Extracellular Vesicles and update of the MISEV2014 guidelines. *J. Extracell. Vesicles* 2018, 7, 1535750, doi:10.1080/20013078.2018.1535750.
- 3) Zhao Z, Fan J, Hsu Y-M, Lyon CJ, Ning B, Hu TY. Extracellular vesicles as cancer liquid biopsies: from discovery, validation, to clinical application. *Lab Chip*. 2019;19(7):1114-1140. doi: 10.1039/c8lc01123k.).
- 4) Banaei, N. et al., Multiplex detection of pancreatic cancer biomarkers using a SERS-based immunoassay. *Nanotechnology* 28(45): 455101 (2017)
- 5) Bishop, C.M. *Pattern Recognition and Machine Learning* (Springer, 2006)
- 6) Breiman, L. Random Forests. *Machine Learning*, 45(1): 5–32 (2001)
- 7) Brusatori, M. et al. Intraoperative Raman Spectroscopy. *Neurosurgery clinics of North America* 28(4): 633–652 (2017)
- 8) Camussi, G. et al., Exosome/Microvesicle-Mediated Epigenetic Reprogramming of Cells. *American Journal of Cancer Research* 1(1): 98–110 (2011)
- 9) Ember, K. et al., Raman spectroscopy and regenerative medicine: a review. *Npj Regenerative Medicine*, 2(1):12 pp.1-12. (2017)

- 10) Abdouh M, Tsering T, Burnier JV, de Alba Graue PG, Arena G, Burnier MN. Horizontal transfer of malignant traits via blood-derived extracellular vesicles of uveal melanoma patients. *Invest. Ophthalmol. Vis. Sci.* 61(7):2835. 2020.
  
- 11) Abdouh M, Floris M, Gao ZH, Arena V, Arena M, Arena GO. Colorectal cancer-derived extracellular vesicles induce transformation of fibroblasts into colon carcinoma cells. *J Exp Clin Cancer Res.* 38(1):257. 2019. / Abdouh M, Gao ZH, Arena V, Arena M, Burnier MN, Arena GO. Oncosuppressor-Mutated Cells as a Liquid Biopsy Test for Cancer-Screening. *Sci Rep.* 9(1):2384. 2019.
  
- 12) Abdouh M, Hamam D, Gao ZH, Arena V, Arena M, Arena GO. Exosomes isolated from cancer patients' sera transfer malignant traits and confer the same phenotype of primary tumors to oncosuppressor-mutated cells. *J Exp Clin Cancer Res.* 36(1):113. 2017.
  
- 13) Arena GO, Arena V, Arena M, Abdouh M. Transfer of malignant traits as opposed to migration of cells: A novel concept to explain metastatic disease. *Med Hypotheses.* 100:82-86. 2017.
  
- 14) Abdouh M, Hamam D, Arena V, Arena M, Alamri H, Arena GO. Novel blood test to predict neoplastic activity in healthy patients and metastatic recurrence after primary tumor resection. *J Circ Biomark.* 5. doi: 10.1177. 2016.


- 15) Hamam D, Abdouh M, Gao ZH, Arena V, Arena M, Arena GO. Transfer of malignant trait to BRCA1 deficient human fibroblasts following exposure to serum of cancer patients. *J Exp Clin Cancer Res.* 35:80. 2016.
- 16) Abdouh M, Zhou S, Arena V, Arena M, Lazaris A, Onerheim R, Metrakos P, Arena GO. Transfer of malignant trait to immortalized human cells following exposure to human cancer serum. *J Exp Clin Cancer Res.* 33:86. 2014
- 17) Géron, A., *Hands-On Machine Learning with Scikit-Learn, Keras, and TensorFlow (Second Ed., O'Reilly Media, Inc, 2019)*
- 18) Gualerzi, A. et al., Raman spectroscopy as a quick tool to assess purity of extracellular vesicle preparations and predict their functionality. *Journal of extracellular vesicles*, 8(1):1568780 (2019)
- 19) Guo, Y. et al., Effects of Exosomes on Pre-metastatic Niche Formation in Tumors. *Molecular Cancer*, 18:39 (2019)
- 20) Hoshino, A. et al., Extracellular Vesicle and Particle Biomarkers Define Multiple Human Cancers. *Cell* 182(4): 1044–1061.e18 (2020)
- 21) Jermyn, M. et al., Intraoperative Brain Cancer Detection with Raman Spectroscopy in Humans. *Science Translational Medicine*, 7(274): 274ra19 (2015)
- 22) Jones, R.R., Hooper, D.C., Zhang, L. et al. Raman Techniques: Fundamentals and Frontiers. *Nanoscale Res Lett* 14, 231 (2019).
- 23) Keklikoglou, I. et al., Chemotherapy elicits pro-metastatic extracellular vesicles in breast cancer models. *Nature Cell Biology* 21(2): 190-202 (2019)

- 24) Larkin, P. Infrared and Raman Spectroscopy: Principles and Spectral Interpretation (Elsevier, 2011)
- 25) Park, J. et al., Exosome classification by Pattern analysis of surface-enhanced Raman spectroscopy data for lung cancer. *Analytical Chemistry* 89(12): 6695–6701 (2017)
- 26) Pedregosa, F. et al., Scikit-learn: Machine Learning in Python, *JMLR* 12: 2825-2830 (2011)
- 27) Raman, C. V., Krishnan, K. S. A New Type of Secondary Radiation. *Nature*, 121(3048): 501–502 (1928)
- 28) Rojalin, T. et al., Nanoplasmonic Approaches for Sensitive Detection and Molecular Characterization of Extracellular Vesicles. *Frontiers in Chemistry* 7:279 (2019)
- 29) Shin, H. et al., Correlation between Cancerous Exosomes and Protein Markers Based on Surface-Enhanced Raman Spectroscopy (SERS) and Principal Component Analysis (PCA). *ACS Sensors* 3(12): 2637–2643 (2018)
- 30) Smith, E. and Dent, G. *Modern Raman Spectroscopy - A Practical Approach* (John Wiley and Sons, Ltd, England, 2005)
- 31) Steinbichler, T.B. et al., Therapy resistance mediated by exosomes. *Molecular Cancer* 18:58 (2019)
- 32) Théry, C. et al. Minimal information for studies of extracellular vesicles 2018 (MISEV2018): a position statement of the International Society for Extracellular Vesicles and update of the MISEV2014 guidelines. *J. Extracell. Vesicles.* 7: 1535750 (2018)
- 33) Uthamacumaran, A. A Review of Complex Systems Approaches to Cancer Networks. *Complex Systems* 29(4): 779-835 (2020)


- 34) Zhao, J. et al., Automated Autofluorescence Background Subtraction Algorithm for Biomedical Raman Spectroscopy. *Applied Spectroscopy* 61(11): 1225-1232 (2007)
- 35) Zhou, S. et al., Reprogramming Malignant Cancer Cells toward a Benign Phenotype following Exposure to Human Embryonic Stem Cell Microenvironment. *PLoS One* 12,1 e0169899 (2017)
- 36) Zlotogorski-Hurvitz, A. et al. FTIR-based spectrum of salivary exosomes coupled with computational-aided discriminating analysis in the diagnosis of oral cancer. *Journal of Cancer Research and Clinical Oncology* 145(3): 685–694 (2019)
- 37) Romana<sup>2</sup>, Sabrina; Di Giacinto, Flavio; Primiano, Aniello; Mazzini, Alberto; Panzetta, Claudia; Papi, Massimiliano; Di Gaspare, Alessandra; Ortolani, Michele; Gervasoni, Jacopo; De Spirito, Marco; Nocca, Giuseppina; Ciasca, Gabriele (2020). Fourier Transform Infrared Spectroscopy as a useful tool for the automated classification of cancer cell-derived exosomes obtained under different culture conditions. *Analytica Chimica Acta*, (), S0003267020309703–. doi:10.1016/j.aca.2020.09.037



## SUPPLEMENTARY INFORMATION




**GRAPH S1. RAMAN SPECTRA OF EVs SAMPLES I.** Representative raw Raman spectra of cancer and healthy control are shown as measured at a 514 nm excitation laser with 10% laser power, 10 accumulations for 30 seconds. EVs drops were air dried on the calcium fluoride slides prior to measurement acquisitions.




**GRAPH S2. RAMAN SPECTRA ON EVs SAMPLES II.** Single raw Raman spectra (without baseline correction) are shown to help visualize the spectral differences in the cancer EVs samples.

### PEAK FIT ASSIGNMENT

Normalized and baseline-corrected spectra. Some curves show peaks which others do not. Only the most important peaks or peak ranges with significant features were discussed in the results section characterizing the differences between the patient groups (cancer vs. healthy).




GRAPH S3. RAMAN SPECTRA OF COLORECTAL CANCER SAMPLE 426.




**GRAPH S4. RAMAN SPECTRA OF HEPATOCELLULAR CANCER SAMPLE 549.**

The following analysis applies to both the Raman spectra of 426 and 549. Three prominent peaks roughly at 930, 1062, and 1124  $\text{cm}^{-1}$ . However, they are present in both groups: control and cancer spectra. The band found within the 2750-3040  $\text{cm}^{-1}$  range denotes the lipid band. Traditionally, the EV Raman spectra can be used to measure the spectroscopic protein-to-lipid (P/L) and nucleic acid-to-lipid (NA/L) ratio by their absorption bands. The P/L and NA/L ratio were estimated by dividing the relative intensity of amide I protein band (1600–1690  $\text{cm}^{-1}$ ) and nucleic acid band (720–800  $\text{cm}^{-1}$ ) by the lipid-band (2750–3040  $\text{cm}^{-1}$ ), respectively (Gualerzi et al., 2019). However, the use of ML algorithms as shown in our study provides a more robust quantitative method to detect patterns which distinguish healthy EVs from cancer-patient derived EVs.




**GRAPH S5. RAMAN SPECTRA OF BREAST CANCER SAMPLE.**


Single peak at 2850-3010  $\text{cm}^{-1}$  denotes the lipid band (C-H stretches). However, there appears to be/there is no obvious difference amidst the cancer vs. healthy groups. A peak at 720  $\text{cm}^{-1}$  may correspond to the presence of nucleic acids. No major differences are observed between the cancer and healthy groups suggesting further the premise that traditional approaches of peak-molecular bond assignment are insufficient in the study of such complex systems. The same analysis applies to pancreatic cancer (160517) spectra.



**GRAPH S6. RAMAN SPECTRA OF PANCREATIC CANCER SAMPLE.**





**GRAPH S7. FTIR SPECTRA OF 426.** Some characteristic peaks are found at:  $1120\text{ cm}^{-1}$  and at  $3070\text{ cm}^{-1}$  (the peaks are only present in the cancer spectra). There are additional peaks such as within the range of  $1225\text{-}1450\text{ cm}^{-1}$  and one at  $1740\text{ cm}^{-1}$  which were stronger (higher in absorption intensity) only in the cancer spectra.




**GRAPH S8. FTIR SPECTRA OF 549.** The peak at 3070 cm<sup>-1</sup> is only present in cancer spectra. The peak at 980 cm<sup>-1</sup> is stronger in intensity in cancer spectra. Whether these distinct peaks are identified by the ML algorithms for the classification of cancer vs. healthy EVs remains unknown.





**GRAPH S9. ROC CURVE FOR MULTILAYER PERCEPTRON(MLP).** As discussed, the simple neural network classifier MLP had poorer performance compared to the classifiers discussed in the report above. Commonly researchers associate neural networks to optimal machine learning performance compared to simpler algorithms like SVM or Random Forests. However, these findings demonstrate this is not necessarily the case when dealing with smaller sample sizes. Neural networks are better suited for larger datasets, while smaller patient samples as in our study could benefit from simpler data mining algorithms. Furthermore, there remains a black box problem of how and which type of ML algorithms are better suited for a given dataset. Our findings clearly indicate decision trees/random forests and SVM are better in predicting vibrational spectra amidst other data mining algorithms such as the MLP or linear regression.



**GRAPH S10. SERS RAMAN SAMPLE SPECTRA.**

The raw SERS (surface enhanced Raman Spectroscopy) spectra on an EVs sample and PBS buffer at 50 X objective are shown. The SERS signal was acquired at 785 nm on an Au (gold) nanoparticle substrate (manufactured by Ocean Optics, Inc). Further runs were not followed since the gold-substrate consisted of a non-uniform paper. A purely homogeneous gold substrate is required to proceed with the SERS method. As shown in the data by Shin et al. (2018), the SERS spectra would provide detailed spectra and hence, allow the ML classifiers to better distinguish the cancer and healthy patient samples. The resolution of the Raman peaks are enhanced many-fold by the SERS technique in comparison to those reported in our study (on calcium fluoride slides). Despite these advantages the method with CaF<sub>2</sub> slides is cheaper and efficient as demonstrated by our findings. The demonstrated statistical results from the ML predictions indicate they are reasonably good in distinguishing the two patient groups, even when presented with a different set of heterogeneous (patient- derived) EVs samples.