
CLASSIFICATION OF ANOMALOUS GAIT USING MACHINE LEARNING TECHNIQUES AND EMBEDDED SENSORS

A PREPRINT

Tiago R. de Sá
Electrical Engineering
University of State of Amazonas
Manaus, AM
trs.eng17@uea.edu.br

Carlos M. S. Figueiredo
PhD, Computer Science
University of State of Amazonas
Manaus, AM
cfigueiredo@uea.edu.br

October 13, 2021

ABSTRACT

Human gait can be a predictive factor for detecting pathologies that affect human locomotion according to studies. In addition, it is known that a high investment is demanded in order to raise a traditional clinical infrastructure able to provide human gait examinations, making them unaffordable for economically vulnerable patients. In face of this scenario, this work proposes an accessible and modern solution composed of a wearable device, to acquire 3D-accelerometer and 3D-gyroscope measurements, and machine learning techniques to classify between distinct categories of induced gait disorders. In order to develop the proposed research, it was created a dataset with the target label being 4 distinct and balanced categories of anomalous gait. The machine learning techniques that achieved the best performances (in terms of accuracy) in this dataset were through the application of Principal Component Analysis algorithm following of a Support Vector Machines classifier (94 %). Further, an architecture based on a Feedforward Neural Network yielded even better results (96 %). Finally, it is also presented computational performance comparison between the models implemented.

Keywords Human Gait · Machine Learning · Internet Of Things · Wearables · Healthcare

1 Introduction

Gait analysis is the systematic study of human locomotion [1, 2]. Thereby, from the point of view of clinical observation, it is impossible to analyze and quantify its degree of deviation from normality without adequate instruments. So, gait analysis, using clinical instruments available in a motion analysis laboratory, provides relevant data for the treatment of pathologies that affect human locomotion [3]. Consequently a high investment is requested in a traditional clinical infrastructure. In parallel to this context, with the continuous advancement of the scientific community regarding Internet of Things and modern techniques of Digital Signal Processing based on Machine Learning approaches, excellent progress has been seen in applications in the healthcare domain [4, 5, 6, 7], especially with regard to the analysis of the gait of individuals based on data provided by low-cost embedded sensors such as accelerometer and gyroscope [8, 9, 10, 11].

The implantation of a motion analysis laboratory demands a high investment with state-of-the-art technologies for a complete analysis of the parameters that characterize human gait, making clinical exams less affordable. In view of the above, according to studies [12, 13, 4, 14, 15], the main components needed in a traditional motion analysis laboratory are: a video system for capturing movement through markers [16, 3]; equipment system for electromyographic analysis; equipment system for static and dynamic capture of plantar pressure data; software for analyzing three-dimensional motion and, finally, physical space. Thereby, the investment to raise such infrastructure is not less than a few tens of millions of dollars [17]. In this context, more economically viable alternatives have been developed to capture the data needed to analyze human gait [2].

Among the alternatives above, the use of wearable sensors such as accelerometer [18] and gyroscope stands out [19], as they allow a three-dimensional kinematic analysis of human movement [20]. In this scenario, briefly, the process for identifying irregularities in a patient's gait is conditioned to the processing of signals data obtained from wearable sensors. Given the processed data, a possible approach is the application of machine learning techniques for feature extraction [21]. Then, with a trained model, it is possible to make inferences about the parameters of a patient's gait [22]. Finally, according to studies [9, 23], among the most common methods for movements patterns classification, stands out Support Vector Machine, Random Forest Classifiers, Multi-Layer Perceptrons and Convolutional Neural Networks.

To exemplify one of the possible clinical scenarios where the analysis of human gait parameters is relevant, it is considered the case of patients with diabetes mellitus (DM). Recent studies reveal that such patients tend to have irregularities in their gait due to a chronic complication known as diabetic peripheral neuropathy from DM [24, 25, 26]. Thus, early detection of this anomaly would prevent future complications in the diabetic patient's condition, such as the reduction in the risk of falls and the consequent reduction in the appearance of ulcers.

Considering the aspects mentioned above, this work seeks (i) to contribute to the creation of a human gait dataset through the implementation of a low-cost application composed of hardware and software, and (ii) to present comparative assessments between different machine learning models applied to this dataset.

In view of these proposed objectives, section 2 presents the related works. Section 3 describes about each component of the proposed application. Section 4, in turn, presents the results obtained as well as performance comparisons for each proposed technique. Finally, section 5 presents the final considerations and suggestions for future work.

2 Related Works

The number of scientific contributions in understanding the parameters that characterize human gait through wearable embedded sensors and Machine Learning techniques has been increasing. Following is presented some related scientific contributions in this scenario.

In [27], it was developed a data collection system composed of an embedded wearable hardware and a software implemented in Java. Then, plantar pressure data of the reference foot and kinematic data were collected to measure the orientations of the reference leg segment. Regarding the system software, it was used to add additional information such as age, weight and height to the database.

Given the work above, to estimate the parameters that characterized the patient's gait, it was implemented a model based on a stacked Long-Term Memory Network (LSTM). Through this model, it was possible to demonstrate how the behavior of a normal gait could be detected. For this detection, the characteristics of error predictions in normal gaits compared with pathological simulated gaits were considered.

In [1], on the other hand, they described how non-invasive wearable sensors could be used in combination with Deep Learning techniques to classify artificially induced gait changes. For this purpose, the plantar pressure of the foot was measured in 12 participants, who each underwent 8 artificially induced gait changes by modifying the lower part of the shoe. Data was recorded at 100 Hz on 2520 data channels. After data collection, these were used to train a model also based on Deep Learning techniques through a stacked Long-Term Memory Network (LSTM).

Many of the related works describe advantages and disadvantages between important materials and methods for building a database. For example, in [3], they presented a comparative view of current quantitative measurement instruments for gait analysis. Thus, it was possible to observe the advantages of using wearable embedded sensors for the following reasons: low computational cost, low financial cost, practicality of use, satisfactory accuracy and precision. In all these points, the choice for the analysis system based on wearable sensors stood out compared to the conventional system as they require an infrastructure that is costly from a financial and computational point of view.

The proposed work, unlike the related works presented, stands out for the following reasons:

- Implementation of a low-cost wearable device to collect kinematic data;
- Creation of a balanced gait dataset based on only two measurement units: 3D-accelerometer and 3D-gyroscope;
- Presentation of performance comparisons among different machine learning and deep learning approaches.

Next, the methods and materials used in this work will be described.

3 Material and Methods

The Fig. 1 below provides an overview of all components that will be part of the development of this work: (1) creation of the dataset and (2) implementation and comparison between machine learning models.




Figure 1: Components presents in the workflow of the proposed methodology

3.1 Gait Dataset

The steps to create the dataset were: (1) definition of collection parameters given the human gait context, (2) implementation of the embedded application and (3) implementation of the web application. The specifications and responsibilities of each step above are described in detail below.

3.1.1 Dataset Description

For proof of concept of the proposed work, 840 instances of human gait data were collected among 20 volunteers. In a balanced way, data were distributed between 4 distinct categories.

The 4 distinct categories of human gait were considered based on four different artificial modes of walking, namely, (i) walking, (ii) marching, (iii) limping and (iv) feint movements. These 4 categories were used as dependent variables to be predicted, namely, target labels.

In order to illustrate the use of the categories above when collecting data from the volunteer gait, it is asked to him to walk by simulating each of the category above. Thus, considering the first category, walking (i), the participant is instructed to walk at his common pace. Yet, considering the second category, march (ii), he is instructed to imitate, as closely as possible, the military march. The same procedure is followed for the other two remaining categories.

In all categories defined above, the following simulation criteria were pre-established: a flat surface, collection period equal to 5 (five) seconds and the reference joint being the right leg, in the ankle region. Furthermore, the data content consisted of two vector physical quantities - the three-dimensional linear acceleration and the three-dimensional angular rotation.

All participants who voluntarily contributed to the construction of the dataset were healthy, without any pathology related to locomotion.

3.1.2 Embedded Application

The Fig. 2 illustrates the prototype of the proposed embedded device for data collection. This consists of a microcontroller (Fig. 2-b), a sensor (Fig. 2-a) and a 3D support case to fit the electrical circuit and use it in the region near the ankle (Fig. 3).

The choice of the microcontroller was made based on a survey of requirements given the context of the proposed work. This should contain, at a minimum, UART and I2C communication protocols, wireless network connectivity, embedded memory, low-power processors, in addition to input/output pins.

Thus, with regard to the UART (Universal Asynchronous Receiver/Transmitter) protocol, this would be responsible for enabling serial communication aiming at loading the embedded software developed on a host platform to the target platform, in addition to enabling the debug of the embedded program. Still, the I2C protocol would be responsible for initializing and reading the motion sensor data through a master and slave relationship. At the same time, wireless




Figure 2: Prototype of the embedded wearable device. Figure a) represents the movement sensor MPU6050, b) illustrates the microcontroller ESP32 DevKit C and c) represents the power supply.

network connectivity would make it possible to send data collected from an embedded device to an external server. Embedded memory, in turn, would provide the space to store the application code and others. Finally, the input/output pins in a minimum quantity sufficient for the peripheral connection of the kinematic sensor.

In order to meet the specifications listed above, it was opted for the ESP32 DevKit C development board as its system-on-a-chip (SoC) meets all the requirements defined above.

The motion sensor chosen was the MPU-6050 module. This enabled the collection of three-dimensional data on physical quantities such as linear acceleration and angular rotation. Then, to enable communication between the sensor and the microcontroller, the `i2cdevlib` library, implemented by Jeff Rowberg, was used. Through this library, the sensor sampling frequency was defined at 100 Hz. From this definition, it was possible to collect 500 samples in each of the three-dimensional axes. That is, for each axis of linear acceleration and angular rotation, 500 samples were simultaneously collected, providing a total of 3000 (6 x 500) samples per file. In summary, these samples represented the behavior of human gait in different categories in terms of linear acceleration and angular rotation discretized in time and amplitude.

Finally, 3D case support was developed through TinkerCad software. The support was designed in a mold so that it had dimensions compatible with the designed hardware, in addition to having the characteristic of making the device wearable on the ankle, as shown in Figs. 2 and 3.




Figure 3: Prototype of the embedded wearable application fixed with velcro tape close to the ankle region.

As will be seen below, after the data is collected by the embedded device, it is published in a broker server through mqtt protocol in order to automate the dataset labeling process.

3.1.3 Web Application

The proposed web application had two main purposes: labeling the data and then storing them in csv (comma-separated-values) format on the server's hard disk. The following figure illustrates one of the interfaces developed for the web application through the Node-RED framework.

Figure 4: Implemented interface for inserting data in the supervised dataset.

Regarding the frontend, two interfaces were implemented. The first interface, as illustrated in Fig. 4, allows the user to enter the following information: name, age, height, weight and the gait category that will be artificially reproduced afterwards. For the second interface, a connection test between the web application and the embedded device was implemented before starting the induced gait reproduction process.

Regarding the backend, on the other hand, the information entered by the user was used to label the file that was generated after the end of the simulation. After the labeling step, the file is properly saved according to the reproduced gait category.

Finally, the communication between the embedded application and the web application was established through the MQTT (Message Queuing Telemetry Transport) protocol and the Eclipse Mosquitto broker on a Linux server.

3.2 Machine Learning Models

A total of 4 models were implemented aiming to provide insights about performance evaluation considering the dataset created. Thereby, the goal of each model is the same: classify a given instance in one of the 4 established classes. In all cases, the models were trained based on the workflow illustrated by Fig. 5 with the split being 20% to the test dataset and the remaining data (80 %) to the train dataset.

In summary, 2 models were trained with different Machine Learning algorithms - namely, Support Vector Machine (SVM) and Random Forests (RF). In addition, a model based on a Feedforward Neural Network (FNN) is presented as well as a deep learning model based on a Convolutional Neural Network (CNN), totaling 4 models. The table 1 provides an overview of the parameters for each chosen model.

Table 1: Summary of training configurations used for models based on Machine Learning and Deep Learning techniques.

Algorithm	Input	Preprocessing	Validation	Tuning
SVM	Raw data (3D-accelerometer and 3D-gyroscope)	Normalization and dimensionality reduction	K-Folder	GridSearchCV
RF	Raw data (3D-accelerometer and 3D-gyroscope)	Normalization and dimensionality reduction	K-Folder	GridSearchCV
FNN	Raw data (3D-accelerometer and 3D-gyroscope)	Normalization and dimensionality reduction	Holdout	Adjustment by trial and error
CNN	Images (espectograms)	Data conversion to ".wav" audio format following of DTFT application	Holdout	Adjustment by trial and error




Figure 5: Defined steps for implementing the models

The hyperparameters (number of hidden layers, perceptrons and activation functions) of the model trained based on a FNN architecture were adjusted by trial and error. The table 1 summarizes the main definitions and procedures performed.

The following hyperparameters were chosen for the FNN architecture:

- Input Layer: 60 neurons with ReLU function activation;
- Hidden Layer 1: 2048 neurons with ReLU function activation;
- Hidden Layer 2: 1024 neurons with ReLU function activation;
- Output Layer: 4 neurons with softmax function.

As is illustrated by table 1, the model trained based on a CNN algorithm had an input considerably different. It performed predictions based on images of spectrograms. They represent a snapshot of 3D-acceleration and 3D-rotation data acquired over the period considered. Fig. 6 illustrates the CNN hyperparameters and spectrograms been used as input to the model.




Figure 6: Neural Convolutional Network implemented using spectrograms.

In order to evaluate this model, it was created a dataset containing spectrograms instead of raw data. Then, to generate them, the following preprocessing steps were executed:

- Concatenation of 3D-acceleration and 3D-rotation raw data for each dataset instance;
- Creation of audios files .wav (main lib used to this: librosa) from the concatenation above;
- Creation of spectrograms by applying discrete-time Fourier transform (DTFT) on each audio file created;

Finally, a total of 840 spectrograms were generated, balanced between the 4 distinct categories defined previously.

Next section presents the results obtained from the models defined above.

4 Results


After building the dataset, efforts were focused on evaluating the models performance. The metrics chosen were accuracy, precision, recall and F1 score. In addition, it were also assessed the computational performance in terms of time required to perform a predict task for each model. The results are presented below.

4.1 Model 1 - Support Vector Machine (SVM)

Among the classic machine learning techniques used, the model based on SVM algorithm yielded the best performance in terms of the evaluation metrics considered. The following hyperparameters were used after applying grid search method::

- Decision function shape: One vs one;
- Regularization parameter: 100;
- Kernel coefficient: 0.01;
- Kernel type: radio basis function.

Fig. 7 illustrates the learning curve for training the model based on the SVM classifier and presents the confusion matrix results.



(a) Model 1 learning curve.

(b) Model 1 confusion matrix.

Figure 7: Training results for the model based on the SVM classifier.

The confusion matrix presented in the Fig 7b shows that some gaits reproduced in category MM (limping gait) are misclassified as MN (normal gait) and vice versa. This observation is coherent from the point of view of the empirical analysis of the reproduced movement, since, in few situations, during the dataset creation process, the difference between a normal gait and a limp gait was not evident.


Finally, for the learning curve of the model based on the SVM algorithm (Fig. 7a), it is possible to infer that the increasing in the number of samples (horizontal axis) does not tend to cause significant increases in accuracy metric (vertical axis). In this way, the cross-validation accuracy is numerically close to its maximum reachable value.

4.2 Model 2 - Random Forest (RF)

The model trained based on Random Forest technique had the advantage of using ensemble methods. In that way, results are yielded by averaging multiple decision tree predictions and, consequently, the variance is reduced. The following hyperparameters were used after applying grid search method:

- Number of trees in the forest: 100;
- Criteria to measure the quality of a split: Gini;
- Minimum number of samples required to split an internal node: 2;
- The minimum number of samples required to be at a leaf node: 1.

Fig. 8 illustrates the learning curve for training the model based on the RF classifier and presents the confusion matrix results.



(a) Model 2 learning curve.

(b) Model 2 confusion matrix.

Figure 8: Training results for the model based on the RF classifier.

The confusion matrix presented in the Fig 8b shows that some gaits reproduced in category MM (limping gait) are also missclassified as MN (normal gait) and vice versa. In addition, missclassifications are produced between MN and MS (military gait).


Finally, considering the Fig. 8a, it is possible to infer that the increasing in the number of samples could yield even better and significant results.

4.3 Model 3 - Feedforward Neural Network (FNN)

So far, models have all been trained with classic Machine Learning techniques. For model 3, however, an architecture based on a Feedforward Neural Network was used (architecture described in section 3.2). Thus, for the definition of the architecture, several tests and manual enhancements were made in the hyperparameters until the best results were reached. Next, Fig. 9 illustrates the learning curve and confusion matrix of model 3.

The learning curve of model 3 reveals that the increasing in the number of training periods does not tend to cause relevant changes in the validation accuracy. However, increasing the number of samples can make the learning curve more consistent and less oscillatory.

In summary, in all the confusion matrices presented, it is observed that the category "MM" (limping gait) had the highest number of instances missclassified.




(a) Model 3 learning curve.

(b) Model 3 confusion matrix.

Figure 9: Training results for the model based on FNN architecture.

4.4 Model 4 - Convolutional Neural Network (CNN)

The learning curve of the trained model according to the architecture illustrated in Fig. 6 for spectrogram classification is shown in Fig. 10.



(a) Model 4 learning curve.

(b) Model 4 confusion matrix.

Figure 10: Training results for the model based on CNN architecture.

The results achieved with this approach were not satisfactory compared to the other applied techniques. Despite this, future studies may be developed in order to reach a conclusive result on the use of Convolutional Neural Networks to classify spectrograms from human gait data.

Finally, the table 2 presents the results generated from each of the models trained. The results are evaluated in terms of 4 following metrics: F1 score, precision, recall and accuracy

The results summarized in the table above shows that the model that achieved the best performance, in terms of accuracy, was the model 3.

Table 2: Summary of the evaluation metrics obtained from the models above.

	Precision	Recall	F1 Score	Accuracy
FNN	0.94	0.94	0.93	0.96
SVM	0.94	0.93	0.93	0.94
RF	0.90	0.90	0.90	0.90
CNN	0.70	0.72	0.69	0.77

4.5 Computational performance comparison between the models

The comparative evaluation of the models in terms of prediction time was carried out with the test set (168 instances). In this way, the mean time (in milliseconds) and standard deviation required by the prediction task on the test set was performed. This procedure was executed for each model 20 times. Finally, this evaluation was performed on a machine with the following configurations:

- GPU: NVIDIA GeForce RTX 2060
- RAM: 16 GB
- SSD: 512 GB
- Processor: Intel Core i7 10th 2.6 GHz

Table 3: Computational performance comparison between the models 1, 2, 3 and 4 in terms of milliseconds with standard deviation.

Models	Prediction time over the test set in milliseconds
Model 1 (SVM)	4.18 ± 0.35
Model 2 (RF)	8.05 ± 0.72
Model 3 (FNN)	28.48 ± 14.60
Model 4 (CNN)	702.73 ± 12.19

5 Conclusion

The application presented in this work contributed to the implementation of a low-cost wearable device to collect kinematic data from human gait. The data were used in the creation of a balanced human gait dataset in 4 distinct categories, composed by only two measurement units: 3D-acceleration and 3D-rotation. Then, as seen, it was presented performance comparisons among different machine learning and deep learning approaches. The models that achieved the best results were model 1 (SVM - 94 %) and model 2 (FNN - 96 %). Furthermore, although the second model had yielded better results than model 1, its computational cost, in terms of prediction time, is much higher than this model. Finally, through these satisfactory results, future research can be carried out using the proposed work as a starting point.

References

- [1] Alexander Turner and Stephen Hayes. The classification of minor gait alterations using wearable sensors and deep learning. *IEEE Transactions on Biomedical Engineering*, 66(11):3136–3145, 2019.
- [2] W. Tao, T. Liu, R. Zheng, and H. Feng. Gait analysis using wearable sensors. *Sensors (Basel)*, 12:2255–83, 2012.
- [3] Chen S., Lach J., Lo B., and Yang GZ. Pervasive gait analysis with wearable sensors: A systematic review. *IEEE journal of biomedical and health informatics*, 20(6):1521–1537, 2016.
- [4] M Jamal Deen Sumit Majumder, Tapas Mondal. Wearable sensors for remote health monitoring. *Sensors (Basel)*, 17:130, 2017.
- [5] Arijit Ukil, Soma Bandyopadhyay, Chetanya Puri, and Arpan Pal. Iot healthcare analytics: The importance of anomaly detection. *2016 IEEE 30th International Conference on Advanced Information Networking and Applications (AINA)*, pages 994–997, 2016.
- [6] Daniela Tarniță. Wearable sensors used for human gait analysis. *Romanian journal of morphology and embryology = Revue roumaine de morphologie et embryologie*, 57(2):373–382, 2016.

- [7] Djordje Slijepcevic, Fabian Horst, Sebastian Lapuschkin, Anna-Maria Raberger, Matthias Zeppelzauer, Wojciech Samek, Christian Breiteneder, Wolfgang I. Schöllhorn, and Brian Horsak. On the explanation of machine learning predictions in clinical gait analysis. *arXiv*, 2020.
- [8] Sasanka Potluri, Arvind Beerjapalli Chandran, Christian Diedrich, and Lutz Schega. Machine learning based human gait segmentation with wearable sensor platform. *2019 41st Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*, pages 588–594, 2019.
- [9] Igor Bisio, Alessandro Delfino, Fabio Lavagetto, and Andrea Sciarrone. Enabling iot for in-home rehabilitation: Accelerometer signals classification methods for activity and movement recognition. *IEEE Internet of Things Journal*, 4(1):135–146, 2017.
- [10] P. Catalfamo, S. Ghoussayni, and D. Erwins. Gait event detection on level ground and incline walking using a rate gyroscope. *Sensors (Basel)*, 10:5683–702, 2010.
- [11] Ryo Takeda, Shigeru Tadano, Akiko Natorigawa, Masahiro Todoh, and Satoshi Yoshinari. Gait posture estimation using wearable acceleration and gyro sensors. *Journal of Biomechanics*, 42(15):2486–2494, 2009.
- [12] Samaneh Davarzani, David Saucier, Preston Peranich, Will Carroll, Alana Turner, Erin Parker, Carver Middleton, Phuoc Nguyen, Preston Robertson, Brian Smith, John Ball, Reuben Burch, Harish Chander, Adam Knight, Raj Prabhu, and Tony Luczak. Closing the wearable gap—part vi: Human gait recognition using deep learning methodologies. *Electronics*, 9(5), 2020.
- [13] Julius Hannink, Thomas Kautz, Cristian F. Pasluosta, Karl-Günter Gaßmann, Jochen Klucken, and Bjoern M. Eskofier. Sensor-based gait parameter extraction with deep convolutional neural networks. *IEEE Journal of Biomedical and Health Informatics*, 21(1):85–93, 2017.
- [14] A. Muro-de-la Herran, B. Garcia-Zapirain, and A. & Mendez-Zorrilla. Gait analysis methods: an overview of wearable and non-wearable systems, highlighting clinical applications. *Sensors (Basel)*, 14:2, 2014.
- [15] Daniel Tik-Pui Fong and Yue-Yan Chan. The use of wearable inertial motion sensors in human lower limb biomechanics studies: a systematic review. *Sensors (Basel)*, 10:11556–65, 2010.
- [16] David Saucier, Samaneh Davarzani, Alana Turner, Tony Luczak, Phuoc Nguyen, Will Carroll, Reuben F. Burch V, John E. Ball, Brian K. Smith, Harish Chander, Adam Knight, and Raj. K. Prabhu. Closing the wearable gap—part iv: 3d motion capture cameras versus soft robotic sensors comparison of gait movement assessment. *Electronics*, 8(12), 2019.
- [17] A. F. F. D. Silva. Projecto de uma clínica de análise biomecânica do movimento. *Faculdade de Engenharia da Universidade do Porto*, pages 1–55, 2011.
- [18] Spyroula Masiala, Willem Huijbers, and Martin Atzmueller. Feature-set-engineering for detecting freezing of gait in parkinson’s disease using deep recurrent neural networks. *arXiv*, 2019.
- [19] Szilárd Nemes and Margit Antal. Feature learning for accelerometer based gait recognition. *arXiv*, 2020.
- [20] A. Muro-de-la Herran, B. Garcia-Zapirain, and A. & Mendez-Zorrilla. Assessing gait in parkinson’s disease using wearable motion sensors: A systematic review. *Diseases (Basel, Switzerland)*, 7:1–18, 2019.
- [21] Imanne El Maachi, Guillaume-Alexandre Bilodeau, and Wassim Bouachir. Deep 1d-convnet for accurate parkinson disease detection and severity prediction from gait. *Expert Systems with Applications*, 143:113075, Apr 2020.
- [22] S. Pardoel, J. Kofman, J. Nantel, and E. D. & Lemaire. Wearable-sensor-based detection and prediction of freezing of gait in parkinson’s disease: A review. *Diseases (Basel, Switzerland)*, 19:5141, 2019.
- [23] Johannes Burdack, Fabian Horst, Sven Giesselbach, Ibrahim Hassan, Sabrina Daffner, and Wolfgang I. Schöllhorn. Systematic comparison of the influence of different data preprocessing methods on the performance of gait classifications using machine learning. *Frontiers in Bioengineering and Biotechnology*, 8, Apr 2020.
- [24] A. R Martinelli and et al. Alterações dos parâmetros da marcha e déficit sensorio-motor associado à neuropatia diabética periférica. *Acta Fisiátrica*, pages 36–40, 2014.
- [25] V. Saura and et al. Fatores preditivos da marcha em pacientes diabéticos neuropáticos e não neuropáticos. *Acta Ortop Bras*, pages 148–151, 2010.
- [26] C. E. P. T. Fregonesi and M. R. D. Camargo. Parâmetros da marcha em portadores de diabetes mellitus. *Rev. bras. cineantropom. desempenho hum.*, pages 155–163, 2010.
- [27] Sasanka Potluri, Srinivas Ravuri, Christian Diedrich, and Lutz Schega. Deep learning based gait abnormality detection using wearable sensor system. *2019 41st Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*, pages 3613–3619, 2019.