

Method for Measuring Absolute Optical Properties of Turbid Samples in a Standard Cuvette

Giles Blaney *¹, Angelo Sassaroli ¹ and Sergio Fantini ¹

Department of Biomedical Engineering, Tufts University, 4 Colby Street, Medford, MA 02155, USA;
giles.blaney@tufts.edu (G.B.); angelo.sassaroli@tufts.edu (A.S.); sergio.fantini@tufts.edu (S.F.)

* Correspondence: giles.blaney@tufts.edu

Featured Application: Quantitative and calibration free determination of the absolute optical properties of turbid samples in a standard cuvette with milliliter scale volume.

Abstract: Many applications seek to measure a sample's absorption coefficient spectrum to retrieve the chemical makeup. Many real-world samples are optically turbid, causing scattering confounds which many commercial spectrometers cannot address. Using diffusion theory and considering absorption and reduced scattering coefficients on the order of 0.01 mm^{-1} and 1 mm^{-1} , respectively, we develop a method which utilizes frequency-domain to measure absolute optical properties of turbid samples in a standard cuvette ($45 \text{ mm} \times 10 \text{ mm} \times 10 \text{ mm}$). Inspired by the self-calibrating method, which removes instrumental confounds, the method uses measurements of the diffuse complex transmittance at two sets of two different source-detector distances. We find: this works best for highly scattering samples (reduced scattering coefficient above 1 mm^{-1}); higher relative error in the absorption coefficient compared to the reduced scattering coefficient; accuracy is tied to knowledge of the sample's index of refraction. Noise simulations with 0.1 % amplitude and $0.1^\circ = 1.7 \text{ mrad}$ phase uncertainty find errors in absorption and reduced scattering coefficients of 4 % and 1 %, respectively. We expect that higher error in the absorption coefficient can be alleviated with highly scattering samples and that boundary condition confounds may be suppressed by designing a cuvette with high index of refraction. Further work will investigate implementation and reproducibility.

Keywords: absolute optical properties; absorption coefficient; reduced scattering coefficient; diffusion theory; turbid samples; optical spectroscopy; sample measurement; cuvette; frequency-domain near-infrared spectroscopy; self-calibration

1. Introduction

Samples in which the propagation of light is dominated by random scattering are considered optically diffuse. Such samples can be characterized by two absolute optical properties, the absorption coefficient (μ_a) and the reduced scattering coefficient (μ'_s) [1]. The μ_a represents chemical information, and its spectral measurement allows for determination of the sample's chemical constituents and concentrations. Meanwhile, the μ'_s describes the micrometer scale structure of diffuse samples. However, in many applications μ'_s is considered a confounding parameter, since measurement of μ_a and chemical makeup is often the end goal. For this reason, even when diffuse sample measurement of only μ_a is sought, μ'_s must also be determined since it significantly impacts the behavior of light and thus the recovered μ_a .

Applications that seek to measure these diffuse optical properties are numerous and span many fields. For example, applications include those within biomedical research and clinical applications [1–3], of food science and quality [4–6], concerning pharmaceutical metrology [7,8], pertaining to art and archaeology [9,10], and within dendrology [11] to name a few. In all cases, one has two options:

1. To make a measurement which retrieves the total attenuation coefficient (μ_t) or the effective attenuation coefficient (μ_{eff}).
2. To make a measurement that can separate both μ_a and μ'_s .

However, only option 2 allows for careful quantitative analysis of the sample properties; since in option 1 one can only measure a coefficient, namely μ_t or μ_{eff} , that couples both the μ_a and μ'_s of the sample. There are few methods capable of achieving option 2. One such technique is Near-InfraRed Spectroscopy (NIRS) implemented in Frequency-Domain (FD) [12] (or Time-Domain (TD) [13]) which can recover μ_a and μ'_s by using temporally modulated light. In the case of FD, photon density waves are generated by using a sinusoidally modulated source on the order of 100 MHz, and the amplitude and phase of these photon density waves are measured from the detected modulated light signal. An example of a commercially available FD instrument capable of this measurement is the ISS Imagent V2 [Champaign, IL USA] (Imagent). A second technique capable of option 2 is the integrating sphere [14,15]. This technique measures total diffuse reflectance and total diffuse transmittance to separate μ_a and μ'_s . Both techniques have their strengths and weaknesses. Common implementations of FD NIRS require large sample volumes to create geometries that are effectively infinite in at least one dimensional extent making simple diffusion theory expressions valid [16]. Other methods, besides diffusion theory, exist to tackle non-simple geometries such as Monte Carlo [17] and the radiative transport equation implemented with higher order spherical harmonics [18–20]. However these methods are computationally costly compared to diffusion theory and would likely be impractical to implement on a broad range of optical wavelengths (λ s). Another weakness of FD regarding the number of λ s, is the typical implementation at only discrete λ s such as with the Imagent. Meanwhile, the integrating sphere requires careful calibration or a reference sample and is easily susceptible to errors induced by the measurement technique (for example, light loss causing an incorrect measurement of total reflectance and transmittance).

Due to these difficulties with option 2 (and relative ease implementing option 1), most commercial spectrometers make a measurement that is based on the retrieving non-diffuse transmittance. Therefore, for quantitative determination of a sample's chemical concentrations (through the μ_a spectrum), samples must be non-scattering or transparent; either innately or through some chemical washing. Whenever this is not possible the measurement will be confounded by scattered light. This implies that the μ_a will be overestimated and its spectral dependence distorted leading to errors in the estimation of a sample's chemical constituents.

One such instrument that shines when samples are transparent and non-scattering is the Perkin Elmer LAMBDA 365+ (Waltham, MA USA), this and instruments like it are the workhorses of many chemical and biological laboratories. However, when diffuse sample measurement is necessary (and quantitative measurement of properties sought), one of the aforementioned techniques capable of option 2 is required. One example is the Gigahertz Optik SphereSpectro 150H (Türkenfeld, Germany) (SphereSpectro), an instrument directly designed for spectroscopic measurement of both μ_a and μ'_s via integrating sphere. Furthermore, integrating spheres may be purchased as attachments to traditional spectrometers, thus adding diffuse functionality. One such example of a spectrometer that has this option is the Perkin Elmer LAMBDA 1050+ (Waltham, MA USA). However, we are not aware of any commercially available instrument that utilizes the FD in such applications.

Because of the apparent gap in the market for instruments which complete diffuse measurement of μ_a and μ'_s , namely implementation with instruments that utilize FD NIRS like techniques, we will focus closer on FD. Measurements of μ_a and μ'_s with temporally modulated light, such as FD NIRS, is actually rather common but typically only in the research setting (using the Imagent for example). However, we know of no FD instruments designed for the sample sizes and form factors of traditional spectrometers which accept a cuvette. In-fact FD NIRS methods typically require large sample volumes on the scale of liters to implement

simple diffusion theory solutions. Two examples of work that considered FD measurements in confined regions were that in the slab [21] or block [22] (which utilized the same diffusion theory model implemented here [23]), however the geometries considered in these works were rather large compared to a cuvette. There are advantages of FD NIRS which would lead one to seek or design and manufacture such an instrument. For example, FD NIRS can leverage existing techniques which eliminate the need for instrumental calibration such as the Self-Calibrating (SC) method [24]. Additionally, despite FD typically being implemented at discrete λ , methods exist to achieve broadband μ_a measurement. This may be done by combining measurements at discrete λ s in FD (or TD) with measurements at broadband λ s with Continuous-Wave (CW) [25–27], implemented with SC and a method dubbed Dual-Slope (DS), respectively, [28–30]. This leverages the fact that SC and DS both rely on a difference type measurement which is capable of subtracting away instrumental confounds.

We see an opportunity to develop a method which leverages the tools available in FD NIRS to measure absolute μ_a and μ'_s in a standard cuvette with milliliter scale volume in an attempt to compete with the existing integrating sphere type devices. Therefore, in this work we present a method that utilizes FD NIRS in a small geometry the size of a standard cuvette (45 mm \times 10 mm \times 10 mm). Our proposal relies on the SC/DS method to remove a majority of the instrumental confounds. To our knowledge no work has yet leveraged SC/DS directly on the cuvette geometry as we propose here. First, we utilize a seldom implemented but still computationally inexpensive diffusion theory derived expression for the box geometry [23] to model our proposed measurement and determine the method's feasibility in theory. Then we further develop ways to retrieve μ_a and μ'_s from the proposed measurement. Lastly, we determine the strengths and weaknesses of the proposed method. Our end goal is to implement the method for broadband λ measurement of μ_a [28], thus computation cost and model simplicity are of importance leading to the choice of a diffusion theory model. In this article we focus only on the FD part since the extension to broadband λ CW will utilize all the same theory.

2. Methods

2.1. Geometry

In this work, we consider a box geometry with the dimensions of a standard cuvette (45 mm \times 10 mm \times 10 mm; Figure 1). A DS/SC arrangement (the word slope in Dual-Slope (DS) is historical [29,30] as no slopes are actually considered in this work) is achieved by placing 2 sources (1 & 2; Figure 1a,b) and 2 detectors (A & B; Figure 1b,c) symmetrically on opposing sides of the cuvette. Using the coordinate system shown in Figure 1, the optodes were considered at the following position vectors (\vec{r} s): $\vec{r}_1 = -17\hat{x}$ mm, $\vec{r}_2 = 17\hat{x}$ mm, $\vec{r}_A = -6\hat{x} + 10\hat{z}$ mm, and $\vec{r}_B = 6\hat{x} + 10\hat{z}$ mm. This forms two possible source-detector distances (ρ s) of 14.9 mm and 25.1 mm (2 each), for 1A & 2B and 2A & 1B, respectively.

2.2. Types of Measurement

The signal obtained between a single temporally modulated source and a single detector recovers the Green's function for the complex Transmittance (\tilde{T}) with FD NIRS. \tilde{T} is a complex number to represent the amplitude and phase of the transmitted photon density waves modulated on the order of 100 MHz. These signals are named: \tilde{T}_{1A} , \tilde{T}_{1B} , \tilde{T}_{2A} , and \tilde{T}_{2B} ; where the first subscript indicates the source and the second the detector. The short ρ measurements ($\rho = 14.9$ mm) are \tilde{T}_{1A} and \tilde{T}_{2B} while the long ρ measurements ($\rho = 25.1$ mm) are \tilde{T}_{2A} and \tilde{T}_{1B} .

From these \tilde{T} measurements, ratios between the short and long ρ measurements may be obtained. Therefore we introduce the Single-Ratio of the \tilde{T} s ($\text{SR}\{\tilde{T}\}$ s) for the geometry in Figure 1 as follows:

$$\text{SR}\{\tilde{T}\}_{1AB} = \frac{\tilde{T}_{1B}}{\tilde{T}_{1A}} \quad (1)$$

$$\text{SR}\{\tilde{T}\}_{2BA} = \frac{\tilde{T}_{2A}}{\tilde{T}_{2B}} \quad (2)$$

and the Dual-Ratio of the \tilde{T} ($\text{DR}\{\tilde{T}\}$) as the geometric mean of the two symmetric $\text{SR}\{\tilde{T}\}$ s.

$$\text{DR}\{\tilde{T}\}_{1AB2} = \sqrt{\text{SR}\{\tilde{T}\}_{1AB} \times \text{SR}\{\tilde{T}\}_{2BA}} = \sqrt{\frac{\tilde{T}_{1B}\tilde{T}_{2A}}{\tilde{T}_{1A}\tilde{T}_{2B}}} \quad (3)$$

This forms a similar type of measurement to DS/SC but replacing the concept of slope with that of ratio (The word slope in Dual-Slope (DS) is historical [29,30] as no slopes are actually considered in this work). We acknowledge that the notation used here is verbose since, that even for $\text{DR}\{\tilde{T}\}$ we utilize subscripts to show all optodes used. However, for this work we have opted to use this notation to distinguish explicitly, the origin of the measurements. This is helpful in observing the differences in $\text{SR}\{\tilde{T}\}$ s, particularly in Section 3.1.2. In other work we opt to utilize a simpler notation with numbered subscripts [30].




Figure 1. Schematic diagram of cuvette geometry measuring $45 \text{ mm} \times 10 \text{ mm} \times 10 \text{ mm}$. Sources and detectors are considered at the following \vec{r} s: $\vec{r}_1 = -17\hat{x}$ mm, $\vec{r}_2 = 17\hat{x}$ mm, $\vec{r}_A = -6\hat{x} + 10\hat{z}$ mm, and $\vec{r}_B = 6\hat{x} + 10\hat{z}$ mm. (a) $y-x$ plane for $z = 0$ mm. (b) Transparent projected view. (c) $y-x$ plane for $z = 10$ mm. Acronyms and Symbols: Position vector (\vec{r}).

We can further expand the expression for $\text{SR}\{\tilde{T}\}$ to consider $\text{SR}\{\tilde{T}\}$ amplitude ($|\text{SR}\{\tilde{T}\}|$) and $\text{SR}\{\tilde{T}\}$ phase ($\angle\text{SR}\{\tilde{T}\}$). For example, with 1AB we have:

$$|\text{SR}\{\tilde{T}\}|_{1AB} = \frac{|\tilde{T}_{1B}|}{|\tilde{T}_{1A}|} \quad (4)$$

$$\angle\text{SR}\{\tilde{T}\}_{1AB} = \angle\tilde{T}_{1B} - \angle\tilde{T}_{1A} \quad (5)$$

and we introduce the final ratio type, the natural logarithm of $|\text{SR}\{\tilde{T}\}|$ ($\ln|\text{SR}\{\tilde{T}\}|$):

$$\ln|\text{SR}\{\tilde{T}\}|_{1AB} = \ln|\tilde{T}_{1B}| - \ln|\tilde{T}_{1A}| \quad (6)$$

The motivation for utilizing the natural logarithm in this way is that it partly linearizes typical expressions for diffuse optical measurements (\tilde{T} in this case) as a function of ρ [31]. Additionally it shows a symmetry between $\ln|\text{SR}\{\tilde{T}\}|$ and $\angle\text{SR}\{\tilde{T}\}$ as they are both differences. This work focuses on utilizing these $\ln|\text{SR}\{\tilde{T}\}|$ and $\angle\text{SR}\{\tilde{T}\}$ ratio types in the development of the proposed method.

Note that similar expressions can also be written for $\text{DR}\{\tilde{T}\}$ amplitude ($|\text{DR}\{\tilde{T}\}|$), $\text{DR}\{\tilde{T}\}$ phase ($\angle\text{DR}\{\tilde{T}\}$), and natural logarithm of $|\text{DR}\{\tilde{T}\}|$ ($\ln|\text{DR}\{\tilde{T}\}|$):

$$|\text{DR}\{\tilde{T}\}|_{1AB2} = \sqrt{|\text{SR}\{\tilde{T}\}|_{1AB} \times |\text{SR}\{\tilde{T}\}|_{2BA}} = \sqrt{\frac{|\tilde{T}_{1B}||\tilde{T}_{2A}|}{|\tilde{T}_{1A}||\tilde{T}_{2B}|}} \quad (7)$$

$$\angle\text{DR}\{\tilde{T}\}_{1AB2} = \frac{\angle\text{SR}\{\tilde{T}\}_{1AB} + \angle\text{SR}\{\tilde{T}\}_{2BA}}{2} = \frac{\angle\tilde{T}_{1B} + \angle\tilde{T}_{2A} - \angle\tilde{T}_{1A} - \angle\tilde{T}_{2B}}{2} \quad (8)$$

$$\begin{aligned} \ln|\text{DR}\{\tilde{T}\}|_{1AB2} &= \frac{\ln|\text{SR}\{\tilde{T}\}|_{1AB} + \ln|\text{SR}\{\tilde{T}\}|_{2BA}}{2} = \\ &= \frac{\ln|\tilde{T}_{1B}| + \ln|\tilde{T}_{2A}| - \ln|\tilde{T}_{1A}| - \ln|\tilde{T}_{2B}|}{2} \end{aligned} \quad (9)$$

From this it can be seen that $|\text{DR}\{\tilde{T}\}|$ is a geometric mean of $|\text{SR}\{\tilde{T}\}|$ s and both $\angle\text{DR}\{\tilde{T}\}$ as well as $\ln|\text{DR}\{\tilde{T}\}|$ are arithmetic means of $\angle\text{SR}\{\tilde{T}\}$ s and $\ln|\text{SR}\{\tilde{T}\}|$ s, respectively.

For theoretical calculations, not considering optode coupling differences and medium heterogeneity, the different $\text{SR}\{\tilde{T}\}$ s and $\text{DR}\{\tilde{T}\}$ s have the same value. This is due to the symmetry shown in Figure 1 considering a homogeneous medium. For this reason, only the set 1AB, and $\text{SR}\{\tilde{T}\}_{1AB}$, is considered for most of the results. Coupling is considered in Section 3.1.2, therefore, discrepancies between the difference measurements are investigated in that section and the distinction between $\text{SR}\{\tilde{T}\}$ and $\text{DR}\{\tilde{T}\}$ becomes important there.

2.3. Analytical Box Model

To generate data for the cuvette geometry (Figure 1) we utilized the following diffusion theory derived analytical expression for the \tilde{T} [23] (The expression used for the Green's function for the complex Transmittance (\tilde{T}) represents the measured transmittance normalized by the source power giving it units of mm^{-2}):

$$\begin{aligned}
\tilde{T}(x_{Det.}, y_{Det.}, z_{Det.} = L_z) = & \frac{1}{4\pi} \sum_{l=-\infty}^{\infty} \sum_{m=-\infty}^{\infty} \sum_{n=-\infty}^{\infty} \left[\right. \\
& \frac{(L_z - z_{1n})(\tilde{\mu}_{eff} + 1/r_1)}{r_1^2} e^{-\tilde{\mu}_{eff}r_1} - \frac{(L_z - z_{2n})(\tilde{\mu}_{eff} + 1/r_2)}{r_2^2} e^{-\tilde{\mu}_{eff}r_2} \\
& - \frac{(L_z - z_{1n})(\tilde{\mu}_{eff} + 1/r_3)}{r_3^2} e^{-\tilde{\mu}_{eff}r_3} + \frac{(L_z - z_{2n})(\tilde{\mu}_{eff} + 1/r_4)}{r_4^2} e^{-\tilde{\mu}_{eff}r_4} \\
& - \frac{(L_z - z_{1n})(\tilde{\mu}_{eff} + 1/r_5)}{r_5^2} e^{-\tilde{\mu}_{eff}r_5} + \frac{(L_z - z_{2n})(\tilde{\mu}_{eff} + 1/r_6)}{r_6^2} e^{-\tilde{\mu}_{eff}r_6} \\
& \left. + \frac{(L_z - z_{1n})(\tilde{\mu}_{eff} + 1/r_7)}{r_7^2} e^{-\tilde{\mu}_{eff}r_7} - \frac{(L_z - z_{2n})(\tilde{\mu}_{eff} + 1/r_8)}{r_8^2} e^{-\tilde{\mu}_{eff}r_8} \right] \quad (10)
\end{aligned}$$

where the optical properties, the μ_a and the μ'_s , are contained within the complex effective attenuation coefficient ($\tilde{\mu}_{eff}$):

$$\tilde{\mu}_{eff} = \sqrt{3\mu'_s \left(\mu_a - \frac{\omega n_i}{c} i \right)} \quad (11)$$

and the remaining non-spatial variables are the angular modulation frequency (ω), the index of refraction (n) inside the medium (n_i), and the speed of light in vacuum (c).

For spatial variables we first have the cuvette dimensions: $L_x = 45$ mm, $L_y = 10$ mm, and $L_z = 10$ mm (Figure 1). Next, we have the source coordinates: $x_{Src.}$ and $y_{Src.}$ (given that a pencil beam impinges on the $z = 0$ mm face so that an isotropic source is placed at $z_{Iso.-Src.} = 1/\mu'_s$); as well as the detector coordinates: $x_{Det.}$ and $y_{Det.}$ (given that the detector is placed on the $z = L_z$ face, thus $z_{Det.} = L_z$). Using these variables and remembering the sum indexes from Equation (10) we can now write the various point source positions (infinite of both positive and negative as per the indexing variables l , m , and n) [23]:

$$x_{1l} = 2lL_x + 4lh + x_{Src.} \quad (12)$$

$$x_{2l} = (2l - 1)L_x + (4l - 2)h - x_{Src.} \quad (13)$$

$$y_{1m} = 2mL_y + 4mh + y_{Src.} \quad (14)$$

$$y_{2m} = (2m - 1)L_y + (4m - 2)h - y_{Src.} \quad (15)$$

$$z_{1n} = 2nL_z + 4nh + 1/\mu'_s \quad (16)$$

$$z_{2n} = 2nL_z + (4n - 2)h - 1/\mu'_s \quad (17)$$

where h is the distance between the extrapolated boundary and the actual box boundary:

$$h = \frac{2a(n_r)}{3\mu'_s} \quad (18)$$

a is the n mismatch parameter [16,32] which is a function of the relative n mismatch ($n_r = n_i/n_o$, where n_o is the n outside). Finally, using these positions, we can define the distances to the point sources:

$$r_1 = \sqrt{(x_{Det.} - x_{1l})^2 + (y_{Det.} - y_{1m})^2 + (z_{Det.} - z_{1n})^2} \quad (19)$$

$$r_2 = \sqrt{(x_{Det.} - x_{1l})^2 + (y_{Det.} - y_{1m})^2 + (z_{Det.} - z_{2n})^2} \quad (20)$$

$$r_3 = \sqrt{(x_{Det.} - x_{1l})^2 + (y_{Det.} - y_{2m})^2 + (z_{Det.} - z_{1n})^2} \quad (21)$$

$$r_4 = \sqrt{(x_{Det.} - x_{1l})^2 + (y_{Det.} - y_{2m})^2 + (z_{Det.} - z_{2n})^2} \quad (22)$$

$$r_5 = \sqrt{(x_{Det.} - x_{2l})^2 + (y_{Det.} - y_{1m})^2 + (z_{Det.} - z_{1n})^2} \quad (23)$$

$$r_6 = \sqrt{(x_{Det.} - x_{2l})^2 + (y_{Det.} - y_{1m})^2 + (z_{Det.} - z_{2n})^2} \quad (24)$$

$$r_7 = \sqrt{(x_{Det.} - x_{2l})^2 + (y_{Det.} - y_{2m})^2 + (z_{Det.} - z_{1n})^2} \quad (25)$$

$$r_8 = \sqrt{(x_{Det.} - x_{2l})^2 + (y_{Det.} - y_{2m})^2 + (z_{Det.} - z_{2n})^2} \quad (26)$$

This diffusion theory derived expression (Equation (10)) was previously presented and validated against Monte Carlo in [23]. The validity of such an expression is dependent on the distance from the source, the value of μ'_s , and the ratio between μ_a and μ'_s . Only locations far enough from the source for scattering to be considered isotropic may be considered, μ'_s must be large enough for isotropic scattering to dominate within the medium, and μ_a must be much less than μ'_s . Given that we consider measurements on the opposite side of a 10 mm thick cuvette the first and second conditions are met for μ'_s values on the order of 1 mm^{-1} (cuvette thickness of 10 isotropic scattering mean free paths). Finally, the last condition may be met by considering values of μ_a on the order of 100 times smaller than μ'_s . For this work, we choose values which encompass the edge of validity of diffusion theory with μ_a from 0 mm^{-1} to 0.05 mm^{-1} and μ'_s from 0.5 mm^{-1} to 5 mm^{-1} for demonstration purposes. Most focus is on the values of 0.01 mm^{-1} for μ_a and 1 mm^{-1} for μ'_s , for which our independent (aside from what is presented in [23]) validation against Monte Carlo [17] found $<5\%$ discrepancy for $|\tilde{T}|$ and 0.005 rad for $\angle\tilde{T}$ over the face of the cuvette opposing the source. For the purposes of this work, a feasibility study of the proposed method, we believe this diffusion theory model is appropriate.

To provide some physical intuition regarding the FD portion of this model, we can discuss the wavelength of the photon density waves (λ_{PDW}), which can be approximated as [1]:

$$\lambda_{PDW} = \frac{2\pi}{\sqrt{\frac{3}{2}\mu_a\mu'_s \left(\sqrt{1 + \left(\frac{\omega n}{c\mu_a}\right)^2} - 1 \right)}} \quad (27)$$

This leads to a λ_{PDW} of about 160 mm using values of 0.01 mm^{-1} for μ_a , 1 mm^{-1} for μ'_s , $2\pi \times 100 \times 10^6 \text{ rad s}^{-1}$ for ω , 1 for n , and $2.99792458 \times 10^{11} \text{ mm s}^{-1}$ for c . This is larger than the cuvette, but not so large that it dwarfs the cuvette scale completely. This indicates that phase measurements are reasonable for this volume, given that the phase will not wrap (λ_{PDW} not too short) but will still change considerably throughout the size of the cuvette (λ_{PDW} not too long).

To demonstrate the implementation of this expression for \tilde{T} (Equation (10)) we show a map of the amplitude ($|\tilde{T}|$) and phase ($\angle\tilde{T}$) on the $z = L_z = 10 \text{ mm}$ face (opposing the source; Figure 1) for $\mu_a = 0.01 \text{ mm}^{-1}$ and $\mu'_s = 1 \text{ mm}^{-1}$ in Figure 2 (considering source 1). This shows the spatial continuum of \tilde{T} which can be simulated with diffusion theory. The positions of the source (1) and detectors (A & B) are also indicated in Figure 2 to show the positions which will be considered throughout this work. For computation based on Equation (10), l , m , and n were

each summed from -3 to 3 , and inclusion of more terms was found to not significantly impact the results.

2.4. Optical Properties Fit

The end goal of this work is to develop a method capable of measuring the absolute μ_a and μ'_s in the geometry of Figure 1 using FD. With this in mind, we define a cost (χ^2) function which can be minimized by varying μ_a and μ'_s , thus creating a fit for μ_a and μ'_s (We acknowledge that the cost (χ^2) function is dependent on parameters beyond absorption coefficient (μ_a) and reduced scattering coefficient (μ'_s) such as index of refraction (n), and investigate this in further sections of this work):

$$\chi^2(\mu_a, \mu'_s) = \kappa \left(\frac{\left[\ln |\text{DR}\{\tilde{T}\}| \right]_{meas} - \left[\ln |\text{DR}\{\tilde{T}\}| \right]_{theo}(\mu_a, \mu'_s)}{\sigma_{\ln |\text{DR}\{\tilde{T}\}|}} \right)^2 + \left(\frac{\left[\angle \text{DR}\{\tilde{T}\} \right]_{meas} - \left[\angle \text{DR}\{\tilde{T}\} \right]_{theo}(\mu_a, \mu'_s)}{\sigma_{\angle \text{DR}\{\tilde{T}\}}} \right)^2 \quad (28)$$

where the *meas* subscript represents the measured difference (In this work, the measurement is simulated using Equation (10), and noise may be added depending on the purpose) and the *theo* subscript represents the value retrieved from Equation (10) considering a particular μ_a and μ'_s (Again we note that the Dual-Ratio of the \tilde{T} ($\text{DR}\{\tilde{T}\}$) and a Single-Ratio of the \tilde{T} ($\text{SR}\{\tilde{T}\}$) are the same when not considering optode coupling (shown in Section 3.1.2)).




Figure 2. Example of the implementation of the diffusion theory derived expression for \tilde{T} in Equation (10) showing \tilde{T} (The expression used for the Green's function for the complex Transmittance (\tilde{T}) represents the measured transmittance normalized by the source power giving it units of mm^{-2}) on the cuvette face opposing the source ($z = L_z = 10 \text{ mm}$) considering the geometry in Figure 1 and source 1. For this simulation the μ_a was 0.01 mm^{-1} , the μ'_s 1 mm^{-1} , n inside 1.3, n outside 1, ω $2\pi 100 \times 10^6 \text{ rad s}^{-1}$, and the cuvette measured $45 \text{ mm} \times 10 \text{ mm} \times 10 \text{ mm}$. The source ($x - y$ position is shown as cross; x position as black dotted line) was placed at $\vec{r}_{\text{Src.1}} = -17\hat{x} \text{ mm}$. Detector positions which are considered in the following work are shown as circles (for $x - y$ position) or dotted lines (for x position) (a) \tilde{T} amplitude ($|\tilde{T}|$) on the $x - y$ plane at $z = L_z = 10 \text{ mm}$. (b) \tilde{T} phase ($\angle\tilde{T}$) on the $x - y$ plane at $z = L_z = 10 \text{ mm}$. (c) $|\tilde{T}|$ and $\angle\tilde{T}$ along the x direction for $z = L_z = 10 \text{ mm}$ and $y = 0 \text{ mm}$. Acronyms and Symbols: Green's function for the complex Transmittance (\tilde{T}), absorption coefficient (μ_a), reduced scattering coefficient (μ'_s), index of refraction (n), angular modulation frequency (ω), and position vector (\vec{r}).

Three further variables are introduced in Equation (28), which we define below. First is the $\ln |\text{DR}\{\tilde{T}\}|$ scaling coefficient (κ) which is discussed further in Section 3.2.1. Second and third are the uncertainties of $\ln |\text{DR}\{\tilde{T}\}|$ ($\sigma_{\ln |\text{DR}\{\tilde{T}\}|}$) and $\angle\text{DR}\{\tilde{T}\}$ ($\sigma_{\angle\text{DR}\{\tilde{T}\}}$) which are expressed based on 1st order error propagation as:

$$\sigma_{\ln |\text{DR}\{\tilde{T}\}|} = \sqrt{\left(\frac{\sigma_{|\tilde{T}|_{\text{long}}}}{|\tilde{T}|_{\text{long}}}\right)^2 + \left(\frac{\sigma_{|\tilde{T}|_{\text{short}}}}{|\tilde{T}|_{\text{short}}}\right)^2} \quad (29)$$

$$\sigma_{\angle\text{DR}\{\tilde{T}\}} = \sqrt{\left(\sigma_{\angle\tilde{T}_{\text{long}}}\right)^2 + \left(\sigma_{\angle\tilde{T}_{\text{short}}}\right)^2} \quad (30)$$

and

$$\sigma_{\ln|\text{DR}\{\tilde{T}\}|} = \frac{\sigma_{\ln|\text{SR}\{\tilde{T}\}|}}{\sqrt{2}} \quad (31)$$

$$\sigma_{\angle\text{DR}\{\tilde{T}\}} = \frac{\sigma_{\angle\text{SR}\{\tilde{T}\}}}{\sqrt{2}} \quad (32)$$

where, $\sigma_{|\tilde{T}|}$ is the uncertainty in $|\tilde{T}|$ and $\sigma_{\angle\tilde{T}}$ is the uncertainty in $\angle\tilde{T}$, and assuming that the uncertainties in the two $\ln|\text{SR}\{\tilde{T}\}|$ s ($\sigma_{\ln|\text{SR}\{\tilde{T}\}|}$) and the two $\angle\text{SR}\{\tilde{T}\}$ s ($\sigma_{\angle\text{SR}\{\tilde{T}\}}$) are each the same. For this work we set $\sigma_{|\tilde{T}|}/|\tilde{T}| = 0.001$ and $\sigma_{\angle\tilde{T}} = 1.7 \text{ mrad} = 0.1^\circ$ which would be typical for a FD NIRS instrument such as the Imagent.

3. Results

3.1. Investigation of Difference Measurements

3.1.1. Variation over Optical Properties

The chief measurements which we consider are $\ln|\text{SR}\{\tilde{T}\}|$ and $\angle\text{SR}\{\tilde{T}\}$ (or $\ln|\text{DR}\{\tilde{T}\}|$ and $\angle\text{DR}\{\tilde{T}\}$ considering coupling; (Again we note that the Dual-Ratio of the \tilde{T} ($\text{DR}\{\tilde{T}\}$) and a Single-Ratio of the \tilde{T} ($\text{SR}\{\tilde{T}\}$) are the same when not considering optode coupling (shown in Section 3.1.2))). Figure 3 shows these measurements (Equations (5) and (6)) over a large range of optical properties, specifically μ_a and μ'_s (Figure 3a,c) or n_i and n_o (Figure 3b,d).

Since the intention is to convert these measurements of $\ln|\text{SR}\{\tilde{T}\}|$ and $\angle\text{SR}\{\tilde{T}\}$ to μ_a and μ'_s , the desire is for the measurements to vary significantly more as μ_a and μ'_s are varied as compared to varying n_i and n_o . The iso-lines (white lines) in Figure 3a,b consider the same values (the same is true for Figure 3c,d). From this we see that varying μ_a and μ'_s varies $\ln|\text{SR}\{\tilde{T}\}|$ across 4 more iso-lines than varying n_i and n_o (and about 2 times more for $\angle\text{SR}\{\tilde{T}\}$). This suggests promise in the goal of retrieving μ_a and μ'_s .

To recover μ_a and μ'_s from $\ln|\text{SR}\{\tilde{T}\}|$ and $\angle\text{SR}\{\tilde{T}\}$ we must also have significantly different information in $\ln|\text{SR}\{\tilde{T}\}|$ and $\angle\text{SR}\{\tilde{T}\}$ so that the recovered variables (μ_a and μ'_s) have a unique solution and little cross-talk. There is also promise along these lines as the iso-lines in Figure 3a versus Figure 3c are qualitatively orthogonal. This suggests a fit to μ_a and μ'_s from $\ln|\text{SR}\{\tilde{T}\}|$ and $\angle\text{SR}\{\tilde{T}\}$ should be possible. This is further investigated in Section 3.2.

One final insight that can be drawn from Figure 3 is the effect of n_r which is constant along diagonal lines with positive slopes in Figure 3b,d. From this we see that $\ln|\text{SR}\{\tilde{T}\}|$ is little effected by n_r . Further, in the upper left portion of the plots where $n_r < 1$ $\angle\text{SR}\{\tilde{T}\}$ is only significantly effected by n_i . Therefore we may be able to optimize the design of the cuvette boundary to reduce cross-talk with the n_s which is discussed further in Section 4.

3.1.2. Optode Coupling and Auto-Calibration

As has been stated above including in (Again we note that the Dual-Ratio of the \tilde{T} ($\text{DR}\{\tilde{T}\}$) and a Single-Ratio of the \tilde{T} ($\text{SR}\{\tilde{T}\}$) are the same when not considering optode coupling (shown in Section 3.1.2)), $\text{SR}\{\tilde{T}\}$ & $\text{DR}\{\tilde{T}\}$ (as well as $|\text{SR}\{\tilde{T}\}|$ & $|\text{DR}\{\tilde{T}\}|$, $\angle\text{SR}\{\tilde{T}\}$ & $\angle\text{DR}\{\tilde{T}\}$, and $\ln|\text{SR}\{\tilde{T}\}|$ & $\ln|\text{DR}\{\tilde{T}\}|$) are equivalent when optode coupling is not considered. For this reason, other sections of this manuscript are not careful to distinguish between them as theoretical calculations are being carried out and coupling is not a consideration. However, in this section we show the effect of optode coupling and the auto-calibration of the $\text{DR}\{\tilde{T}\}$ which is inherited/inspired by the SC method [24].




Figure 3. How the measurements of the $\ln|\text{SR}\{\tilde{T}\}|$ and $\angle\text{SR}\{\tilde{T}\}$ (Equations (6) and (5)) are effected by optical parameters, namely the μ_a , the μ'_s , the n inside (n_i) and the n outside (n_o). Simulation geometry and parameters not explicitly shown here are stated in detail in Figures 1 and 2. (a) $\ln|\text{SR}\{\tilde{T}\}|$ versus μ_a and μ'_s . (b) $\ln|\text{SR}\{\tilde{T}\}|$ versus n_i and n_o . (c) $\angle\text{SR}\{\tilde{T}\}$ versus μ_a and μ'_s . (d) $\angle\text{SR}\{\tilde{T}\}$ versus n_i and n_o . Note: (a,b) have the same iso-line and color-map values/scales; as do (c,d). Acronyms and Symbols: Green's function for the complex Transmittance (\tilde{T}), Single-Ratio of the \tilde{T} ($\text{SR}\{\tilde{T}\}$), $\text{SR}\{\tilde{T}\}$ amplitude ($|\text{SR}\{\tilde{T}\}|$), natural logarithm of $|\text{SR}\{\tilde{T}\}|$ ($\ln|\text{SR}\{\tilde{T}\}|$), $\text{SR}\{\tilde{T}\}$ phase ($\angle\text{SR}\{\tilde{T}\}$), absorption coefficient (μ_a), reduced scattering coefficient (μ'_s), and index of refraction (n).

To do this, first we define a complex optical Coupling, power, and/or efficiency factor (\tilde{C}) for each optode: $\tilde{C}_1, \tilde{C}_2, \tilde{C}_A$, and \tilde{C}_B . Physically, $|\tilde{C}|$ represents a multiplicative factor (attenuation or amplification) on the amplitude of \tilde{T} and $\angle\tilde{C}$ represents a phase shift on the phase of \tilde{T} . \tilde{C} s applied to sources (number subscripts) have units of mW since their amplitude also includes source power; while \tilde{C} s for detectors (letter subscripts) are unit-less. Therefore, adding *coupl* subscripts to our measurements when they are confounded by coupling (opposed to the theoretical value without the *coupl* subscript) we have the following signals considering coupling:

$$\tilde{T}_{1A,coupl} = \tilde{C}_1 \tilde{C}_A \tilde{T}_{1A} \quad (33)$$

$$\tilde{T}_{1B,coupl} = \tilde{C}_1 \tilde{C}_B \tilde{T}_{1B} \quad (34)$$

$$\tilde{T}_{2A,coupl} = \tilde{C}_2 \tilde{C}_A \tilde{T}_{2A} \quad (35)$$

$$\tilde{T}_{2B,coup} = \tilde{C}_2 \tilde{C}_B \tilde{T}_{2B} \quad (36)$$

Now, let us revisit Equations (1)–(3) but with optode coupling considered:

$$\text{SR}\{\tilde{T}\}_{1AB,coup} = \frac{\tilde{C}_1 \tilde{C}_B \tilde{T}_{1B}}{\tilde{C}_1 \tilde{C}_A \tilde{T}_{1A}} = \frac{\tilde{C}_B \tilde{T}_{1B}}{\tilde{C}_A \tilde{T}_{1A}} = \frac{\tilde{C}_B}{\tilde{C}_A} \text{SR}\{\tilde{T}\}_{1AB} \quad (37)$$

$$\text{SR}\{\tilde{T}\}_{2BA,coup} = \frac{\tilde{C}_2 \tilde{C}_A \tilde{T}_{2A}}{\tilde{C}_2 \tilde{C}_B \tilde{T}_{2B}} = \frac{\tilde{C}_A \tilde{T}_{2A}}{\tilde{C}_B \tilde{T}_{2B}} = \frac{\tilde{C}_A}{\tilde{C}_B} \text{SR}\{\tilde{T}\}_{2BA} \quad (38)$$

$$\text{DR}\{\tilde{T}\}_{1AB2,coup} = \sqrt{\frac{\tilde{C}_B \tilde{T}_{1B} \tilde{C}_A \tilde{T}_{2A}}{\tilde{C}_A \tilde{T}_{1A} \tilde{C}_B \tilde{T}_{2B}}} = \sqrt{\frac{\tilde{T}_{1B} \tilde{T}_{2A}}{\tilde{T}_{1A} \tilde{T}_{2B}}} = \text{DR}\{\tilde{T}\}_{1AB2} \quad (39)$$

showing that the measured $\text{DR}\{\tilde{T}\}$ is the same as the theoretical values regardless of the various optode couplings \tilde{C} . Notice that the same follows for $|\text{DR}\{\tilde{T}\}|$, $\angle\text{DR}\{\tilde{T}\}$, and $\ln|\text{DR}\{\tilde{T}\}|$.

This property of auto-calibration (coming from the SC method [24]) is demonstrated in Figure 4. However, in this case, unlike SC, the symmetry requirements are not as strict since ratios instead of slopes are used as the measurement. In this case a random \tilde{C} was applied for each optode and the difference measurements both averaged and not were simulated. From Figure 4 one can see that the $\ln|\text{DR}\{\tilde{T}\}|$ and $\angle\text{DR}\{\tilde{T}\}$ measurements are the same as the theoretical values. This is significant since it shows that the proposed measurement method would be insensitive to optode coupling, and further, optode coupling would not effect the recovered μ_a and μ'_s . Therefore, the instrument would not need to be calibrated in terms of coupling, reducing possible systematic errors and making the method simpler to implement.




Figure 4. Demonstration of the cancellation of coupling factors when considering the $\ln|\text{DR}\{\tilde{T}\}|$ and $\angle\text{DR}\{\tilde{T}\}$ measurements. For this simulation, random \tilde{C} s were generated for each optode and Equations (33)–(36) implemented. All other simulation parameters are the same as Figures 1 and 2. The expected theoretical value for the measured differences is shown as a dashed line. (a) The two symmetric $\ln|\text{SR}\{\tilde{T}\}|$ measurements and the $\ln|\text{DR}\{\tilde{T}\}|$ measurement. (b) The two symmetric $\angle\text{SR}\{\tilde{T}\}$ measurements and the $\angle\text{DR}\{\tilde{T}\}$ measurement. Acronyms and Symbols: Green's function for the complex Transmittance (\tilde{T}), Single-Ratio of the \tilde{T} ($\text{SR}\{\tilde{T}\}$), $\text{SR}\{\tilde{T}\}$ amplitude ($|\text{SR}\{\tilde{T}\}|$), natural logarithm of $|\text{SR}\{\tilde{T}\}|$ ($\ln|\text{SR}\{\tilde{T}\}|$), $\text{SR}\{\tilde{T}\}$ phase ($\angle\text{SR}\{\tilde{T}\}$), Dual-Ratio of the \tilde{T} ($\text{DR}\{\tilde{T}\}$), $\text{DR}\{\tilde{T}\}$ amplitude ($|\text{DR}\{\tilde{T}\}|$), natural logarithm of $|\text{DR}\{\tilde{T}\}|$ ($\ln|\text{DR}\{\tilde{T}\}|$), $\text{DR}\{\tilde{T}\}$ phase ($\angle\text{DR}\{\tilde{T}\}$), and complex optical Coupling, power, and/or efficiency factor (\tilde{C}).

3.2. Development of Fit for Absolute Optical Properties

3.2.1. Optimization of Cost Space Shape

In order to fit for the absolute optical properties μ_a and μ'_s we consider the χ^2 function in Equation (28). This function contains the scaling parameter κ which balances the scale of $\ln |\text{SR}\{\tilde{T}\}|$ versus $\angle\text{SR}\{\tilde{T}\}$. The intention of such a parameter is to modify the χ^2 space to be as circular as possible. This circularity can be quantitatively defined by considering iso-lines in cost space and their perimeter (P) as well as area (A). A circle has the minimum ratio of P to A of all 2-Dimensional (2D) shapes. Therefore, the dimensionless metric P^2/A was minimized by varying κ (note that P^2/A has a minimum theoretical value of 4π , for a circle) [33].

The effect of the κ value on χ^2 space shape is shown in Figure 5 using the same parameters as Figures 1 and 2 (where the optical properties are the *true* values). Figure 5b shows the optimal κ of 1.2×10^{-3} , which is the case where P^2/A was minimized. The resulting P^2/A , for this optimal κ was 36 about 3 times worse than the value of $4\pi \approx 12.6$ for a ideal circular cost space. This can be seen by how oblique the χ^2 iso-lines are in the μ_a direction suggesting a higher relative uncertainty in μ_a . We investigate this further in Section 3.3. Figure 5a,c show the effect of favoring either the $\angle\text{SR}\{\tilde{T}\}$ or $\ln |\text{SR}\{\tilde{T}\}|$ term in the χ^2 expression (Equation (28)). In either case μ_a and μ'_s are correlated and the space is spread more in μ_a . However, when $\ln |\text{SR}\{\tilde{T}\}|$ is favored (Figure 5c) μ_a and μ'_s have a negative correlation while the correlation is positive when $\angle\text{SR}\{\tilde{T}\}$ is favored (Figure 5a).

Note that the optimal $\kappa = 1.2 \times 10^{-3}$ was found for the μ_a of 0.01 mm and μ'_s of 1 mm and a different optimal κ s may be found elsewhere for different *true* μ_a and μ'_s . Despite this we have opted to utilize this one κ value for the rest of this work to reduce computation time, in the future a map of optimal κ could be found as a function of μ_a and μ'_s .




Figure 5. Three examples of the χ^2 (Equation (28)) space shape for different κ s. The shape of cost space is determined by the shape of an iso-line and its ratio of P squared divided by A (P^2/A) which has a minimum possible value of 4π (in the case of a circle). Iso-lines are the 5th quantile of all χ^2 values in each image and are meant to represent the overall shape of χ^2 space. For all axes, χ^2 , the μ_a , and the μ'_s are normalized. (a) $\kappa = 3 \times 10^{-5}$ and $P^2/A = 106$. (b) Optimal value of κ for the *true* optical properties used here found by minimizing P^2/A , resulting in $\kappa = 1.2 \times 10^{-3}$ and $P^2/A = 36$. (c) $\kappa = 3 \times 10^{-2}$ and $P^2/A = 110$. Acronyms and Symbols: Cost (χ^2), scaling factor (κ), Perimeter (P), Area (A), absorption coefficient (μ_a), and reduced scattering coefficient (μ'_s).

3.2.2. Cost Space Shape for Various Optical Properties

Now that the entire cost (χ^2 ; Equation (28)) function including $\kappa = 1.2 \times 10^{-3}$ has been determined, we can plot some example cost spaces for various *true* μ_a s and μ'_s s. This is shown for 9 cases in Figure 6. For the 9 cases all combinations of the following optical properties were

used: $\mu_a = 0.005 \text{ mm}^{-1}$, 0.010 mm^{-1} and 0.020 mm^{-1} combined with $\mu'_s = 0.5 \text{ mm}^{-1}$, 1.0 mm^{-1} and 2.0 mm^{-1} .

Examining Figure 6 we notice that in general μ_a will likely have more error or has a less unique solution compared to μ'_s . This is evident by the spreading of the low values of χ^2 along the μ_a direction near the local minimum and *true* value. This result is an extension of what was seen for one set of optical properties in Figure 5b. We also notice that this oblique χ^2 space shape is worse for small μ'_s (0.5 mm^{-1}), which is somewhat expected since diffusion theory is not meant to be used in the low scattering regime. For this reason, finding optimal κ as a function of μ_a and μ'_s may help alleviate this problem. Regardless, from this result we should expect the fit to work less well when attempting to retrieve μ_a when μ'_s is low.




Figure 6. 9 examples of χ^2 (Equation (28); $\kappa = 1.2 \times 10^{-3}$) space for different sets of the *true* μ_a and μ'_s . (a) $\mu_{a,true} = 0.005 \text{ mm}^{-1}$ and $\mu'_{s,true} = 0.5 \text{ mm}^{-1}$. (b) $\mu_{a,true} = 0.005 \text{ mm}^{-1}$ and $\mu'_{s,true} = 1.0 \text{ mm}^{-1}$. (c) $\mu_{a,true} = 0.005 \text{ mm}^{-1}$ and $\mu'_{s,true} = 2.0 \text{ mm}^{-1}$. (d) $\mu_{a,true} = 0.010 \text{ mm}^{-1}$ and $\mu'_{s,true} = 0.5 \text{ mm}^{-1}$. (e) $\mu_{a,true} = 0.010 \text{ mm}^{-1}$ and $\mu'_{s,true} = 1.0 \text{ mm}^{-1}$. (f) $\mu_{a,true} = 0.010 \text{ mm}^{-1}$ and $\mu'_{s,true} = 2.0 \text{ mm}^{-1}$. (g) $\mu_{a,true} = 0.020 \text{ mm}^{-1}$ and $\mu'_{s,true} = 0.5 \text{ mm}^{-1}$. (h) $\mu_{a,true} = 0.020 \text{ mm}^{-1}$ and $\mu'_{s,true} = 1.0 \text{ mm}^{-1}$. (i) $\mu_{a,true} = 0.020 \text{ mm}^{-1}$ and $\mu'_{s,true} = 2.0 \text{ mm}^{-1}$. Acronyms and Symbols: Cost (χ^2), scaling factor (κ), absorption coefficient (μ_a), and reduced scattering coefficient (μ'_s).

3.2.3. Fit Initial Guess

We finish our development of the fit for μ_a and μ'_s with a demonstration of exact retrieval when the same inverse and forward models are used for \tilde{T} without noise (Equation (10)). In doing so we also investigate the effect of different initial guesses on μ_a and μ'_s to show that convergence is not dependent on this initial guess (The result is not dependent of initial guess given that the initial guess is of reasonable optical properties). For this, the fit was implemented by using the MathWorks MATrix LABoratory [Natick, MA USA] (MATLAB) function `fmincon` to minimize χ^2 (Equation (28); $\kappa = 1.2 \times 10^{-3}$) function. For `fmincon`, the algorithm `interior-point` was used and the minimum constraints on μ_a and μ'_s set to $[0, 0]$, respectively, with all other bound types unconstrained.

Using this optimization setup, the fit was run with the $\mu_{a,true} = 0.010 \text{ mm}^{-1}$ and the $\mu'_{s,true} = 1.0 \text{ mm}^{-1}$ using 4 different initial guesses:

- $\mu_{a,guess} = 0.005 \text{ mm}^{-1}$ & $\mu'_{s,guess} = 0.5 \text{ mm}^{-1}$.
- $\mu_{a,guess} = 0.005 \text{ mm}^{-1}$ & $\mu'_{s,guess} = 2.0 \text{ mm}^{-1}$.
- $\mu_{a,guess} = 0.020 \text{ mm}^{-1}$ & $\mu'_{s,guess} = 0.5 \text{ mm}^{-1}$.
- $\mu_{a,guess} = 0.020 \text{ mm}^{-1}$ & $\mu'_{s,guess} = 2.0 \text{ mm}^{-1}$.

The results from these fits and the fit trajectory (shown as dotted lines with circles) are shown in Figure 7. In all cases the fit converged to the *true* optical properties regardless of start point. A second observation that can be made from Figure 7 is what trajectory the fit follows during convergence. Acknowledging that this is highly dependent on algorithm choice, we still note that the fit spent most of its time traversing in the μ_a direction, converging close to the correct μ'_s comparatively fast. This is a consequence of the shape of cost space, having a longer trough in the μ_a direction than the μ'_s .




Figure 7. Example trajectories of `fmincon` minimization of χ^2 (Equation (28); $\kappa = 1.2 \times 10^{-3}$) to fit for the μ_a and the μ'_s . Results from 4 different initial guesses shown: (Red) $\mu_{a,guess} = 0.005 \text{ mm}^{-1}$ & $\mu'_{s,guess} = 0.5 \text{ mm}^{-1}$. (Green) $\mu_{a,guess} = 0.005 \text{ mm}^{-1}$ & $\mu'_{s,guess} = 2.0 \text{ mm}^{-1}$. (Cyan) $\mu_{a,guess} = 0.020 \text{ mm}^{-1}$ & $\mu'_{s,guess} = 0.5 \text{ mm}^{-1}$. (Purple) $\mu_{a,guess} = 0.020 \text{ mm}^{-1}$ & $\mu'_{s,guess} = 2.0 \text{ mm}^{-1}$. Acronyms and Symbols: Cost (χ^2), scaling factor (κ), absorption coefficient (μ_a), and reduced scattering coefficient (μ'_s).

3.3. Confounds to Fit Retrieved Absolute Optical Properties

3.3.1. Propagation of Noise to Optical Property Uncertainty

To test how noise propagates through the recovery of μ_a and μ'_s , when using the fit developed in Section 3.2, we simulated $\sigma_{|\tilde{T}|}/|\tilde{T}| = 0.01$ and $\sigma_{\angle\tilde{T}} = 1.7 \text{ mrad} = 0.1^\circ$ as mentioned in Section 2.4. This was done by simulating measured $\ln|\text{DR}\{\tilde{T}\}|$ and $\angle\text{DR}\{\tilde{T}\}$ 101 times and each time adding Gaussian noise with the σ s stated above. For each of the 101, the fit was run to recover some μ_a and μ'_s . This was done for all 9 of the sets of *true* μ_a and μ'_s shown in Figure 6.

The results from this exercise are shown in Figure 8 and Table 1, from these three main observations can be drawn:

- A The fractional error in μ_a is always larger compared to μ'_s suggesting the system can more precisely recover μ'_s .
- B Errors in μ_a are much larger for small μ'_s and slightly larger for small μ_a (with small μ_a and μ'_s together being the worst case).
- C That μ_a and μ'_s are highly negatively correlated (as suggested by Figure 5b,c).

Observation A again expounds upon what has been expected from the shape of χ^2 space presented in previous sections. Furthermore, observations B and C are typical for such diffusion theory based problems.




Figure 8. Result from 101 simulations of noise ($\sigma_{|\tilde{T}|}/|\tilde{T}| = 0.01$ and $\sigma_{\angle\tilde{T}} = 1.7 \text{ mrad} = 0.1^\circ$) for the 9 different *true* sets of the μ_a and the μ'_s shown in Figure 6. (a) Marginal histograms for recovered μ'_s values of the 9 sets of optical properties and 101 noise simulations. (b) Scatter plot of recovered μ_a and μ'_s for the 9×101 noise simulations. (c) Marginal histograms for recovered μ_a values of the 9 sets of optical properties and 101 noise simulations. Acronyms and Symbols: Green's function for the complex Transmittance (\tilde{T}), uncertainty (σ), absorption coefficient (μ_a), and reduced scattering coefficient (μ'_s).

To draw some quantitative values for this exercise, we can closely examine Table 1. It is helpful to extract the worst case (for the type of simulations we have done), and typical (con-

sidering typical being the case when $\mu_{a,true} = 0.010 \text{ mm}^{-1}$ and $\mu'_{s,true} = 1.0 \text{ mm}^{-1}$) fractional errors in μ_a and μ'_s . These are as follows:

- For μ_a :
 - Typical error of 4 %.
 - Worst case error of 20 % (for low μ_a and μ'_s).
- For μ'_s :
 - Typical error of 1 %.
 - Worst case error of 3 % (for high μ_a and low μ'_s).

Of course these values are dependent on the simulated measurement errors of $\sigma_{|\tilde{T}|}/|\tilde{T}| = 0.01$ and $\sigma_{\angle\tilde{T}} = 1.7 \text{ mrad} = 0.1^\circ$ which may be different for different instruments.

Table 1. Errors for 9 sets of *true* optical properties and 101 noise simulations using $\sigma_{|\tilde{T}|}/|\tilde{T}| = 0.01$ and $\sigma_{\angle\tilde{T}} = 1.7 \text{ mrad} = 0.1^\circ$.

$\mu_{a,true}$ (mm^{-1})	$\mu'_{s,true}$ (mm^{-1})	σ_{μ_a} (mm^{-1})	$\sigma_{\mu_a}/\bar{\mu}_a$	$\sigma_{\mu'_s}$ (mm^{-1})	$\sigma_{\mu'_s}/\bar{\mu}'_s$	r_{μ_a,μ'_s}
0.005	0.5	0.0008	0.2	0.01	0.02	−0.9986
0.005	1.0	0.0003	0.06	0.01	0.01	−0.9970
0.005	2.0	0.0001	0.02	0.02	0.008	−0.9931
0.010	0.5	0.001	0.1	0.01	0.03	−0.9986
0.010	1.0	0.0004	0.04	0.01	0.01	−0.9983
0.010	2.0	0.0002	0.02	0.02	0.009	−0.9971
0.020	0.5	0.001	0.07	0.02	0.03	−0.9990
0.020	1.0	0.0007	0.04	0.02	0.02	−0.9988
0.020	2.0	0.0003	0.01	0.02	0.009	−0.9981

Symbols: Absorption coefficient (μ_a), reduced scattering coefficient (μ'_s), Green's function for the complex Transmittance (\tilde{T}), uncertainty (σ), and correlation coefficient (r).

3.3.2. Assumption of Index of Refraction

Finally, we examine how the assumption of n (and by extension the model boundary conditions) affects the recovered μ_a and μ'_s . We have done this by running the fit *assuming* sets of n_i and n_o , but generating forward data with different *true* n s in the range 1 to 2 (we do not co-vary n_i and n_o for simplicity). 2 sets of n s *assumed* in the fit were investigated:

- $n_{i,assumed} = 1.3$ & $n_{o,assumed} = 1.0$ (Figure 9 solid lines).
- $n_{i,assumed} = 1.3$ & $n_{o,assumed} = 2.0$ (Figure 9 dashed lines).

This exercise was done for all 9 sets of μ_a and μ'_s shown in Figure 6.

Figure 9 shows these recovered μ_a and μ'_s for the 2 *assumed* cases while varying $n_{i,true}$ and $n_{o,true}$. First, we note that n_i has a larger effect on the recovered μ_a and μ'_s compared to n_o , with μ_a having a negative, and μ'_s a positive, correlation with $n_{i,true}$ (Figure 9a,c). Furthermore, μ_a is much more strongly affected by $n_{i,true}$ than μ'_s , with recovered values being up to about 7 times greater than the *true* value when there is a low $n_{i,true}$ value. For high $n_{i,true}$ the recovered μ_a often approaches 0 mm^{-1} (hitting the `fmincon` constraint). All of this suggests that the method's ability to accurately recover μ'_s and particularly μ_a is dependent on knowledge of n_i .




Figure 9. Effect of recovered μ_a and μ'_s on the true n inside (n_i) and outside (n_o) when fixed values of n_s are assumed. Shown for the 9 sets of optical properties used in Figure 6. (Solid Lines) $n_i = 1.3$ and $n_o = 1.0$ assumed in fit. (Dashed Lines) $n_i = 1.3$ and $n_o = 2.0$ assumed in fit. (a) Recovered μ_a while varying $n_{i,true}$ and fixing $n_{o,true}$ to the assumed value. (b) Recovered μ_a while varying $n_{o,true}$ and fixing $n_{i,true}$ to the assumed value. (c) Recovered μ'_s while varying $n_{i,true}$ and fixing $n_{o,true}$ to the assumed value. (d) Recovered μ'_s while varying $n_{o,true}$ and fixing $n_{i,true}$ to the assumed value. Acronyms and Symbols: Absorption coefficient (μ_a), reduced scattering coefficient (μ'_s), and index of refraction (n).

Now, focusing on Figure 9b,d, we see the effect of n_o . In this case μ'_s is almost not affected at all by $n_{o,true}$ and μ_a is much more significantly affected when $n_{o,true} < n_{i,true}$ or $n_{r,true} > 1$. In this case the correlation between μ_a and n_o is positive (opposite to that for n_i), suggesting a connection to the dependence on n_r . These results further re-enforce the idea that recovered μ_a would be highly effected by the true n_s , thus control or knowledge of the n_s for this method is critical.

Lastly, by comparing the two $n_{i,assumed}$ and $n_{o,assumed}$ sets (solid versus dashed lines) we notice that the recovered μ_a and μ'_s vary less for the dashed lines (Figure 9). The dashed line is the case where $n_{i,assumed} = 1.3$ and $n_{o,assumed} = 2.0$ (In Figure 9 the *assumed* index of refraction (n), is equal to the *true* when the other n is varied, for example in Figure 9a the true n outside ($n_{o,true}$) is 1 for the solid line and 2 for the dashed line). This tells us that the incorrect recovery of μ_a and μ'_s can be partially alleviated when n_o is large, even if n_i is unknown. Since when this method is implemented it would be more practical to control n_o than n_i , it would be advantageous to design a cuvette with high n to take advantage of this reduction of the effect of the assumption of n_i seen by comparing dashed to solid lines in Figure 9a.

4. Discussion

The method presented appears to be feasible in measuring absolute μ_a and μ'_s in a standard cuvette (45 mm \times 10 mm \times 10 mm). This is significant given that typical/traditional measurements of μ_a and μ'_s with diffuse optical methods require large sample volumes (on the order of liters) and careful instrumental calibration. In this case, small sample volumes may be used (on the order of 1 mL to 10 mL) without the need for calibration of optode coupling (as described in Section 3.1.2).

To summarize, we started the development of this measurement method by choosing which data we intended to collect from the cuvette, namely $\ln |\text{DR}\{\tilde{T}\}|$ and $\angle\text{DR}\{\tilde{T}\}$, and determining how these data vary in respect to the desired recovered properties, namely μ_a and μ'_s (Figure 3). This leads to the development of a fit for μ_a and μ'_s and a careful examination of χ^2 space (Section 3.2). From this examination, one major result was discovered, this being that μ_a is less determined (has a broad local minimum area in χ^2 space) compared to μ'_s . This is the first potential limitation of this method since often μ_a is in-fact the targeted property of interest while μ'_s may be considered a confound. Despite this it appears that this weakness mainly occurs when μ'_s is small ($<1 \text{ mm}^{-1}$; Figure 6) telling us that the method has its main strength when the sample is highly scattering. Given that most commercial spectrometers designed for cuvette measurement require a non-scattering sample.

We also simulated two types of confounds that may lead to incorrect recovered μ_a and μ'_s . First, we investigated how instrumental noise would propagate through the measurement to the recovered μ_a and μ'_s (Section 3.3.1). Here we confirmed what was expected when the χ^2 space was examined, specifically that μ_a has a higher relative error compared to μ'_s (Figure 8 and Table 1). However, this error becomes comparable when μ'_s is high. For example, with $\sigma_{|\tilde{T}|}/|\tilde{T}| = 0.01$ and $\sigma_{\angle\tilde{T}} = 1.7 \text{ mrad} = 0.1^\circ$:

- If $\mu_a = 0.005 \text{ mm}^{-1}$ & $\mu'_s = 0.5 \text{ mm}^{-1}$ then μ_a has an error of 20% and μ'_s of 2%.
- If $\mu_a = 0.020 \text{ mm}^{-1}$ & $\mu'_s = 2.0 \text{ mm}^{-1}$ then μ_a has an error of 1% and μ'_s of 0.9%.

Therefore, we see that this method really shines when the sample is very diffuse, which is another way of saying highly scattering.

Investigation of the effect of incorrectly assumed boundary conditions on the fit results was also done (Section 3.3.2). Many different boundary conditions have been extensively studied and modeled in the past [34–38], however just because conditions can be modeled does not mean that there is knowledge of them in practice which can be corrected for. Given that a diffusion theory model was used, we varied boundary conditions in terms of n_i and n_o . Again, we found that μ_a is more likely to be incorrectly recovered compared to μ'_s . However, further we found that n_i had the largest effect on the recovered μ_a and μ'_s (Figure 9). This in principle is a short-coming of the method since one could argue that if the method were implemented, n_o could be controlled through instrument design but n_i would be unknown. However, Figure 9 shows that the effect of n_i is suppressed when n_o is large, suggesting a relationship to n_r . Therefore, we expect that an instrumental design for this method would include a cuvette designed for high n_o . Additionally, the current model considers a cuvette

closed on all sides so that n_o is the same on all six faces. This is unrealistic for a typical cuvette which would have one side open to the air. When this method is implemented in practice, either a lid with the same material as the cuvette would need to be incorporated or the air boundary on one side considered. If a top air boundary is considered the $\text{SR}\{\tilde{T}\}$ s would not be the same in theory, but the SC/DS coupling cancellation would still apply. The model would need to be more complex since $\text{DR}\{\tilde{T}\}$ would no longer equal the theoretical $\text{SR}\{\tilde{T}\}$ but instead the average of the two since theoretical $\text{SR}\{\tilde{T}\}$ s would not be equal, but given the correct model an inversion is still expected to work. Further, we also note that examining the expression for $\tilde{\mu}_{eff}$ (Equation (11)) ones sees that μ_a and μ'_s are coupled to n_i . This means that any diffuse measurement using such theory would actually measure $\mu'_s n_i$ and μ_a/n_i . Therefore, cross-talk with n_i is a necessary consequence of the theory and can only really be suppressed, not removed. This is seen by examining the data-sheet for the SphereSpectro which utilizes the integrating sphere measurement method [14,15]. The SphereSpectro states that with a n uncertainty of 0.06 one should expect a μ_a uncertainty of 12% and a μ'_s of 7% [39]. Therefore, the method presented here is, at least in theory, comparable to existing instruments.

Finally, we revisit the idea of calibration. In Section 3.1.2, we showed that this method takes the advantages of SC [24]/DS [29,30] meaning that the measurements of $\ln|\text{DR}\{\tilde{T}\}|$ and $\angle\text{DR}\{\tilde{T}\}$ are insensitive to instrumental coupling. However, due to the fact that this method utilizes a small geometry and is highly affected by boundary conditions, other instrumental calibration may be required. First, since this diffusion theory solution (Equation (10)) is for such a small geometry, calibration of the inverse model maybe necessary if factors exist which are not modeled by \tilde{C} (Section 3.1.2). Three options are available for creating an inverse model:

- I Diffusion theory based cost minimization (shown here).
- II Look-up table with Monte-Carlo generated data.
- III Look-up table with instrumental measurements of known samples.

Option I is the most elegant which is why it was chosen here, but option III (being the most brute-force and likely infeasible in practice due to the extensive calibration phantom preparation and measurement needed for each unique instrument) would almost definitely work, and allows for correction of systematic confounds, provided that the measurements are repeatable. The auto-calibration in Section 3.1.2 is expected to significantly help with this repeatability, but the biggest secondary factor is repeatable boundary conditions. The investigation here showed promise to alleviate the boundary conditions issue by using high n_o , but future work will investigate the repeatability of measurements of cuvettes with various boundary conditions experimentally. Of course, this future work would involve experimental implementation of the measurement. For this we plan to utilize the Imagent which utilizes fiber bundles that will be coupled to the sides of a standard cuvette. Regardless of the coupling method, we expect the SC/DS method to compensate for optode and coupling losses. However, these optodes will not act as true pencil beams or point detectors as is modeled here. This is another condition which could cause errors in the measurement and would need to be modeled or calibrated for. Area detectors may be modeled with diffusion theory by integrating over the area of the detector while various types of sources could be modeled using Monte-Carlo. If such realistic forward models are still not enough to account for the practicals of the optodes themselves then option III may be needed. However, we emphasize that we expect that these considerations will be partially alleviated by the SC/DS method and its measurement symmetry.

5. Conclusions

The purpose of this article is to present and determine the feasibility as well as strengths and weaknesses of a method to measure diffuse absolute optical properties in a standard cuvette. The strengths of this method lie in the way it is posited, a way to measure absolute

diffuse optical properties in small samples for which no commercial instruments which utilize frequency-domain type measurements exist to our knowledge. Our intention is to expand this method to spectral measurements of absorption to recover chemical concentrations of a diffuse sample [28]. Two main limitations were found in this method: first, higher error in absorption properties compared to scattering; second, high dependence on the knowledge of the index of refraction of the sample. However, the investigation lead to possible methods to address or alleviate these limitations. That being, the measurement of strongly scattering samples to address the first limitation, and the use of a cuvette with high index of refraction to address the second. Future work will move beyond theoretical development of the method to experimental implementation, and an investigation of boundary conditions and repeatability which can really only be done in experimental practice.

Author Contributions: Conceptualization, G.B., A.S. and S.F.; methodology, G.B. and A.S.; software, G.B. and A.S.; validation, G.B. and A.S.; formal analysis, G.B.; investigation, G.B.; resources, A.S. and S.F.; data curation, G.B.; writing—original draft preparation, G.B.; writing—review and editing, G.B., A.S. and S.F.; visualization, G.B.; supervision, A.S. and S.F.; project administration, S.F.; funding acquisition, S.F. All authors have read and agreed to the published version of the manuscript.

Funding: This research was funded by National Institutes of Health (NIH) grant numbers R01-NS095334 and R01-EB029414.

Institutional Review Board Statement: Not applicable

Informed Consent Statement: Not applicable

Data Availability Statement: <https://github.com/DOIT-Lab/DOIT-Public/tree/master/OpticalPropertiesInCuvette> (25 October 2022).

Conflicts of Interest: The authors declare a current patent application regarding the method presented in this article.

Symbols

The following symbols are used in this manuscript:

References

1. Bigio, I.J.; Fantini, S. *Quantitative Biomedical Optics: Theory, Methods, and Applications*; Cambridge University Press: Cambridge, UK, 2016.
2. Quaresima, V.; Ferrari, M. Functional Near-Infrared Spectroscopy (fNIRS) for Assessing Cerebral Cortex Function During Human Behavior in Natural/Social Situations: A Concise Review. *Organ. Res. Methods* **2019**, *22*, 46–68. <https://doi.org/10.1177/1094428116658959>.
3. Rahman, M.A.; Siddik, A.B.; Ghosh, T.K.; Khanam, F.; Ahmad, M. A Narrative Review on Clinical Applications of fNIRS. *J. Digit. Imaging* **2020**, *33*, 1167–1184. <https://doi.org/10.1007/s10278-020-00387-1>.
4. Ozaki, Y.; McClure, W.F.; Christy, A.A. (Eds.) *Near-Infrared Spectroscopy in Food Science and Technology*; John Wiley & Sons: Hoboken, NJ, USA, 2006.
5. Johnson, J.B. An Overview of Near-Infrared Spectroscopy (NIRS) for the Detection of Insect Pests in Stored Grains. *J. Stored Prod. Res.* **2020**, *86*, 101558. <https://doi.org/10.1016/j.jspr.2019.101558>.
6. Kademi, H.I.; Ulusoy, B.H.; Hecer, C. Applications of Miniaturized and Portable near Infrared Spectroscopy (NIRS) for Inspection and Control of Meat and Meat Products. *Food Rev. Int.* **2019**, *35*, 201–220. <https://doi.org/10.1080/87559129.2018.1514624>.
7. Razuc, M.; Grafia, A.; Gallo, L.; Ramirez-Rigo, M.V.; Romañach, R.J. Near-Infrared Spectroscopic Applications in Pharmaceutical Particle Technology. *Drug Dev. Ind. Pharm.* **2019**, *45*, 1565–1589. <https://doi.org/10.1080/03639045.2019.1641510>.
8. Stranzinger, S.; Markl, D.; Khinast, J.G.; Paudel, A. Review of Sensing Technologies for Measuring Powder Density Variations during Pharmaceutical Solid Dosage Form Manufacturing. *TrAC Trends Anal. Chem.* **2021**, *135*, 116147. <https://doi.org/10.1016/j.trac.2020.116147>.
9. Chen, Z.; Gu, A.; Zhang, X.; Zhang, Z. Authentication and Inference of Seal Stamps on Chinese Traditional Painting by Using Multivariate Classification and Near-Infrared Spectroscopy. *Chemom. Intell. Lab. Syst.* **2017**, *171*, 226–233. <https://doi.org/10.1016/j.chemolab.2017.10.017>.

10. Trant, P.L.K.; Kristiansen, S.M.; Sindbæk, S.M. Visible Near-Infrared Spectroscopy as an Aid for Archaeological Interpretation. *Archaeol. Anthropol. Sci.* **2020**, *12*, 280. <https://doi.org/10.1007/s12520-020-01239-3>.
11. Tsuchikawa, S.; Kobori, H. A Review of Recent Application of near Infrared Spectroscopy to Wood Science and Technology. *J. Wood Sci.* **2015**, *61*, 213–220. <https://doi.org/10.1007/s10086-015-1467-x>.
12. Fantini, S.; Sassaroli, A. Frequency-Domain Techniques for Cerebral and Functional Near-Infrared Spectroscopy. *Front. Neurosci.* **2020**, *14*, 300. <https://doi.org/10.3389/fnins.2020.00300>.
13. Torricelli, A.; Contini, D.; Pifferi, A.; Caffini, M.; Re, R.; Zucchelli, L.; Spinelli, L. Time Domain Functional NIRS Imaging for Human Brain Mapping. *NeuroImage* **2014**, *85*, 28–50. <https://doi.org/10.1016/j.neuroimage.2013.05.106>.
14. Foschum, F.; Bergmann, F.; Kienle, A. Precise Determination of the Optical Properties of Turbid Media Using an Optimized Integrating Sphere and Advanced Monte Carlo Simulations. Part 1: Theory. *Appl. Opt.* **2020**, *59*, 3203–3215. <https://doi.org/10.1364/AO.386011>.
15. Bergmann, F.; Foschum, F.; Zuber, R.; Kienle, A. Precise Determination of the Optical Properties of Turbid Media Using an Optimized Integrating Sphere and Advanced Monte Carlo Simulations. Part 2: Experiments. *Appl. Opt.* **2020**, *59*, 3216–3226. <https://doi.org/10.1364/AO.385939>.
16. Contini, D.; Martelli, F.; Zaccanti, G. Photon Migration through a Turbid Slab Described by a Model Based on Diffusion Approximation. I. Theory. *Appl. Opt.* **1997**, *36*, 4587–4599. <https://doi.org/10.1364/AO.36.004587>.
17. Fang, Q.; Boas, D.A. Monte Carlo Simulation of Photon Migration in 3D Turbid Media Accelerated by Graphics Processing Units. *Opt. Express* **2009**, *17*, 20178–20190. <https://doi.org/10.1364/OE.17.020178>.
18. Hielscher, A.H.; Klose, A.D.; Scheel, A.K.; Moa-Anderson, B.; Backhaus, M.; Netz, U.; Beuthan, J. Sagittal Laser Optical Tomography for Imaging of Rheumatoid Finger Joints. *Phys. Med. Biol.* **2004**, *49*, 1147–1163. <https://doi.org/10.1088/0031-9155/49/7/005>.
19. Klose, A.D.; Ntziachristos, V.; Hielscher, A.H. The Inverse Source Problem Based on the Radiative Transfer Equation in Optical Molecular Imaging. *J. Comput. Phys.* **2005**, *202*, 323–345. <https://doi.org/10.1016/j.jcp.2004.07.008>.
20. Klose, A.D.; Larsen, E.W. Light Transport in Biological Tissue Based on the Simplified Spherical Harmonics Equations. *J. Comput. Phys.* **2006**, *220*, 441–470. <https://doi.org/10.1016/j.jcp.2006.07.007>.
21. Ho, J.H.; Chin, H.L.; Dong, J.; Lee, K. Multi-Harmonic Homodyne Approach for Optical Property Measurement of Turbid Medium in Transmission Geometry. *Opt. Commun.* **2012**, *285*, 2007–2011. <https://doi.org/10.1016/j.optcom.2011.12.018>.
22. Taniguchi, J.; Murata, H.; Okamura, Y. Light Diffusion Model for Determination of Optical Properties of Rectangular Parallelepiped Highly Scattering Media. *Appl. Opt.* **2007**, *46*, 2649–2655. <https://doi.org/10.1364/AO.46.002649>.
23. Kienle, A. Light diffusion through a turbid parallelepiped. *J. Opt. Soc. Am. A* **2005**, *22*, 1883–1888. <https://doi.org/10.1364/JOSAA.22.001883>.
24. Hueber, D.M.; Fantini, S.; Cerussi, A.E.; Barbieri, B.B. New Optical Probe Designs for Absolute (Self-Calibrating) NIR Tissue Hemoglobin Measurements. In *Optical Tomography and Spectroscopy of Tissue III*; San Jose, CA, USA; International Society for Optics and Photonics: Bellingham, WA, USA, 1999; Volume 3597, pp. 618–631. <https://doi.org/10.1117/12.356784>.
25. Bevilacqua, F.; Berger, A.J.; Cerussi, A.E.; Jakubowski, D.; Tromberg, B.J. Broadband Absorption Spectroscopy in Turbid Media by Combined Frequency-Domain and Steady-State Methods. *Appl. Opt.* **2000**, *39*, 6498–6507. <https://doi.org/10.1364/AO.39.006498>.
26. O'Sullivan, T.D.; Cerussi, A.E.; Cuccia, D.J.; Tromberg, B.J. Diffuse Optical Imaging Using Spatially and Temporally Modulated Light. *J. Biomed. Opt.* **2012**, *17*, 0713111. <https://doi.org/10.1117/1.JBO.17.7.071311>.
27. Vasudevan, S.; Forghani, F.; Campbell, C.; Bedford, S.; O'Sullivan, T.D. Method for Quantitative Broadband Diffuse Optical Spectroscopy of Tumor-Like Inclusions. *Appl. Sci.* **2020**, *10*, 1419. <https://doi.org/10.3390/app10041419>.
28. Blaney, G.; Curtsmith, P.; Sassaroli, A.; Fernandez, C.; Fantini, S. Broadband Absorption Spectroscopy of Heterogeneous Biological Tissue. *Appl. Opt.* **2021**, *60*, 7552–7562. <https://doi.org/10.1364/AO.431013>.
29. Sassaroli, A.; Blaney, G.; Fantini, S. Dual-Slope Method for Enhanced Depth Sensitivity in Diffuse Optical Spectroscopy. *J. Opt. Soc. Am. A* **2019**, *36*, 1743–1761. <https://doi.org/10.1364/JOSAA.36.001743>.
30. Blaney, G.; Sassaroli, A.; Pham, T.; Fernandez, C.; Fantini, S. Phase Dual-Slopes in Frequency-Domain near-Infrared Spectroscopy for Enhanced Sensitivity to Brain Tissue: First Applications to Human Subjects. *J. Biophotonics* **2020**, *13*, e201960018. <https://doi.org/10.1002/jbio.201960018>.
31. Fantini, S.; Hueber, D.; Franceschini, M.A.; Gratton, E.; Rosenfeld, W.; Stubblefield, P.G.; Maulik, D.; Stankovic, M.R. Non-Invasive Optical Monitoring of the Newborn Piglet Brain Using Continuous-Wave and Frequency-Domain Spectroscopy. *Phys. Med. Biol.* **1999**, *44*, 1543–1563. <https://doi.org/10.1088/0031-9155/44/6/308>.
32. Martelli, F.; Contini, D.; Taddeucci, A.; Zaccanti, G. Photon Migration through a Turbid Slab Described by a Model Based on Diffusion Approximation II Comparison with Monte Carlo Results. *Appl. Opt.* **1997**, *36*, 4600–4612. <https://doi.org/10.1364/AO.36.004600>.
33. Yang, L.; Yang, L.; Wabnitz, H.; Gladysz, T.; Macdonald, R.; Macdonald, R.; Grosenick, D. Spatially-Enhanced Time-Domain NIRS for Accurate Determination of Tissue Optical Properties. *Opt. Express* **2019**, *27*, 26415–26431. <https://doi.org/10.1364/OE.27.26415>.
34. Patterson, M.S.; Chance, B.; Wilson, B.C. Time Resolved Reflectance and Transmittance for the Noninvasive Measurement of Tissue Optical Properties. *Appl. Opt.* **1989**, *28*, 2331–2336. <https://doi.org/10.1364/AO.28.002331>.
35. Aronson, R. Boundary Conditions for Diffusion of Light. *JOSA A* **1995**, *12*, 2532–2539. <https://doi.org/10.1364/JOSAA.12.002532>.

-
36. Popescu, G.; Mujat, C.; Dogariu, A. Evidence of Scattering Anisotropy Effects on Boundary Conditions of the Diffusion Equation. *Phys. Rev. E* **2000**, *61*, 4523–4529. <https://doi.org/10.1103/PhysRevE.61.4523>.
 37. Ripoll, J.; Nieto-Vesperinas, M.; Arridge, S.R.; Dehghani, H. Boundary Conditions for Light Propagation in Diffusive Media with Nonscattering Regions. *JOSA A* **2000**, *17*, 1671–1681. <https://doi.org/10.1364/JOSAA.17.001671>.
 38. El-Wakil, S.A.; Degheidy, A.R.; Machali, H.M.; El-Depsy, A. Radiative Transfer in a Spherical Medium. *J. Quant. Spectrosc. Radiat. Transf.* **2001**, *69*, 49–59. [https://doi.org/10.1016/S0022-4073\(00\)00061-3](https://doi.org/10.1016/S0022-4073(00)00061-3).
 39. Optik, G. SphereSpectro 150H. Available online: <https://www.gigahertz-optik.com/en-us/product/spherespectro%20150h/getpdf/> (25 October 2022).