A ROBUST PIPELINE FOR CLASSIFICATION AND DETECTION OF BLEEDING FRAMES IN WIRELESS CAPSULE ENDOSCOPY USING SWIN TRANSFORMER AND RT-DETR

Sasidhar Alavala, Anil Kumar Vadde, Aparnamala Kancheti, Subrahmanyam Gorthi

Department of Electrical Engineering, IIT Tirupati, India

ansr2510@gmail.com, {ee21b050, ee21b023, s.gorthi}@iittp.ac.in

ABSTRACT

In this paper, we present our approach to the Auto WCEBleedGen Challenge V2 2024. Our solution combines the Swin Transformer for the initial classification of bleeding frames and RT-DETR for further detection of bleeding in Wireless Capsule Endoscopy (WCE), enhanced by a series of image preprocessing steps. These steps include converting images to Lab colour space, applying Contrast Limited Adaptive Histogram Equalization (CLAHE) for better contrast, and using Gaussian blur to suppress artefacts. The Swin Transformer utilizes a tiered architecture with shifted windows to efficiently manage self-attention calculations, focusing on local windows while enabling cross-window interactions. RT-DETR features an efficient hybrid encoder for fast processing of multi-scale features and an uncertainty-minimal query selection for enhanced accuracy. The class activation maps by Ablation-CAM are plausible to the model's decisions. On the validation set, this approach achieves a classification accuracy of 98.5% (best among the other state-of-theart models) compared to 91.7% without any pre-processing and an AP₅₀ of 66.7% compared to 65.0% with state-of-the-art YOLOv8. On the test set, this approach achieves a classification accuracy and F1 score of 87.0% and 89.0% respectively.

Index Terms— Wireless Capsule Endoscopy, Gastrointestinal Bleeding, Swin Transformer, Real-Time Detection Transformer, Interpretability

1. INTRODUCTION

The large volume of data generated in a WCE video and the nuances of the pathological features, such as bleeding, pose considerable challenges for manual analysis by a gastroenterologist. Therefore there is a need for automated, accurate, and efficient diagnostic tools [1]. The task in this challenge [2] is to develop an automatic pipeline which can classify each frame in a WCE video into bleeding or non-bleeding and then subsequently detect and segment the bleeding regions in that frame. Our approach to this challenge focused on two aspects:

- Data pre-processing and augmentation: To overcome some problems due to artefacts in WCE frames and also limited and repetitive images, we used some efficient preprocessing and different augmentation techniques.
- Pipeline design: To be able to accurately classify, detect and segment the bleeding frames we have used SOTA models SwinV2 [3, 4], RT-DETR [5] and SwinUNETR [6]. To the

best of our knowledge, this seems to be the first work to explore the usage of RT-DETR in medical imaging. We also incorporate Ablation-CAM [7] to provide visual explanations for the classification model.

According to the official evaluation of the challenge, our proposed pipeline achieved an accuracy of 0.88, an AP₅₀ of 70.0 and a dice score of 0.58 on test-2 data. The results show that our pipeline is robust and accurate in classifying and detecting bleeding frames and stands third among the participants.

2. METHODOLOGY

2.1. Data Preprocessing and Augmentation

These steps include converting RGB images to Lab colour space, applying Contrast Limited Adaptive Histogram Equalization (CLAHE) for better contrast and minimising the artefacts due to bubbles, and using Gaussian blur to suppress the sharp features in G and B channels to mitigate the issues associated with high values of component a that does not correspond to the bleeding in the frame [1]. We add several data augmentations to ensure the model's robustness since the data is limited. These include random horizontal and vertical flips, random rotations, gaussian blurring, random affine transformations, random perspective distortions, and MixUp as these augmentations.

2.2. Pipeline



Fig. 1. Proposed bleeding detection pipeline.

Our approach includes initially classifying the frame into bleeding or non-bleeding and then detecting and segmenting the bleeding regions in that frame.

SwinV2 and RT-DETR are employed for their robust performance in detecting and classifying intricate details within images.

The Auto-WCEBleedGen Challenge V1 2023 is part of the 8th International Conference on Computer Vision & Image Processing (CVIP-2023). Our solution stood first in that challenge.



Fig. 2. Block level RT-DETR architecture [5].

Ablation-CAM [7] in this pipeline introduces a layer of interpretability, offering visual explanations behind the classification decisions made by the model. While SwinUNETR [6] is used to segment the bleeding frames, it combines patch-based processing and transformer encoders with a U-Net style decoder

Real-Time Detection Transformer, RT-DETR [5], the current state-of-the-art end-to-end detection model, incorporates an efficient hybrid encoder that processes multi-scale features to speed up inference and introduces uncertainty-minimal query selection to improve initial query quality for the decoder, boosting overall accuracy. Attention-based Intra-scale Feature Interaction (AIFI) and the CNN-based Cross-scale Feature Fusion (CCFF) make up the encoder block optimizing the processing of multi-scale features.

2.3. Training Details

The input image size is 224x224 pixels trained with a single NVIDIA RTXA4000.

- 1. For SwinV2-T: Model is trained using Cross-Entropy loss and the AdamW optimizer with an initial learning rate of 0.0001, adjusted by a Cosine Annealing scheduler.
- 2. For RT-DETR: Model is trained using a combination of focal loss and a mix of L1 and generalized IoU losses for bounding box regression. The uncertainty in the object queries is also integrated into the loss function [5]. The training employs the AdamW optimizer with a specific learning rate strategy that includes a warm-up and subsequent decay phase.
- 3. For SwinUNETR: Model is trained using a combination of Cross-Entropy, Dice losses (with a weightage of 0.3 and 0.7 respectively) and the Adam optimizer with a learning rate of 0.0001.

3. RESULTS

Figure 3 shows the accuracy of the classification and detection model in predicting the bleeding regions and class activation maps also support our visual reasoning. Table 1 presents a comparison of classification results on the validation set, highlighting the impact of our preprocessing on model performance. The results are listed for several models, including EfficientNet-B5, ViT-B, DeIT-S, Swin-T, and SwinV2-T, with a clear indication of accuracy improvements. SwinV2-T model outperforms the others, achieving a remarkable accuracy of 98.5% with our preprocessing, compared to its initial 91.7% accuracy.

Table 2 presents a comparison of detection results on the validation set for some methods, comparing them based on the Average Precision (AP) metric at different Intersections over Union (IoU) thresholds. RT-DETR-R101 model demonstrates superior performance with an AP₅₀ of 66.7% and an AP_{50:95} of 81.0%, outpacing



Fig. 3. Visual results of our approach on test-1 marked images (top row) and test-2 unmarked images (bottom row).

the state-of-the-art YOLOv8-L. RT-DETR model not only has detection accuracy but also better convergence compared to the traditional DETR model. For segmentation, SwinUNETR yielded a dice score of 0.62 and an IoU of 0.48 on the validation set.

 Table 1. Classification results on validation set showing accuracy with and without our pre-processing.

Method	#param.	Accuracy%
EfficientNet-B5 [8]	25M	85.2/95.3
ViT-B [9]	40M	86.1/96.5
DeIT-S [10]	20M	86.2/96.9
Swin-T [3]	27M	88.7/97.8
SwinV2-T [4]	37M	91.7/98.5

Table 2.	Detection	results or	1 validation	set.

Method	#param.	Epochs	AP ₅₀ /AP _{50:95}
YOLOv5-L [11]	45M	150	63.2/75.6
YOLOv8-L [12]	43M	150	65.0/79.7
DETR-DC5-R101 [13]	58M	500	61.2/72.3
RT-DETR-R101 [5]	75M	150	66.7/81.0

4. CONCLUSION

In conclusion, our approach to this Auto-BleedGEN V2 2024 challenge, has been able to accurately classify, detect and segment the bleeding regions in WCE frames. The class activation maps also corroborate the model's decision-making process, aligning closely with the identified areas of interest. Future work includes addressing the cases where the model failed (including when blood vessels are prominent, small bleeding regions, and illumination conditions) and also addressing the multiple false positives in detection.

5. ACKNOWLEDGEMENT

This research has received partial support through project grant number 2022-IRP-26694272 from the Semiconductor Research Corporation (SRC) under the India Research Program (IRP).

6. REFERENCES

- [1] Jorge Bernal and Aymeric Histace, *Computer-Aided Analysis* of *Gastrointestinal Videos*, Springer, 2021.
- [2] Palak Handa, Divyansh Nautiyal, Deepti Chhabra, Manas Dhir, Anushka Saini, Shreshtha Jha, Harshita Mangotra, Nishu Pandey, Advika Thakur, et al., "Auto-wcebleedgen version v1 and v2: Challenge, datasets and evaluation," *Authorea Preprints*, 2024.
- [3] Ze Liu, Yutong Lin, Yue Cao, Han Hu, Yixuan Wei, Zheng Zhang, Stephen Lin, and Baining Guo, "Swin transformer: Hierarchical vision transformer using shifted windows," in *Proceedings of the IEEE/CVF international conference on computer vision*, 2021, pp. 10012–10022.
- [4] Ze Liu, Han Hu, Yutong Lin, Zhuliang Yao, Zhenda Xie, Yixuan Wei, Jia Ning, Yue Cao, Zheng Zhang, Li Dong, et al., "Swin transformer v2: Scaling up capacity and resolution," in Proceedings of the IEEE/CVF conference on computer vision and pattern recognition, 2022, pp. 12009–12019.
- [5] Wenyu Lv, Shangliang Xu, Yian Zhao, Guanzhong Wang, Jinman Wei, Cheng Cui, Yuning Du, Qingqing Dang, and Yi Liu, "Detrs beat yolos on real-time object detection," *arXiv preprint arXiv:2304.08069*, 2023.
- [6] Ali Hatamizadeh, Vishwesh Nath, Yucheng Tang, Dong Yang, Holger R Roth, and Daguang Xu, "Swin unetr: Swin transformers for semantic segmentation of brain tumors in mri images," in *International MICCAI Brainlesion Workshop*. Springer, 2021, pp. 272–284.
- [7] Harish Guruprasad Ramaswamy et al., "Ablation-cam: Visual explanations for deep convolutional network via gradient-free localization," in *proceedings of the IEEE/CVF winter conference on applications of computer vision*, 2020, pp. 983–991.
- [8] Mingxing Tan and Quoc Le, "Efficientnet: Rethinking model scaling for convolutional neural networks," in *International conference on machine learning*. PMLR, 2019, pp. 6105–6114.
- [9] Alexey Dosovitskiy, Lucas Beyer, Alexander Kolesnikov, Dirk Weissenborn, Xiaohua Zhai, Thomas Unterthiner, Mostafa Dehghani, Matthias Minderer, Georg Heigold, Sylvain Gelly, et al., "An image is worth 16x16 words: Transformers for image recognition at scale," *arXiv preprint arXiv:2010.11929*, 2020.
- [10] Hugo Touvron, Matthieu Cord, Matthijs Douze, Francisco Massa, Alexandre Sablayrolles, and Hervé Jégou, "Training data-efficient image transformers & distillation through attention," in *International conference on machine learning*. PMLR, 2021, pp. 10347–10357.
- [11] Glenn Jocher, "Ultralytics yolov5," in https://github.com/ultralytics/yolov5, 2020.
- [12] Glenn Jocher, Ayush Chaurasia, and Jing Qiu, "Ultralytics yolov8," in https://github.com/ultralytics/ultralytics, 2023.

[13] Nicolas Carion, Francisco Massa, Gabriel Synnaeve, Nicolas Usunier, Alexander Kirillov, and Sergey Zagoruyko, "End-toend object detection with transformers," in *European conference on computer vision*. Springer, 2020, pp. 213–229.