Bayesian Model Parameter Learning in Linear Inverse Problems: Application in EEG Focal Source Imaging

Alexandra Koulouri¹ and Ville Rimpiläinen^{2,3}

¹ Tampere University, Centre of Mathematics, Faculty of Information Technology and Communication Sciences, Tampere, Finland

² University of Liverpool, Department of Physics, Liverpool, United Kingdom

³ University of Bristol, School of Physics, Bristol, United Kingdom

E-mail: a.koulouri84@gmail.com, vrimpila@gmail.com

Abstract. Inverse problems are often described as limited-data problems in which the signal of interest cannot be observed directly. Therefore, a physics-based forward model that relates the signal with the observations is typically needed. Unfortunately, unknown model parameters and imperfect forward models can still undermine the signal recovery. Even though supervised machine learning techniques offer promising avenues to improve the robustness of the solutions, we have to rely on modelbased learning when there is no access to ground truth for the training. In this work, we studied a linear inverse problem that included an unknown non-linearly related model parameter and utilized a Bayesian model-based learning approach that allowed reliable signal recovery and subsequently estimation of the unknown model parameter. This approach, often referred to as Bayesian Approximation Error approach, employed a simplified model of the physics of the problem augmented with an approximation error term that compensated for the simplification. An error subspace was spanned with the help of the eigenvectors of the approximation error covariance matrix which allowed, alongside the primary signal, simultaneous estimation of the induced error. The estimated error and signal were then used to determine the unknown model parameter. For the model parameter estimation, we tested several different approaches: a conditional Gaussian regression, an iterative (model-based) optimization, and a Gaussian process that was modeled with the help of physicsinformed learning. In addition, alternating optimization was used as a reference method. As an example application, we focused on the problem of reconstructing brain activity from EEG recordings (a.k.a. EEG source imaging) under the condition that the electrical conductivity of the patient's skull was unknown in the model. Poorly selected conductivity values cause well-documented artifacts in the EEG source imaging results, and the determination of patient-specific head tissue conductivities is a significant technical problem. Our results demonstrated clear improvements in EEG source localization accuracy and provided feasible estimates for the unknown model parameter, skull conductivity.

skull conductivity

1. Introduction

1.1. Model Uncertainties in Inverse Problems

In many biomedical imaging, seismic imaging, remote sensing, and tomographic applications, the features of interest cannot be observed directly but must be inferred from other measurable quantities. This is known as an inverse problem [1]. The limited, complicated, or distorted nature of observations, coupled with inherent limitations such as large null spaces in model operators, often make inverse problems ill-posed [2]. To address this, standard regularization techniques are often used [3, 4, 5]. Strategies to learn effective regularizers directly from the data are a promising field of research [6, 7, 8, 9]. However, reliable and accurate computational models that relate the features of interest and the observations are also needed. Unfortunately, uncertainties are usually present in these models due to the modeling formulation itself (such as discretization that is required to achieve numerical solutions), imperfect sensing, or lack of knowledge of suitable model parameter values, which can significantly affect the final solution.

In applications like computerized tomography and magnetic resonance imaging [10, 11], well-established methods are employed to tackle unknown model parameters. In particular, for the estimation of signals of interest and model parameters, the corresponding bilinear optimization problems can be approached by convexifying the problem using methods such as linearization (lifting) [12, 13, 11] or by alternating minimization algorithms [14, 15]. However, optimizing both the primary unknowns and the model parameters can be time consuming, convergence is not always guaranteed, or several reinitializations may be required, and approximations for parameters that are deep in the model may not always work. Techniques that handle model uncertainties with the help of machine learning (ML) and data-driven approaches [16, 17, 18] represent a new direction to accelerate computations and improve estimates in inverse problems [19, 20, 21]. However, unlike in many image and pattern recognition problems [22, 23], in inverse problems there is often a limited or no access to ground truth data to train ML algorithms and deep neural networks [10, 23]. In this study, we revisit the application of machine learning to severely ill-posed inverse problems when no access to ground truth is possible. Specifically, we generate training data stemming from physics-based models and integrate this domain-specific knowledge into the inverse problem formulation and recover both the primary unknown and extract an estimate for the (unknown) model parameter. We approach the problem from a Bayesian perspective, which facilitates the design of optimization algorithms under model uncertainty [24, 25].

As an application, we will focus on imaging brain sources from EEG recordings, due to its importance as a low-cost diagnostic tool and the inherent mathematical challenges and associated model uncertainties. In particular, we study imaging under the influence of erroneous skull conductivity (too high or too low), which has been shown to significantly affect the imaging solution [26, 27, 28, 29, 30, 31, 32]. Subsequently, the solved conductivity value could be used in further EEG studies and in the optimization/planning of transcranial electrical brain stimulation treatment [33, 34, 35]. We remark that the developments of this paper are not limited to EEG source imaging but can also be used, for example, in other imaging and deconvolution problems.

1.2. Bayesian Approximation Error Approach and Our Contributions

In this work, we aim to bring Bayesian statistics, which is not often explicitly used or contributed, into machine learning and provide methodological ways to efficiently solve linear inverse problems with model parameter uncertainties. In particular, our developments are based on the well-known Bayesian approximation error (BAE) approach [24, 36] in which a *standard* (approximate) observation model is used in the imaging accompanied with an additive error term that takes into account the effects of the approximation. Similarly as in [37, 36], we first derive a posterior distribution which simultaneously predicts the primary unknown signal and a low-order representation of the model-induced error. For the representation, we employ the top-eigenvectors of the modeling error covariance, obtained through Monte Carlo simulations. In addition to this, we derive an analytical expression for the modeling error covariance to investigate the connections between the error, model parameter, and primary signal. The provided insight allowed us to propose other Bayesian strategies to infer the model parameter.

In the application domain, we evolve our previous work, in which we demonstrated that BAE can improve source localization when skull conductivity was unknown [31], and our preliminary study on simultaneous estimation of focal source activity and skull conductivity from EEG recordings [38]. Here, with the help of simulated EEG data that corresponded to focal source activity in the somatosensory brain area, we demonstrate the potential of our approach to reconstruct the underlying focal sources and low-order estimates for the errors induced by unknown skull conductivity. Subsequently, the estimated errors are used to approximate the skull conductivity.

Our contributions are twofold. First, we propose a Bayesian model-based learning technique for linear inverse problems with unknown model parameters; second, we present different ways to estimate both the primary unknown and the (unknown) model parameter. We demonstrate the approach by imaging the source configuration (primary unknown) of the EEG brain imaging problem and estimating the unknown skull conductivity (model parameter) in the observation model.

Although the BAE approach has been studied in a variety of settings [39, 40, 41, 42, 43, 44, 45, 46, 47, 48], its use in the estimation of model parameters has been rather limited [37]. Therefore, in this paper, we concentrate on the model parameter learning aspect of BAE, offer insight on the conditions when it works, and propose two new ways to improve the model parameter estimates. Namely, we first propose to replace the previously used linear regression, that is based on a conditional Gaussian

(CG) approximation [37], with an iterative algorithm, and second, we propose to use a Gaussian process (GP) to infer the unknown model parameter. We show that these new developments are superior to the CG as the boundary data depend non-linearly on the model parameter. Finally, we compare our methodology with other standard approaches dealing with blind inverse problems (in which model parameters are partly or fully unknown) and show that the proposed methodologies, which only rely on precomputed training data and off-line statistical learning, can outperform alternating optimization methods when unknown parameters are deep in the model.

2. Background

2.1. Bayesian Inference in Linear Inverse Problems

Let us assume the discrete observation model

$$v = A(\sigma)x + e,\tag{1}$$

where $v \in \mathbb{R}^m$ are the measurements, m is the number of measurements, $A(\sigma) \in \mathbb{R}^{m \times n}$ $(m \ll n)$ is the system matrix (discrete forward model) that depends on a model parameter $\sigma \in \mathbb{R}^{N_{\sigma}}$, $x \in \mathbb{R}^n$ is the distributed signal (e.g. amplitudes in n locations) that we wish to solve, and e is the measurement noise, which is modeled as Gaussian $e \sim \mathcal{N}(e_*, \Gamma_e)$ with mean e_* and covariance Γ_e . The model parameter σ is usually (unrealistically) considered known. However, in this paper, we treat σ as unknown, and subsequently estimate it alongside our *primary* unknown x.

In the Bayesian framework, the inverse solution is the posterior density

$$\pi(x|v) \propto \pi(v|x)\pi(x),\tag{2}$$

where $\pi(v|x)$ is the likelihood and $\pi(x)$ the prior.

For the model (1), the likelihood can be written as

$$\pi(v|x) \propto \exp\Big(-\frac{1}{2}(v - A(\sigma)x - e_*)^{\mathrm{T}}\Gamma_e^{-1}(v - A(\sigma)x - e_*)\Big).$$
(3)

The model parameter σ is usually unknown and we want to obtain information on it, in addition to x. These problems are often called blind inverse problems. Such problems are much harder to solve than standard inverse problems. One approach is to estimate the parameters of the model and the signal by alternating optimizations, but this has certain limitations; namely, convergence is not always reached, and the computations can be time-consuming. Ideally, we would like to marginalize out the unknown parameter σ ; however this is not usually tractable. Hence, instead of the accurate matrix $A(\sigma)$, an approximate (standard) matrix $A_0 = A(\sigma_0)$ with fixed parameters σ_0 is often used in the inversion.

2.2. Bayesian Machine Learning

Machine learning in inverse problems relies on training a model with known data, $D = \{(x_s, y_s)\}_{s=1:S}$, in order to predict underlying signals x given a new set of observations, y. In Bayesian ML, model parameter σ is a random variable with distribution $\pi(\sigma|D)$ and the likelihood is expressed as

$$\pi(y|x) = \int \pi(y|x, e, \sigma) \pi(e) \pi(\sigma|D) d\sigma de.$$
(4)

To estimate the primary unknown x, we have to estimate the previous likelihood; however, marginalizing out directly the unknown σ parameter is not always tractable (especially for non-linear parameters). Even if we approximate $\pi(\sigma|D) = \delta(\sigma - \hat{\sigma})$, the question on how we estimate $\hat{\sigma}$ still remains open. Furthermore, in such inverse problems where the access to ground truth distribution for x is (most of the time) impossible, we have to rely on training techniques that employ model-based learning and sampling distributions. However, to guarantee that feasible model parameters have been selected, the training data D (obtained from model distributions) has to be produced under carefully designed simulations.

3. Bayesian Approximation Error Approach

3.1. Overview

To ease the model parameter estimation, we split the problem in two parts. First, we rewrite the observation model (1) with the help of an approximate model $A_0 = A(\sigma_0)$, where σ_0 has fixed values, in such a way that we remove the non-linearity with respect to the model parameter as

$$v = A_0 x + \varepsilon + e. \tag{5}$$

Here, we added an error term (to compensate for the simplification),

$$\varepsilon(x,\sigma) = A(\sigma)x - A_0x,\tag{6}$$

which is referred to as approximation error, $\varepsilon \in \mathbb{R}^m$ [24, 36]. This linear error term enables the simplification of the likelihood function (linearization of the problem) under the assumption that this error term can be expressed as a linear combination of few basis functions, i.e. $\varepsilon = W\alpha + \varepsilon_*$, where ε_* is the mean of ε , and $W \in \mathbb{R}^{m \times p}$ (p < m) are basis functions (obtained as in [37, 36]), we can infer x and α based on Bayes' rule, $\pi(x, \alpha | v) \propto \pi(v | x, \alpha) \pi(x, \alpha)$.

Then, one way to obtain parameter estimates is by maximizing the conditional probability distribution $\hat{\sigma} := \max_{\sigma} \pi(\sigma | \alpha)$, considering a joint Gaussian for (σ, α) as has been done previously [37]. However, this approach has limitations, since it considers that σ and α are sensibly related through a linear relationship (or at least monotonic relationship), which may not be valid, and the approach also ignores the fact that α may

depend on the signal x. This approximation can work usually well when the difference $\sigma - \sigma_0$ is small.

In a more general case, we should consider that there is a relationship $f(\alpha, x) = \sigma$. Since the analytical expression of the function f is often unknown (or too complicated to be derived), we can use non-parametric methods, such as a Gaussian process, for the Bayesian inference of the model parameter. Overall, as we will see later by approximating (6) and obtaining statistical knowledge through training data, we can recover meaningful estimates $\hat{\sigma}$ under certain conditions.

3.2. Approximation Error through Basis Functions

A feasible set of basis functions that assigns most of the variance of ε in the first few terms can be chosen based on the eigenvalue decomposition of the approximation error covariance $\Gamma_{\varepsilon} = \mathbb{E}[(\varepsilon - \varepsilon_*)(\varepsilon - \varepsilon_*)^T]$, where ε_* is the mean of the modeling error. In particular, Γ_{ε} can be decomposed according to

$$\Gamma_{\varepsilon} = \Sigma_{k=1}^{m} \lambda_k w_k w_k^{\mathrm{T}} = W \Lambda W^{\mathrm{T}}, \tag{7}$$

where $\lambda_k \in \mathbb{R}^m$ are the eigenvalues and $w_k \in \mathbb{R}^m$ are the eigenvectors of Γ_{ε} [37, 36]. Based on this decomposition, we have that

$$\varepsilon - \varepsilon_* \in \operatorname{span}\{w_1, \dots, w_k\}.$$
 (8)

Hence, the *approximation error* can be expressed as

$$\varepsilon = \varepsilon_* + \varepsilon' + \varepsilon'',\tag{9}$$

where $\varepsilon' = \sum_{k=1}^{p} \alpha_k w_k$, $\varepsilon'' = \sum_{j=p+1}^{m} \beta_j w_j$, with $\mathbb{E}[\varepsilon'] = 0$, $\mathbb{E}[\varepsilon''] = 0$ and the coefficients α_k and β_j are given by the inner products $\alpha_k = \langle \varepsilon - \varepsilon_*, w_k \rangle$ and $\beta_j = \langle \varepsilon - \varepsilon_*, w_j \rangle$ [37, 36]. The observation model (5) can be rewritten as

$$v = A_0 x + \varepsilon_* + W_p \alpha + \varepsilon'' + e, \tag{10}$$

where $\varepsilon' = W_p \alpha$, $W_p = [w_1, w_2, ..., w_p] \in \mathbb{R}^{m \times p}$ contains the first p eigenvectors and $\alpha = (\alpha_1, \alpha_2, ..., \alpha_p)^{\mathrm{T}} \in \mathbb{R}^p$ [37, 36]. Note that $\varepsilon'' = Q\beta = QQ^{\mathrm{T}}\varepsilon$ where $Q = [w_{p+1}, ..., w_m] \in \mathbb{R}^{(m-p) \times m}$ and $\operatorname{var}(\alpha_p) = \lambda_p$. In practice, the coefficients of W can be learned from training data obtained from model distributions $\pi(x)$ and $\pi(\sigma)$ or semi-analytically as we will explain next.

3.2.1. Semi-Analytical Covariance Matrix for Approximation Error The approximation error (6) depends on both, $\sigma \in \mathbb{R}^+$ and signal x. Therefore, to understand how the linear and non-linear effects of σ and x are incorporated in the basis functions W, we express the error covariance semi-analytically with the help of the Taylor expansion of the error (6). In particular, the Taylor expansion around the mean σ_* is

$$\varepsilon(x,\sigma) = A(\sigma_*)x - A_0x + \sum_k \frac{\partial^k A(\sigma_*)}{\partial \sigma^k} (\sigma - \sigma_*)^k x = \sum_k J_k(\sigma_*) z_k(\sigma) x, \quad (11)$$

where $J_k = \frac{\partial^k A(\sigma)}{\partial \sigma^k}|_{\sigma=\sigma_*} \in \mathbb{R}^{m \times n}$, $z_k(\sigma) = (\sigma - \sigma_*)^k \mathbf{1}_n$, $\mathbf{1}_n$ is a vector of ones of size n, and we have assumed $\sigma_* = \sigma_0$ for simplicity. Since z_k and x are statistically independent, according to [49], we have that the covariance matrix of a product of two independent variables is

$$\operatorname{cov}[xz_k] = \operatorname{cov}[x]\operatorname{cov}[z_k] + \operatorname{cov}[x]\mathbb{E}[z_k]\mathbb{E}[z_k]^{\mathrm{T}} + \operatorname{cov}[z_k]\mathbb{E}[x]\mathbb{E}[x]^{\mathrm{T}}.$$
 (12)

Therefore, the error covariance can be expressed as

$$\Gamma_{\varepsilon} = \sum_{k} J_k \text{cov}[xz_k] J_k^{\mathrm{T}}.$$
(13)

If we assume $x \sim \mathcal{N}(0, \gamma I_n)$, then we have that

$$\Gamma_{\varepsilon} = \gamma \sum_{k} J_{k}(\operatorname{cov}[z_{k}]) + \mathbb{E}[z_{k}]\mathbb{E}[z_{k}]^{\mathrm{T}})J_{k}^{\mathrm{T}}.$$
(14)

Based on the previous expression of the error covariance, we can observe that if an i.i.d Gaussian prior for the signal x is assumed, then x only has a scaling effect on the error covariance.

In general, we need to analyze the properties of each studied problem to determine whether the matrix J_k implicitly encodes information about x (potentially coming from the modeling, e.g., linearization of a nonlinear problem). Thus, the inference of the model parameter σ from the estimated error ε' (as in [37]) is feasible if the basis functions W primarily reflect the error induced by the unknown model parameter σ . However, this is not always the case, and in some cases also the signal x must be considered when estimating σ .

3.3. Bayesian Inference of Primary Signal and Approximation Error

For the simultaneous estimation of x and α (and $\varepsilon \approx W\alpha + \varepsilon_*$), we have to construct the posterior model $\pi(x, \alpha | v)$. To obtain a computationally efficient solution, we make the technical approximation that $(x, \alpha, e, \varepsilon'')$ are mutually Gaussian and uncorrelated [37, 36]. Then, we obtain the approximate likelihood

$$\tilde{\pi}(v|x,\alpha) \propto \exp\left(-\frac{1}{2}(v-A_0x-\varepsilon_*-W_p\alpha-e_*)^{\mathrm{T}}\right)$$
$$(\Gamma_{\varepsilon''}+\Gamma_e)^{-1}(v-A_0x-\varepsilon_*-W_p\alpha-e_*),$$
(15)

where $\Gamma_{\varepsilon''} = \sum_{p=1}^{m} \lambda_j w_j w_j^{\mathrm{T}}$. Thus, the posterior density becomes

$$\tilde{\pi}(x,\alpha|v) \propto \exp\left(-\frac{1}{2}L_{\varepsilon''+e}\|v-A_0x-W\alpha-\varepsilon_*-e_*\|_2^2\right) \pi(x,\alpha),$$
(16)

where the Cholesky factors are $(\Gamma_{\varepsilon''} + \Gamma_e)^{-1} = L^{\mathrm{T}}_{\varepsilon''+e} L_{\varepsilon''+e}$. Then, we can estimate the maximum a posteriori (MAP) estimate

$$(\hat{x}, \hat{\alpha}) \leftarrow \min_{x, \alpha} \left\{ \|\frac{1}{2} L_{\varepsilon''+e} (A_0 x + W_p \alpha + e_* + \varepsilon_* - v)\|_2^2 - \ln(\pi(x, \alpha)) \right\}.$$
(17)

[†]We note that even though in reality approximation errors usually depend on x, this approximation often leads to very similar inverse solutions [24, 25, 36].

3.4. Inference of the Model Parameter

3.4.1. Conditional Gaussian Approximation In Bayesian framework, we expect to obtain model parameter estimates by maximizing the conditional probability distribution

$$\hat{\sigma} := \max_{\sigma} \pi(\sigma | \alpha). \tag{18}$$

In particular, the conditional probability distribution $\pi(\sigma|\alpha)$ can be expressed as

$$\pi(\sigma|\alpha) = \int \pi(\sigma, x|\alpha) \, \mathrm{d}x,\tag{19}$$

where σ and x are mutually independent, and $\pi(\sigma)$ can be modeled as a Gaussian distribution with mean σ_* and variance γ_{σ} . Even when both $\pi(\sigma)$ and $\pi_x(.)$ are modeled as Gaussian, the relationship between x, σ and α can be complicated and cause difficulties for Bayesian inference.

A simple approach is to approximate $\pi(\sigma|\alpha) \propto \pi(\sigma,\alpha)$ and rely on Monte Carlo simulations [37]. In particular, a joint Gaussian distribution for the pair (α, σ) can be a good approximation for small perturbations $\sigma - \sigma_*$. In other words, a first order Taylor approximation with only the linear term in (11) can be sufficient to describe the modeling error. Therefore, given an estimate for the error coefficients $\hat{\alpha}$, the mean of $\pi(\sigma|\hat{\alpha})$ is given by

$$\hat{\sigma}_{*|\alpha} = \sigma_* + \Gamma_{\sigma\alpha} \Gamma_{\alpha}^{-1} \hat{\alpha}, \tag{20}$$

where σ_* is the mean value of the postulated model parameter distribution, $\Gamma_{\sigma\alpha}$ is the cross-covariance between σ and α (estimated using samples $\alpha^{(s)}$ and $\sigma^{(s)}$) and $\Gamma_{\alpha} = \text{diag}\{\lambda_1, \ldots, \lambda_p\}.$

As we shall see later (in Section 4.2.2), in some cases the CG-based parameter estimation can be accompanied with an iterative algorithm in order to improve the estimates.

3.4.2. Gaussian Process In more general cases, variables (x, α) and σ can be nonlinearly related, and therefore we should consider $\sigma = f(\alpha, x)$, where f is an unknown function. To infer σ , we can model f to be distributed as a Gaussian process, $f \sim \mathcal{GP}(m, K)$, with mean function $m(\alpha, x) = \mathbb{E}[f(\alpha, x)]$ and covariance (or kernel) function $K_{(\alpha,x)} = \mathbb{E}[(f(\alpha,x)) - m(\alpha,x))((f(\alpha',x')) - m(\alpha',x'))^{\mathrm{T}}]$ [23]. In practice, based on training input-output data, $\mathcal{D} = \{x^{(s)}, \alpha^{(s)}\}_{s=1}^{N_s}$ and $f_{\mathcal{D}} = \{\sigma^{(s)}\}_{s=1}^{N_s}$, respectively, we can infer a new model parameter value σ given a new input estimated from (16). Thus, we can predict $\hat{\sigma}$ given $\hat{\mathcal{D}} = (\hat{\alpha}, \hat{x})$ from (17) and the training data \mathcal{D} . The predictive distribution is $\pi(f_{\hat{\mathcal{D}}}|\hat{\mathcal{D}}, \mathcal{D}, f_{\mathcal{D}}) \sim \mathcal{N}(\hat{f}_{\hat{\mathcal{D}}|\mathcal{D}}, K_{\hat{\mathcal{D}}|\mathcal{D}})$ where

$$\hat{f}_{\hat{\mathcal{D}}|\mathcal{D}} = m(\hat{D}) + K_{\hat{\mathcal{D}},\mathcal{D}} K_{\mathcal{D}}^{-1} (f_{\mathcal{D}} - m(\mathcal{D}))$$
(21)

is the conditional mean based on the training set, $K_{\hat{\mathcal{D}}|\mathcal{D}} = K_{\hat{\mathcal{D}}} - K_{\hat{\mathcal{D}},\mathcal{D}} K_{\mathcal{D}}^{-1} K_{\hat{\mathcal{D}},\mathcal{D}}^{\mathrm{T}}$ is the conditional variance, $K_{\hat{\mathcal{D}},\mathcal{D}}$ is a vector of covariances between every training case and

4. Application in EEG Source Imaging

4.1. Simultaneous Approximation Error and Source Estimation using Model-based Learning and Dipole Scan

In this work, we apply the described Bayesian framework for the reconstruction of the unknown skull conductivity (model parameter) in the EEG source imaging problem alongside the source activity. Here, we consider the distributed source modeling [50] where the primary unknown $x \in \mathbb{R}^{3n}$ is the distributed (electrical) current dipole source configuration (or field) in a source space that consists of n discrete locations, $|x_i| = \sqrt{x_{i1}^2 + x_{i2}^2 + x_{i3}^2}$ is the amplitude of the source at location i, and (x_{i1}, x_{i2}, x_{i3}) are the components of x_i along the coordinate axes. The observation model is linear, as Equation (1), where $v \in \mathbb{R}^m$ are the EEG recordings, and the system matrix $A(\sigma) \in \mathbb{R}^{m \times 3n}$ is called the *leadfield* matrix, which depends on the model parameter, electrical conductivity of the skull $\sigma \in \mathbb{R}^+$.

To obtain estimates for the source and approximation error, we first compute the maximum a posteriori (MAP) estimates of the posterior (16). Here, we utilize the dipole scan solver which is a very commonly used approach in EEG studies [51]. In the dipole scan algorithm, the main assumption is that only a single dipole source is active at a time.

Hence, the EEG observation model can be written as

$$v = A_0^i x_i + W_i \alpha_i + \varepsilon_{i*} + \varepsilon_i'' + e.$$
⁽²²⁾

The subscript *i* denotes a specific dipole location and the superscript in $A^i(\sigma)$ denotes the *i*1, *i*2, and *i*3 columns of the leadfield matrix. The approximation error covariance matrix related to the dipole at location *i* is denoted by $\Gamma_{\varepsilon_i} = \mathbb{E}[(\varepsilon_i - \varepsilon_{i*})(\varepsilon_i - \varepsilon_{i*})^T]$ where ε_{i*} is the mean approximation error at location *i*. Then, the primary error is $\varepsilon'_i = W_i \alpha_i$, where W_i are the first *p* eigenvectors corresponding to the highest eigenvalues obtained from the eigenvalue decomposition of Γ_{ε_i} .

Based on the model (22), we can compute a MAP estimate of the pair (x_i, α_i) at each location *i*, and then select as a solution the pair that maximizes the posterior (16). Therefore, given location *i*, we solve the linear system

$$\begin{bmatrix} \hat{x}_i \\ \hat{\alpha}_i \end{bmatrix} := \begin{bmatrix} L_{\varepsilon_i''+e} A_0^i & L_{\varepsilon_i''+e} W_i \\ \mathbf{0} & L_{\alpha_i} \end{bmatrix}^{-1} \begin{bmatrix} L_{\varepsilon_i''+e} (v - \varepsilon_{i*} - e_*) \\ 0 \end{bmatrix}, \quad \text{for } i = 1, \dots, n, \qquad (23)$$

where *n* is the total number of source locations, \hat{x}_i is the dipole source reconstructed at location *i*, $L_{\varepsilon_i''+e}$ comes the Cholesky factorization of $(\Gamma_{\varepsilon_i''} + \Gamma_e)^{-1}$, $W_i \in \mathbb{R}^{m \times p}$ are the eigenvectors corresponding to the *p* largest eigenvalues obtained from the eigenvalue decomposition of the error covariance $\Gamma_{\varepsilon_i} \in \mathbb{R}^{m \times m}$, $\alpha_i \in \mathbb{R}^p$ are the error coefficients at this location, and $L_{\alpha_i} = \frac{1}{2} \text{diag}\{\lambda_{i1}^{1/2}, \dots, \lambda_{ip}^{1/2}\}$, where λ_{ip} are the *p* largest eigenvalues of the covariance Γ_{ε_i} . The solution is the pair

$$(\hat{x}_{l}, \hat{\alpha}_{l}) \leftarrow \min_{i=1:n} \{ \|L_{\varepsilon_{i}''+e+\delta_{i}}(v - A_{0}^{i}\hat{x}_{i} - \varepsilon_{i*} - W_{i}\hat{\alpha}_{i} - e_{*})\|_{2}^{2} + \|L_{\alpha_{i}}\hat{\alpha}_{i}\|_{2}^{2} \},$$
(24)

where index l refers to the location that gives the lowest value for the functional (24), and the corresponding approximation error is given by $\hat{\varepsilon}'_l = W_l \hat{\alpha}_l$. In the following subsections, we analyze the strategies for estimating σ given the pair $(\hat{x}_l, \hat{\alpha}_l)$.

4.2. Skull Conductivity Estimation

To obtain sensible estimates for the skull conductivity (model parameter), the estimated $\hat{\alpha}$ (and the corresponding modeling error) has to be clearly correlated with σ . To analyze this, we inspect first the covariance matrix of the approximation error at location l that is analytically given by

$$\Gamma_{\varepsilon_l} = \gamma \left(\operatorname{var}(\sigma) J_1^l J_1^{l^{\mathrm{T}}} + \sum_{k=2} (\operatorname{cov}[z^k]) + \mathbb{E}[z^k] \mathbb{E}[z^k]^{\mathrm{T}} J_k^l J_k^{l^{\mathrm{T}}} \right),$$
(25)

where $z_k = (\sigma - \sigma_*)^k$, $J_k^l = \frac{\partial^k A^l(\sigma_*)}{\partial \sigma^k} \in \mathbb{R}^m$ is the *k*th derivative of the leadfield $A^l(\sigma)$ and γ is the dipole variance. Here, without a loss of generality, we have considered that the standard conductivity is equal to the mean conductivity of the training data $\pi(\sigma)$, i.e. $\sigma_0 \approx \sigma_*$, to ease our analysis.

Based on the previous expression for the error covariance, we see that x_i has a scaling effect on the error covariance, and the basis W_i from the eigenvalue decomposition of the approximation error covariance matrix are primarily related to the model parameter σ . So, we can expect that the error coefficients α will be proportional to the amplitude of x_i . Therefore, we can suspect that the methods that employ only the estimated error $\hat{\alpha}_l$ to infer σ , as the conditional Gaussian approximation described below, will not perform as well as the methods that include both, $\hat{\alpha}_l$ and the estimated source \hat{x}_l as inputs.

4.2.1. Conditional Gaussian Approximation Given $\hat{\alpha}_l$, the conditional mean of the skull conductivity distribution $\pi(\sigma|\hat{\alpha}_l)$ is

$$\hat{\sigma}_{\rm CG} = \sigma_* + \Gamma_{\sigma\alpha_l} \Gamma_{\alpha_l}^{-1} \hat{\alpha}_l, \tag{26}$$

where σ_* is the mean value of the postulated skull conductivity distribution, $\Gamma_{\sigma\alpha_l}$ is the cross-covariance between σ and α_l (estimated from the samples of $\alpha_l^{(s)}$ and $\sigma^{(s)}$, as will be described in Section 5.4), and $\Gamma_{\alpha_l} = \text{diag}\{\lambda_{1l}, \ldots, \lambda_{pl}\}$ where λ_{1p} are the first peigenvalues of covariance Γ_{ε_l} .

 \ddagger In practice, we noticed that the derivatives of order $k \ge 2$ were negligible.

This Gaussian approximation approach can provide fast results, since it relies on offline precomputed statistics based on training data. However, in our EEG source imaging case, the approximation error and the model parameter are non-linearly related, and the approximation error depends also on the dipole values. Hence, the CG approximation may not give sufficiently accurate results further away from the linearization point σ_* .

4.2.2. Conditional Gaussian with Iterations In some cases, it is possible to improve the CG-based model parameter estimates by iteratively updating the linearization point. Algorithm 1 presents an iterative approach that uses the CG result as an initialization and also considers the estimated value of the source.

4.2.3. Gaussian Process Another way to estimate model parameters is to employ a Gaussian process. In GP, we can predict $\hat{\sigma}$ given $\hat{\mathcal{D}} = (\hat{\alpha}_l, \hat{x}_l)$ as estimated from (17) and the training data \mathcal{D} . Hence, we have

$$\hat{\sigma}_{\rm GP} = m(\hat{D}) + K_{\hat{\mathcal{D}},\mathcal{D}} K_{\mathcal{D}}^{-1} (f_{\mathcal{D}} - m(\mathcal{D})).$$
(27)

Different kernel K and mean functions m(.) can model complicated relationships between variables, and the way they were chosen in this case is described in Section 5.4.2.

5. Materials and Methods

5.1. Meshes

For our study, we built the two used meshes (a fine and a coarse one) with the help of the MRI data of the so-called ernie subject and SimNIBS 4 software§ [53]. Four different tissue compartments (scalp, skull, cerebrospinal fluid, and brain) were considered, and 76 electrodes were placed around the head according to the international 10-10 system (see, Fig.1). The fine mesh consisted of 2,103,623 elements joined in 377,150 nodes, and the coarse mesh of 743,575 elements joined in 136,868 nodes.

§https://simnibs.github.io/simnibs/build/html/index.html

5.2. Leadfield Models

The leadfield matrices used in this study were constructed with the help of custom made software that exploited Finite Element Method with linear basis functions, as in [54].

First, we created K = 200 leadfield matrices with skull conductivity samples $\sigma^{(k)}$ drawn from a bounded Gaussian distribution $\pi(\sigma)$ with mean $\sigma_* = 0.0103 \,\mathrm{S/m}$ and standard deviation $0.0035 \,\mathrm{S/m}$. The skull conductivity distribution ranged from $0.0041 \,\mathrm{S/m}$ to $0.033 \,\mathrm{S/m}$ according to the values reported in [55, 56, 57]. The rest of the tissue electric conductivity values were $0.43 \,\mathrm{S/m}$ for the scalp, $1.79 \,\mathrm{S/m}$ for cerebrospinal fluid, and $0.33 \,\mathrm{S/m}$ for the brain (gray matter and white matter) [58]. We refer to these leadfield matrices as *sample* models, $A(\sigma)$. Out of these, 150 *sample* models were used to estimated the skull conductivity related approximation error statistics (in Section 5.4) and the remaining 50 *sample* models were used for the training of the GP (in Section 5.4.2).

For the final testing, we created two *accurate* leadfield matrices $A(\sigma)$ with the help of the fine mesh and skull conductivity values 0.0061 S/m and 0.0139 S/m, that were not included in the sample set. In addition, we created a *standard* head model, A_0 , with the help of the coarse mesh and skull conductivity $\sigma_0 = 0.0103$ S/m.

5.3. Source Spaces

The dipole source space was restricted to an approximately 30 mm thick cross sectional area close to the sensory area (as shown in the left image of Fig. 1). This area was selected because previous studies had found it suitable for model calibration purposes using EEG recordings [59, 60]. The source space of the standard model (in the coarse mesh) is marked with blue circles in the right image of Fig. 1, and it uniformly covers the gray matter of the brain. The red dots in the right image of Fig. 1 indicate the source locations (in the fine mesh) that were used in the estimation of the approximation error statistics, as described in Section 5.4. The number of points (source locations) in both source spaces was 560. For the final testing, to avoid over-fitting, we produced observations from simulated dipole sources placed in different locations (in the fine mesh, marked with yellow in the right image of Fig. 1) than the sources that were used to produce the statistics.

5.4. Approximation Error Statistics

The samples for the estimation of the approximation error statistics at location i were created by first choosing randomly one of the *sample* models and then evaluating both the sample model and the standard model with the same single dipole source $x_i^{(j)}$ at location i as

$$\varepsilon_i^{(s)} = A(\sigma^{(k)}) x_i^{(j)} - A_0 x_i^{(j)}.$$
(28)

A set of J = 100 single radial dipole samples $x_i^{(j)}$ with amplitudes drawn from a Rayleigh distribution $\pi(|x_i|)| = \text{Rayleigh}(\sqrt{2\gamma})$ with parameter $\gamma = 1.85$ and mode 2.55, and 150



Figure 1. Left: The brain area that was studied and positions of electrodes denoted by red dots. Right: A zoom-in of the area of the brain that was studied to illustrate the candidate source locations. The red dots denote the source locations of the simulated dipoles used for training and production of the approximation error statistics. The blue circles indicate the candidate source locations for the reconstructions. To avoid over-fitting issues, we used different source locations in the final testing marked with yellow circles. The axes are in meters.

different sample leadfield models were used to estimate the approximation error statistics at each location *i*. The superscripts are as follows: s = j + J(k-1) where j = 1, ..., J, and k = 1, ..., K, and s = 1, ..., S. The mean ε_{i*} and covariance matrix Γ_{ε_i} of the corresponding approximation error at location *i* were

$$\varepsilon_{i*} = \frac{\sum_{s=1}^{S} \varepsilon_i^{(s)}}{S} \text{ and } \Gamma_{\varepsilon_i} = \frac{1}{S-1} \sum_{s=1}^{S} (\varepsilon_i^{(s)} - \varepsilon_{i*}) (\varepsilon_i^{(s)} - \varepsilon_{i*})^{\mathrm{T}}.$$
 (29)

Samples $\alpha_i^{(s)}$ and $\varepsilon_i^{''(s)}$ were evaluated according to $\alpha_i^{(s)} = W_i^{\mathrm{T}}(\varepsilon_i^{(s)} - \varepsilon_{i*})$ and $\varepsilon_i^{''(s)} = Q_i Q_i^{\mathrm{T}}(\varepsilon_i^{(s)} - \varepsilon_{i*})$, where W_i and Q_i were matrices with columns $1, \ldots, p$ and $p+1, \ldots, m$ of the eigenvectors of Γ_{ε_i} , respectively. Furthermore, by using these samples we estimated numerically Γ_{α_i} and the cross-covariance $\Gamma_{\sigma\alpha_i}$.

5.4.1. Selection of p Eigenvectors The number of eigenvectors p required to describe the primary error ε' depends on the discrepancies that produce the approximation errors. In general, it is preferable if only few eigenvectors can be used. In the current setup, when the source location was fixed, the source amplitude had only a scaling effect on the covariance matrix Γ_{ε_i} . Furthermore, since we were looking for a single model parameter value (skull conductivity $\sigma > 0$), we could set p = 1 and use only the eigenvector corresponding to the largest eigenvalue. Thus, we had a single error coefficient $\alpha_i \in \mathbb{R}$.

5.4.2. Inference of Skull Conductivity from Training Data In this section, we investigate the relationships between the variables $\{x_i^{(s)}, \alpha_i^{(s)}, \sigma^{(s)}\}_{s=1}^{N_s}$ for a single dipole location *i* in order to better understand the skull conductivity estimates that could be obtained from CG (26) and GP (27).

Conditional Gaussian: The red line in Fig. 2 shows the conditional skull conductivity predictions for different α_i values using $\sigma_{\text{CG}} = \sigma_* + \Gamma_{\sigma\alpha_i} \Gamma_{\alpha_i}^{-1} \alpha_i^{(s)}$. In addition, in the left image of Fig. 2 we have visualized the pairs $\{\alpha_i^{(s)}, \sigma^{(s)}\}$ for different source amplitudes. The horizontal scatter of the error coefficients α_i for a fixed $\sigma^{(s)}$ shows the effect of varying source amplitudes. Furthermore, the green cross shows the sample mean σ_* and the sample mean of the error coefficient. We can observe that for skull conductivity values close to the mean $\sigma_*=0.0103 \text{ S/m}$ (which is equal to the standard conductivity σ_0), the conditional conductivity predictions from (26) are accurate. Also, we notice that the red curve (CG conductivity prediction curve) passes through the error coefficients that correspond to the mode of the dipole amplitude distribution $\pi(|x_i|)$. Thus, we can expect feasible conductivity predictions only when the source amplitude is close to the mode. When the underlying skull conductivity is further away from σ_* the predictions deteriorate. A more detailed explanation for this is given in Appendix A1.

The right image of Fig. 2 shows the sample pairs $\{\alpha_i^{(s)}, \sigma^{(s)}\}\$ for few fixed amplitudes $|x_i|$. We observe monotonic relationships between the conductivity and the error coefficient α_i . Also, we can see that when $\sigma > \sigma_0$, the error coefficient α_i is negative, and when $\sigma < \sigma_0$ the error coefficient α_i is positive. Hence, the sign of the error coefficient could indicate whether the true skull conductivity is lower/higher than σ_0 .



Figure 2. Left: For a given source location *i*, we visualize here the sample pairs $\{\alpha_i^{(s)}, \sigma^{(s)})\}$ for different source amplitudes. The red line shows the conditional conductivity $\sigma_{*|\alpha_i} = \sigma_* + \Gamma_{\sigma\alpha_i}\Gamma_{\alpha_i}^{-1}\alpha_i^{(s)}$ (26). The green cross shows the sample mean σ_* and the mean of the error coefficient. Right: Here, we show how the error coefficient α_i is related to the skull conductivity value σ for a fixed source amplitude. In this case, we can observe monotonic relationships between conductivity and α_i .

Gaussian Process: To allow more complex relationships (instead of just linear) and to take into account the effect of the source amplitude, we employed a Gaussian process to infer the skull conductivity. For this purpose, we needed to define the mean and covariance functions, specified by hyperparameters. In the current problem, the statistics of the underlying (unknown) function $\sigma = f(\alpha_i, x_i), f \sim \mathcal{N}(m(\alpha_i, x_i), K_{\alpha_i, x_i})$, could be approximated based on the expected relationships between the parameters and simulated training data. In particular, the mean function $m(\alpha_i, x_i)$ could be modeled as a polynomial of the following form

$$m(\alpha_i, x_i) = \sum_{0 \le t \le T} c_t \left(\frac{\alpha_i + k}{|x_i|} \right)^t,$$
(30)

where $k = w_{i1}^{\mathrm{T}} \varepsilon_{i*}$. The covariance kernel was modeled using the following exponential function

$$K_{\alpha_i,x_i} = s_f^2 \exp\left(-\frac{1}{L^2} \left[\frac{\alpha_i}{|x_i|} - \frac{\alpha_i'}{|x_i'|}\right] \left[\frac{\alpha_i}{|x_i|} - \frac{\alpha_i'}{|x_i'|}\right]^T\right),\tag{31}$$

where the scaling factor $s_f = 0.001$ has the same order of magnitude as the standard deviation of the conductivity values (vertical scaling) and the length scaling L was set equal to 10 based on the range of $\frac{\alpha_i}{|x_i|}$ values (aiming to avoid under-fitting and over-fitting). The selected mean and covariance functions and their parameters encapsulated our beliefs on the data and our understanding of the related physics. A qualitative explanation for these choices can be found in the Appendix A2.



Figure 3. The blue dots are the target values of the conductivities, the red line is the mean of the GP prediction, the gray color shows the 95% confidence intervals, and the red crosses depict the training data.

In Fig. 3, the blue dots show the simulated (target) values that were not used in the estimation of the c_t coefficient, the red line shows the mean prediction of the Gaussian

process regression given by (30), the shaded gray areas represent the (pointwise) two standard deviation limits corresponding to each input, and the red crosses show the observations / training data. Here, we can see that the order T = 2 is able to describe the relationships between the three parameters. Higher order T did not improve the fit.

In practice to apply the GP, we divided the artificially generated training data set $\{x_i^{(s)}, \alpha_i^{(s)}, \sigma^{(s)}\}_{s=1}^{N_s}$ into two subsets. The first subset was used to estimate the coefficients c_t through least squares, and the second subset was used later as input-output training data when solving the problem (27). It is worth noting that the triplets were from additional samples that were not used in the estimation of the approximation error statistics Γ_{ε_i} , ε_{i*} nor W_i in Section 5.4.

5.5. Testing Data and Different Estimates

For the final testing, the EEG recordings were generated using one of the accurate leadfield models (in the fine mesh) as

$$v = A(\sigma_{\text{true}})x + e, \tag{32}$$

where the value of σ_{true} was either 0.00601 S/m or 0.0139 S/m that were not included in the training sets, and e denotes the additive (white) measurement noise.

5.5.1. Dipole Source Estimates Since we studied only single dipole sources, we used the single dipole scan algorithm [51] to estimate the sources.

• When using the *standard* leadfield model (in the coarse mesh) with standard skull conductivity $\sigma_0 = 0.0103 \text{ S/m}$, we solved the following optimization problem

$$\hat{x}_{l} \leftarrow \min_{i=1:n} \{ \min_{x_{i}} \| L_{e}(v - A_{0}^{i}x_{i} - e_{*}) \|_{2}^{2} \},$$
(33)

where L_e is a matrix square root of $\Gamma_e^{-1} = L_e^T L_e$. In practice, this functional was minimized for each source space node l considering that the dipoles in other locations were zero, and the solution was the dipole source that minimized the residual.

• When using the proposed Bayesian Approximation Error algorithm, we found the pair that minimized

$$(\hat{x}_{l}, \hat{\alpha}_{l}) \leftarrow \min_{i=1:n} \{ \|L_{\varepsilon_{i}''+e}(v - A_{0}^{i}\hat{x}_{i} - \varepsilon_{i*} - w_{i1}\alpha_{i} - e_{*})\|_{2}^{2} + \|L_{\alpha_{i}}\alpha_{i}\|_{2}^{2} \},$$
(34)

where $(\Gamma_{\varepsilon_i''} + \Gamma_e)^{-1} = L_{\varepsilon_i''+e}^{\mathrm{T}} L_{\varepsilon_i''+e}$, and the estimated $\hat{\alpha}_l$ was subsequently used for the skull conductivity estimation. Note that also this dipole source estimate was computed using the *standard* leadfield model A_0 (i.e. without knowledge of the *true* skull conductivity). 5.5.2. Skull Conductivity Estimates Then, the skull conductivity estimation was carried out either by using only the solved $\hat{\alpha}_l$ from the previous step, or the pair $(\hat{\alpha}_l, \hat{x}_l)$.

• CG: The first approach was to use the conditional Gaussian estimation in which the model parameter was solved simply by using the cross-covariance as

$$\hat{\sigma}_{\rm CG} = \sigma_* + \Gamma_{\sigma\alpha_l} \Gamma_{\alpha_l}^{-1} \hat{\alpha}_l, \tag{35}$$

where Γ_{α_l} and Γ_{σ,α_l} were computed from the samples $\{\alpha_l^{(s)}, \sigma^{(s)}\}$.

- CG+Iter: To improve the above CG solution, we employed the iterative approach presented in Section 4.2.2 (Algorithm 1) that took the estimated source also as an input.
- GP: Skull conductivity estimation using a Gaussian Process

$$\hat{\sigma}_{\rm GP} = m(\hat{\alpha}_l, \hat{x}_l) + K_{(\hat{\alpha}_l, \hat{x}_l), \mathcal{D}} K_{\mathcal{D}}^{-1} (f_{\mathcal{D}} - m(\mathcal{D})),$$
(36)

where $\mathcal{D} = \{\alpha_l^{(s)}, x_l^{(s)}\}, f_{\mathcal{D}} = \{\sigma^{(s)}\}$ and the functions as described in Section 5.4.2.

5.5.3. Reference Method: Alternating Dipole Scan As a reference, we compared our dipole source and skull conductivity estimates with the corresponding estimates of the following alternating approach

$$\hat{x}_{l}^{(k+1)} \leftarrow \min_{i=1:n} \{ \min_{x_{i}} \| L_{e}(v - A(\sigma^{(k)})^{i} x_{i} - e_{*}) \|_{2}^{2} \}, \\ \hat{\sigma}^{(k+1)} \leftarrow \min_{\sigma} \| L_{e}(v - A(\sigma)^{l} \hat{x}_{l}^{(k+1)} - e_{*}) \|_{2}^{2},$$
(37)

where $A(\sigma)^{l} \approx A(\sigma^{(k)})^{l} + \frac{\partial A^{l}}{\partial \sigma}|_{\sigma = \sigma^{(k)}} (\sigma - \sigma^{(k)}).$

6. Results and Discussion

We first study the effects of varying source amplitude on source localization accuracy and skull conductivity estimates. Subsequently, we compared the proposed algorithms with a fixed source amplitude. In all the test cases, we used single radial dipole sources. We randomly picked 88 source locations (that were not used in the training) from the area of the brain shown in Fig. 1. The EEG measurements v were computed using the *accurate* leadfield model (32) in the fine mesh that had either skull conductivity 0.00601 S/m or 0.0139 S/m (that were not included in the training sets).

Random white noise was added to the measurements. In Section 6.1, a low noise level, $SNR = 40 \, dB$, was used in order to study the effects of the varying source amplitude on the results. In Section 6.2, a higher noise level, $SNR = 30 \, dB$, was used.

All the dipole source reconstructions were carried out using the *standard* leadfield model (in the coarse mesh) that had fixed skull conductivity $\sigma_0 = 0.0103 \text{ S/m}$; the alternating dipole scan obviously updated the leadfield model during the iterations and used the standard leadfield model only in the initial step. For the model parameter (skull conductivity) estimation, CG+Iter also updated the leadfield model in the iterations.

To evaluate the dipole source reconstructions, we calculated localization errors (Euclidean distances in milli meters) between the true and reconstructed source location. In the following, we denote $X_{\rm st}$, $X_{\rm BAE}$, and $X_{\rm alt}$ as the localization errors of the standard solution (33), BAE solution (34), and alternating algorithm (37), respectively. Subsequently, we estimated the improvements in source localization ΔX of the algorithms defined as $X_{\rm st} - X_{\rm BAE}$ and $X_{\rm st} - X_{\rm alt}$.

6.1. Effects of Varying Source Amplitude in Source Localization Improvement and Skull Conductivity Estimates

In the first test case, we simulated boundary measurements resulting from 88 radial dipoles (one at a time) with increasing source amplitude (from 0.3 to 4.2, in normalized source amplitude units), and then computed source reconstructions and skull conductivity estimates using the proposed approaches, CG (35), CG+Iter. (Algorithm 1), and GP (36).

Figure 6.1 shows the dipole localization errors, $X_{\rm st}$ and $X_{\rm BAE}$, of the standard solution (33) and the BAE solution (34), respectively, for four different source intensities. As can be seen, the localization errors of the BAE solutions are much more often smaller than the localization errors of the standard solution in all the cases. The benefits of BAE are more evident in the case with $\sigma_{\rm true} = 0.00601 \,{\rm S/m} < \sigma_0$ than in the case with $\sigma_{\rm true} = 0.0139 \,{\rm S/m} > \sigma_0$. Moreover, the localization errors do not seem to have any particularly strong trend with respect to the source intensity.



Figure 4. Localization errors of the standard solution and the BAE solution in the two studied cases, $\sigma_{\text{true}} = 0.00601 \text{ S/m}$ and $\sigma_{\text{true}} = 0.0139 \text{ S/m}$, when different source intensities were used. The standard model used in all the reconstructions assumed (erroneously) skull conductivity $\sigma_0 = 0.0103 \text{ S/m}$.



Figure 5. Comparison of localization improvements ΔX and skull conductivity estimates across various source amplitudes. The boundary measurements v were simulated by using single dipoles and two testing skull conductivities, $\sigma_{\text{true}} = 0.00601$ S/m and $\sigma_{\text{true}} = 0.0139$ S/m, while the standard model used in the reconstructions assumed (erroneously) skull conductivity $\sigma_0 = 0.0103$ S/m. The left image illustrates the localization improvements as box plots in these two cases as a function of the varying source amplitude. The images in the middle and right show the estimated skull conductivity values as a function of source amplitude using the three methods (CG, CG+Iter, and GP). The dashed horizontal lines represents the standard conductivity σ_0 and the blue lines shows the true skull conductivity.

In Figure 5, we show with box plots the source localization improvements ($\Delta X = X_{\rm st} - X_{\rm BAE}$, in milli meters) and the estimated skull conductivities in the two test cases as a function of the source amplitude. The medians ar represented by the horizontal lines inside the boxes, the edges of the boxes denote the 25th and 75th percentiles, and the whiskers and outliers indicate the range of values. It is worth emphasizing that the source reconstructions were carried out by first, using the standard leadfield model that assumed (erroneously) skull conductivity value $\sigma_0 = 0.0103$ S/m, and second, using the same standard model accompanied with the statistics of the BAE modeling. Higher ΔX values in the first (left) image of Fig. 5 indicate greater improvements using the proposed BAE dipole scan. Light gray refers to the $\sigma_{\rm true} < \sigma_0$ case and dark gray to the proposed BAE dipole scan in both cases, when compared to using the standard dipole scan, and that the BAE dipole scan particularly improves the results in the $\sigma_{\rm true} < \sigma_0$ case.

The second (middle) and third (right) image in Fig. 5 show box plots of the estimated skull conductivities in the test cases with $\sigma_{\text{true}} = 0.00601 \,\text{S/m} < \sigma_0$ and $\sigma_{\text{true}} = 0.0139 \,\text{S/m} > \sigma_0$, respectively. The skull conductivities were estimated using

the proposed techniques, CG, CG+Iter and GP, across various source amplitudes. The horizontal solid blue line represents the standard conductivity σ_0 , and the dashed black line shows the true conductivity used in the test cases.

Based on these results, the CG-based conductivity estimation performs well only when the source amplitude is close to the mode of the source amplitude distribution $\pi(|x_i|)$ (see Appendix A1 for an explanation). Furthermore, the solution tends to be close to the standard value (σ_0) when the source amplitude is small. This is because for small source amplitudes, the reconstructed error coefficient (α_i) is usually small (due to the scaling effect, see Section 4.2) and thus, favors conductivity values close to σ_0 (as was expected based on the red curve in Fig. 2). Furthermore, we can see that the skull conductivity estimates for increasing dipole amplitude either decrease (in the $\sigma_{true} < \sigma_0$ case), or increase (in the $\sigma_{true} > \sigma_0$ case). Therefore, it is clear that the conductivity solutions of CG depend on the amplitude of the source.

As a first improvement to CG, we proposed an iterative algorithm (Algorithm 1) that takes as an input the estimated source \hat{x} and uses the CG estimated conductivity as an initialization. As can be seen, the accuracy of the iterative algorithm is much better compared to using only CG, and the solution is much less dependent on the source amplitude. However, the iterative algorithm is computationally much more effortful, e.g. because of the repetitive computations of the Jacobian matrix.

As an alternative to CG+Iter., we proposed to use a GP for the conductivity estimation. GP is computationally much less demanding, the effort being essentially the same as in the non-iterative CG. Based on the results, we can observe that GP performs even better than CG+Iter, with medians closer to the true skull conductivity and shorter whiskers.

6.2. Source Localization Improvements and Skull Conductivity Estimates with the Proposed Algorithms

In the following two case studies, we fixed the source amplitude to 1.3 (normalized units). We estimated the sources and their localization accuracy, as before. For the model parameter (skull conductivity) estimation, we used only the proposed CG+Iter and GP, as it was shown in the previous section that these two were superior to the simple CG approach. We ran the algorithms for each location with 5 different (measurement) noise realizations at SNR = 30 dB, and we estimated the corresponding localization improvements, $\Delta X = X_{\rm st} - X_{\rm BAE}$, and the errors of the conductivity estimates as percentages, $|\Delta \sigma| = 100 \times \frac{|\hat{\sigma} - \sigma_{\rm true}|}{\sigma_{\rm true}}$. The presented results are the average values of ΔX and $|\Delta \sigma|$ for each tested location. As a reference method, we computed the source reconstructions and skull conductivity estimates by using the alternating dipole scan (37).

Fig. 6 depicts the results of the proposed Bayesian approach for the $\sigma_{\text{true}} < \sigma_0$ case applied over selected source locations in MRI slices 72, 74, 76, and 78. The top row presents the source localization improvements resulting from the BAE solution with

respect to the standard model. The white triangles depict the locations in which the source localization improvements were the highest ($\Delta X > 6 \text{ mm}$), and the red and orange triangles correspond to smaller localization improvements, ranging from 2 mm to 6 mm. In locations marked with white squares, the localization errors were negligibly small ($\leq 2 \text{ mm}$). The yellow (upside down) triangles depict negative localization improvements (between -6 mm and -2 mm), which means that in these cases the standard model (with erroneous skull conductivity value) performed better than the proposed BAE approach. The note below the top row gives the percentage of cases in which the localization improvements were greater than 2 mm.

The middle and bottom rows present data related to conductivity errors, comparing the two different methods, CG+Iter. in the middle row and GP in the bottom row. Because the conductivity values were estimated alongside the sources, the conductivity estimation errors are presented with the help of the (true) locations of the sources that generated the corresponding EEG testing data. The symbols (squares and circles) illustrate different conductivity estimation errors: the white squares indicate the smallest errors ($|\Delta\sigma| \leq 5\%$), and the circles (with different shades of blue and black) indicate higher estimation errors, from 5% to more than 20%. In Fig. 7, we show the results of the reference method, the alternating dipole scan set in (37), in the $\sigma_{true} < \sigma_0$ case. The localization improvements and conductivity estimation errors are presented using the same notations and symbols as in Fig. 6. Figs. 8 and 9 present the corresponding Bayesian and reference results in the $\sigma_{true} > \sigma_0$ case.

6.3. Discussion

6.3.1. Source Estimates The Bayesian approximation error approach showed substantial improvements in source localization accuracy when compared to both, using the standard model with fixed skull conductivity and the alternating dipole scan that was used as a reference method. When compared to the standard model, the BAE improved source localization by $\geq 2 \text{ mm}$ in 75% of the tests in the $\sigma_{\text{true}} < \sigma_0$ case and in 51% of the tests in the $\sigma_{\text{true}} > \sigma_0$ case. In the rest of the tests, the performances of the BAE and the standard model were similar, with the exception of only two source locations in which the standard model outperformed the BAE. The alternating dipole scan performed in most tests equally to the standard model, outperforming the standard model only in 37% of the tests in the $\sigma_{\text{true}} < \sigma_0$ case and in 19% of the tests in the $\sigma_{\text{true}} > \sigma_0$ case.

The improvements provided by the BAE modeling were particularly clear in the case where the true skull conductivity was lower than σ_0 . This is because the EEG measurements, and thus the approximation errors, are non-linearly proportional to the skull conductivity, roughly $\varepsilon \propto 1/\sigma$. Therefore, the absolute values of the approximation errors in the $\sigma_{\text{true}} < \sigma_0$ case were (on average) higher than in the $\sigma_{\text{true}} > \sigma_0$ case. In the tests where the approximation errors were high (causing poor performance for the standard model), the BAE approach could effectively alleviate the source



In 61.4% of test cases conductivity error less than or equal to 5%

Figure 6. Results of the proposed Bayesian approaches in terms of source localization improvement and conductivity estimation error across MRI slices (72, 74, 76, and 78) when $\sigma_{true} < \sigma_0$ and SNR = 30 dB. The top row displays localization improvements, with triangles indicating different magnitudes of the improvements (in milli meters). As the conductivity values were estimated together with the sources, both the localization improvements and conductivity estimation errors are presented with the help of the (true) locations of the sources that generated the corresponding EEG testing data. The middle and bottom rows show the errors of the conductivity estimates when CG+Iter. (middle) and GP (bottom) were used. The symbols represent different levels of conductivity errors (in percentages): white squares denote errors less than 5%, and circles with shades of blue and black higher skull conductivity estimation errors. Here, GP achieved conductivity errors below 5% in 71.6% of cases, outperforming CG+Iter. that achieved the same only in 48.9% of cases.

localization errors with the help of the embedded model-based training statistics of the approximation errors. The other way around, when the approximation errors were smaller in the $\sigma_{\text{true}} > \sigma_0$ case, also the potential improvements in the localization accuracy that could be gained with the help of the BAE approach were more modest. These findings are in line with our previously published results [31].

6.3.2. Model Parameter Estimates The conditional Gaussian regression-based model parameter estimation did not work robustly for two main reasons. First, the model



Alternating optimization when $\sigma_{\rm true} < \sigma_0$ and ${\rm SNR} = 30 \, {\rm dB}$

In 8.0% of test cases conductivity error less than or equal to 5%

Figure 7. Results of the reference method, the alternating minimization algorithm (37), for the $\sigma_{\rm true} < \sigma_0$ case. The top row shows the localization improvements similarly as the top row of Fig. 6, and the bottom row shows the errors of the estimated skull conductivities in percentages similarly as the middle and bottom rows of Fig. 6.

parameter cannot be solved using only the low-rank estimator for ε , because the solution additionally depends on the source amplitude (as seen in Figs. 2 and 5). Second, the model parameter depends non-linearly on ε (and α , as seen in Fig. 2) which can cause inaccuracies in the estimates particularly when the *true* conductivity values are relatively far from the linearization point, σ_0 .

Thus, the iterative algorithm that used the CG solution as initialization performed better. The estimates were more accurate because the (estimated) source intensity was taken into account and because the problem was iteratively linearized. The trade-off was the computational burden that significantly increased.

GP offered an effective alternative to the above as it did not require iterations. However, the prerequisite was that a suitable kernel function was needed to describe the relationship between the model parameter and ε . In our case, this relationship could be approximated based on the underlying physics (see Appendix A2). From the tested approaches, GP gave the best model parameter estimates.

The skull conductivity estimates of the alternating optimization (37) were poor. Furthermore, we want to stress that the computation of a new lead field matrix and Jacobian matrix, both required by this algorithm, are highly demanding steps, which makes the alternating optimization unfavorable. In addition, we observed that the alternating optimization was sensitive to the initialization of the conductivity parameter. In some cases, the method did not converge with the initialization σ_0 , and another initialization point close to σ_0 was required.



In 55.7% of test cases conductivity error less than or equal to 5%

Figure 8. Results of the proposed Bayesian approaches in terms of source localization improvement and conductivity estimation error in the $\sigma_{\text{true}} > \sigma_0$ case. The notations and symbols are as in Fig. 6.

6.4. Model-based Learning in Inverse Problems

Inverse problems are challenging due to their ill-posed nature, but also because model uncertainties are often present. The lack of access to ground truth data makes it difficult (or impossible) to apply standard and well-known supervised learning techniques for inferring model parameters. Instead, we must rely on meticulously designed simulations to generate sampling distributions for the model parameters, based on evidence, physics, and carefully made assumptions on the problem. This presents a complex challenge, making the task of designing and creating realistic simulations more critical than ever for effective training and testing, and for avoiding such issues as data snooping and overfitting.

7. Conclusion and Future Work

In this paper, we employed the Bayesian Approximation Error approach to solve a blind linear inverse problem with a non-linear (unknown) model parameter in the forward model and tested different approaches to estimate both the primary unknown and the



Alternating optimization when $\sigma_{\rm true} > \sigma_0$ and ${\rm SNR} = 30\,{\rm dB}$

In 10.2% of test cases conductivity error less than or equal to 5%

Figure 9. Results of the reference method, the alternating minimization algorithm (37), for the $\sigma_{\text{true}} > \sigma_0$ case. The top row shows the localization improvements similarly as the top row of Fig. 6, and the bottom row shows the errors of the estimated skull conductivities in percentages similarly as the middle and bottom rows of Fig. 6.

model parameter. To do that, we produced model-based training data that gave us samples of the approximation error. We used these samples to numerically estimate the mean and covariance matrix of the approximation errors, and subsequently a subspace spanned by the top eigenvector of this covariance was used to give a low-rank estimate of the approximation errors in the test cases.

Based on the solved primary unknown and approximation error estimate, we inferred the (unknown) model parameter using three approaches. The first approach was based on conditional Gaussian regression, the second iteratively updated the linearization point, and the third utilized a Gaussian process that was modeled with the help of physics-informed learning. In addition, alternating optimization was used as a reference method.

As an application, we studied the EEG source imaging problem in which the (forward) leadfield model contained a non-linear unknown model parameter, the skull conductivity. The utilization of the BAE approach clearly improved the source localization accuracy of the imaging and the GP provided superior estimates for the skull conductivity when compared to the other tested approaches.

In the future, the BAE approach with GPs will be tested with experimental EEG data, and the estimated skull conductivities will be compared to results from other (experimental) skull conductivity calibration techniques. In EEG imaging, the BAE approach could be tested with more complicated priors, such as ℓ_1 and group-lasso priors, and in scenarios where several tissue conductivities are recovered simultaneously with the source configuration. Finally, as the proposed developments are not limited

to the EEG source imaging, other inverse problems such as kernel estimation in blind deconvolution problems could be studied.

Acknowledgements

This was supported by the Research Council of Finland: Flagship of Advanced Mathematics for Sensing Imaging and Modeling (359185) and PROFI 6–TAU Imaging Research Platform (336357).

Appendix A1: Predicting σ under Linear Approximation

In this paper, we studied the conditional Gaussian $\pi(\sigma|\alpha_l)$ and the MAP estimates $\sigma_{\text{MAP}} = \max_{\sigma>0} \log \pi(\sigma|\alpha_l)$. We observed that these estimates were accurate only when the true source amplitude at location l was close to the mode of the distribution of the source amplitude $|x_l|$. Here, the source is described by $\mathbf{x}_l = |x_l|\vec{n}$, considering a fixed orientation \vec{n} .

In this section, we try to explain why the estimates performed well only under these circumstances. First, we try to derive an explicit expression for the posterior $\pi(\sigma|\alpha_l)$ using the sampling distributions and a linear approximation of the *approximation error*.

We start with

$$\pi(\sigma|\alpha_l) \propto \pi(\sigma)\pi_{|x_l|}(\alpha_l|\sigma) = \pi(\sigma) \int \pi(\alpha_l|\sigma, |x_l|)\pi(|x_l|) \, \mathrm{d}|x_l|.$$

The conditional probability can be approximated as $\pi(\alpha_l | \sigma, x_l) \approx \delta(\alpha_l - G_l(\sigma) |x_l| - c)$, where $\delta(.)$ is the Dirac-delta function and G_l is a function of σ .

For small perturbations $\sigma - \sigma_*$, the function $G_l(\sigma)$ can be estimated with the help of the 1st order Taylor series of the *approximation error* around $\sigma_* = \sigma_0$. In particular, we can write

$$\varepsilon_l = w_{l1}\alpha_l + \varepsilon_{l*} = (\sigma - \sigma_*) \frac{\partial A^l(\sigma_*)}{\partial \sigma} \vec{n} |x_l|,$$

where $\frac{\partial A^l(\sigma_*)}{\partial \sigma} \in \mathbb{R}^{m \times 3}$. Now we can multiply both sides with the eigenvector $w_{l1}^{\mathrm{T}} \in \mathbb{R}^m$. This results in

$$\alpha_l + c = (\sigma - \sigma_*) w_{l1}^{\mathrm{T}} \frac{\partial A^l(\sigma_*)}{\partial \sigma} \vec{n} |x_l|,$$

where $c = w_{l_1}^{\mathrm{T}} \varepsilon_{l*} \in \mathbb{R}$. Hence, we can define

$$G_l(\sigma) = (\sigma - \sigma_*) w_l^{\mathrm{T}} \frac{\partial A^l(\sigma_*)}{\partial \sigma} \vec{n}.$$

Also, we can write

$$|x_l| = G_l^{-1}(\sigma)(\alpha_l + c) = \frac{k_l^{-1}(\alpha_l + c)}{\sigma - \sigma_*},$$

where $k_l = w_{l1}^{\mathrm{T}} \frac{\partial A^l(\sigma_*)}{\partial \sigma} \vec{n} \in \mathbb{R}$ and where $k_l^{-1}(\alpha_l + c)$ and $\sigma - \sigma_*$ will have the same sign.

When $x_l \in \mathbb{R}^3$ with $x_l \sim \mathcal{N}(0, \gamma^2 I_3)$, the distribution of the amplitude $|x_l|$ is $|x_l| \sim \text{Rayleigh}(\sqrt{2\gamma})$, i.e. $\pi(|x_l|) = \frac{|x_l|}{2\gamma^2} \exp\left(-\frac{|x_l|^2}{4\gamma^2}\right)$ with $|x_l| > 0$.

Therefore, the posterior becomes

$$\pi(\sigma|\alpha_l) = \pi(\sigma)\pi_{|x_l|}(G_l^{-1}(\sigma)(\alpha_l+c)).$$

Here, the distribution $\pi(\sigma)$ can be approximated as a uniform distribution in the area around the mean σ_* and thus $\pi(\sigma|\alpha_l) \propto \pi_{|x_l|}(G_l^{-1}(\sigma)(\alpha_l+c))$.

Based on the previous, the log of $\pi(\sigma|\alpha_l)$ is

$$\log \pi(\sigma | \alpha_l) \propto \log \left(G_l^{-1}(\sigma) \right) - \frac{(G_l^{-1}(\sigma)(\alpha_l + c))^2}{4\gamma^2}.$$

To find the MAP estimate, we solve $\frac{d \log \pi(\sigma | \alpha_l)}{d\sigma} = 0$ which gives

$$\sigma_{\rm MAP} = \sigma_* + k_l^{-1} (\alpha_l + c) \frac{1}{\sqrt{2}\gamma},$$

where $\sqrt{2\gamma}$ is the mode of the source amplitude. Therefore, the skull conductivity is predicted based on the estimated α_l and the mode of the source amplitude (instead of the estimated amplitude).

Hence, this modeling demonstrates good predictive capabilities only when the source amplitude is close to the mode (as was shown with simulations in Section 6.1). Therefore, in general, more advanced methods that take into account the estimated source amplitude work better (as was also shown in Section 6.1).

Appendix A2: Choice of the GP Function

From Ohm's law, we approximate that the electric potential v is proportional to the electric conductivity and a current source as $v \propto \frac{1}{\sigma}|x|$. Furthermore, we approximate that the approximation error (difference of two potentials), $\varepsilon = w_1 \alpha + \varepsilon_*$, is proportional to these variables as

$$\alpha + w_1^{\mathrm{T}} \varepsilon_* \propto \frac{1}{\sigma} |x| - \frac{1}{\sigma_0} |x|.$$
(38)

From this, we write

$$\sigma \propto \left(\frac{\alpha + w_1^{\mathrm{T}}\varepsilon_*}{|x|} + \frac{1}{\sigma_0}\right)^{-1}.$$
(39)

Now, if we denote $y = \frac{\alpha + w_1^{\mathrm{T}} \varepsilon_*}{|x|}$ and define $\tilde{\sigma}(y) = (y + \frac{1}{\sigma_0})^{-1}$, then by applying Taylor expansion around 0, we get $\tilde{\sigma}(y) = \sum_{i=0}^{\infty} c_i y^i$.

References

- Kirsch A 1996 An introduction to the mathematical theory of inverse problems (NY, USA: Springer-Verlag New York.) ISBN 0-387-94530-X
- [2] Tikhonov A 1943 Doklady Akademii nauk SSSR 39 195–198

- [3] Plato R and Vainikko G 1990 Numerische Mathematik 57(1) 63–79
- [4] Engl H W, Hanke M and Neubauer A 1996 Regularization of Inverse Problems (Kluwer Academic Publishers)
- [5] Hansen P C 1998 Rank-deficient and Discrete Ill-posed Problems: Numerical Aspects of Linear Inversion (Philadelphia, PA, USA: Society for Industrial and Applied Mathematics) ISBN 0-89871-403-6
- [6] Benning M and Burger M 2018 Acta Numerica 27 1–111 ISSN 1474-0508
- [7] Afkham B M, Chung J and Chung M 2021 Inverse Problems 37 105017 ISSN 1361-6420
- [8] Mattsson P, Zachariah D and Stoica P 2023 IEEE Transactions on Signal Processing 71 1175–1183 ISSN 1941-0476
- [9] Burger M and Kabri S 2013 arXiv (Preprint 2312.09845)
- [10] Campisi P and Egiazarian K (eds) 2007 Blind Image Deconvolution: Theory and Applications (Boca Raton, FL: CRC Press) ISBN 9780849390725
- [11] Ni Y and Strohmer T 2024 Auto-calibration and biconvex compressive sensing with applications to parallel mri
- [12] Ahmed A, Recht B and Romberg J 2014 IEEE Transactions on Information Theory 60 1711–1732 ISSN 1557-9654
- [13] Ling S and Strohmer T 2015 Inverse Problems **31** 115002 ISSN 1361-6420
- [14] Bolte J, Combettes P L and Pesquet J C 2010 Alternating proximal algorithm for blind image recovery 2010 IEEE International Conference on Image Processing (IEEE)
- [15] Li X, Ling S, Strohmer T and Wei K 2019 Applied and Computational Harmonic Analysis 47 893–934 ISSN 1063-5203
- [16] Adler J and Oktem O 2017 Inverse Problems 33 124007 ISSN 1361-6420
- [17] Adler J and Oktem O 2018 IEEE Transactions on Medical Imaging 37 1322–1332 ISSN 1558-254X
- [18] Jin K H, McCann M T, Froustey E and Unser M 2017 IEEE Transactions on Image Processing 26 4509–4522 ISSN 1941-0042
- [19] Arridge S, Maass P, Öktem O and Schönlieb C B 2019 Acta Numerica 28 1–174 ISSN 1474-0508
- [20] Lunz S, Hauptmann A, Tarvainen T, Schönlieb C B and Arridge S 2021 SIAM Journal on Imaging Sciences 14 92–127 ISSN 1936-4954
- [21] Arridge S, Hauptmann A and Korolev Y 2023 Inverse problems with learned forward operators
- [22] LeCun Y, Bengio Y and Hinton G 2015 Nature 521 436-444 ISSN 1476-4687
- [23] Murphy K P 2022 Probabilistic Machine Learning (MIT Press Ltd) ISBN 0262046822 URL https: //www.ebook.de/de/product/41791478/kevin_p_murphy_probabilistic_machine_learning.html
- [24] Kaipio J P and Somersalo E 2004 Statistical and Computational Inverse Problems Applied Mathematical Series (Springer)
- [25] Kaipio J and Somersalo E 2007 J. Comput. Appl. Math. 198 493–504
- [26] Vanrumste B, Hoey G V, de Walle R V, D'Have M, Lemahieu I and Boon P 2000 Med. Biol. Eng. Comput. 38 528–534
- [27] Dannhauer M, Lanfer B, Wolters C and Knösche T 2011 Hum. Brain Mapp. 32 1383–1399
- [28] Lew S, Sliva D D, Choe M, Grant P E, Okada Y, Wolters C H and Hämäläinen M S 2013 NeuroImage 76 282–293
- [29] Ollikainen J O, Vauhkonen M, Karjalainen P A and Kaipio J P 1999 Med. Eng. Phys. 21(3) 143–154
- [30] Montes-Restrepo V, van Mierlo P, Strobbe G, Staelens S, Vanderberghe S and Hallez H 2014 Brain Topogr. 27(1) 95–111
- [31] Rimpilainen V, Koulouri A, Lucka F, Kaipio J P and Wolters C H 2019 NeuroImage 188 252–260
- [32] Vorwerk J, Aydin U, Wolters C H and Butson C R 2019 Frontiers in Neuroscience 13 ISSN 1662-453X
- [33] Schmidt C, Wagner S, Burger M, Rienen U v and Wolters C H 2015 Journal of Neural Engineering 12 046028 ISSN 1741-2552
- [34] Saturnino G B, Thielscher A, Madsen K H, Knösche T R and Weise K 2019 NeuroIm-

age 188 821–834 ISSN 1053-8119 URL https://www.sciencedirect.com/science/article/pii/S1053811918322031

- [35] McCann H and Beltrachini L 2021 Biomedical Physics & Engineering Express 7 045018 URL https://dx.doi.org/10.1088/2057-1976/ac0547
- [36] Kaipio J and Kolehmainen V 2013 Approximate marginalization over modeling errors and uncertainties in inverse problems *Bayesian Theory and Applications* ed Damien P, Polson N and Stephens D (Oxford University Press)
- [37] Nissinen A, Kolehmainen V and Kaipio J P 2011 International Journal for Uncertainty Quantification 1 203–222 ISSN 2152-5080
- [38] Koulouri A and Rimpiläinen V 2020 Simultaneous skull conductivity and focal source imaging from EEG recordings with the help of bayesian uncertainty modelling 8th European Medical and Biological Engineering Conference (Springer International Publishing) pp 1019–1027
- [39] Lipponen A, Seppänen A and Kaipio J P 2011 Meas. Sci. Technol. 22 104013
- [40] Nissinen A, Heikkinen L M, Kolehmainen V and Kaipio J P 2009 Meas. Sci. Technol. 20 105504
- [41] Arridge S R, Kaipio J P, Kolehmainen V, Schweiger M, Somersalo E, Tarvainen T and Vauhkonen M 2006 Inverse Problems 22(1) 175–195
- [42] Kolehmainen V, Schweiger M, Nissilä I, Tarvainen T, Arridge S R and Kaipio J P 2009 J. Opt. Soc. Am. A 26 2257–2268
- [43] Tarvainen T, Kolehmainen V, Pulkkinen A, Vauhkonen M, Schweiger M, Arridge S R and Kaipio J P 2010 Inverse Problems 26 015005
- [44] Koulouri A, Rimpiläinen V, Brookes M and Kaipio J P 2016 Appl. Num. Math. 106 24–36
- [45] Mozumder M, Tarvainen T, Arridge S R, Kaipio J and Kolehmainen V 2013 Biomed. Opt. Express 4 2015–2031 URL http://www.opticsinfobase.org/boe/abstract.cfm?URI=boe-4-10-2015
- [46] Alexanderian A, Nicholson R and Petra N 2024 Inverse Problems 40 095001 URL https://dx.doi. org/10.1088/1361-6420/ad602e
- [47] Candiani V, Hyvönen N, Kaipio J P and Kolehmainen V 2021 Inverse Problems 37 125008 URL https://dx.doi.org/10.1088/1361-6420/ac346a
- [48] Nicholson R, Petra N, Villa U and Kaipio J P 2023 Inverse Problems 39 054001 URL https: //dx.doi.org/10.1088/1361-6420/acc129
- [49] Bohrnstedt G W and Goldberger A S 1969 Journal of the American Statistical Association 64(328) 1439—-1442
- [50] Grech R, Cassar T, Muscat J, Camilleri K, Fabri S, Zervakis M, Xanthopoulos P, Sakkalis V and Vanrumste B 2008 J. Neuroeng. Rehabil. 5 25
- [51] Mosher J C, Lewis P S and m Leahy R 1992 IEEE Transactions on Biomedical Engineering 39(6) 541–557
- [52] Vauhkonen M 1997 Electrical impedance tomography and prior information PhD thesis
- [53] Puonti O, Van Leemput K, Saturnino G B, Siebner H R, Madsen K H and Thielscher A 2020 NeuroImage 219 117044 ISSN 1053-8119 URL https://www.sciencedirect.com/science/article/ pii/S1053811920305309
- [54] Wolters C H, Grasedyck L and Hackbusch W 2004 Inverse Problems 20 1099–1116
- [55] Hoekema R, Wieneke G H, van Veelen C W, van Rijen P C, Huiskamp G J, Ansems J and van Huffelen A C 2003 Brain Topogr. 16 29–38
- [56] Homma S, Musha T, Nakajima Y, Okomoto Y, Blom S, Flink R and Hagbarth K E 1995 Neurosci. Res. 22 51–55
- [57] Aydin U, Vorwerk J, Küpper P, Heers M, Kugel H, Galka A, Hamid L, Wellmer J, Kellinghaus C, Rampp S and Wolters C H 2014 PLoS ONE 9(3) e93154
- [58] Ramon C, Schimpf P H and Haueisen J 2006 Biomed. Eng. Online 5
- [59] Schrader S, Antonakakis M, Rampp S, Engwer C and Wolters C H 2020 Phys Med Biol. 65 245043 ISSN 1361-6560
- [60] Antonakakis M, Schrader S, Aydin U, Khan A, Gross J, Zervakis M, Rampp S and Wolters C H 2020 NeuroImage 223 117353 ISSN 1053-8119