On the Bias, Fairness, and Bias Mitigation for a Wearable-based Freezing of Gait Detection in Parkinson's Disease

TIMOTHY ODONGA, Department of Biomedical Informatics, Emory University, USA CHRISTINE D. ESPER, Department of Neurology, Emory University, USA STEWART A. FACTOR, Department of Neurology, Emory University, USA J. LUCAS MCKAY, Department of Biomedical Informatics, Department of Neurology, Emory University, USA HYEOKHYEN KWON, Department of Biomedical Informatics, Emory University, USA and Department of Biomedical Engineering, Emory University & Georgia Institute of Technology, USA

Freezing of gait (FOG) is a debilitating feature of Parkinson's disease (PD) which is a common cause of injurious falls among PD patients. Recent advances in wearable-based human activity recognition (HAR) technology have enabled the detection of various FOG subtypes, such as tremulous freezing, across benchmark datasets using different sensor placement configurations. Since the manifestation of FOG is heterogeneous, it is important to develop a model that can quantify FOG consistently across patients with varying demographics, FOG types, and PD conditions. We defined bias as a systematic error in decision-making leading to unequal outcomes, and fairness as consistent model performance across groups. In the HAR community, bias and fairness in FOG models have been understudied, with most research solely focused on FOG detection using single benchmark datasets collected from single study sites with specific protocols. In this work, we systematically evaluated the bias and fairness of state-of-the-art HAR models for wearable-based FOG detection across varying demographics and PD conditions, using multiple benchmark datasets. Additionally, we evaluated the effectiveness of transfer learning as a potential bias mitigation approach. Our fairness evaluation, using the demographic parity ratio (DPR) and equalized odds ratio (EOR), demonstrated that the HAR model is biased (DPR & EOR < 0.8) for all stratified demographic variables, including age, sex, and disease duration. Our experiments empirically demonstrated that transfer learning from multi-site datasets and generic human activity representations significantly improved the fairness (average change in DPR of +0.027 and +0.039, respectively) and performance (average change in F1-score of +0.026 and +0.018, respectively) of the model across all attributes and datasets. This supported the hypothesis that generic human activity representations would likely learn fairer representations encoding rich information about diverse human activity contexts, which could then be transferred to health analytics. This highlights the importance of considering fairness and multi-site studies in developing HAR models for health applications.

CCS Concepts: • Human Centered Computing; • Ubiquitous Computing; • Computing Methodologies; • Artificial Intelligence;

Additional Key Words and Phrases: Bias, Fair Machine Learning, Freezing of Gait, Parkinson's Disease, Wearable Sensors, Human Activity Recognition

Authors' Contact Information: Timothy Odonga, Department of Biomedical Informatics, Emory University, Atlanta, Georgia, USA, timothy. odonga@emory.edu; Christine D. Esper, Department of Neurology, Emory University, Atlanta, Georgia, USA, cedoss@emory.edu; Stewart A. Factor, Department of Neurology, Emory University, Atlanta, Georgia, USA, sfactor@emory.edu; J. Lucas McKay, Department of Biomedical Informatics, Department of Neurology, Emory University, Atlanta, Georgia, USA, lucas@dbmi.emory.edu; Hyeokhyen Kwon, Department of Biomedical Informatics, Emory University, Atlanta, Georgia, USA and Department of Biomedical Engineering, Emory University & Georgia Institute of Technology, Atlanta, Georgia, USA, hyeokhyen.kwon@emory.edu.

© 2025 Copyright held by the owner/author(s). Publication rights licensed to ACM. ACM XXXX-XXXX/2025/2-ART https://doi.org/XXXXXXXXXXXXXXX

Permission to make digital or hard copies of all or part of this work for personal or classroom use is granted without fee provided that copies are not made or distributed for profit or commercial advantage and that copies bear this notice and the full citation on the first page. Copyrights for components of this work owned by others than the author(s) must be honored. Abstracting with credit is permitted. To copy otherwise, or republish, to post on servers or to redistribute to lists, requires prior specific permission and/or a fee. Request permissions from permissions@acm.org.

ACM Reference Format:

1 INTRODUCTION

Freezing of gait (FOG) is one of the most debilitating symptoms of Parkinson's disease (PD), often leading to falls and fall-related injuries among patients [64]. FOG assessment in clinical practice relies on coarse and subjective evaluation by movement disorder specialists using a 0-4 scale in one field of the motor examination section, known as Part III of the Movement Disorder Society Unified Parkinson's Disease Rating Scale (MDS-UPDRS-III) [25, 29]. Although useful, the specialist rating of FOG using the MDS-UPDRS-III scale occurs within a clinical setting, which does not fully capture the occurrences and manifestations of FOG that a patient might experience in daily living [59]. Developing a system that can continuously measure FOG episodes is essential to implementing personalized intervention plans for patients with PD and developing clinical trials for potential therapeutic measures [59].

Recently, multiple studies explored using wearable sensors and machine learning (ML) to quantify FOG across clinical and daily living environments [59]. These models were mainly evaluated for their performance in detecting FOG episodes on benchmark datasets collected from single study sites with specific protocols. However, evaluating a FOG detection model also requires considering the biases in model behavior relating to demographic factors, such as age and sex, due to the disparities in PD care and heterogeneity in FOG symptoms. In current clinical practice, a few studies have noted the differences in PD care across demographics that include disparities affecting women such as diagnosis and referral delays [80], infrequent and delayed access to specialized care services [27], and under-representation in PD trials [86, 90]. Other studies have outlined disparities related to demographic variables such as race that include differences in PD incidence and prevalence between black and white communities [6, 17, 96]. FOG may be challenging for clinicians to assess due to its unpredictable and episodic nature, as well as the heterogeneity of its manifestations [59, 81]. These manifestations arise from different behavioral phenotypes that include alternate leg trembling, very small shuffling steps, and the akinetic form where no leg movement is observed [81]. Given the different manifestations of FOG and existing disparities in PD care across different patient demographics, wearable-based FOG analysis models need to serve the wide range of patients suffering from FOG. Consequently, it is essential to consider the fairness of the FOG analysis models across the different patient populations.

In this work, we adopted a widely used definition of bias in automated decision systems as a systematic error in decision-making processes that results in unequal outcomes [26]. We also defined fairness in terms of demographic groups as outlined by Varshney [91], where fairness is satisfied when the average classifier behavior is the same across subject groups defined by protected subject attributes such as age, sex, etc. To the best of our knowledge, bias and fairness of wearable-based HAR models for FOG analysis, along with bias mitigation approaches, remain understudied. We systematically analyzed the bias and fairness of state-of-the-art HAR models for FOG episode detection [10, 40, 66] and applied bias mitigation approaches to the model. To quantify classifier behavior across subject groups, two standard group fairness metrics from fairness in machine learning (Fair ML), i.e., demographic parity ratio (DPR) and equalized odds ratio (EOR) [2, 23, 94] were used. We applied the four-fifths rule [16] as the threshold for determining fairness based on the DPR and EOR values for each attribute. Subjects from four benchmark wearable-based FOG datasets (Daphnet [76], De Souza *et al.* [18], tDCS FOG [43], and DeFOG [43]) were stratified according to age, sex, and disease duration. The HAR models were evaluated using both group fairness metrics (DPR and EOR) [2, 94] and standard model performance metrics like F1-score.

Our analysis showed that the HAR models exhibited bias across all datasets with respect to protected subject attributes. Also, the standard bias mitigation approaches [35, 103] proved inadequate in addressing the biases across multiple attributes, in some cases even decreasing the DPR and EOR values. To mitigate these biases, transfer learning was implemented using both multi-site FOG datasets and generic human activity representations trained on large-scale wearable datasets collected during daily living. Our underlying hypothesis was that pretraining models with FOG data from other sites or generic activity representations would create fairer models by encoding the rich information present in diverse human activities across various demographic contexts, which will then be effectively applied to health analytics. We found that both transfer learning approaches, multi-site transfer and generic feature activity transfer, significantly and consistently improved the models' DPR and EOR values across all the subject attributes (average change in DPR of +0.027 and +0.039, respectively). This research emphasizes the importance of collecting and evaluating HAR datasets with respect to bias in diverse demographic variables and disease conditions, especially for health-related applications, and the importance of training models using multi-site datasets.

2 RELATED WORK

In this section, we provide an overview of prior research relevant to our work on using ML for FOG detection, as well as research on Fair ML and bias mitigation in HAR.

2.1 Machine Learning for Parkinson's Disease

With the adoption of ML in PD research, most research has focused on developing models to differentiate patients with PD from healthy subjects [87]. This was done using models trained on data from different modalities, which were linked to specific symptoms of PD. For other related symptoms of PD like dystonia and dysphonia [55, 71], existing studies [38, 55, 62] implemented ML models to classify healthy subjects from PD patients, using statistical features extracted from voice data such as average, minimum, and maximum vocal fundamental frequencies. Similarly, other works [15, 56] applied brain imaging techniques to magnetic resonance imaging (MRI) data to extract features from the gray and white matter tissue volumes in the MRI data, train ML models to learn the differences in brain structure between healthy subjects and PD patients, and classify them. With motor-related PD symptoms such as bradykinesia, tremor, and muscular rigidity [44], different data modalities, such as tabular gait cycle data and wearable sensor data, have been used to develop ML models. Different studies [46, 69] explored learning the gait dynamics of PD patients by training ML models directly on spatiotemporal gait variables, such as swing interval, stride length, and stride speed, present in the tabular gait cycle data. In studies that used wearable sensor data like inertial measurement unit (IMU) data [13, 82], spatiotemporal gait variables such as stride interval, swing interval, and heel strike were computed from the IMU readings and used to train ML models to learn PD gait dynamics.

Another important area where ML has been used in PD research involves quantifying the severity of PD using clinician rating scales such as the MDS-UPDRS-III scale [29] and classifying different subtypes of PD. ML approaches have been applied to the MDS-UPDRS-III scale that quantifies PD severity related to motor symptoms on a scale of 0 (normal) to 4 (severe) [29]. Different studies have used shallow ML models [1, 11, 84] and deep learning models [105] trained on wearable-based data (i.e. IMU data) to predict the scores for motor-related items on the MDS-UPDRS-III scale. For the studies that used shallow ML models, time-based features such as mean, variance, kurtosis, and frequency-based features such as peak frequency and moments of power spectral density were extracted from the IMU data to train the models [1, 11]. Other approaches used 3D motion sequences extracted from different input modalities such as videos [57] and 3D optical motion capture data [50]. With the extracted 3D motion sequences, different deep learning-based architectures were applied, e.g., convolutional neural network model (CNN) [57], and adaptive graph convolutional neural network (AGCN) model [50] to

automatically learn motion-based features that were used to classify the MDS-UPDRS-III score. In classifying the different subtypes of PD, prior works have focused on using ML models to classify two PD motor subtypes, i.e., tremor dominant and postural instability gait difficulty, using 3D kinematic data [30], wearable sensor data [97], and resting-state functional magnetic resonance imaging data [68].

2.2 Quantifying Freezing of Gait in Parkinson's Disease

Within the context of FOG research in PD, existing ML work has primarily focused on three tasks, namely, FOG detection, FOG prediction, and scoring FOG severity. FOG detection involves identifying FOG episodes within an analysis window [60]. FOG prediction refers to identifying when a patient is about to experience a FOG episode in a pre-FOG period to enable preemptive cues that can help prevent the occurrence of an episode [60]. Researchers have developed FOG detection models using data collected from laboratory settings [60, 61] and home settings [75]. These models utilized temporal and frequency-domain features extracted from multiple on-body sensor signals within specified analysis windows. Regarding FOG prediction, prior works [9, 60, 67] have implemented ML prediction through analysis of temporal windows preceding FOG onset by 2-5 seconds. To score the severity of FOG, prior works [50] used adaptive convolutional neural networks (AGCN) models to automatically learn motion-based features from 3D motion sequences extracted from 3D optical motion capture data.

The existing literature on FOG has also focused on characterizing the heterogeneity of its manifestations and developing protocols to provoke freezing in different settings. Different manifestations of FOG related to movement types have been documented, such as alternate trembling of legs, small shuffling of steps, and the akinetic form when no movement of legs is observed [81]. In addition, the different manifestations related to the trigger events of FOG have also been investigated, including motor impairments such as difficulty turning on the spot and increased anxiety levels, e.g., more freezing when in a hurry [24]. Consequently, different studies have devised tasks to evoke FOG based on these triggering factors, such as the Timed Up and Go [83], single and dual motor tasks, turning clockwise and counterclockwise, and walking through hallways [106]. While these tasks have been effective in provoking FOG within controlled laboratory or clinical environments, other studies have noted the challenges in extensively assessing FOG in routine clinical practice due to reasons that include the limited time for gait analysis and the tendency to improve when patients consciously focus on walking [33, 59, 74]. To address these limitations, a few studies, such as the work by Mancini *et al.* [59] have explored the potential of wearable sensors for continuous monitoring of PD patients at home to better capture the diverse manifestations of FOG during daily activities.

While prior research has significantly contributed to developing ML models for FOG detection, a few gaps remain in the existing literature. A key limitation is that most studies developed and evaluated their models within the context of single datasets, often derived from single study environments and specific FOG-provoking protocols. Although these studies have demonstrated the efficacy of FOG detection, there is insufficient evidence to support whether the models would generalize across diverse datasets with different study environments and FOG-provoking protocols. Furthermore, few studies [50] have reported the performance of implementing ML models across the different patient demographics represented in the available datasets. While overall model performance is important, given the heterogeneity of FOG manifestations in PD and the demographic differences in PD, it is vital to systematically evaluate model performance across the diverse patient populations who may exhibit varying expressions of FOG and thereby prevent further potential disparities in care. Our work aimed to address these gaps in the existing research. We systematically evaluated the bias and fairness of a state-of-the-art FOG detection model across different patient demographics represented in the dataset. Additionally, we assessed the model's bias and fairness, and performance across four datasets derived from varying study environments, FOG-provoking protocols, and on-body sensor locations to examine the generalizability of our approach.

[,] Vol. 1, No. 1, Article . Publication date: February 2025.

2.3 Demographic Differences in Parkinson's Disease and Fair ML in Human Activity Recognition

Demographic factors, including sex, race, and age, have also been studied as determinants of healthcare disparities in PD. Epidemiological studies have documented significant sex-based differences in PD incidence rates, with the ratio of male to female prevalence of approximately 1.4 to 1 [21]. Notable investigations into sex-age interactions have demonstrated that women typically develop PD at a later age than men [4, 32]. Furthermore, female PD patients generally experience delayed referrals to movement disorder specialists [80] and remain underrepresented in PD trials [90]. Race has also been identified as a significant determinant of healthcare access inequities, with documented variations in prevalence between black and white communities [6, 17, 96], and research indicating that black women demonstrate notably lower rates of specialist consultation [95]. Given these documented demographic disparities in PD care, it is imperative to evaluate model performance across diverse demographic groups to prevent perpetuating existing disparities in PD care.

Fair ML has become critical with the increased adoption of ML models in automated decision-making processes, including HAR. Fairness considerations are needed to ensure that automated decision-making systems do not place individuals or demographic groups at a systematic disadvantage [8]. The Fair ML community developed two primary notions of fairness that include group fairness, which requires average classifier performance to be the same across groups defined by protected subject attributes [23, 91], and individual fairness, which ensures that individuals with similar features receive similar model predictions [102]. Prior works have also explored bias mitigation approaches such as preprocessing methods like correlation remover [94], in-processing techniques like adversarial debiasing [103], post-processing techniques such as threshold optimizer [35]. These bias mitigation techniques have primarily been applied to tabular and text datasets. Within ML for health, researchers have studied bias and fairness across various medical domains. Studies have investigated model bias in dermatology [47], medical imaging [28], Alzheimer's disease (AD) [100], and multimodal remote mental health assessment [45]. Prior work has included determining models' ability to predict latent protected attributes like race in medical images [28]. Researchers have evaluated AD progression classifiers using demographic parity across multi-site datasets, examining attributes such as sex, age, and race [100]. Regarding research on depression, studies have also assessed bias and fairness in a multimodal remote mental health assessment system across protected attributes like race, gender, education level, and age using demographic parity ratio and equalized odds ratios [45].

Despite the extensive research on bias and fairness in ML, the research addressing bias and fairness in HAR is sparse [99]. Among the limited existing studies, Nair *et al.* [63] showed that reducing the sampling bias with respect to physical characteristics such as sex, age, and height increases the performance of HAR models for daily activity recognition tasks such as walking, jogging and sitting. Consequently, with the limited works on bias and fairness in HAR and ubiquitous computing, there are efforts to develop frameworks to incorporate fairness, such as Zhang *et al.* [104] who proposed frameworks for incorporating fairness into system design to guide the development of fair and unbiased ubiquitous computing systems, ensuring equal treatment and positive societal impact. To the best of our knowledge, no existing works in HAR or the Fair ML community have explored applying transfer learning from multi-site datasets or generic feature activity representations as bias mitigation approaches. Consequently, in this work, we analyzed the effectiveness of conventional bias mitigation approaches in Fair ML for FOG detection and explored the effectiveness of transfer learning (i.e., transfer learning from multi-site datasets or generic feature activity approaches.

3 METHODS

We introduced the benchmark datasets used in this study, HAR models, bias mitigation approaches, and experiment settings to evaluate FOG detection performance and model biases. Figure 1 shows the overall framework of our evaluation.

6 • Odonga, et al.



Fig. 1. Model bias evaluation pipeline for FOG detection tasks. In this work, we analyze the group fairness of the model based on sensitive attributes in FOG and PD, including age, sex, and disease duration.

Table 1. Dataset characteristics including subject demographic distribution, label distribution, and sensor locations	T 1 1 1 D 1 1 1 1 1 1	• • • • •		1	11	1 1 1
Table 1. Dataset characteristics menualing subject demographic distribution, laber distribution, and sensor locations	Table I Dataset characteristic	rs including subjec	t demographic	distribution label	distribution an	d sensor locations
	Table 1. Dataset characteristic	.5 menuumg subjee	t ucinographic	uistribution, laber	uistribution, an	a sensor rocations.

Dataset	N	Sex	Age (years)	Disease Duration (years)	Frequency (Hz)	Sensor locations	Label distribution	Study Protocol
Daphnet	10	M=7 F=3	66±5	13.7±9.7	64	Ankle, Lower back, Thigh	No FOG=0.81 FOG=0.19	Daily walking tasks e.g. walking in straight line, 360 degree turns
DeFOG	38	M=24 F=14	67±9	9.5±5.5	100	Lower back	No FOG=0.95 FOG=0.05	TUG, TUG dual task, turning task, Hotspot door task, Personalized Hotspot task
De Souza	35	M=19 F=16	65±10	$8.0{\pm}4.1$	128	Shank	No FOG=0.80 FOG= 0.20	Turning task
tDCS FOG	62	M=50 F=12	69±8	$12.4{\pm}6.7$	128	Lower back	No FOG=0.68 FOG=0.32	Single task, dual motor task motor cognitive task

3.1 Wearable Datasets for FOG Episodes

We used four publicly available wearable sensor-based datasets for FOG: DeFOG and tDCS FOG [43], Daphnet [76], and the De Souza *et al.* dataset (De Souza) [18]. These datasets were selected for the variability of FOG conditions, the availability of subject demographics and disease duration metadata, the study environments, including both home and clinic/lab, the different on-body sensor locations, and the different study protocols for invoking FOG (walking or turning). A table summarizing the dataset characteristics of each dataset is shown in Table 1.

3.1.1 DeFOG. This dataset was provided in the Kaggle Parkinson's Freezing of Gait Prediction 2023 challenge [43]. The dataset was collected during two visits by a specialist to a subject's home environment, during which subjects were evaluated in on- and off-medication states. Subjects wore Ax3 by Axivity 3D accelerometer sensors placed at the lower back with a sampling rate of 100Hz. The study protocol consisted of a number of FOG-provoking tests that included a four-meter walk test, the Timed Up & Go (TUG) test as a single task, the Timed Up & Go dual task (TUG dual task), turning task, Hotspot Door task, and Personalized Hotspot task [43]. The

TUG test consists of a series of tasks, including rising from a chair, walking a distance of 3 meters, walking back to the chair, and sitting down [70] The TUG dual task required subjects to perform the TUG test while mentally subtracting numbers [43]. The turning task included the subject performing four 360-degree turns, alternating the rotation direction with each turn [43]. The Hotspot Door task is a walking trial that involves opening a door, entering another room, turning, and returning to the start point [43]. The Personalized Hotspot task involves walking through an area in the house that the subject described as FOG-provoking [43].

3.1.2 tDCS FOG. This dataset was also included in the Kaggle Parkinson's Freezing of Gait Prediction 2023 challenge [43], which was collected at a clinic. Subjects wore Opals by APDM 3D accelerometer sensors placed at the lower back, with the data recorded at a sampling rate of 128Hz. Subject underwent FOG-provoking tests described in Ziegler *et al.* [106] that included a subject performing the following series of actions defined as a single task; sit on a chair, rise up & start walking, turn 360 clockwise and counterclockwise, walk 2 meters, open a door, go through, turn 180 degrees and then go back an sit on the chair. The test was performed at three levels of difficulty that included a single task, a dual motor task (i.e., performing the trial while carrying a tray with a bottle on it), and a motor cognitive task (i.e., performing the dual motor task while also performing serial seven subtractions) [73].

3.1.3 Daphnet. Ten subjects underwent a series of tests that included tasks to simulate daily walking (i.e., walking back and forth in a straight line with several 180-degree turns, random walking with a series of initiated stops and 360-degree turns) in the lab [5]. Eight subjects were evaluated in an off-medication state, and two subjects were evaluated in an on-medication state. The data was measured from 3D accelerometer sensors placed in the shank, thigh, and lower back of each subject and transmitted via a Bluetooth link at a rate of 64Hz for online processing.

3.1.4 De Souza. 35 subjects in a clinicre evaluated in an anti-Parkinsonian medication state over three sessions [18]. The subjects wore inertial measurement units (IMU) sensors (Physiolog 5 by Gait Up CH) on the shank of the most affected side, with the data recorded at a sampling rate of 128 Hz. This study mainly focused on provoking FOG during turning tasks, which consisted of the following actions: turning in place, alternating 360-degree turns to the right and then to the left, and repeating the turning at a self-selected pace for 2 minutes [18].

3.2 HAR Models and Pipeline

For developing the HAR model, we followed the standard activity recognition chain using the sliding window approach [12]. As part of the preprocessing, we applied Min Max scaling to each sensor channel across the readings from all patients [77]. For analysis frames, we used a 3-second time window with no overlap for the DeFOG, tDCS FOG, and De Souza datasets. The 3-second window size was selected particularly for the tDCS FOG dataset as it was noted as the optimal window size for FOG detection in Reches *et al.* [73]. Since no previous report on optimal window sizes for DeFOG and De Souza existed, we applied a 3-second window size, as the sampling frequencies were closer to those of the tDCS FOG dataset. For the Daphnet dataset, we used a 4.5-second window length with no overlap as this was noted as the optimal window size in Bächlin *et al.* [5]. We used two models for HAR approaches: Random Forest (RF) and DeepConvLSTM model [66]. Previous HAR studies used RF and DeepconvLSTM models to classify complex activities in daily living and locomotion [49]. For Random Forest (RF), we extracted ECDF features [34] from the analysis frame. Following Ordóñez and Roggen [66], the raw sensor time series was provided to the DeepConvLSTM model. Each analysis window was labeled as having FOG episodes or normal movements based on the annotations in the last timestep of the analysis window. This was done to enable real-time prediction of the onset or offset of FOG episodes during subject movement [98].

3.3 Bias Mitigation

We explored four different approaches for mitigating bias in HAR models for FOG detection. Two approaches include bias mitigation widely used in previous work in Fair ML: threshold optimizer [35, 94] and adversarial debiasing [103]. We additionally explored transfer learning approaches for bias mitigation using two different approaches. We first used the multi-site FOG training to understand the utility of learning FOG patterns across multiple sites. We also studied the utility of generic wearable representation learned from large-scale unlabeled wearable datasets collected from various healthy individuals to understand the utility of generic human activity features in mitigating biases in clinical populations like FOG patients.

3.3.1 Threshold Optimizer. We employed the threshold optimizer postprocessing bias mitigation technique using the FairLearn Python package [35, 94]. This approach uses different thresholds for each attribute (sex, age, and disease duration) to improve DPR or EOR when making predictions with a trained model, finding a balance between the model's performance and fairness. This method required finding multiple thresholds targeted for each combination of attributes and fairness metrics. It lacked the capability to improve fairness encompassing all attributes at the same time.

3.3.2 Adversarial Debiasing. Adversarial debiasing is an in-process bias mitigation technique aiming to learn attribute-agnostic feature representations framed as an adversarial learning problem [103]. This approach uses two models, predictor and adversary models, that are trained simultaneously for the predictor model to detect FOG episodes (classification loss) and the adversary model to be confused with the subject attribute from the learned representation (adversary loss). The adversarial debiasing minimize $\nabla_W L_P - proj_{\nabla_W L_A} \nabla_W L_P - \alpha \nabla_W L_A$, where L_P is classification loss for predictor model, L_A is adversary loss for adversary model, and α is a hyperparameter that controls the strength with which the fairness constraint is enforced. We used the DeepConvLSTM model as the predictor model and a multi-layer perceptron (MLP) with a single hidden layer as the adversary model, following the standard approach for adversarial debiasing [94, 103]. We applied adversarial debiasing in two different ways. First, we trained separate models to debias each attribute independently, following the standard approaches [103]. We also explored multi-task adversarial debiasing, by modifying adversary loss, L_A , to decrease bias in all attributes (age, sex, and disease duration) simultaneously, using multi-head MLP. We hypothesized that the multi-head adversarial debiasing would derive a fairer model across all attributes by modeling the interplay between those attributes at the same time.

3.3.3 Transfer Learning from Multi-site Datasets. We studied the effect of diversifying training datasets through multi-site datasets to debias the model. We hypothesize that the increased diversity of training subjects will significantly improve fairness across all attributes and improve the model performance at the same time. To the best of our knowledge, this is the first work studying transfer learning in the context of model debiasing in FOG. Our datasets contained readings from different sensor locations, e.g., lower back and ankle. For transfer learning, we paired the sensor locations of source and target datasets from similar on-body sensor locations. For example, when Daphnet and De Souza datasets are source and target datasets, respectively, we used sensor channel readings from the ankle, thigh, and shank sensor locations denoting lower extremities for pre-training with Daphnet and fine-tuning the model on shank location with De Souza. When both Daphnet and tDCS FOG are the source datasets, and DeFOG is the target dataset, we only used lower back sensor locations for pre-training and fine-tuning the model. Specifically, in our experiments, we explored the following pairs of source and target datasets: Daphnet \rightarrow De Souza; De Souza \rightarrow Daphnet; DeFOG + Daphnet \rightarrow tDCS FOG; tDCS FOG + Daphnet \rightarrow DeFOG. During fine-tuning the DeepConvLSTM model, we froze the weights in the first two convolutional layers and only fine-tuned the last two convolutional layers and LSTM layer with target datasets, following standard practice in transfer learning [22, 41]. In performing transfer learning on the RF model, we retrained the pre-trained model with the target dataset, following the previous work [51]. We also matched the sampling

Attribute	Daphnet	De Souza	tDCS FOG	DeFOG
Age (years)	66	69	69	69
Disease Duration (years)	12.5	7	9	13

Table 2. Median values for continuous subject attributes

frequency between the source and target datasets by downsampling the dataset with a higher sampling frequency to match the other dataset with a lower sampling frequency. For example, when using tDCS FOG (128 Hz) and Daphnet (64 Hz) datasets as the source datasets and DeFOG (100Hz) as the target datasets, we downsampled tDCS FOG and DeFOG to 64Hz before pre-training and finetuning the model. When using De Souza (128Hz) and Daphnet (64 Hz) as the source and target datasets, we downsampled the De Souza dataset to 64 Hz before pre-training the model.

3.3.4 Transfer Learning from Generic Human Activity Representation. Recently, with the advances in selfsupervised [36, 37, 78, 79, 88, 89] and generative training [3, 52, 54], the HAR community has been exploring models, so-called foundation HAR models, that encode generic wearable-based human activity representations trained from large-scale wearable datasets that can be efficiently fine-tuned for various downstream tasks. We hypothesized that generic activity representations will likely learn fairer representations encoding rich information about diverse human activity contexts from various demographic backgrounds that can also be transferred to health analytics. To this end, we applied transfer learning from a pretrained HAR model [101] that was trained through self-supervised learning on a large unlabeled dataset i.e. UK-Biobank activity tracker dataset [20] containing 700,000 person days (100,000 participants, seven days of wear) of unlabeled tri-axial accelerometer data from wrist sensors. We hypothesized that the scale of the wrist-sensor dataset would provide sufficient activity representations for FOG detection, irrespective of sensor placement. Specifically, we used the ResNet model trained on 5-second data [101], whose window size was similar to the time window length (3-5 seconds) for detecting FOG in our study. The only modification made to the ResNet model architecture of the pretrained model was changing the number of output nodes in the classifier part of the architecture (i.e., fully connected linear layer) for binary classification of FOG episodes. The rest of the feature extractor architecture and weights were reused for transfer learning. During finetuning on the four FOG datasets, we finetuned all the layers in the model architecture as this was reported to give the maximum F1-score in Yuan *et al.* [101].

3.4 Model Performance and Bias Evaluation

3.4.1 FOG Detection. Our HAR model was evaluated with 5-fold user-independent cross-validation for DeFOG, tDCS FOG, and De Souza datasets and 3-fold user-independent cross-validation for the Daphnet dataset. A 3-fold user-independent cross-validation was used on the Daphnet dataset because the dataset was relatively small (only ten subjects), and it was important to ensure that the test split of the data in each fold contained more than one sensitive group for group fairness evaluation. For each fold, the training dataset was further split into train (80%) and validation (20%) splits for hyperparameter tuning, following the nested cross-validation approach. Following the standard model evaluation in HAR [49], we used macro F1-scores for evaluating the HAR model in the presence of substantial label imbalance [65]. To consider the randomness of data split and model training, we ran 10 trials of cross-validation and reported an average F1-score across all test fold scores and 95% confidence interval for the statistical significance of model comparisons.

3.4.2 Model Fairness. For evaluating the bias and fairness of the trained models, we follow the standard approach for group fairness evaluation [23, 35]. We used a definition of fairness from the group fairness perspective given by Varshney [91], where fairness is satisfied when the average classifier behavior is the same across subject groups defined by protected attributes. As shown in Figure 1, the subjects in each dataset were grouped into two categories through stratifying sensitive attributes. For sex, we used the self-reported labels provided in the datasets (i.e., male and female). Continuous attributes (age and disease duration) were dichotomized using the median split procedure following standard biostatistics analysis [31], as shown in Table 2. PD research considers subjects with younger age categories (years < 66 for Daphnet and years < 69 for tDCS FOG, DeFOG, and De Souza) and shorter disease duration (years <12.5 for Daphnet, years < 7 for De Souza, years < 9 for tDCS FOG, years < 13 for DeFOG) as the group of individuals more likely to have milder symptoms of PD and males to better utilize PD care [17, 80, 96] (These categories was referred to as privileged groups, based on the Fair ML literature [42, 91]).

We evaluated model fairness using the Demographic Parity Ratio (DPR) and Equalized Odds Ratio (EOR), the key Fair ML metrics [2]. DPR checks for consistent model performance (F1-score) across stratified groups, while EOR compares true and false positive rates among them [45].

$$DPR^{k} = \frac{min_{g}S_{g}^{k}}{max_{g}S_{g}^{k}}$$
$$EOR^{k} = min\left(\frac{min_{g}TPR_{g}^{k} + \delta}{max_{g}TPR_{g}^{k} + \delta}, \frac{min_{g}FPR_{g}^{k} + \delta}{max_{g}FPR_{g}^{k} + \delta}\right)$$

where $g \in G_k$ for sensitive attribute k with G different groups, S is the selection rate of the classifier, i.e., the ratio of positive classification, δ is a value set to avoid division by zero. Similarly to the F1-score, we reported the average DPR and EOR and 95% confidence interval from 10 trials of cross-validation to validate the statistical significance of model bias and our bias mitigation approaches. In our experiments, the threshold optimization was only applied to RF. Adversarial debiasing and transfer learning for generic activity features were only applied to the deep learning models, DeepConvLSTM and ResNet, respectively. Transfer learning with the multi-site dataset was applied to the RF and DeepConvLSTM models.

3.4.3 The Four-Fifths Rule. For the threshold of fairness, we applied the four-fifths rule (threshold = 0.8) from the U.S. Equal Employment Opportunity Commission [16], where DPR = 1 and EOR = 1 are indicative of perfect fairness. Given that there is no existing threshold of fairness specific to FOG, PD, or ML for health research, we applied the four-fifths rule since it is widely used in the Fair ML community.

4 RESULTS

In this section, we first report the performance and bias of the models without bias mitigation, then present them after applying bias mitigation techniques. Table 3 shows overall results for model performance (F1-score) and biases (DPR and EOR for each sensitive attribute) across all our benchmark datasets.

4.1 Model Performance and Bias without Bias Mitigation

Model Performance. RF models showed the macro F1-scores of 0.490, 0.463, 0.722, and 0.513 for DeFOG, tDCS FOG, De Souza, and Daphnet dataset, respectively (rows 1, 9, 17, 25 in F1-score column of Table 3). DeepConvLSTM model showed macro F1-scores of 0.582, 0.722, 0.778, and 0.613 for DeFOG, tDCS FOG, De Souza, and Daphnet dataset, respectively (rows 4, 12, 20, 28 in F1-score column of Table 3). Compared to RF, the DeepConvLSTM model showed an average F1-score increase of +0.127 across all four datasets.

[,] Vol. 1, No. 1, Article . Publication date: February 2025.

Table 3. Model bias and performance of the trained models for the FOG detection task. **Bold** text indicates the best value in each column for each dataset. Non-overlapping confidence intervals indicate statistically significant differences ($p \le 0.05$). Multi-site transfer occurs when models are pretrained on source datasets (in the Pretrain dataset column). Generic activity feature transfer uses a model pretrained on the unlabeled UK Biobank dataset [20] and finetuned on the target dataset (in the Dataset column).

					Demographic Parity Ratio (DPR)			Equalized Odds Ratio (EOR)		
Dataset	Model	Bias mitigation	Pretrain dataset	F1-score	Sex	Age	Disease duration	Sex	Age	Disease duration
		×	-	0.490 ± 0.002	0.360 ± 0.112	0.318 ± 0.101	0.313 ± 0.096	0.065 ± 0.048	0.054 ± 0.045	0.092 ± 0.076
		Threshold optimizer	-	0.498 ± 0.003	0.255 ± 0.074	0.323 ± 0.076	0.354 ± 0.083	0.138 ± 0.054	0.171 ± 0.062	0.155 ± 0.063
	RF	Multi-site Transfer	tDCS FOG + Daphnet	0.495 ± 0.004	0.260 ± 0.108	0.271 ± 0.083	0.251 ± 0.105	0.033 ± 0.045	0.044 ± 0.041	0.022 ± 0.029
		×	-	0.582 ± 0.010	0.402 ± 0.068	0.528 ± 0.076	0.569 ± 0.077	0.300 ± 0.054	0.320 ± 0.058	0.382 ± 0.064
DeFOG		Adversarial debiasing (Single attribute)	-	0.575 ± 0.009	0.391 ± 0.090	0.406 ± 0.101	0.461 ± 0.086	0.287 ± 0.075	0.270 ± 0.074	0.348 ± 0.089
	DeepConv LSTM	Adversarial debiasing (Multi-Head MLP)	-	0.569 ± 0.014	0.379 ± 0.064	0.467 ± 0.081	0.569 ± 0.072	0.272 ± 0.055	0.304 ± 0.068	0.399 ± 0.069
		Multi-site Transfer	tDCS FOG + Daphnet	0.644 ± 0.012	0.487 ± 0.073	$\textbf{0.556} \pm \textbf{0.068}$	0.562 ± 0.071	0.356 ± 0.055	$\textbf{0.429} \pm \textbf{0.069}$	0.465 ± 0.068
	ResNet	Generic Activity Feature Transfer	-	$\textbf{0.659} \pm \textbf{0.012}$	$\textbf{0.522} \pm \textbf{0.068}$	0.539 ± 0.063	$\textbf{0.592} \pm \textbf{0.073}$	0.366 ± 0.058	0.395 ± 0.065	$\textbf{0.475} \pm \textbf{0.071}$
		×	-	0.463 ± 0.018	0.624 ± 0.075	0 693 + 0 054	0.676 ± 0.053	0.409 ± 0.086	0.573 ± 0.054	0.455 ± 0.055
		Threshold	-	0.469 ± 0.017	0.427 ± 0.077	0.163 ± 0.025	0.164 ± 0.021	0.477 ± 0.081	0.624 ± 0.053	0.553 ± 0.054
	RF	Multi-site	DeFOG + Daphnet	0.463 ± 0.019	0.592 ± 0.078	0.648 ± 0.056	0.660 ± 0.065	0.349 ± 0.092	0.481 ± 0.056	0.466 ± 0.066
		X	-	0.722 ± 0.019	0.614 ± 0.081	0.598 ± 0.064	0.591 ± 0.060	0.543 ± 0.083	0.658 ± 0.054	0.624 ± 0.053
tDCS FOG		Adversarial debiasing (Single attribute)	-	0.718 ± 0.013	0.654 ± 0.091	0.644 ± 0.077	0.605 ± 0.064	0.597 ± 0.097	0.665 ± 0.056	0.594 ± 0.053
	DeepConv LSTM	Adversarial debiasing (Multi-Head MLP)	-	0.699 ± 0.023	0.599 ± 0.074	0.616 ± 0.053	0.618 ± 0.056	0.532 ± 0.080	0.640 ± 0.047	0.621 ± 0.051
		Multi-site Transfer	DeFOG + Daphnet	0.752 ± 0.022	$\textbf{0.658} \pm \textbf{0.078}$	0.578 ± 0.054	0.626 ± 0.056	$\textbf{0.636} \pm \textbf{0.072}$	$\textbf{0.704} \pm \textbf{0.055}$	$\textbf{0.680} \pm \textbf{0.044}$
	ResNet	Generic Activity Feature Transfer	-	$\textbf{0.763} \pm \textbf{0.024}$	0.584 ± 0.086	0.567 ± 0.074	0.573 ± 0.074	0.523 ± 0.088	0.615 ± 0.060	0.563 ± 0.057
		×	-	0.722 ± 0.027	0.343 ± 0.088	0.337 ± 0.067	0.318 ± 0.073	0.206 ± 0.067	0.274 ± 0.072	0.203 ± 0.064
		Threshold optimizer	-	0.707 ± 0.027	0.323 ± 0.076	0.371 ± 0.079	0.300 ± 0.087	0.302 ± 0.068	0.325 ± 0.061	0.259 ± 0.069
	RF	Multi-site Transfer	Daphnet	0.722 ± 0.027	0.343 ± 0.088	0.337 ± 0.067	0.318 ± 0.073	0.206 ± 0.067	0.274 ± 0.072	0.203 ± 0.064
		×	-	$\textbf{0.778} \pm \textbf{0.033}$	$\textbf{0.464} \pm \textbf{0.080}$	0.423 ± 0.078	0.481 ± 0.082	0.317 ± 0.060	0.362 ± 0.073	0.404 ± 0.075
De Souza		Adversarial debiasing (Single attribute)	-	0.777 ± 0.022	0.455 ± 0.082	0.446 ± 0.080	0.472 ± 0.084	0.328 ± 0.074	$\textbf{0.391} \pm \textbf{0.086}$	0.416 ± 0.082
	LSTM	Adversarial debiasing (Multi-Head MLP)	-	0.773 ± 0.033	0.438 ± 0.071	0.439 ± 0.071	0.509 ± 0.075	0.327 ± 0.070	0.372 ± 0.068	$\textbf{0.419} \pm \textbf{0.073}$
		Multi-site Transfer	Daphnet	0.766 ± 0.035	0.455 ± 0.083	0.393 ± 0.076	0.455 ± 0.076	$\textbf{0.361} \pm \textbf{0.072}$	0.277 ± 0.058	0.349 ± 0.066
	ResNet	Generic Activity Feature Transfer	-	0.712 ± 0.028	0.430 ± 0.077	$\textbf{0.494} \pm \textbf{0.077}$	$\textbf{0.509} \pm \textbf{0.071}$	0.360 ± 0.066	0.333 ± 0.075	0.381 ± 0.065
		×	-	0.513 ± 0.021	0.205 ± 0.128	0.216 ± 0.092	0.277 ± 0.111	0.024 ± 0.030	0.052 ± 0.033	0.058 ± 0.034
		Threshold optimizer	-	0.509 ± 0.012	0.093 ± 0.077	0.041 ± 0.060	0.041 ± 0.032	0.136 ± 0.094	0.037 ± 0.048	0.180 ± 0.099
	RF	Multi-site Transfer	De Souza	0.489 ± 0.011	0.252 ± 0.148	0.251 ± 0.173	0.252 ± 0.157	0.116 ± 0.111	0.013 ± 0.019	0.103 ± 0.099
- Daphnet	DeepConv LSTM	×	-	0.613 ± 0.037	0.330 ± 0.114	0.383 ± 0.101	0.401 ± 0.105	0.185 ± 0.090	0.312 ± 0.107	0.285 ± 0.108
		Adversarial debiasing (Single Attribute)	-	0.602 ± 0.022	0.378 ± 0.165	0.368 ± 0.101	0.377 ± 0.131	0.223 ± 0.137	0.306 ± 0.123	0.241 ± 0.155
		Adversarial debiasing (Multi MLP Head)	-	0.620 ± 0.030	$\textbf{0.456} \pm \textbf{0.142}$	0.418 ± 0.099	$\textbf{0.489} \pm \textbf{0.116}$	$\textbf{0.248} \pm \textbf{0.130}$	0.321 ± 0.104	0.307 ± 0.122
		Multi-site Transfer	De Souza	$\textbf{0.638} \pm \textbf{0.034}$	0.301 ± 0.124	$\textbf{0.595} \pm \textbf{0.106}$	0.443 ± 0.118	0.177 ± 0.105	$\textbf{0.431} \pm \textbf{0.105}$	$\textbf{0.292} \pm \textbf{0.116}$
	ResNet	Generic Activity Feature Transfer	-	0.636 ± 0.034	0.450 ± 0.152	0.385 ± 0.120	0.440 ± 0.111	0.121 ± 0.077	0.199 ± 0.078	0.155 ± 0.065

Model Biases. On all datasets and attributes, both RF and DeepConvLSTM models did not meet the four-fifths threshold for fairness (DPR < 0.8 and EOR < 0.8). Across four datasets, the RF model showed the average DPR of

0.383, 0.391, and 0.396, and the average EOR of 0.176, 0.238, and 0.202 for sex, age, and disease duration attributes, respectively. The DeepConvLSTM model showed the average DPR of 0.453, 0.483, and 0.511, and the average EOR of 0.336, 0.413, and 0.423 for sex, age, and disease duration attributes, respectively, across four datasets. Overall, compared to the RF model, DeepConvLSTM showed an increase in the average DPR of 0.070, 0.092, and 0.115 and an increase in the average EOR of 0.160, 0.175, and 0.222 for sex, age, and disease duration, respectively.

4.2 Model Performance and Bias with Bias Mitigation

4.2.1 Threshold Optimizer.

Model Performance. Across all datasets, an average change of -0.001 in the F1-score was observed after applying the threshold optimizer.

Model Biases. Interestingly, the threshold optimizer reduced the average DPR for the RF model across all datasets, showing -0.108, -0.167, and -0.181 for sex, age, and disease duration attributes, respectively. However, on average, EOR showed an increase of +0.087, +0.051, and +0.085 for sex, age, and disease duration attributes, respectively. Applying the threshold optimizer to the RF model showed mixed results for the DPR and EOR, with an average change in DPR of -0.152 and EOR of +0.074 across all datasets and attributes.

4.2.2 Adversarial Debiasing for Each Attribute Independently.

Model Performance. The change in the F1-score from applying adversarial debiasing for each attribute independently was not statistically significant, with an average change in F1-score of -0.018 across all four datasets.

Model Biases. Across datasets, adversarial debiasing for each attribute showed average changes in the DPR of +0.017, -0.017, and -0.032 and EOR of +0.022, -0.005, and -0.024 for sex, age, and disease duration, respectively. There was an overall decrease in EOR and DPR, with an average change in DPR of -0.011 and EOR of -0.002 across all datasets and attributes.

4.2.3 Adversarial Debiasing using Multi-head MLP.

Model Performance. When simultaneously debiasing for all attributes using a multi-head MLP, F1-scores for DeepConvLSTM were statistically similar to those when not applying adversarial debiasing, with an average change in F1-score of -0.009 across the datasets.

Model Biases. Multi-head debiasing showed an overall increase in DPR and EOR across datasets for all attributes. The average changes in DPR across all datasets were +0.017, +0.003, and +0.036 for sex, age, and disease duration attributes, respectively. The average changes in EOR across all datasets were +0.009, -0.004, and +0.014, for sex, age, and disease duration attributes, respectively. Overall, the average change in DPR was +0.018, and EOR was +0.006 across all datasets and attributes.

4.2.4 Transfer Learning from Multi-site Datasets.

Model Performance. Applying transfer learning from multi-site datasets produced different results for the RF and DeepConvLSTM models. For the RF model, multi-site transfer learning did not significantly change the F1-score, with an average change of -0.005 across the four datasets. For the DeepConvLSTM model, applying multi-site transfer learning led to an average change in F1-score of +0.026 across the four datasets.

Model Biases. When using the RF model, multi-site transfer learning decreased the DPR and EOR across all attributes in the four datasets, with average changes in DPR of -0.021, -0.014, and -0.026, and average changes in EOR of +0.000, -0.035, and -0.004 for the sex, age, and disease duration attributes respectively. Unlike the RF model, the DeepConvLSTM model showed an average increase in DPR and EOR for all attributes across all

[,] Vol. 1, No. 1, Article . Publication date: February 2025.

datasets. DPR showed an average increase of +0.023, +0.048, and +0.011, and EOR showed an average increase of +0.046, +0.047, and +0.023 for sex, age, and disease duration, respectively. There was an overall change in DPR of -0.020 and EOR of -0.013 across all datasets and attributes for the RF model, and an overall average change in DPR of +0.027 and EOR of +0.039 across all datasets and attributes for the DeepConvLSTM model.

4.2.5 Transfer Learning from Generic Human Activity Representation.

Model Performance. When transferring generic activity representation for FOG representation, the ResNet model showed an average change in the F1-score of +0.018 across the four datasets.

Model Biases. Finetuning generic activity representation improved the average DPR across the datasets for all the attributes, with an average change in DPR of +0.035, +0.013, and +0.019 for sex, age, and disease duration, respectively. However, it showed mixed results for EOR with changes with +0.006, -0.028, and -0.030 for sex, age, and disease duration, respectively. Overall, the average change in DPR was +0.022, and EOR was -0.017 across all datasets and attributes.

5 DISCUSSION

5.1 Model Performance and Bias without Bias Mitigation

5.1.1 Model Performance. Prior to applying any bias mitigation algorithms, the DeepConvLSTM model performed better than the RF model, with an average performance increase of +0.127 across all four datasets. This was consistent with prior work by Kwon *et al.* [49] that showed that deep learning models like the DeepConvLSTM model effectively detected FOG compared to shallow machine learning models. The highest model performance was observed with the De Souza dataset with a macro F1-score of 0.778, which could be attributed to the low complexity of the study protocol, which was mainly based on turning [18], compared to other datasets that included both turning, walking, and other activities, such as sit, stand, or opening doors.

5.1.2 Model Bias. Without bias mitigation, no sensitive attribute was found fair across the four datasets (meeting the 0.8 threshold for DPR and EOR). The tDCS FOG dataset had the highest DPR and EOR values, with DPRs of 0.624, 0.693, and 0.676 and EORs of 0.409, 0.573, and 0.455 for sex, age, and disease duration, respectively. We consider this came from the dataset's diversity of subjects (N=62), which was significantly larger than others from Table 1. This shows that the data representation and bias are essential factors in mitigating the bias in the developed models. Across all datasets, the sex attribute showed the lowest average values for DPR and EOR of 0.418 and 0.277, as opposed to average values of DPR and EOR values of 0.450 and 0.373 for age and average values DPR and EOR of 0.459 and 0.347 for disease duration. We suspect this comes from the limited representation of females in PD research in general due to lower incidence of PD in women and more delayed referrals to movement disorders specialists [7, 80, 90].

5.2 Model Performance and Bias with Bias Mitigation

5.2.1 Model Performance. For the shallow RF model, the decrease in F1-score across all datasets on average was observed when using the bias mitigation approaches: -0.001 for the threshold optimizer and -0.005 for multi-site transfer. For the DeepConvLSTM model, the average change in F1-score across all datasets was -0.018 for adversarial debiasing for single attribute, and -0.009 for adversarial debiasing using the multi-head MLP. The decrease in F1-score (average change in F1-score of -0.001) after applying the threshold optimizer was consistent with work by Hardt *et al.*[35], which showed that finding a threshold that optimized for DPR or EOR for all groups was sub-optimal in overall performance. Similarly, the decrease in F1-score after applying adversarial debiasing (average change of -0.018 for single attribute adversarial debiasing and -0.009 for adversarial debiasing using

multi-head MLP) was consistent with prior work [103] that showed that by adding fairness constraints during training, the model converges with lower performance.

On the contrary, there was an increase in F1-score with transfer learning from multi-site datasets and from generic human activity representation, with an average change in F1-score across all datasets of +0.026 and +0.019, respectively. The improvement in the F1-score of +0.019 for the generic human activity representation was consistent with the claim by Yuan *et al.* [101] that their pretrained model was useful even in downstream health applications. It is also worth noting that when using transfer learning from generic human activity representation, the most significant increase in F1-score was observed with the DeFOG dataset i.e. +0.077 increase in F1-score. This could be attributed to the fact that the model was pretrained on a dataset from a free-living setting, better-modeled FOG on the downstream target DeFOG dataset that was collected from a subject's home setting as opposed to the other datasets that were collected in a lab setting.

5.2.2 Model Bias. Overall, the conventional bias mitigation techniques (threshold optimizer and single attribute adversarial debiasing) underperformed in improving DPR and EOR, with mixed results across attributes, e.g., -0.152 for DPR and +0.074 for EOR with the threshold optimizer, -0.011 for DPR and -0.002 for EOR with single attribute adversarial debiasing. On the other hand, DPR and EOR improved for DeepConvLSTM when using transfer learning with a multi-site dataset (+0.027 for DPR and +0.039 for EOR on average across all attributes and datasets) and generic human activity representation (+0.022 for DPR and -0.017 for EOR on average across all attributes and datasets). The increase in DPR and EOR with multi-site transfer and generic feature activity transfer supported our initial hypotheses that multi-site and rich human activity datasets will help train less biased models effectively in improving the DPR and EOR values with respect to the subject attributes. With generic feature activity transfer, the largest increase in DPR and EOR was observed with the DeFOG dataset, i.e., +0.040 average increase in DPR and +0.078 average increase in EOR across all attributes. This could also be attributed to the similar free-living setting in which the pretrain dataset and the target DeFOG dataset were collected and the diversity of subjects in the pretrain dataset.

5.3 Limitations of Existing Fair ML methods for HAR

In the context of HAR and FOG detection models specifically, existing approaches from the Fair ML literature did not consistently increase the DPR and EOR values, as shown by the mixed results in the average change in DPR and EOR across all attributes and datasets. For approaches such as adversarial debiasing, we experienced common problems related to adversarial learning such as mode collapse, increased sensitiveness to hyperparameters like the learning rate, and vanishing gradients since two networks were trained concurrently [94]. More importantly, we found that bias mitigation approaches that focused on individual protected attributes at a time (threshold optimizer and single attribute adversarial debiasing) did not capture the association between protected attributes, such as age, sex, and disease duration, that were shown in previous studies [27, 58] as associated predictors of FOG in PD patients.

5.4 Importance of Multi-site Transfer Learning

Our results showed that multi-site transfer learning with the DeepConvLSTM model was effective at increasing the trained model's F1-score and reducing the model biases with respect to each protected subject attribute. It showed the highest average change in F1-score across datasets(+0.026) and the highest average change in DPR and EOR across all datasets and attributes (+0.027 in DPR and +0.039 in EOR). The increase in F1-score, DPR, and EOR with the multi-site transfer learning is due to the model being exposed to more diverse data from diverse patient demographics, different study protocols for invoking FOG, and different study settings, such as home and lab settings. This enables the model to learn richer feature representations of FOG, which helps the model overcome any inherent sampling biases in the target datasets while achieving better generalization.

[,] Vol. 1, No. 1, Article . Publication date: February 2025.

On the Bias, Fairness, and Bias Mitigation for a Wearable-based Freezing of Gait Detection in Parkinson's Disease • 15

5.5 Limitations and Future Work

Our results showed that transfer learning from multi-site datasets and from generic human activity representation increased the model performance (F1-score) and decreased the model biases, as shown by the increase in EOR and DPR with respect to the protected attributes. Yet, there exist a few limitations in this study, outlined below:

5.5.1 Biases in Demographic Representation. In all of the datasets, females were underrepresented, representing an inherent sampling bias and limited diversity in the data. Consequently, our conclusions on the model bias with respect to the sex attribute are potentially associated with the imbalanced representation of the sex attribute among subjects. To address the limited diversity in our datasets with respect to some protected attributes such as sex, we plan to increase our training data and its diversity. We plan on exploring techniques such as cross-modal HAR techniques such as IMUGPT [52] and IMUTube [51] to create synthetic data either from video [18, 48] and motion capture datasets [85, 92] capturing FOG episodes from more diverse demographic backgrounds (age, sex, race, etc.) in the pretraining process for our FOG detection models.

5.5.2 Intersectionality. Our work focused on evaluating the bias and fairness of the models with respect to single attributes such as sex, age, and disease duration separately. However, a few studies have highlighted the association between multiple protected attributes (or intersectionality) in the representation of FOG and PD among patients, e.g., women developing PD later than men [27, 58]. We plan on studying bias and fairness evaluation with respect to multi-attribute subject groups using approaches such as intersectional fairness [53, 93] to account for the different associations between protected attributes, such as age and sex, that contribute to the PD and FOG representation among patients.

5.5.3 Scope of Fairness Evaluation. The scope of fairness evaluation in our work was restricted to demographic and PD condition variables present in the datasets used. Future work needs to expand the scope of fairness evaluation to include evaluation with respect to FOG types (i.e., tremulous, akinetic) and other factors, such as the geographic locations of the subjects and the type of wearable devices used.

5.5.4 Temporal Window Size for FOG Detection. We employed the standard activity recognition chain with the sliding window approach [12], using the different optimal temporal window sizes reported in the papers describing the FOG wearable-based datasets. Future work will focus on developing a data-driven method to determine a single temporal optimal window size that generalizes across multiple FOG benchmark wearable-based datasets.

5.5.5 Clinical Validity of Four-fifths Rule. We used the four-fifths rule (0.8 threshold) commonly used in the Fair ML literature and AI Fairness toolkits as a rule of thumb to determine whether a model achieved fairness. However, recent studies [72] have found that the four-fifths rule can only measure effect size and not statistical significance, in addition to being unreliable with small sample sizes. To address the limitations of applying the four-fifths rule as a threshold for determining fairness in healthcare applications, a potential area for future work involves working with clinicians to determine a context-specific fairness threshold for FOG and PD models in general.

5.5.6 Advanced Transfer Learning Techniques. In this work, we only applied basic transfer learning techniques such as freezing selected parts of the model architecture and finetuning other parts of the model architecture. In addition, when applying transfer learning, we downsampled the sampling frequencies of the source and target datasets, which reduced the resolution of the wearable sensor readings. Our future work will explore different transfer learning methods, such as domain adaptation for time series data [39], which account for the difference in time and frequency representations of features extracted from source and target datasets.

5.5.7 Advanced HAR Techniques for FOG Detection. Lastly, in our work, we only experimented with a few ML models, i.e., RF with ECDF features [10, 34], DeepConvLSTM [66], and ResNet [40]. We also plan on evaluating the bias and fairness of a more extensive set of models that will include transformer-based models [19] that have been used on wearable sensor datasets in HAR. In addition, we plan on experimenting with different feature extraction techniques in training shallow models like RF such as time series features extracted from packages such as TSFresh [14] that provides more time series features and automated feature engineering.

6 CONCLUSION

FOG, a major cause of falls in PD [64], requires continuous monitoring for personalized interventions [59]. While existing FOG detection models trained on wearable sensor data can detect FOG episodes, no studies have systematically analyzed the bias and fairness of these models with respect to patient demographics and disease conditions. Our analysis of the state-of-the-art FOG detection models [10, 40, 66] using standard group fairness metrics (DPR and EOR) [2, 23] showed that the FOG detection models failed to meet the four-fifths fairness threshold [16] (DPR and EOR below 0.8) across protected attributes like age, sex, and disease duration in the four public datasets used [5, 18, 43]. Our results also showed that conventional bias mitigation approaches like threshold optimizer [35] and adversarial debiasing [103] were insufficient to consistently improve fairness metrics (DPR and EOR) across datasets and attributes. In contrast, the transfer learning methods (multi-site transfer and generic feature activity transfer) were more effective with the deep learning model [40, 66] at increasing overall DPR or EOR across datasets and attributes (average change in DPR and EOR of +0.027 and +0.039 for multi-site transfer, +0.022 and -0.017 for generic feature activity transfer) while increasing the performance of the model (average change in F1-score of +0.026 and +0.018 for multi-site transfer and generic feature activity transfer, respectively). Our findings supported the hypothesis that FOG detection models could benefit from diverse representations across human activity contexts and patient demographics. A few limitations in our work included the underrepresentation of females due to inherent sampling bias in the datasets [86, 90] and the need for bias and fairness evaluation of multi-attribute subject groups. Consequently, our future will address these limitations through applying data augmentation methods [51, 52], and intersectional fairness analysis across multiple attribute groups [53, 93]. In conclusion, we hope this work will motivate future research incorporating the bias and fairness evaluation of models developed for FOG, PD, and HAR applications. In the broader sense, we hope this research will lead to the development of HAR-based technologies that perform consistently across the demographics where they are deployed.

REFERENCES

- Hamza Abujrida, Emmanuel Agu, and Kaveh Pahlavan. 2017. Smartphone-based gait assessment to infer Parkinson's disease severity using crowdsourced data. In 2017 IEEE Healthcare Innovations and Point of Care Technologies (HI-POCT). IEEE, IEEE, Bethesda, MD, USA, 208–211.
- [2] Alekh Agarwal, Alina Beygelzimer, Miroslav Dudik, John Langford, and Hanna Wallach. 2018. A Reductions Approach to Fair Classification. In Proceedings of the 35th International Conference on Machine Learning (Proceedings of Machine Learning Research, Vol. 80), Jennifer Dy and Andreas Krause (Eds.). PMLR, Vancouver, Canada, 60–69. https://proceedings.mlr.press/v80/agarwal18a.html
- [3] Fayez Alharbi, Lahcen Ouarbya, and Jamie A Ward. 2020. Synthetic Sensor Data for Human Activity Recognition. In 2020 International Joint Conference on Neural Networks (IJCNN). IEEE, Virtual, 1–9. https://doi.org/10.1109/IJCNN48605.2020.9206624
- [4] Guido Alves, Bernd Müller, Karen Herlofson, Ineke HogenEsch, Wenche Telstad, Dag Aarsland, Ole-Bjørn Tysnes, and Jan Petter Larsen. 2009. Incidence of Parkinson's disease in Norway: the Norwegian ParkWest study. *Journal of Neurology, Neurosurgery & Psychiatry* 80, 8 (2009), 851–857.
- [5] Marc Bächlin, Meir Plotnik, Daniel Roggen, Inbal Maidan, Jeffrey M. Hausdorff, Nir Giladi, and Gerhard Tröster. 2010. Wearable assistant for Parkinson's disease patients with the freezing of gait symptom. *Trans. Info. Tech. Biomed.* 14, 2 (Mar 2010), 436–446. https://doi.org/10.1109/TITB.2009.2036165
- [6] Meagan Bailey, Sharlet Anderson, and Deborah A Hall. 2020. Parkinson's disease in African Americans: A review of the current literature. Journal of Parkinson's Disease 10, 3 (2020), 831–841.
- , Vol. 1, No. 1, Article . Publication date: February 2025.

On the Bias, Fairness, and Bias Mitigation for a Wearable-based Freezing of Gait Detection in Parkinson's Disease • 17

- [7] Marzia Baldereschi, A Di Carlo, Walter A Rocca, P Vanni, S Maggi, E Perissinotto, F Grigoletto, L Amaducci, and Domenico Inzitari. 2000. Parkinson's disease and parkinsonism in a longitudinal study: two-fold higher incidence in men. *Neurology* 55, 9 (2000), 1358–1363.
- [8] R. K. E. Bellamy, K. Dey, M. Hind, S. C. Hoffman, S. Houde, K. Kannan, P. Lohia, J. Martino, S. Mehta, A. Mojsilović, S. Nagar, K. Natesan Ramamurthy, J. Richards, D. Saha, P. Sattigeri, M. Singh, K. R. Varshney, and Y. Zhang. 2019. AI Fairness 360: An extensible toolkit for detecting and mitigating algorithmic bias. *IBM Journal of Research and Development* 63, 4/5 (2019), 4:1–4:15. https://doi.org/10.1147/JRD.2019.2942287
- [9] Luigi Borzi, Ivan Mazzetta, Alessandro Zampogna, Antonio Suppa, Gabriella Olmo, and Fernanda Irrera. 2021. Prediction of freezing of gait in Parkinson's disease using wearables and machine learning. Sensors 21, 2 (2021), 614.
- [10] Leo Breiman. 2001. Random forests. Machine Learning 45 (2001), 5-32.
- [11] Rene Peter Bremm, Lukas Pavelka, Maria Moscardo Garcia, Laurent Mombaerts, Rejko Krüger, and Frank Hertel. 2024. Sensor-Based Quantification of MDS-UPDRS III Subitems in Parkinson's Disease Using Machine Learning. Sensors 24, 7 (2024), 2195. https://doi.org/10.3390/s24072195
- [12] Andreas Bulling, Ulf Blanke, and Bernt Schiele. 2014. A tutorial on human activity recognition using body-worn inertial sensors. ACM Computing Surveys (CSUR) 46, 3 (2014), 1–33.
- [13] Carlotta Caramia, Diego Torricelli, Maurizio Schmid, Adriana Muñoz-Gonzalez, Jose Gonzalez-Vargas, Francisco Grandas, and Jose L. Pons. 2018. IMU-Based Classification of Parkinson's Disease From Gait: A Sensitivity Analysis on Sensor Location and Feature Selection. IEEE Journal of Biomedical and Health Informatics 22, 6 (2018), 1765–1774. https://doi.org/10.1109/JBHI.2018.2865218
- [14] Maximilian Christ, Nils Braun, Julius Neuffer, and Andreas W. Kempa-Liehr. 2018. Time Series FeatuRe Extraction on basis of Scalable Hypothesis tests (tsfresh – A Python package). *Neurocomputing* 307 (2018), 72–77. https://doi.org/10.1016/j.neucom.2018.03.067
- [15] Ozkan Cigdem, Hasan Demirel, and Devrim Unay. 2019. The Performance of Local-Learning Based Clustering Feature Selection Method on the Diagnosis of Parkinson's Disease Using Structural MRI. In 2019 IEEE International Conference on Systems, Man and Cybernetics (SMC). IEEE, Bari, Italy, 1286–1291. https://doi.org/10.1109/SMC.2019.8914611
- [16] U.S. Equal Employment Opportunity Commission. 2023. Select Issues: Assessing Adverse Impact in Software, Algorithms, and Artificial Intelligence Used in Employment Selection Procedures Under Title VII of the Civil Rights Act of 1964. https://www.eeoc.gov/laws/ guidance/select-issues-assessing-adverse-impact-software-algorithms-and-artificial#_ednref14.
- [17] Nabila Dahodwala, Andrew Siderowf, Ming Xie, Elizabeth Noll, Matthew Stern, and David S Mandell. 2009. Racial differences in the diagnosis of Parkinson's disease. *Movement Disorders* 24, 8 (2009), 1200–1205.
- [18] Claudia Eunice Neves de Oliveira, Caroline Ribeiro de Souza, Renata de Castro Treza, Sandy Mikie Hondo, Emanuele Los Angeles, Claudionor Bernardo, Thiago Kenzo Fujioka Shida, Luana dos Santos de Oliveira, Thayna Magalhães Novaes, Débora da Silva Fragoso de Campos, Emerson Gisoldi, Margarete de Jesus Carvalho, and Daniel Boari Coelho. 2022. A Public Data Set With Ground Reaction Forces of Human Balance in Individuals With Parkinson's Disease. Frontiers in Neuroscience 16 (2022). https://doi.org/10.3389/fnins.2022.865882
- [19] Iveta Dirgová Luptáková, Martin Kubovčík, and Jiří Pospíchal. 2022. Wearable Sensor-Based Human Activity Recognition with Transformer Model. Sensors 22, 5 (2022), 18 pages. https://doi.org/10.3390/s22051911
- [20] Aiden Doherty, Dan Jackson, Nils Hammerla, Thomas Plötz, Patrick Olivier, Malcolm H Granat, Tom White, Vincent T Van Hees, Michael I Trenell, Christoper G Owen, et al. 2017. Large scale population assessment of physical activity using wrist worn accelerometers: the UK biobank study. PLOS One 12, 2 (2017), e0169649.
- [21] E Ray Dorsey, Alexis Elbaz, Emma Nichols, Nooshin Abbasi, Foad Abd-Allah, Ahmed Abdelalim, Jose C Adsuar, Mustafa Geleto Ansha, Carol Brayne, Jee-Young J Choi, et al. 2018. Global, regional, and national burden of Parkinson's disease, 1990–2016: a systematic analysis for the Global Burden of Disease Study 2016. *The Lancet Neurology* 17, 11 (2018), 939–953.
- [22] Xin Du, Katayoun Farrahi, and Mahesan Niranjan. 2019. Transfer learning across human activities using a cascade neural network architecture. In Proceedings of the 2019 ACM International Symposium on Wearable Computers (London, United Kingdom) (ISWC '19). Association for Computing Machinery, New York, NY, USA, 35–44. https://doi.org/10.1145/3341163.3347730
- [23] Cynthia Dwork, Moritz Hardt, Toniann Pitassi, Omer Reingold, and Richard Zemel. 2012. Fairness through awareness. In Proceedings of the 3rd Innovations in Theoretical Computer Science Conference (Cambridge, Massachusetts) (ITCS '12). Association for Computing Machinery, New York, NY, USA, 214–226. https://doi.org/10.1145/2090236.2090255
- [24] Kaylena A Ehgoetz Martens, James M Shine, Courtney C Walton, Matthew J Georgiades, Moran Gilat, Julie M Hall, Alana J Muller, Jennifer YY Szeto, and Simon JG Lewis. 2018. Evidence for subtypes of freezing of gait in Parkinson's disease. *Movement Disorders* 33, 7 (2018), 1174–1178.
- [25] Luc JW Evers, Jesse H Krijthe, Marjan J Meinders, Bastiaan R Bloem, and Tom M Heskes. 2019. Measuring Parkinson's disease over time: The real-World Within-Subject Reliability of the MDS-UPDRS. *Movement Disorders* 34, 10 (2019), 1480–1487.
- [26] Emilio Ferrara. 2023. Fairness and bias in artificial intelligence: A brief survey of sources, impacts, and mitigation strategies. Sci 6, 1 (2023), 3.
- [27] Michelle E Fullard, Dylan P Thibault, Veronica Todaro, Susan Foster, Lori Katz, Robin Morgan, Drew S Kern, Jason M Schwalb, Enrique Urrea Mendoza, Nabila Dahodwala, et al. 2018. Sex disparities in health and health care utilization after Parkinson diagnosis: Rethinking PD associated disability. *Parkinsonism & related disorders* 48 (2018), 45–50.

- 18 Odonga, et al.
- [28] Judy Wawira Gichoya, Imon Banerjee, Ananth Reddy Bhimireddy, John L Burns, Leo Anthony Celi, Li-Ching Chen, Ramon Correa, Natalie Dullerud, Marzyeh Ghassemi, Shih-Cheng Huang, et al. 2022. AI recognition of patient race in medical imaging: a modelling study. *The Lancet Digital Health* 4, 6 (2022), e406–e414.
- [29] Christopher G Goetz, Barbara C Tilley, Stephanie R Shaftman, Glenn T Stebbins, Stanley Fahn, Pablo Martinez-Martin, Werner Poewe, Cristina Sampaio, Matthew B Stern, Richard Dodel, et al. 2008. Movement Disorder Society-sponsored revision of the Unified Parkinson's Disease Rating Scale (MDS-UPDRS): scale presentation and clinimetric testing results. *Movement Disorders* 23, 15 (2008), 2129–2170.
- [30] N Jabin Gong, Gari D Clifford, Christine D Esper, Stewart A Factor, J Lucas McKay, and Hyeokhyen Kwon. 2023. Classifying Tremor Dominant and Postural Instability and Gait Difficulty Subtypes of Parkinson's Disease from Full-Body Kinematics. *Sensors* 23, 19 (2023), 8330.
- [31] N. Jabin Gong, Gari D. Clifford, Christine D. Esper, Stewart A. Factor, J. Lucas McKay, and Hyeokhyen Kwon. 2023. Classifying Tremor Dominant and Postural Instability and Gait Difficulty Subtypes of Parkinson's Disease from Full-Body Kinematics. *Sensors* 23 (10 2023), 8330. Issue 19. https://doi.org/10.3390/s23198330
- [32] Charlotte A Haaxma, Bastiaan R Bloem, George F Borm, Wim JG Oyen, Klaus L Leenders, Silvia Eshuis, Jan Booij, Dean E Dluzen, and Martin WIM Horstink. 2007. Gender differences in Parkinson's disease. *Journal of Neurology, Neurosurgery & Psychiatry* 78, 8 (2007), 819–824.
- [33] Mark Hallett. 2008. The intrinsic and extrinsic aspects of freezing of gait. Movement Disorders 23, S2 (2008), S439-S443.
- [34] Nils Y. Hammerla, Reuben Kirkham, Peter Andras, and Thomas Ploetz. 2013. On preserving statistical characteristics of accelerometry data using their empirical cumulative distribution. In *Proceedings of the 2013 International Symposium on Wearable Computers* (Zurich, Switzerland) (*ISWC '13*). Association for Computing Machinery, New York, NY, USA, 65–68. https://doi.org/10.1145/2493988.2494353
- [35] Moritz Hardt, Eric Price, and Nathan Srebro. 2016. Equality of opportunity in supervised learning. In Proceedings of the 30th International Conference on Neural Information Processing Systems (Barcelona, Spain) (NIPS'16). Curran Associates Inc., Red Hook, NY, USA, 3323–3331.
- [36] Harish Haresamudram, Apoorva Beedu, Varun Agrawal, Patrick L. Grady, Irfan Essa, Judy Hoffman, and Thomas Plötz. 2020. Masked reconstruction based self-supervision for human activity recognition. In *Proceedings of the 2020 ACM International Symposium* on *Wearable Computers* (Virtual Event, Mexico) (ISWC '20). Association for Computing Machinery, New York, NY, USA, 45–49. https://doi.org/10.1145/3410531.3414306
- [37] Harish Haresamudram, Irfan Essa, and Thomas Plötz. 2021. Contrastive predictive coding for human activity recognition. Proceedings of the ACM on Interactive, Mobile, Wearable and Ubiquitous Technologies 5, 2 (2021), 1–26.
- [38] M. Hariharan, Kemal Polat, and R. Sindhu. 2014. A new hybrid intelligent system for accurate detection of Parkinson's disease. Computer Methods and Programs in Biomedicine 113, 3 (2014), 904–913. https://doi.org/10.1016/j.cmpb.2014.01.004
- [39] Huan He, Owen Queen, Teddy Koker, Consuelo Cuevas, Theodoros Tsiligkaridis, and Marinka Zitnik. 2023. Domain adaptation for time series under feature and label shifts. In *Proceedings of the 40th International Conference on Machine Learning* (Honolulu, Hawaii, USA) (*ICML'23*). JMLR.org, Honolulu, Hawaii, USA, Article 518, 29 pages.
- [40] Kaiming He, Xiangyu Zhang, Shaoqing Ren, and Jian Sun. 2016. Deep Residual Learning for Image Recognition. In 2016 IEEE Conference on Computer Vision and Pattern Recognition (CVPR). IEEE, Las Vegas, NV, USA, 770–778. https://doi.org/10.1109/CVPR.2016.90
- [41] Alexander Hoelzemann and Kristof Van Laerhoven. 2020. Digging deeper: towards a better understanding of transfer learning for human activity recognition. In Proceedings of the 2020 ACM International Symposium on Wearable Computers (Virtual Event, Mexico) (ISWC '20). Association for Computing Machinery, New York, NY, USA, 50–54. https://doi.org/10.1145/3410531.3414311
- [42] Max Hort and Federica Sarro. 2022. Privileged and unprivileged groups: an empirical study on the impact of the age attribute on fairness. In *Proceedings of the 2nd International Workshop on Equitable Data and Technology* (Pittsburgh, Pennsylvania) (*FairWare '22*). Association for Computing Machinery, New York, NY, USA, 17–24. https://doi.org/10.1145/3524491.3527308
- [43] Addison Howard, Amit Salomon, Eran Gazit, Jeff Hausdorff, HCL Jevster, Leslie Kirsch, Pieter Ginis Maggie, Ryan Holbrook, and Karim F Yasir. 2023. Parkinson's Freezing of Gait Prediction. https://kaggle.com/competitions/tlvmc-parkinsons-freezing-gait-prediction
- [44] Joseph Jankovic. 2008. Parkinson's disease: clinical features and diagnosis. Journal of Neurology, Neurosurgery & Psychiatry 79, 4 (2008), 368–376.
- [45] Zifan Jiang, Salman Seyedi, Emily Griner, Ahmed Abbasi, Ali Bahrami Rad, Hyeokhyen Kwon, Robert O Cotes, and Gari D Clifford. 2024. Evaluating and mitigating unfairness in multimodal remote mental health assessments. PLOS Digital Health 3, 7 (2024), e0000413.
- [46] Deepak Joshi, Aayushi Khajuria, and Pradeep Joshi. 2017. An automatic non-invasive method for Parkinson's disease classification. Computer Methods and Programs in Biomedicine 145 (2017), 135–145. https://doi.org/10.1016/j.cmpb.2017.04.007
- [47] Newton M Kinyanjui, Timothy Odonga, Celia Cintas, Noel CF Codella, Rameswar Panda, Prasanna Sattigeri, and Kush R Varshney. 2020. Fairness of classifiers across skin tones in dermatology. In *International Conference on Medical Image Computing and Computer-Assisted Intervention*. Springer, Springer, Virtual, 320–329.
- [48] Yuki Kondo, Kyota Bando, Ippei Suzuki, Yuta Miyazaki, Daisuke Nishida, Takatoshi Hara, Hideki Kadone, and Kenji Suzuki. 2024. Video-Based Detection of Freezing of Gait in Daily Clinical Practice in Patients With Parkinsonism. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 32 (2024), 2250–2260. https://doi.org/10.1109/TNSRE.2024.3413055

On the Bias, Fairness, and Bias Mitigation for a Wearable-based Freezing of Gait Detection in Parkinson's Disease • 19

- [49] Hyeokhyen Kwon, Gregory D. Abowd, and Thomas Plötz. 2019. Handling annotation uncertainty in human activity recognition. In Proceedings of the 2019 ACM International Symposium on Wearable Computers (London, United Kingdom) (ISWC '19). Association for Computing Machinery, New York, NY, USA, 109–117. https://doi.org/10.1145/3341163.3347744
- [50] Hyeokhyen Kwon, Gari D Clifford, Imari Genias, Doug Bernhard, Christine D Esper, Stewart A Factor, and J Lucas McKay. 2023. An Explainable Spatial-Temporal Graphical Convolutional Network to score Freezing of Gait in Parkinsonian Patients. Sensors 23, 4 (2023), 1766.
- [51] Hyeokhyen Kwon, Catherine Tong, Harish Haresamudram, Yan Gao, Gregory D Abowd, Nicholas D Lane, and Thomas Ploetz. 2020. Imutube: Automatic extraction of virtual on-body accelerometry from video for human activity recognition. Proceedings of the ACM on Interactive, Mobile, Wearable and Ubiquitous Technologies 4, 3 (2020), 1–29.
- [52] Zikang Leng, Hyeokhyen Kwon, and Thomas Ploetz. 2023. Generating Virtual On-body Accelerometer Data from Virtual Textual Descriptions for Human Activity Recognition. In *Proceedings of the 2023 ACM International Symposium on Wearable Computers* (Cancun, Quintana Roo, Mexico) (*ISWC '23*). Association for Computing Machinery, New York, NY, USA, 39–43. https://doi.org/10.1145/3594738. 3611361
- [53] Elle Lett and William G La Cava. 2023. Translating intersectionality to fair machine learning in health sciences. Nature Machine Intelligence 5, 5 (2023), 476–479.
- [54] Xi'ang Li, Jinqi Luo, and Rabih Younes. 2020. ActivityGAN: generative adversarial networks for data augmentation in sensor-based human activity recognition. In Adjunct Proceedings of the 2020 ACM International Joint Conference on Pervasive and Ubiquitous Computing and Proceedings of the 2020 ACM International Symposium on Wearable Computers (Virtual Event, Mexico) (UbiComp/ISWC '20 Adjunct). Association for Computing Machinery, New York, NY, USA, 249–254. https://doi.org/10.1145/3410530.3414367
- [55] Max A. Little *, Patrick E. McSharry, Eric J. Hunter, Jennifer Spielman, and Lorraine O. Ramig. 2009. Suitability of Dysphonia Measurements for Telemonitoring of Parkinson's Disease. *IEEE Transactions on Biomedical Engineering* 56, 4 (2009), 1015–1022. https://doi.org/10.1109/TBME.2008.2005954
- [56] Luyan Liu, Qian Wang, Ehsan Adeli, Lichi Zhang, Han Zhang, and Dinggang Shen. 2016. Feature Selection Based on Iterative Canonical Correlation Analysis for Automatic Diagnosis of Parkinson's Disease. In Medical Image Computing and Computer-Assisted Intervention – MICCAI 2016: 19th International Conference, Athens, Greece, October 17-21, 2016, Proceedings, Part II (Athens, Greece). Springer-Verlag, Berlin, Heidelberg, 1–8. https://doi.org/10.1007/978-3-319-46723-8_1
- [57] Mandy Lu, Qingyu Zhao, Kathleen L. Poston, Edith V. Sullivan, Adolf Pfefferbaum, Marian Shahid, Maya Katz, Leila Montaser-Kouhsari, Kevin Schulman, Arnold Milstein, Juan Carlos Niebles, Victor W. Henderson, Li Fei-Fei, Kilian M. Pohl, and Ehsan Adeli. 2021. Quantifying Parkinson's disease motor severity under uncertainty using MDS-UPDRS videos. *Medical Image Analysis* 73 (2021), 102179. https://doi.org/10.1016/j.media.2021.102179
- [58] Michael Macht, Yvonne Kaussner, Jens Carsten Möller, Karin Stiasny-Kolster, Karla Maria Eggert, Hans-Peter Krüger, and Heiner Ellgring. 2007. Predictors of freezing in Parkinson's disease: a survey of 6,620 patients. *Movement disorders* 22, 7 (2007), 953–956.
- [59] Martina Mancini, Bastiaan R Bloem, Fay B Horak, Simon JG Lewis, Alice Nieuwboer, and Jorik Nonnekes. 2019. Clinical and methodological challenges for assessing freezing of gait: future perspectives. *Movement Disorders* 34, 6 (2019), 783–790.
- [60] Sinziana Mazilu, Alberto Calatroni, Eran Gazit, Daniel Roggen, Jeffrey M Hausdorff, and Gerhard Tröster. 2013. Feature learning for Detection and Prediction of Freezing of Gait in Parkinson's disease. In Machine Learning and Data Mining in Pattern Recognition: 9th International Conference, MLDM 2013, New York, NY, USA, July 19-25, 2013. Proceedings 9. Springer, Springer, New York, NY, USA, 144–158.
- [61] Sinziana Mazilu, Michael Hardegger, Zack Zhu, Daniel Roggen, Gerhard Tröster, Meir Plotnik, and Jeffrey M Hausdorff. 2012. Online detection of freezing of gait with smartphones and machine learning techniques. In 2012 6th International Conference on Pervasive Computing Technologies for Healthcare (PervasiveHealth) and Workshops. IEEE, IEEE, San Diego, CA, USA, 123–130.
- [62] Salama A. Mostafa, Aida Mustapha, Mazin Abed Mohammed, Raed Ibraheem Hamed, N. Arunkumar, Mohd Khanapi Abd Ghani, Mustafa Musa Jaber, and Shihab Hamad Khaleefah. 2019. Examining multiple feature evaluation and classification methods for improving the diagnosis of Parkinson's disease. *Cognitive Systems Research* 54 (2019), 90–99. https://doi.org/10.1016/j.cogsys.2018.12.004
- [63] Nilah Ravi Nair, Lena Schmid, Fernando Moya Rueda, Markus Pauly, Gernot A. Fink, and Christopher Reining. 2023. Dataset Bias in Human Activity Recognition. arXiv:2301.10161 [eess.SP] https://arxiv.org/abs/2301.10161
- [64] John G Nutt, Bastiaan R Bloem, Nir Giladi, Mark Hallett, Fay B Horak, and Alice Nieuwboer. 2011. Freezing of gait: moving forward on a mysterious clinical phenomenon. *The Lancet Neurology* 10, 8 (2011), 734–744.
- [65] Juri Opitz and Sebastian Burst. 2021. Macro F1 and Macro F1. arXiv:1911.03347 [cs.LG] https://arxiv.org/abs/1911.03347
- [66] Francisco Javier Ordóñez and Daniel Roggen. 2016. Deep Convolutional and LSTM Recurrent Neural Networks for Multimodal Wearable Activity Recognition. Sensors 16, 1 (2016), 115. https://doi.org/10.3390/s16010115
- [67] Natasa K Orphanidou, Abir Hussain, Robert Keight, Paulo Lishoa, Jade Hind, and Haya Al-Askar. 2018. Predicting freezing of gait in Parkinsons disease patients using machine learning. In 2018 IEEE Congress on Evolutionary Computation (CEC). IEEE, IEEE, Rio de Janeiro, Brazil, 1–8.

- [68] HuiZe Pang, ZiYang Yu, HongMei Yu, JiBin Cao, YingMei Li, MiaoRan Guo, ChengHao Cao, and GuoGuang Fan. 2021. Use of machine learning method on automatic classification of motor subtype of Parkinson's disease based on multilevel indices of rs-fMRI. *Parkinsonism & Related Disorders* 90 (2021), 65–72.
- [69] Tuan D Pham and Hong Yan. 2017. Tensor decomposition of gait dynamics in Parkinson's disease. IEEE Transactions on Biomedical Engineering 65, 8 (2017), 1820–1827.
- [70] Diane Podsiadlo and Sandra Richardson. 1991. The timed "Up & Go": a test of basic functional mobility for frail elderly persons. Journal of the American geriatrics Society 39, 2 (1991), 142–148.
- [71] WH Poewe, AJ Lees, and GM Stern. 1988. Dystonia in Parkinson's disease: clinical and pharmacological features. Annals of Neurology: Official Journal of the American Neurological Association and the Child Neurology Society 23, 1 (1988), 73–78.
- [72] Manish Raghavan and Pauline T Kim. 2024. Limitations of the" Four-Fifths Rule" and Statistical Parity Tests for Measuring Fairness. Geo. L. Tech. Rev. 8 (2024), 93.
- [73] Tal Reches, Moria Dagan, Talia Herman, Eran Gazit, Natalia A Gouskova, Nir Giladi, Brad Manor, and Jeffrey M Hausdorff. 2020. Using wearable sensors and machine learning to automatically detect freezing of gait during a FOG-provoking test. *Sensors* 20, 16 (2020), 4474.
- [74] Peter Redgrave, Manuel Rodriguez, Yoland Smith, Maria C Rodriguez-Oroz, Stephane Lehericy, Hagai Bergman, Yves Agid, Mahlon R DeLong, and Jose A Obeso. 2010. Goal-directed and habitual control in the basal ganglia: implications for Parkinson's disease. *Nature Reviews Neuroscience* 11, 11 (2010), 760–772.
- [75] Daniel Rodríguez-Martín, Albert Samà, Carlos Pérez-López, Andreu Català, Joan M Moreno Arostegui, Joan Cabestany, Àngels Bayés, Sheila Alcaine, Berta Mestre, Anna Prats, et al. 2017. Home detection of freezing of gait using support vector machines through a single waist-worn triaxial accelerometer. *PloS one* 12, 2 (2017), e0171764.
- [76] Daniel Roggen, Meir Plotnik, and Jeff Hausdorff. 2013. Daphnet Freezing of Gait. UCI Machine Learning Repository. DOI: https://doi.org/10.24432/C56K78.
- [77] Samanta Rosati, Gabriella Balestra, and Marco Knaflitz. 2018. Comparison of Different Sets of Features for Human Activity Recognition by Wearable Sensors. Sensors 18, 12 (2018), 4189. https://doi.org/10.3390/s18124189
- [78] Aaqib Saeed, Tanir Ozcelebi, and Johan Lukkien. 2019. Multi-task self-supervised learning for human activity detection. Proceedings of the ACM on Interactive, Mobile, Wearable and Ubiquitous Technologies 3, 2 (2019), 1–30.
- [79] Aaqib Saeed, Victor Ungureanu, and Beat Gfeller. 2021. Sense and learn: Self-supervision for omnipresent sensors. *Machine Learning with Applications* 6 (2021), 100152.
- [80] Rachel Saunders-Pullman, Cuiling Wang, Kaili Stanley, and Susan B Bressman. 2011. Diagnosis and Referral Delay in Women with Parkinson's disease. Gender Medicine 8, 3 (2011), 209–217.
- [81] Schaafsma, JD and Balash, Y and Gurevich, T and Bartels, AL and Hausdorff, Jeffrey M and Giladi, Nir. 2003. Characterization of freezing of gait subtypes and the response of each to levodopa in Parkinson's disease. *European Journal of Neurology* 10, 4 (2003), 391–398.
- [82] Johannes CM Schlachetzki, Jens Barth, Franz Marxreiter, Julia Gossler, Zacharias Kohl, Samuel Reinfelder, Heiko Gassner, Kamiar Aminian, Bjoern M Eskofier, Jürgen Winkler, et al. 2017. Wearable sensors objectively measure gait parameters in Parkinson's disease. PLOS One 12, 10 (2017), e0183989.
- [83] J.M. Shine, S.T. Moore, S.J. Bolitho, T.R. Morris, V. Dilda, S.L. Naismith, and S.J.G. Lewis. 2012. Assessing the utility of Freezing of Gait Questionnaires in Parkinson's Disease. *Parkinsonism & Related Disorders* 18, 1 (2012), 25–29. https://doi.org/10.1016/j.parkreldis.2011. 08.002
- [84] Charalampos Sotirakis, Zi Su, Maksymilian A Brzezicki, Niall Conway, Lionel Tarassenko, James J FitzGerald, and Chrystalina A Antoniades. 2023. Identification of motor progression in Parkinson's disease using wearable sensors and machine learning. npj Parkinson's Disease 9, 1 (2023), 142.
- [85] Joke Spildooren, Sarah Vercruysse, Kaat Desloovere, Wim Vandenberghe, Eric Kerckhofs, and Alice Nieuwboer. 2010. Freezing of gait in Parkinson's disease: the impact of dual-tasking and turning. *Movement Disorders* 25, 15 (2010), 2563–2570.
- [86] Indu Subramanian, Soania Mathur, Annelien Oosterbaan, Richelle Flanagan, Adrienne M Keener, and Elena Moro. 2022. Unmet needs of women living with Parkinson's disease: gaps and controversies. *Movement Disorders* 37, 3 (2022), 444–455.
- [87] Thasina Tabashum, Robert Cooper Snyder, Megan K O'Brien, and Mark V Albert. 2024. Machine Learning Models for Parkinson Disease: Systematic Review. JMIR Med Inform 12 (17 May 2024), e50117. https://doi.org/10.2196/50117
- [88] Chi Ian Tang, Ignacio Perez-Pozuelo, Dimitris Spathis, Soren Brage, Nick Wareham, and Cecilia Mascolo. 2021. SelfHAR: Improving human activity recognition through self-training with unlabeled data. *Proceedings of the ACM on Interactive, Mobile, Wearable and Ubiquitous Technologies* 5, 1 (2021), 1–30.
- [89] Chi Ian Tang, Ignacio Perez-Pozuelo, Dimitris Spathis, and Cecilia Mascolo. 2021. Exploring Contrastive Learning in Human Activity Recognition for Healthcare. arXiv:2011.11542 [cs.LG] https://arxiv.org/abs/2011.11542
- [90] Anouk Tosserams, Rui Araújo, Tamara Pringsheim, Bart Post, Sirwan KL Darweesh, Joanna IntHout, and Bastiaan R Bloem. 2018. Underrepresentation of women in Parkinson's disease trials. *Movement Disorders* 33, 11 (2018), 1825–1826.

On the Bias, Fairness, and Bias Mitigation for a Wearable-based Freezing of Gait Detection in Parkinson's Disease • 21

- [91] Kush R. Varshney. 2022. Trustworthy Machine Learning. Independently published, Chappaqua, NY, USA. http://trustworthymachinelearning.com
- [92] Griet Vervoort, Aniek Bengevoord, Carolien Strouwen, Esther MJ Bekkers, Elke Heremans, Wim Vandenberghe, and Alice Nieuwboer. 2016. Progression of postural control and gait deficits in Parkinson's disease and freezing of gait: A longitudinal study. *Parkinsonism & Related Disorders* 28 (2016), 73–79.
- [93] Angelina Wang, Vikram V Ramaswamy, and Olga Russakovsky. 2022. Towards intersectionality in machine learning: Including more identities, handling underrepresentation, and performing evaluation. In *Proceedings of the 2022 ACM Conference on Fairness, Accountability, and Transparency*. Association for Computing Machinery, New York, NY, USA, 336–349.
- [94] Hilde Weerts, Miroslav Dudík, Richard Edgar, Adrin Jalali, Roman Lutz, and Michael Madaio. 2023. Fairlearn: Assessing and Improving Fairness of AI Systems. , 8 pages. http://jmlr.org/papers/v24/23-0389.html
- [95] AW Willis, M Schootman, BA Evanoff, JS Perlmutter, and BA Racette. 2011. Neurologist care in Parkinson disease: a utilization, outcomes, and survival study. *Neurology* 77, 9 (2011), 851–857.
- [96] Allison Wright Willis, Bradley A Evanoff, Min Lian, Susan R Criswell, and Brad A Racette. 2010. Geographic and ethnic variation in Parkinson disease: a population-based study of US Medicare beneficiaries. *Neuroepidemiology* 34, 3 (2010), 143–151.
- [97] Zhuang Wu, Min Zhong, Xu Jiang, Bo Shen, Jun Zhu, Yang Pan, Jingde Dong, Jun Yan, Pingyi Xu, Wenbin Zhang, et al. 2020. Can quantitative gait analysis be used to guide treatment of patients with different subtypes of Parkinson's disease? *Neuropsychiatric Disease and Treatment* 16 (2020), 2335–2341.
- [98] Rui Yao, Guosheng Lin, Qinfeng Shi, and Damith C Ranasinghe. 2018. Efficient dense labelling of human activity sequences from wearables using fully convolutional networks. *Pattern Recognition* 78 (2018), 252–266.
- [99] Sofia Yfantidou, Marios Constantinides, Dimitris Spathis, Athena Vakali, Daniele Quercia, and Fahim Kawsar. 2023. Beyond Accuracy: A Critical Review of Fairness in Machine Learning for Mobile and Wearable Computing. arXiv:2303.15585 [cs.CY] https://arxiv.org/ abs/2303.15585
- [100] Chenxi Yuan, Kristin A Linn, and Rebecca A Hubbard. 2023. Algorithmic Fairness of Machine Learning Models for Alzheimer Disease Progression. JAMA Network Open 6, 11 (2023), e2342203–e2342203.
- [101] Hang Yuan, Shing Chan, Andrew P Creagh, Catherine Tong, Aidan Acquah, David A Clifton, and Aiden Doherty. 2024. Self-supervised learning for human activity recognition using 700,000 person-days of wearable data. NPJ Digital Medicine 7, 1 (2024), 91.
- [102] Rich Zemel, Yu Wu, Kevin Swersky, Toni Pitassi, and Cynthia Dwork. 2013. Learning Fair Representations. In Proceedings of the 30th International Conference on Machine Learning (Proceedings of Machine Learning Research, Vol. 28), Sanjoy Dasgupta and David McAllester (Eds.). PMLR, Atlanta, Georgia, USA, 325–333. https://proceedings.mlr.press/v28/zemel13.html
- [103] Brian Hu Zhang, Blake Lemoine, and Margaret Mitchell. 2018. Mitigating Unwanted Biases with Adversarial Learning. In Proceedings of the 2018 AAAI/ACM Conference on AI, Ethics, and Society (New Orleans, LA, USA) (AIES '18). Association for Computing Machinery, New York, NY, USA, 335–340. https://doi.org/10.1145/3278721.3278779
- [104] Han Zhang, Leijie Wang, Yilun Sheng, Xuhai Xu, Jennifer Mankoff, and Anind K. Dey. 2023. A Framework for Designing Fair Ubiquitous Computing Systems. In Adjunct Proceedings of the 2023 ACM International Joint Conference on Pervasive and Ubiquitous Computing & the 2023 ACM International Symposium on Wearable Computing (Cancun, Quintana Roo, Mexico) (UbiComp/ISWC '23 Adjunct). Association for Computing Machinery, New York, NY, USA, 366–373. https://doi.org/10.1145/3594739.3610677
- [105] Rana Zia Ur Rehman, Lynn Rochester, Alison J. Yarnall, and Silvia Del Din. 2021. Predicting the Progression of Parkinson's Disease MDS-UPDRS-III Motor Severity Score from Gait Data using Deep Learning. In 2021 43rd Annual International Conference of the IEEE Engineering in Medicine & Biology Society (EMBC). IEEE, Virtual, 249–252. https://doi.org/10.1109/EMBC46164.2021.9630769
- [106] Kerstin Ziegler, Frauke Schroeteler, Andres O Ceballos-Baumann, and Urban M Fietzek. 2010. A new rating instrument to assess festination and freezing gait in Parkinsonian patients. *Movement Disorders* 25, 8 (2010), 1012–1018.