

# Modeling a nanocantilever based biosensor using a stochastically perturbed harmonic oscillator

Patrick Snyder and Amitabh Joshi  
*Department of Physics, Eastern Illinois University*  
*Charleston, Illinois 61920, USA*  
*joshi@eiu.edu*

Juan D. Serna  
*School of Mathematical and Natural Sciences, University of Arkansas at Monticello*  
*Monticello, Arkansas 71656, USA*  
*serna@uamont.edu*

Nanoscale biosensors are devices designed to detect analytes by combining biological components and physicochemical detectors. One of the well known methods of constructing these sensors is by using nanocantilevers. These microscopic diving boards are coated with binding probes that have an affinity to a particular amino acid, enzyme or protein in living organisms. When these probes attract some target particles, such as biomolecules, they change the vibrating frequency of the cantilever. This process is random in nature and produces fluctuations in the frequency and damping of the cantilever. In this paper, we studied the effect of these fluctuations using a stochastically perturbed classical harmonic oscillator.

*Keywords:* Nanocantilever, biosensor, stochastically perturbed harmonic oscillator.

## 1. Introduction

Biosensors are composite devices consisting of biological sensing elements and transducer systems. The working principle of these devices involves the binding of the desired analyte to the biorecognition element fixed on a suitable support matrix connected to a transducer. The binding of analytes causes changes in the physical and/or chemical properties of the bioreceptive elements together with the support matrix, which can then be sensed by a transducer to generate an electrical signal. This generated signal quantifies the amount of analyte deposited on the system. Classification of biosensors can be based on either biorecognition mechanisms or the methodology of signal transduction.<sup>1</sup>

Nanobiosensors utilizing nanocantilevers can provide extreme sensitivity in the detection of biomolecules (analytes) down to a single-particle level.<sup>2,3</sup> Detecting particular biomolecules can help researchers to recognize pathogens and diseases dur-

ing clinical monitoring. Like many other detectors, nanoscale biosensors are characterized by a quantity called the *dynamic range*, determined by the minimum mass detection limit to the saturation limit in the detector.<sup>3,4</sup>

Nanocantilever biosensors are based on the mechanical motion of cantilevers. A cantilever-based biosensor works as a tiny mechanical device whose mass continually changes as biological analytes attach to it. The attachment of analytes leads to a change in the resonant frequency of the device. The amount of mass deposited on such detectors can be estimated by measuring the shift in the resonance frequency of the resonator. These detectors possess high sensitivity because of their intrinsic mass is small but the sticking of analyte molecules causes significant mass change and hence produces a very large change in the resonant frequency (typically in the range of 15 to 20 %). The quality factor of such nanodevices is also large and that further adds to their sensitivity of detection.<sup>4,5</sup>

Recently, a slightly different kind of nanocantilever has emerged. This kind of nanocantilever-based biosensor device contains nanofluidic channels that are used to detect small mass species, e.g., cells, proteins, etc. The nanofluidic channels allow one to operate under a low pressure background which enhances the quality factor of the detector.<sup>6</sup> In order to measure the change in frequency of nanocantilevers, it is required that the system parameters remain unaltered. Ideally speaking, once the analyte is attached to the nanocantilever, it should not detach or move from it.<sup>7</sup> Such an attachment-detachment or adsorption-desorption of the analyte particles,<sup>8</sup> including their random striking on the detector,<sup>7</sup> which involves momentum transfer to the nanocantilevers,<sup>3</sup> are the main factors for the spread or broadening of the resonant frequency along with shifts in the frequency.

In order to develop a theoretical model related to the detection process for such generic nanocantilevers (a few examples are shown in Fig. 1), it is important to keep all the macroscopic and microscopic factors involved in the detection process in mind.<sup>4</sup> For such a tiny system having dimensions of several tens to hundred nanometers, it is indeed possible to analyze some variables macroscopically, but the proper recognition of such variables is quite critical, which is less obvious and would be negligible if one is not working on the nanoscale. Because amino acids are one of the main bioreceptors used in the nanocantilever-based biosensors, the affinity and specificity of amino acid sequences (peptides, proteins, etc) are a very important factor in such analysis.<sup>4</sup> These amino acids are critically dependent on ligands to which they are attached and responsible for the determination of the geometric dimensions of amino acid sequences. A change in the geometry could cause receptor amino acid sequences to become inactive or even initiate other events like the nonlinear expansion or the stochastic motion of the cantilever. These irregular events (noise) need to be taken into account when nanocantilever biosensors are modeled mathematically.

This paper is organized as follows: In section 2, we give a brief description of how mass deposition causes frequency shifts in the harmonic oscillator. In section 3, we model nanocantilevers mathematically using a stochastically perturbed harmonic oscillator. We explain how noise introduces measurable random frequency fluctuations in the oscillating nanocantilever that are proportional to the mass

deposited on its surface. This feature is used by biosensors as the detecting mechanism.<sup>1,4</sup> Finally, some concluding remarks are presented in section 4.




Fig. 1. Nanocantilever based biosensors in (a) vacuum, (b) gas and (c) solution environments.

## 2. Mass Deposition and Frequency Shift

Nanocantilevers can be simply modeled as damped, driven harmonic oscillators. The oscillating mechanism we consider in this work consists of a mass-spring system, and its equation of motion will be used to describe the *deterministic* behavior of nanocantilevers. As it is well known, this equation includes damping, intrinsic frequency and driving terms. Random damping and random frequency terms due to fluctuation mechanisms in the oscillator change its intrinsic resonance frequency. This is further discussed in sections 3.1 and 3.2. If these frequency changes can be measured, then the oscillators (nanocantilevers) can be used to detect simple physical phenomena such as mass deposition on them.

For a nanocantilever, it is common to measure changes in its frequency of resonance. This frequency shift is directly proportional to the mass deposited on it as we discuss in the following. For a mass-spring system with mass  $m$  and spring constant  $k$ , the characteristic angular frequency under simple harmonic motion is given by  $\omega_0 = \sqrt{k/m}$ . A small change in the mass of the system due to mass

deposition modifies the frequency of oscillation of the system. Taking finite differences of the angular frequency  $\Delta\omega_0$  with respect to mass and solving for  $\Delta m$ , we get

$$\Delta m = -\frac{2m}{\omega_0}\Delta\omega_0. \quad (1)$$

From this equation it is clear that the deposited mass  $\Delta m$  can be estimated by measuring  $\Delta\omega_0$ . For this equation to make physical sense, we assume that the elastic and geometric properties of the nanocantilever remain unchanged after a small mass deposition.

The resonant frequency measurement of the nanocantilever can be done with a microscopic scanning laser Doppler vibrometer setup.<sup>9,10</sup> The setup involves a collimated laser beam passing through a polarizing beam splitter and focused onto a cantilever with the help of a microscope objective. The reflected beam passes through the beam splitter to fall on a photo detector. The typical spot size of the laser beam is about 1–2  $\mu\text{m}$ . The frequency response of the nanocantilever can be measured using a fixed amplitude function generator along with a lock-in amplifier signal detection technique. The signal from the function generator (the external driving force) on the nanocantilever can be applied with the help of an additional electrode situated near to it and keeping the nanocantilever grounded so that electrostatic actuation can be achieved.

When a nanocantilever with physical dimensions of length  $\sim 5 \mu\text{m}$ , width  $\sim 1.5 \mu\text{m}$ , and thickness  $\sim 30 \text{ nm}$  is used, the resonant frequency of the nanocantilever falls in the range 1–2 MHz. Also, the spring constant of the nanocantilever is 0.005–0.010 N/m. If the analyte is a single vaccinia virus particle, its mass is about 5–8 femtograms. The concentration of such virus particles in an aqueous solution is  $\sim 10^9$  PFU/ml, where PFU means plaque forming in the virus sample. The mass of DNA in a single cell is of the order of 6–7 picograms and the size (Plasmid DNA) is about 10–20 nanometers.

When analyte molecules are deposited on a nanocantilever, there is a momentum exchange between the oscillator and the particles colliding with it. This is called a process of random collisions.<sup>3</sup> These collisions bring a random dragging force that affects the nanocantilever motion. If the concentration of analyte molecules is low, then this dragging force can be attributed to a molecular drag proportional to the molecular velocity and contributing to the dissipative part of the equation of motion of

the oscillator. This dragging force can be written as  $F_d = b(dx/dt)$  and corresponds to the damping term of the oscillator's equation (e.g., Eq. (2) of section 3, in which  $b$  is the damping constant and  $dx/dt$  represents the speed of the particle). When the concentration of analyte molecules is low, we can say that  $F_d$  is proportional to the number of analyte particles  $N_a$  striking the nanocantilever per second. However, if the dragging force is random, it can be described by the term  $\xi(t)(dx/dt)$ , in which the parameter  $\xi(t)$  characterizes the randomness. The term brings random damping to the nanocantilever oscillator.<sup>3</sup> This analysis is applicable for a generic nanocantilever system. Another possible type of dragging is called inertial dragging, and it is proportional to the molecular acceleration. We will not further discuss this case here.

Other phenomenon taking place in this system is the adsorption-desorption of the analyte molecules on the nanocantilever surface.<sup>3,8</sup> Such thermally driven effect produces random frequency fluctuations in the oscillator.<sup>3,8,11</sup> The process of adsorption-desorption can be modeled by a molecular flux-dependent adsorption rate and a thermally activated rate of desorption.<sup>3,8</sup> Another interesting mechanism that produces frequency fluctuations is the one occurring with nanocantilevers having nanofluidic channels. In that case, the fluctuations are produced by the diffusion of adsorbed particles along the nanofluidic channel inside the vibrating nanocantilever. As the particle diffuses, the resonance frequency changes according to the relative amplitude of the vibrating mode at the location of the particle.<sup>7</sup> All of these fluctuations or random stochastic processes mentioned above can be incorporated as delta-correlated or exponentially correlated noise processes in the modeling equations.

### 3. Stochastically Perturbed Harmonic Oscillator

Let us consider the simple case of a nanocantilever described by a damped harmonic oscillator of mass  $m$  and spring constant  $k$ , driven by a sinusoidal external force. The differential equation describing the motion is

$$m\frac{d^2x}{dt^2} + b\frac{dx}{dt} + kx = F_0 \sin(\omega t), \quad (2)$$

where  $F_0$  and  $\omega$  are the amplitude and frequency of the external driving force, respectively, and  $b$  is a positive damping coefficient.

If a system is subjected to both random and periodic forces, the well-known phenomenon of stochastic resonance (SR) may emerge.<sup>12,13</sup> In the following, we will describe the effects of stochastic fluctuations in the damping and frequency terms for the equation governing the dynamics of the nanocantilever. These effects can be incorporated in (2) as *multiplicative* noise  $\xi(t)$ , such that the new equations of motion can be written as the stochastic differential equations (SDE)

$$\frac{d^2x}{dt^2} + 2\beta[1 + \xi(t)]\frac{dx}{dt} + \omega_0^2x = A \sin(\omega t) \quad (3)$$

for the random damping, and

$$\frac{d^2x}{dt^2} + 2\beta\frac{dx}{dt} + \omega_0^2[1 + \xi(t)]x = A \sin(\omega t) \quad (4)$$

for the random frequency. Here  $\beta \equiv b/2m$  is the damping parameter,  $\omega_0 = \sqrt{k/m}$  is the characteristic angular frequency in the absence of damping, and  $A = F_0/m$ .

As discussed above, the nanocantilever can be used to estimate the weight of individual nanoparticles, viruses, or protein molecules in an aqueous medium. However, there is a fast dissipation of the nanocantilever energy due to viscous dragging. If this aqueous medium is placed inside the cantilever using nanofluidic channels, then the viscous dragging could be almost eliminated.<sup>10</sup> As an example, the molar concentration of water is 55.5 mol/L but the molar concentration of proteins in an E. Coli bacteria is about 100 nmol/L. Protein size varies from 1 nm to 5 nm, and the mass from 5,000 amu to 500,000 amu. The typical molecular separation goes from 1.18 nm (1 M solution) to 1.18  $\mu\text{m}$  (1 nM solution).<sup>14</sup> In this situation, both water molecules and analyte particles  $N_a$  are randomly striking at the nanocantilever. At thermal equilibrium, the translational kinetic energy of water molecules is comparable to that of analytes (larger molecules). Since the molar concentration of water is larger, then the random fluctuations due to particles striking on the cantilever are dominated by water molecules, which eventually produce the damping.

The random variable  $\xi(t)$  will be considered as both Gaussian white noise with the correlator

$$\langle \xi(t)\xi(t') \rangle = D\delta(t - t'), \quad (5)$$

and colored noise with exponential correlator

$$\langle \xi(t)\xi(t') \rangle = \alpha^2 e^{-\lambda|t-t'|}. \quad (6)$$

The parameters  $D$  and  $\alpha$  represent the white and colored noise strengths, respectively, and  $\lambda$  the correlational decay rate. A special case of the colored noise is the symmetrical dichotomous noise (random telegraph signal) for which  $\xi(t)$  can take one of the values  $\xi = \pm\alpha$ , and the average waiting time for each of these states is  $\lambda^{-1}$ . Under the limiting conditions:  $\alpha^2 \rightarrow \infty$  and  $\lambda \rightarrow \infty$ , Eq. (6) reduces to (5), provided  $\alpha^2/\lambda = D$  (see Ref. 13).

### 3.1. *Nanocantilever with random damping*

In this subsection we study the effect of random fluctuations produced by particles striking on the cantilever in terms of the damping produced by them. We call this phenomenon random damping. In the absence of an external driving force ( $A = 0$ ) the equation of motion (3) takes the form

$$\hat{O}_D \{x\} = -2\beta \xi(t) \frac{dx}{dt}, \quad (7)$$

where the operator  $\hat{O}_D$  is defined by the following expression

$$\hat{O}_D \equiv \frac{d^2}{dt^2} + 2\beta \frac{d}{dt} + \omega_0^2. \quad (8)$$

In multiplicative stochastic processes involving terms like  $\xi(t)x(t)$ , the selection of  $t$  to find the appropriate  $x(t)$  depends on the characteristic noise correlation time for the variable  $\xi(t)$ . This gives rise to the Ito-Strantonovich dilemma. If we take the asymptotic limit for the correlation time, the solutions of the SDEs (3) and (4) are of the Strantonovich form. On the other hand, to get solutions of the Ito form, it is mandatory to have a correlation time of exactly zero. Also, it should be noted that for any external noise source in the system, the Strantonovich formulation works well.<sup>15</sup>

To get solutions of the Strantonovich form, Eq. (8) is recasted into an integro-differential equation using the technique described in Ref. 13. For this purpose, we apply the inverse operator  $O_D^{-1}$  to (7) to get

$$x = -\hat{O}_D^{-1} \left\{ 2\beta \xi(t) \frac{dx}{dt} \right\}, \quad (9)$$

The inverse operator  $\hat{O}_D^{-1}$  is an integral operator and can be determined using (8), so that

$$\hat{O}_D^{-1} \{g\} \equiv \frac{1}{\omega_0'} \int_0^t e^{-\beta(t-t')} \sin[\omega_0'(t-t')] g(t') dt', \quad (10)$$

with  $\omega'_0 = \sqrt{\omega^2 - \beta^2}$ . If we identify the function  $g$  as the one given inside the curly braces in (9), then it is not difficult to obtain

$$x(t) = -\frac{2\beta}{\omega'_0} \int_0^t \left\{ e^{-\beta(t-t')} \sin[\omega'_0(t-t')] \xi(t') \frac{dx(t')}{dt'} \right\} dt', \quad (11)$$

and

$$\frac{d}{dt}x(t) = \frac{2\beta}{\omega'_0} \int_0^t e^{-\beta(t-t')} \xi(t') \frac{dx(t')}{dt'} \left\{ \beta \sin[\omega'_0(t-t')] - \omega'_0 \cos[\omega'_0(t-t')] \right\} dt'. \quad (12)$$

Now, upon substituting (7) and (8) into (11) and (12), we may get

$$\left\{ \frac{d^2}{dt^2} + 2\beta \frac{d}{dt} + \omega_0^2 \right\} x(t) = -\frac{4\beta^2}{\omega'_0} \int_0^t e^{-\beta(t-t')} \xi(t) \xi(t') \frac{dx(t')}{dt'} \left\{ \beta \sin[\omega'_0(t-t')] - \omega'_0 \cos[\omega'_0(t-t')] \right\} dt'. \quad (13)$$

If the random damping is produced by the delta-correlated white noise (5), then (13) reduces to

$$\frac{d^2}{dt^2} \langle x \rangle + 2\beta(1 - 2\beta D) \frac{d}{dt} \langle x \rangle + \omega_0^2 \langle x \rangle = 0. \quad (14)$$

Here, we have use the fact that averages split in the form<sup>16</sup>

$$\left\langle \xi(t) \xi(t') \frac{dx(t')}{dt'} \right\rangle = \langle \xi(t) \xi(t') \rangle \left\langle \frac{dx(t')}{dt'} \right\rangle. \quad (15)$$

Equation (14) models the *collisional* damping produced by analytes (in an aqueous medium, which could be within the cantilever in a micro/nanofluidic channel) getting stuck on the nanocantilever. If  $2\beta D < 1$ , the presence of white noise will lead to a damping decrease (weak noise). On the other hand, if  $2\beta D > 1$ , the effective damping turns negative increasing the amplitude of oscillation  $\langle x \rangle$  and leading to instabilities in the dynamics of the nanocantilever (strong noise).<sup>13</sup>

When the nanocantilever is driven by a sinusoidal external force  $A \sin(\omega t)$  (see Eq. (3)), a solution to (14) can be written as

$$\langle x \rangle = B \sin(\omega t + \varphi). \quad (16)$$

Upon substituting (16) into (14), we can solve for the amplitude  $B$  and get the following expression

$$B = \frac{A}{[(\omega^2 - \omega_0^2)^2 + 4\beta^2 \omega^2 (1 - 2\beta D)^2]^{1/2}}. \quad (17)$$

The presence of noise in the system brings fluctuations in the nanocantilever frequency, resulting in spectral broadening. Hence it becomes necessary to determine the minimum measurable frequency shift that can be observed in this noisy environment. The spread in the frequency  $\delta\omega_0$  can be obtained by integrating the spectral density of fluctuations  $S(\omega)$

$$\delta\omega_0 \approx \left[ \int_{\omega_0 - \Delta\omega_0}^{\omega_0 + \Delta\omega_0} S(\omega) d\omega \right]^{1/2}. \quad (18)$$

Here, we assume that a measurement of the nanocantilever frequency was done with a square-shaped transfer function over the bandwidth  $2\Delta\omega_0$ , and centered at  $\omega_0$ . Equation (18) is an estimate for any real system. The spectral density  $S(\omega)$  is determined by the nature of the noise present in the system, and can be determined by taking the Fourier transform of the white noise correlator<sup>3</sup>

$$S(\omega) = \int_{-\infty}^{\infty} \langle \xi(t) \xi(0) \rangle e^{-i\omega t} dt. \quad (19)$$

At equilibrium and under resonance, the second moment of the nanocantilever displacement,  $\langle x^2 \rangle$ , satisfies the relation<sup>3</sup>

$$\frac{1}{2} m \omega_0^2 \langle x^2 \rangle = \frac{1}{2} \kappa T, \quad (20)$$

where  $\kappa$  is the Boltzmann constant and  $T$  is the absolute temperature at resonance. By using (19), we can find the spectral density corresponding to these displacement fluctuations

$$S_{\langle x^2 \rangle} = \frac{\text{const.} \times S(\omega)}{[(\omega^2 - \omega_0^2)^2 + 4\beta^2 \omega^2 (1 - 2\beta D)^2]^{1/2}}. \quad (21)$$

There are two damping mechanisms present in (21): The thermo-mechanical fluctuations governed by the damping parameter  $\beta$ , and the momentum exchanged with white noise, whose spectral density is proportional to  $\beta D$ . In particular, that momentum exchange is responsible for taking the system into resonant states. The plot in Fig. 1 is for  $S_{\langle x^2 \rangle}/S(\omega)$  vs.  $\omega$  and five curves are drawn for different values of  $\beta$ . The peculiarity in these resonance curves is that at resonance ( $\omega = \omega_0$ ), the peak height increases with increasing noise strength  $D$ , which is counterintuitive. The peaks are appearing at the resonance condition ( $\omega = \omega_0$ ) but their heights are dependent on the value  $D$ . Alternatively, if we keep  $(\omega - \omega_0)$  as a fixed quantity in (21) and plot the expression with respect to  $D$ , we get resonance peaks located at  $D = (2\beta)^{-1}$ . This behavior resembles the stochastic resonance phenomenon<sup>17</sup>

in which the height of a peak (intensity) goes up as noise strength increases up to a certain value and then goes down, a counterintuitive observation (see Fig. 2).




Fig. 2. Plot of  $S_{\langle x^2 \rangle} / S(\omega)$  as a function of the driving field frequency  $\omega$ , with parameter conditions:  $\omega_0 = 50$  and  $\beta = 1.4$  (see Eq. (21)). Curves A, B, C, D, and E are for the diffusion parameter  $D = 0.27, 0.25, 0.20, 0.15$  and  $0.10$ , respectively.

This counterintuitively phenomenon of stochastic resonance is present due to the fact that the “unwanted” noise coherently adds up to the external force signal, increasing the signal to noise ratio instead of reducing it.<sup>12</sup> From (1), it is clear that, for these nanocantilever based biosensors, the resonance peak is not sharp but has a finite width. The peak width is determined by the intrinsic damping of the nanocantilever,  $\beta$ , and the stochastic fluctuations caused by analyte molecules sticking on it. To measure the precise shift in the resonance frequency, the sharpest possible peak from the figure should be the appropriate choice. Therefore, knowing the values of  $\beta$  and  $D$ , and estimating the mass deposition with accuracy become critical in designing suitable nanocantilever based biosensors.

The response curve obtained from (21) and depicted in Fig. 2 can be reproduced experimentally by measuring the frequency response curve with and without the presence of fluctuations (introduced by analytes in an aqueous medium). A brief discussion about such an experiment is given in section 2. The damping parameter  $\beta$  of the nanocantilever can be measured by observing the decay of oscillations of the nanocantilever after it has been excited by the driving force and then disconnected. The value of  $D$  can also be figured out at some known temperature  $T$  using widths of measured frequency curves (like in Fig. 2) and the  $\beta$  value.

### 3.2. Nanocantilever with random fluctuations in frequency

In the absence of an external driving force ( $A = 0$ ) equation (4) takes the form

$$\frac{d^2x}{dt^2} + 2\beta\frac{dx}{dt} + \omega_0^2[1 + \xi(t)]x = 0. \quad (22)$$

Here, the multiplicative noise  $\xi(t)$  introduces random fluctuations in the nanocantilever frequency. It can be shown that, if the noise is white noise with correlator (5), the first moment of the oscillator is not affected by the noise and  $\langle x(t) \rangle = x(t)$  (see Ref. 13). On the other hand, for the exponentially correlated noise (6), the first moment  $\langle x \rangle$  equation can be obtained by defining

$$\hat{O}_E \{x\} = -\omega_0^2 \xi(t)x, \quad (23)$$

where the operator  $\hat{O}_E$  stands for

$$\hat{O}_E \equiv \frac{d^2}{dt^2} + \omega_0^2. \quad (24)$$

By applying the inverse operator  $O_E^{-1}$  to (22) we obtain

$$x = -\hat{O}_E^{-1} \{ \omega_0^2 \xi(t) dx \}, \quad (25)$$

Again,  $O_E^{-1}$  is an integral operator that can be determined using (24)

$$\hat{O}_E^{-1} \{g\} \equiv \frac{1}{\omega_0} \int_0^t \sin[\omega_0(t-t')]g(t')dt'. \quad (26)$$

Identifying function  $g$  as the one given inside the curly braces in (23), and using  $t' = t - \tau$  for convenience, it is easy to obtain

$$x(t) = -\omega_0 \int_0^t \sin[\omega_0(\tau)] \sin[\omega_0(t-\tau)] \xi(t-\tau) d\tau. \quad (27)$$

We use Eqs. (23) and (24) together with (27) to get

$$\left\{ \frac{d^2}{dt^2} + \omega_0^2 \right\} x(t) = \omega_0^2 \int_0^t \xi(t)\xi(t-\tau) \left\{ \omega_0 \frac{\sin(2\omega_0\tau)}{2} x(t) - \frac{1 - \cos(2\omega_0\tau)}{2} \frac{dx(t)}{dt} \right\} d\tau, \quad (28)$$

which can be rearranged as

$$\frac{d^2}{dt^2} \langle x \rangle - \omega_0^2 q_1 \frac{d}{dt} \langle x \rangle + \omega_0^2 (1 - \omega_0 q_2) \langle x \rangle = 0, \quad (29)$$

where the parameters  $q_1$  and  $q_2$  are given by<sup>11,15</sup>

$$q_1 = 2 \int_0^\infty \langle \xi(t)\xi(t-\tau) \rangle [1 - \cos(2\omega_0\tau)] d\tau, \quad (30)$$

$$q_2 = 2 \int_0^\infty \langle \xi(t)\xi(t-\tau) \rangle \sin(2\omega_0\tau) d\tau. \quad (31)$$

At this point, we add the term  $2\beta d\langle x \rangle/dt$  to (29) and assuming  $\beta \ll \omega_0^2$ , we may obtain

$$\frac{d^2}{dt^2}\langle x \rangle + (2\beta - \omega_0^2 q_1) \frac{d}{dt}\langle x \rangle + \omega_0^2 [1 - \omega_0 q_2] \langle x \rangle = 0, \quad (32)$$

The above equation can also be obtained employing the cumulative expansion of van Kampen (see Ref. 15).

In the limit of white noise both parameters  $q_1$  and  $q_2$  vanish. In other words, white noise does not change the width of the resonance profile of the frequency. So the rate at which the frequency of the nanocantilever changes due to mass deposition remains simple to calculate.

With random frequency, the mathematical expression used to find the spectral densities of noise fluctuations becomes a bit more complicated than that used in random damping (see Eq. (19)). Hence, the analytical solution (16) for the first moment in random damping can not be used to obtain a solution for random frequency. However, for dichotomous colored noise, the amplitude associated with the first moment shows a stochastic resonance-like feature in  $\alpha$  and  $\lambda$  similar to that observed by one of us in an experiment (see Ref. 17). Therefore, a similar analysis can be applied to get the resonance profile and find the change in frequency of the nanocantilever due to analyte deposition.

The adsorption-desorption of noise by the nanocantilever gives rise to fluctuations in its frequency. This is caused mostly by the constant bombardment of analyte molecules on its surface. This noise mechanism can be understood from the following perspective. The analyte molecules are adsorbed due to their affinity to the nanocantilever substrate and are desorbed because of a finite temperature change. This creates some *fractional* frequency noise. It is interesting to note that this process of adsorption-desorption of analyte molecules does not produce a damping mechanism per se. The randomness in the sticking and releasing particles on the cantilever does not contribute to the average change in the energy, but it changes the frequency of the nanocantilever in a nondeterministic manner.<sup>3,8</sup> Thus, this process introduces a different parametric noise that does not culminate into dissipation.

The nanoscale cantilevers are quite sensitive to the adsorption-desorption of noise when compared with conventional scale cantilevers. The reason is the difference between the surface to volume ratio for each type. This explains why the number of ad-

sorption locations in nanocantilevers is bigger than those on the counterpart.

The frequency fluctuations caused by noise can also be described using other methods different from that used in obtaining (32). One of these methods follows closely Refs. 3 and 8. Let the adsorption rate be  $R_a$ , which is dependent on the sticking coefficient of the nanocantilever surface, and  $R_d$  the temperature dependent desorption rate. The probability of molecular occupation in a particular area is given by  $p = R_a/(R_a + R_d)$ , and the corresponding variance in the occupation probability  $\sigma_v = \sqrt{R_a R_d}/(R_a + R_d)$ . The correlation time  $\tau_c$  of absorption-desorption noise is also given by  $\tau_c = (R_a + R_d)^{-1}$ . The spectral density of noise in this case can be written as

$$S_a(\omega) = \frac{2\pi\omega_0^2 N_a \sigma_v^2 \tau_c}{[1 + (\omega - \omega_0)^2 \tau_c^2]} \left( \frac{\Delta m}{m} \right)^2, \quad (33)$$

where  $\Delta m$  is the mass of the molecules attached on the cantilever surface. The noise variance is a maximum when the probability of occupation is 1/2, that is the adsorption rate equals the desorption rate. On the other hand, the noise is a minimum when the occupation probability is either 0 or 1. This noise will be superimposed on the frequency change of the biosensor and hence becoming critical in estimating the analyte mass deposition as we discuss in the following.

Integration over the spectral density  $S_a(\omega)$  provides some change in the nanocantilever frequency

$$\begin{aligned} \Delta\omega_0 &\approx \left[ \int_{\omega_0 - \pi\Delta f}^{\omega_0 + \pi\Delta f} S_a(\omega) d\omega \right]^2 \\ &= -\frac{\omega_0 \sigma_v \Delta m}{2\pi m} [N_a \tan^{-1}(2\pi\Delta f \tau_c)]^{1/2}, \end{aligned} \quad (34)$$

where  $\Delta f$  defines the width of passband. Hence, the change in mass on the nanocantilever is given by

$$\delta m \approx \Delta m \sigma_v [N_a \tan^{-1}(2\pi\Delta f \tau_c)]^{1/2}. \quad (35)$$

Note that this expression for the frequency shift (due to mass deposition) may be more realistic than that obtained in (1).

## 4. Summary

In this work, we presented a realistic model for a nanocantilever-based biosensor using a description of stochastically perturbed harmonic oscillator. These biosensors work by mass sensing the analytes through shifts in the characteristic resonance

frequency of the oscillator. When analytes are deposited on the nanocantilever, they bring a variety of noises into the vibrating system which give rise to fluctuations in the damping as well as the frequency of the cantilever. The estimation of such fluctuations on the spectral response of the cantilever is important to find out the exact amount of analyte deposition. This analysis could be used for clinical diagnostic purposes, for example, early detection of proteins present in cancerous cells. In summary, this work provides models for damping and frequency fluctuations in a nanocantilever according to different types of noise and their effect on the mass deposition.

### Acknowledgments

Funding supports from RCSA and URSCA are gratefully acknowledged.

### References

1. J. Kumar and S. F. D'Souza, *BARC Newsletter* **324**, 34 (2012).
2. D.R. Balset, G. U. Lee and R. J. Colton, *J. Vac. Sci. Technol. B* **14**, 789 (1996).
3. K. L. Ekinici, Y. T. Yang and M. L. Roukes, *J. Appl. Phys.* **95**, 2682 (2004).
4. L. M. Bellan, D. Wu and R. S. Langer, *NANOMED-NANOTECHNOL* **3**, 229 (2011).
5. J. Fritz, *Analyst* **133**, 855 (2008).
6. T. P. Burg, M. Godin, S. M. Knudsen, W. Shen, G. Carlson, J. S. Foster, K. Babcock and S. R. Manalis, *Nature* **446**, 1066 (2007).
7. J. Atalaya, A. Isacsson and M. I. Dykman, *Phys. Rev. B* **83**, 045419 (2011).
8. A. N. Cleland and M. L. Roukes, *J. Appl. Phys.* **92**, 2758 (2002).
9. A. Gupta, D. Akin and R. Bashir, *Appl. Phys. Lett.* **84**, 1976 (2004).
10. J. Lee, W. Shen, K. Payer, T. P. Burg and S. R. Manalis, *Nano Lett.* **10**, 2537 (2010).
11. M. I. Dykman, M. Khasin, J. Portman and S. W. Shaw, *Phys. Rev. Lett.* **105**, 230601 (2010).
12. L. Gammaitoni, P. Hänggi, P. Jung and F. Marchesoni, *Rev. Mod. Phys.* **70**, 223 (1998).
13. M. Gittermann, *Physica A* **352**, 309 (2005).
14. H. P. Erickson, *Biol. Proced. Online* **11**, 32 (2009).
15. N. G. van Kampen, *Stochastic Processes in Physics and Chemistry* (North Holland, Hungary, 2007).
16. R. C. Bourret, U. Frisch and A. Pouquet, *Physica* **65**, 303 (1973).
17. A. Joshi and M. Xiao, *Phys. Rev. A* **74**, 013817 (2006).