

Contactless Polysomnography: What Radio Waves Tell Us about Sleep

Hao He^{1*}, Chao Li^{1*}, Wolfgang Ganglberger^{2,3,5}, Kaileigh Gallagher⁵, Rumen Hristov⁶, Michail Ouroutzoglou¹, Haoqi Sun^{2,5}, Jimeng Sun⁴, Brandon Westover^{2,3,5}, Dina Katabi^{1,6}

¹Department of Electrical Engineering and Computer Science, Massachusetts Institute of Technology, Cambridge, MA, USA

²McCance Center for Brain Health, Massachusetts General Hospital, Boston, MA, USA

³Division of Sleep Medicine, Harvard Medical School, Boston, MA, USA

⁴Computer Science Department, University of Illinois Urbana-Champaign, Urbana, IL, USA

⁵Department of Neurology, Beth Israel Deaconess Medical Center, Boston, MA, USA

⁶Emerald Innovations Inc., Cambridge, MA 02142, USA

* These authors contributed equally to this work.

Corresponding author. Hao He, Department of Electrical Engineering and Computer Science, Massachusetts Institute of Technology, Cambridge, MA 02139, USA. Email haohe@mit.edu;
Corresponding author. Chao Li, Department of Electrical Engineering and Computer Science, Massachusetts Institute of Technology, Cambridge, MA 02139, USA. Email chaoli@mit.edu;
Corresponding author. Dina Katabi, Department of Electrical Engineering and Computer Science, Massachusetts Institute of Technology, Cambridge, MA 02139, USA. Email dk@mit.edu;

Abstract

The ability to assess sleep at home, capture sleep stages, and detect the occurrence of apnea (without on-body sensors) simply by analyzing the radio waves bouncing off people's bodies while they sleep is quite powerful. Such a capability would allow for longitudinal data collection in patients' homes, informing our understanding of sleep and its interaction with various diseases and their therapeutic responses, both in clinical trials and routine care. In this article, we develop an advanced machine learning algorithm for passively monitoring sleep and nocturnal breathing from radio waves reflected off people while asleep. Validation results in comparison with the gold standard (i.e., polysomnography) (n=849) demonstrate that the model captures the sleep hypnogram (with an accuracy of 81% for 30-second epochs categorized into Wake, Light Sleep, Deep Sleep, or REM), detects sleep apnea (AUROC = 0.88), and measures the patient's Apnea-Hypopnea Index (ICC=0.95; 95% CI = [0.93, 0.97]). Notably, the model exhibits equitable performance across race, sex, and age. Moreover, the model uncovers informative interactions between sleep stages and a range of diseases including neurological, psychiatric, cardiovascular, and immunological disorders. These findings not only hold promise for clinical practice and interventional trials but also underscore the significance of sleep as a fundamental component in understanding and managing various diseases.

Keywords: polysomnography, sleep hypnogram, apnea, machine learning, artificial intelligence, contactless at-home sleep monitoring

Statement of Significance

Our research presents an important advancement in sleep medicine, achieving accurate contactless sleep staging and detection of respiratory events (compared to PSG), through contactless radio wave measurements. Demonstrating robust and consistent performance across diverse age groups, genders, races, and health conditions in a large population of sleep clinic patients, our methodology facilitates daily, passive, at-home sleep assessment. This advancement holds promise for enhancing clinical care and clinical trials by providing longitudinal, at-home, cost-effective alternative to traditional in-laboratory sleep studies. Our findings have the potential to transform the long-term monitoring of sleep-related health issues, contributing to a more personalized, accessible, and data-driven approach to healthcare.

Introduction

Sleep monitoring is proven pivotal in understanding diseases and drug response. Sleep is implicated in almost all neurological diseases and inflammatory conditions [1, 2]. Sleep alterations in patients with depression include reduced REM latency and increased REM density [3, 4], and popular depression drugs, such as SSRI and SNRI, counter these effects increasing REM latency and reducing REM density [5]. Slow waves during deep sleep correlate with high activity of the glymphatic system and reduced slow waves have been associated with a greater β -amyloid burden, a hallmark of Alzheimer's disease [6]. Past work has also shown that disrupted sleep can aggravate inflammatory pathways present in conditions like rheumatoid arthritis and inflammatory bowel disease [7]. These examples emphasize the importance of sleep monitoring and assessment for both clinical trials and patient care.

The gold standard for sleep assessment is polysomnography (PSG) [8], which requires patients to sleep in the clinic with a plethora of electrodes and sensors on their bodies (Figure 1A). A key output of a PSG is the sleep hypnogram, which reports for each 30-second epoch the patient's sleep stage (Figure 1B). PSG studies are also used to assess irregular breathing patterns like apnea (cessation of breathing for more than 10 seconds) and hypopnea (shallow breathing for more than 10 seconds) (Figure 1B). However, PSG is costly and incurs a high patient burden [9, 10]. Furthermore, sleeping in an unfamiliar environment with many on-body sensors (as shown in Fig 1A) can bias the assessment [11, 12]. These challenges are exacerbated in longitudinal studies where changes in sleep and nocturnal breathing are tracked over months or years. As a result, both clinical trials and clinical care rarely use PSG but rather rely on subjective and insensitive self-reports of sleep quality (e.g., via Sleep NRS) by patients [13, 14].

Both clinical trials and patient care would benefit from an objective, sensitive, and low burden solution for in-home assessment of sleep hypnogram and nocturnal breathing. Indeed, the last decade has witnessed increasing use of wearable devices [15-18] and in-mattress sensors for measuring sleep at home in clinical trials [19-21]. However, wearable solutions typically require multiple on-body sensors to concurrently track sleep and respiratory signals [16-18], which imposes a relatively high burden in longitudinal trials, and can be cumbersome to wear during sleep. Mattress sensors can similarly affect a person's sleep and have relatively low accuracy [19-21].

Recently, an alternative completely passive approach was proposed (by our lab and others [22-26]): a radio device similar to a WiFi router is placed in the bedroom of the patient. It transmits very low power radio waves (1000 times lower power than home WiFi) and analyzes their reflections using neural networks to infer the sleep hypnogram and the respiration signal (Figure 1D). This approach is referred to as RF-based sleep monitoring (as it leverages Radio Frequency (RF) signals) and is as safe as having a WiFi router at home.




Figure 1. Illustration of traditional and RF-based sleep assessment. (A) PSG sensors during a sleep study. **(B)** Illustration of the output of sleep studies showing an example sleep hypnogram with 5 distinct sleep stages, and examples of respiratory signals with hypopnea events marked in cyan, and apnea events marked in yellow. **(D)** Illustration of our RF-based sleep and apnea monitoring system. Our sensor operates like a low power radar. It transmits wireless signals that are 1/1000 the power of a home Wi-Fi router and collects their reflections from nearby people. The breathing signal is extracted from such reflections, and the sleep hypnogram and respiratory events are then predicted from each night of the nocturnal breathing signal.

RF-based sleep monitoring promises to benefit both clinical trials and clinical care by allowing longitudinal sleep and apnea monitoring of patients in their own homes for months or even years, at zero overhead to patients. Furthermore, it concurrently measures sleep metrics and nocturnal breathing signals. Yet, achieving this potential depends on whether RF-based monitoring works well for patients across diverse demographics and different health statuses. Past work on RF-based sleep monitoring focuses on wellness as opposed to clinical use and hence is validated on a small number of healthy young users, with no consideration to demographics or pre-existing

diseases. Hence, the validation of RF-based sleep monitoring for clinical purposes remains an unmet need.

In this article, we explore the feasibility and accuracy of using RF-based monitoring of sleep and sleep apnea for clinical purposes. We focus on important clinical questions such as: does RF-based monitoring work well for the types of patients evaluated in sleep clinics? Is the performance equitable for different races, sexes, and age groups? Can we trust it for people who have comorbidities? Is the performance impacted by certain comorbidities more than others? And finally, is RF-based monitoring sensitive enough to capture the interplay between sleep and various diseases? Answering these questions is essential for understanding the clinical value of RF-based sleep monitoring.

To investigate the above questions, we developed a new advanced machine learning algorithm for monitoring sleep and nocturnal breathing from radio waves reflected off people while asleep. The model first extracts nocturnal breathing from radio signals; then breathing is analyzed by two neural networks- the first of which outputs the sleep hypnogram and the second outputs apnea/hypopnea events (Figure 1C). By using breathing signals as an intermediate representation, the model can be trained and tested, not only on data from radio waves but also on nocturnal breathing from repositories of public sleep datasets. Furthermore, the model is explicitly designed and trained to ensure equitable performance on different demographics. A detailed description of the model is provided in the Methods section.

This new AI model allows us to positively answer all of the aforementioned questions. Specifically, the model was evaluated on a large dataset comprising 14,068 individuals, using data extracted from radio waves (n=849) and breathing belts commonly used in sleep studies (n = 13,186). Our study reveals the following findings:

- Sleep hypnograms generated by the AI model solely from radio waves have an average accuracy of 81.6% and a standard deviation of 7.9%. Accurate sleep metrics are extracted from radio waves including sleep efficiency ($r=0.93$, $p<0.001$), total sleep time ($r=0.91$, $p<0.001$), sleep latency ($r=0.83$, $p<0.001$), and wake after sleep onset (WASO) ($r=0.89$, $p<0.001$).
- The same model can also generate a sleep hypnogram using data from a breathing belt with an average accuracy of 82.4% and a standard deviation of 7.9%. This accuracy persists on external datasets unseen during training.
- The model demonstrates equitable performance, where the performance gap stays small across races (1.37%), sexes (0.72%), and age groups (1.49%).
- The model can also detect sleep apnea from radio signals (AUC = 0.88) and assess the patient's Apnea-Hypopnea Index ($r=0.87$, $p<0.001$).
- The model continues to deliver high performance even in the presence of existing diseases, including neurological, psychiatric, cardiovascular, and immunological disorders. Notably, though the model relies on breathing signals as an intermediate representation from which it generates sleep stages, the model's performance is not sensitive to whether the person has pulmonary diseases (e.g., COPD, bronchitis, or emphysema).
- Interestingly, the model also reveals differences in sleep architecture in depression, hypertension, rheumatoid arthritis, and diabetes, underscoring the interplay between sleep and such conditions ($p < 0.05$).

Methods

Datasets

We use a large and diverse population created by pooling several datasets from different sources, including Massachusetts General Hospital (MGH) sleep lab, University of Massachusetts (UMASS) sleep lab, and public sleep datasets from the National Sleep Research Resource including the Sleep Heart Health Study (SHHS), Wisconsin Sleep Cohort (WSC), Multi-Ethnic Study of Atherosclerosis (MESA), and MrOS Sleep Study (MROS). The combined dataset contains 18,692 nights with over 149,726 hours of nocturnal breathing signals from 14,068 individuals aged 65.6 (standard deviation of 11.9), of whom 40.1% are females and 80.6% are white participants. Table 1 summarizes the datasets.

	Cohort/Dataset Type	Breathing Signal Source	No. of Participants	No. of Nights	Total Hours	Sex (%)	Age [years] mean (SD)	Race (%)			
						Males		Asian	Black	White	Others
MGH	MGH Sleep Lab Population	Wireless & Belts	849	849	6412	56.9	54.1 (16.7)	4.3	7.9	78.9	8.9
UMASS	UMASS Sleep Lab Population	Wireless	31	31	273	-	-	-	-	-	-
SHHS (visit 1 & 2)	Sleep Disorders and Cardiovascular Outcomes	Breathing Belts	5777	7515	64288	47.7	63.1 (11.2)	0.0	8.8	84.6	6.6
WSC	Longitudinal Study of Sleep Disorders	Breathing Belts	2514	2514	20038	54.0	59.7 (8.5)	0.9	1.4	95.8	1.9
MESA	Sleep Disorders and Cardiovascular Disease	Breathing Belts	2021	2021	21221	46.4	69.3 (9.1)	12.0	27.9	36.3	23.8
MROS (visit 1 & 2)	Sleep Disorders, Fractures and Vascular Disease	Breathing Belts	2875	3925	32279	100	76.4 (5.5)	3.0	3.4	91.0	2.6
Dashes, unavailable data.											

Table 1. Demographics of the datasets in the study.

The datasets can be categorized into two groups. The first group contains nocturnal breathing collected in a contactless manner using wireless radio signals. The radio sensor was deployed in the sleep lab; it analyzed the radio reflections from the environment to extract the person's breathing signal, which was further processed to infer the person's hypnogram and apnea events. This group of datasets allows us to confirm that our model can operate passively using only the radio signals reflected off a person's body during their sleep. The second group contains nocturnal breathing signals collected using (abdominal and thoracic) breathing belts during PSG sleep studies. This group of datasets allows us to confirm that the model works well independent of whether the breathing signals were captured using radio waves or a wearable device. It also allows us to further confirm the accuracy of the model for large datasets collected in different studies and with diverse populations. (Further description of the datasets and the corresponding IRBs is provided in the Supplementary Material.)

Deep Learning Models

As mentioned earlier, our models operate on breathing signals extracted from either radio reflections or wearable breathing belts. In this paper, we focus on the machine learning models and their clinical validations and leverage past work for extracting accurate breathing signals from radio reflections. As in past work [22, 27], the wireless signals were captured using a standard multi-antenna frequency-modulated continuous waves (FMCW) radio. We extract the person's breathing from the wireless signal using the method developed by Yue et al., which has been shown to achieve high accuracy [22]. (More information about the radio device can be found in the Supplementary Material.)

Our sleep staging model is built upon a deep neural network architecture. It takes as input an entire night's worth of breathing signals and generates a sleep hypnogram. This hypnogram represents a time series of sleep stages, with each stage assigned every 30 seconds. The model's architecture comprises 11 convolutional residual blocks, which are responsible for identifying local patterns within the breathing signals. Additionally, it incorporates a long short-term memory (LSTM) layer, facilitating the aggregation of information across the entire sequence of signals. Finally, the model employs three convolutional layers at the end to predict sleep stages based on the extracted latent features. (Refer to Supplementary S1 for model architecture.)

To improve the accuracy of our sleep staging model, we leverage the electroencephalogram (EEG) signals from sleep datasets using a technique called knowledge distillation [28, 29]. Knowledge distillation involves two models: student and teacher. In this work, our student model is the model that takes in breathing signals to predict sleep stages, and our teacher model is the model that takes in EEG signals to predict sleep stages. During training, the features learned by the teacher are distilled into the student model in two ways. First, we applied an L2 consistency loss between the features from the student and teacher models to encourage the student model to learn sleep staging features associated with EEG. Next, we applied a cross-entropy loss between the student and teacher model's outputs. Most importantly, the teacher is used only during training. During inference, only the student model is used and hence there is no need for EEG signals.

To improve the sleep staging model's performance across different subpopulations, we employ a state-of-the-art technique called classifier re-training [30]. This method involves adjusting the final layers of the neural network, which serve as a classifier, by retraining on datasets where the impact of the minority population's loss is adjusted based on their distribution. In our study, we observed a notable imbalance among racial groups, with a substantially smaller number of Black participants compared to White and other races. To address this, we retrain the last three layers of our model using a resampled training set where Black participants are given eight times more weight than non-Black participants.

In the Supplementary Material, we present ablation results that demonstrate the benefits of knowledge distillation (Figure S2(A)) and classifier retraining (Figure S2(B)) in improving the sleep model's accuracy and equitable performance.

Our apnea model is also a deep neural network that takes as input a full night of breathing signals and predicts the likelihood of an apnea/hypopnea happening at each moment. The model comprises 7 convolutional layers with residual connections. The output of the neural network is post-processed to eliminate apnea/hypopnea events shorter than 10 seconds, and merge events that are within 1 second because they are part of the same apnea/hypopnea event.

Training and evaluating protocols

We train a single model that can operate on breathing signals from radio waves or breathing belts. The model is trained using data from MGH, SHHS, WSC, and MESA, while the UMASS and MROS datasets are external datasets that are only used for testing. For the training datasets, we randomly split the patients in them into four folds. We trained the model on three folds and tested on the left-out testing fold. This procedure was repeated four times so that the four testing folds covered the whole dataset. All the performance metrics and statistical analysis are based on the pooled predictions across the four testing folds. (Additional information about training such as hyper-parameters is provided in the Supplementary Material.)

Statistical analysis and evaluation metrics

We evaluated our models using standard metrics. In particular, the accuracy of the sleep staging classifier is computed by comparing the prediction for every 30-second epoch with the ground truth stage, and reporting the average and standard deviation across subjects in each dataset. We used Cohen's Kappa [31] to assess agreement between predictions made using different breathing channels (thorax belt, abdominal belt, and wireless reflections). We also used the Pearson correlation and the corresponding p-value (derived from a one-sided Wald Test with a t-distribution of the test statistic) [32]. A p-value smaller than 0.05 is considered statistically significant.

To evaluate the effectiveness of the apnea model, we calculated the intraclass correlation coefficient (ICC) [33] between the predicted Apnea-Hypopnea Index (AHI) and the ground truth, along with its corresponding 95% confidence interval. We also assessed the model's ability to distinguish between apnea severity categories (Normal, Mild, Moderate, Severe) using the one-sided Mann-Whitney U rank test [34].

To investigate whether our predicted sleep metrics capture changes in sleep architecture associated with diseases and drug effects, we used a linear regression model [35] as follows:

$$\text{Sleep Metric} = \beta_0 + \beta_1 \cdot \text{Variable} + \beta_2 \cdot \text{Age} + \beta_3 \cdot \text{Sex},$$

where Sex is 0 for females and 1 for males, Variable is 1 if the participants have the disease or take the drug, and 0 otherwise. We examined the p-value and the magnitude of the coefficient of the Variable (β_1) to determine its effect on sleep metric. For all the linear models used, we verified that there is no strong collinearity between the covariates by checking that the Variance Inflation Factor (VIF) for each covariate is below 10.

The statistical analysis was conducted using Python version 3.8 (Python Software Foundation).

Results

Evaluation of Predicted Sleep Hypnogram and Sleep Metrics

In this section, we evaluated its accuracy in predicting the person's sleep hypnogram. We differentiate between the accuracy of extracting the sleep hypnogram directly from radio signals (where breathing is extracted as an intermediate stage) and extracting it from breathing signals collected using breathing belts. As shown in Figure 2A, the model predicts the sleep hypnogram from radio signals with an accuracy of $81.6\% \pm 7.9\%$ for the MGH dataset ($n=849$) and $84.1\% \pm 4.0\%$ for the UMass dataset ($n=33$), (for 30-second epochs categorized into Wake, Light Sleep, Deep Sleep, or REM). For comparison, the model predicts the sleep hypnogram from the abdominal breathing belt in the MGH dataset ($n=849$) with an accuracy of $82\% \pm 7.9\%$, (for 30-second epochs categorized into Wake, Light Sleep, Deep Sleep, or REM). Figure 2A further lists the accuracy of all datasets. The differences in accuracy can be attributed to differences in the quality of the sensors in older datasets (e.g. SHHS visit 1 collected from 1995 to 1998), which exhibit more noisy measurements and hence lower accuracy. Interestingly, the accuracy of the AI model is on par with the agreement between different sleep technicians who label the same sleep data (which is about 83% [36, 37]). Additionally, we tested our sleep staging model on datasets that were not used in training the model. On these datasets (MROS and UMass) our model similarly showed high performance, demonstrating its ability to generalize to external datasets.

The sleep hypnogram that the model generates from the radio signals is in high agreement with the hypnogram it generates from the breathing collected by the thorax or abdominal breathing belts worn by the participant on the same night. Specifically, Figure 2B shows Cohen's Kappa between predicted hypnograms (4 stages) from two different channels. For any pair of two channels, Cohen's Kappa is around 0.8 indicating a high level of agreement.

We further analyze the model's performance for different sleep stages. Figure 2C shows the confusion matrix for the AI model for classifying every 30-second epoch into one of the following sleep stages: Wake, Light Sleep, Deep Sleep, and REM. It shows that all sleep stages have a relatively high accuracy, with the most common misclassification involving identifying deep sleep as light sleep. This is not surprising since sleep technicians typically differentiate light sleep from deep sleep using subjective thresholds in the delta power band, and these thresholds can vary from one labeler to another [38]. Figure 2B further differentiates light sleep into two stages N1 and N2, the former is an unstable transition state, and the latter is the more stable light sleep. The figure shows that the model can differentiate these two modes for light sleep but faces challenges in predicting the N1 stage due to its brief and transitional nature. This difficulty in predicting N1 is acknowledged in previous studies [39-42], as even sleep technicians struggle with accurately labeling this stage, resulting in a low interrater agreement of 0.24 [43].




Figure 2. Performance of our sleep staging model. (A) Accuracy of staging model across breathing channels and datasets. (B) Agreement of model's prediction from different breathing channels, evaluated by Cohen's Kappa. Each entry shows the mean and 95% confidence interval. (C) Confusion matrices of predicted sleep stages from radio waves in the MGH dataset. Left is 4 class staging; Right is 5 class staging. (D) Example of a ground truth hypnogram from PSG data and the corresponding hypnogram output by our model based on radio signals. The 4 sleep cycles during the night are separated by the dashed lines. (E) Scatter plots of predicted sleep metrics from radio signals versus the ground truth from PSG data. Metrics are Total Sleep Time (TST), Sleep Efficiency (SE), Sleep Onset Latency (SOL), and Wake After Sleep Onset (WASO). Pearson correlations and p-values from the two-tailed Wald Test with t-distribution are reported.

Figure 2D shows an example of the predicted sleep hypnogram and the corresponding PSG-labeled sleep hypnogram. They highlight the model's ability to capture the overarching sleep architecture and distinct sleep cycles of the patient.

Finally, we consider changes in sleep quality assessed in terms of four sleep metrics: (1) total sleep time; (2) sleep efficiency (i.e., fraction of the sleep time in which the person is asleep); (3) sleep latency (i.e., time spent awake before sleep onset); (4) WASO (i.e., amount of wake time after sleep onset). Figure 2E shows that the AI model extracts accurate sleep metrics from radio waves including sleep efficiency ($r=0.93$, $p<0.001$), total sleep time ($r=0.91$, $p<0.001$), sleep latency ($r=0.83$, $p<0.001$), and WASO ($r=0.89$, $p<0.001$). The figure is for the MGH dataset. The results from other datasets are similar regardless of whether the input to the model is collected via radio signals or breathing belts (as shown in Figure S4 in the Supplementary Material).

Equitable Performance Across Demographics

Machine learning models often perform poorly on minority groups since they are under-represented in the training data [46]. To address this issue, our machine learning model uses "Classifier Retraining" [30], a state-of-the-art design that addresses data imbalance and improves minority performance. In this section, we show that the model performs well on minority groups and the performance (or accuracy) gap for different demographics is negligible.

The performance gap between the two groups is defined as the absolute difference between the model's average accuracy on the two groups. Specifically, for sex, we examined the performance gap between males and females. For race, we categorized individuals into four groups: Asian, Black, White, and Others, and calculated the performance gaps between each pair of groups. For age, we defined three categories: Young (under 40 years), Middle-aged (40 to 59 years), and Senior (60 years and over), and computed the average performance gap between them.

Table 3A reports the mean and standard deviation of the performance gap for different sexes, ages, and races, in the studied datasets. The table shows that the performance gap is negligible ($<1.5\%$), demonstrating that the model maintains good performance independent of the person's sex, race, and age. Among the three demographics, the performance across different sexes has the smallest gap (0.72%) which is attributed to the fact that the training datasets have balanced numbers of males and females. We also note that the gap between different age groups is the largest in datasets that span a wide range of ages (i.e., SHHS1, WSC), which is expected.

To further check that our model achieves equitable performance across demographics when using wireless signals, we zoom in on performance for the MGH wireless dataset. Table 3B lists the accuracy and Cohen's kappa for 4-class sleep staging in each demographic group. Notably, the disparity between the highest and lowest performing groups is marginal, at 2.9% for accuracy and 0.057 for Cohen's kappa. We further conduct hypothesis testing to show that such gaps are not statistically significant. We set the null hypothesis to be that the two groups have the same mean accuracy. The resulting p-value hence indicates the significance of the accuracy gap. Across the different sexes, statistical testing reveals negligible differences in mean accuracy ($p =$

0.173). For age, the result shows young and middle-aged groups have close accuracy ($p = 0.233$). However, the performance of senior participants is observed to be lower. This is explainable as the EEG signals (used in labeling sleep stages in a PSG study) among older participants are less reliable, leading to more uncertainty in sleep stage labels [45]. For race, the p values show that the gap is not significant ($p = 0.200$ for White versus Black and $p = 0.093$ for White versus Asians).

Performance gap of the sleep stage model for different demographics			
Dataset	Accuracy gap (%) (mean \pm std)		
	for Sex	for Age	for Race
MGH	1.08	1.21 \pm 0.43	1.29 \pm 0.58
SHHS1	0.49	2.14 \pm 0.76	1.63 \pm 0.59
SHHS2	0.17	0.66 \pm 0.00	0.68 \pm 0.30
WSC	1.42	2.16 \pm 0.87	1.10 \pm 0.63
MESA	0.15	0.37 \pm 0.00	1.45 \pm 0.97
MROS1	NA	NA	1.70 \pm 0.95
MROS2	NA	NA	1.28 \pm 0.67
Average	0.72	1.31	1.30

NA means "not applicable". All MROS participants are males with age ≥ 60 .
Sex groups: Male, Female
Age groups: Young (age < 40), Middle (40 \leq age < 60), Senior (age ≥ 60)
Race groups: Asian, Black, White, Others

Performance across demographics, on MGH dataset, using wireless signals as inputs			
Demographic	No. of samples	Accuracy (%)	Cohen's Kappa
All	849	81.6 \pm 7.9	0.697 \pm 0.139
<i>Gender</i>			
Male	480	81.3 \pm 7.8	0.690 \pm 0.140
Female	363	81.8 \pm 8.0	0.705 \pm 0.137
<i>Age</i>			
Young (<40)	202	82.3 \pm 8.6	0.717 \pm 0.142
Middle (40,60)	287	82.2 \pm 7.4	0.710 \pm 0.131
Senior (≥ 60)	360	80.6 \pm 7.7	0.676 \pm 0.140
<i>Race</i>			
Asian	35	79.4 \pm 8.5	0.660 \pm 0.152
Black	64	80.9 \pm 7.3	0.687 \pm 0.128
White	639	81.8 \pm 7.7	0.701 \pm 0.137
Others	111	81.0 \pm 8.8	0.694 \pm 0.147

Table 2. Performance across demographics. (A) Our sleep staging model's accuracy gap between subgroups that differ in age, sex, and race in various datasets. Mean and standard deviation are calculated for the pairwise performance gaps for the subgroups defined by each demographic. **(B)** Performance of the sleep staging model on the MGH wireless dataset for different demographics.

Evaluation of Predicted Respiratory Event

According to the American Academy of Sleep Medicine (AASM) guidelines, respiratory events are broadly categorized into two types: hypopnea and apnea [38]. Hypopneas involve shallow breathing, while apneas are characterized by a pause in breathing for at least 10 seconds. Both types of events require a minimum of 3% oxygen desaturation and exclusively occur during sleep. The severity of such respiratory events is measured via the Apnea-Hypopnea Index (AHI) defined as the average number of apnea or hypopnea events per hour. Based on the AASM guidelines, there are 4 general categories of sleep apnea severity: Normal: $AHI \leq 5$; Mild: $5 < AHI \leq 15$; Moderate: $15 < AHI \leq 30$; Severe: $AHI > 30$ [38].




Figure 3. Performance of the apnea model. (A) Performance across different channels and datasets. **(B)** Comparison of predicted AHI from wireless signals (in MGH wireless dataset) across the 4 severity categories of sleep apnea. **(C-D)** Examples of the PSG labeled respiratory events with the predicted respiratory events for two participants associated with different sleep apnea severity (“Normal” **(C)** and “Severe” **(D)**).

Here, we validate that our model can detect respiratory events from breathing signals. The results, in Figure 3A, show that the model accurately assesses AHI independently whether the breathing is collected using a breathing belt or the radio signals reflected off the person’s body. For example, for the MGH dataset, the intraclass correlation coefficient (ICC) for predicting the AHI is 0.951 (95% CI=[0.93,0.97]) from radio signals and 0.956 (95% CI=[0.94,0.97]) from a breathing belt ($p < 0.001$). A similar high accuracy is observed for the other datasets as listed in Figure 3. The performance difference between the AI model and the ground truth is similar to the performance difference between two sleep technicians who label the same data (in which case the ICC is also 0.95 [46]). We also note that the model achieves such high accuracy even on external datasets (i.e., MROS1 and MROS2) not included in its training.

We also evaluate the clinical utility of using radio signals to assess the severity of someone’s sleep apnea. Figure 3B shows that the model can assess the severity of a person’s sleep apnea using solely radio signals ($p < 0.001$). It is worth noting that the model occasionally misclassifies participants between the Normal and Mild categories more frequently than other categories, which is reflected in the overlap in the boxes between the two categories. This can be attributed to the

relatively narrower distinction in AHI values within these less severe categories. Furthermore, it is important to consider that the recommended treatment approaches for both Normal and Mild patients often overlap, as CPAP treatment is recommended only for patients with moderate to severe sleep apnea [47].

Finally, in Figure. 3C-D, we present examples of PSG-labeled respiratory events and the corresponding predictions made by our model for two patients with varying levels of sleep apnea severity (normal vs. severe). Our model successfully detects the occurrence of respiratory events in both instances, accurately capturing the presence of these events across varying degrees of sleep apnea severity.

Robustness to Pre-existing Conditions

A practical sleep staging model must perform well not just for healthy individuals, but also for people with pre-existing health conditions. We leverage that some of the datasets in our study have annotations of existing morbidities to assess our model's performance in four principal disease categories: cardiovascular, respiratory, immune, and neurological diseases. Note that the morbidities are not mutually exclusive, patients can have multiple diseases at the same time. Figure 4A shows that the model has good performance across all these categories for various datasets. Specifically, the average accuracy is 81%, 80%, 80%, and 79% for individuals with cardiovascular, respiratory, immune, and neurological diseases respectively. We make two additional observations related to Figure 4A. First, although our model relies on respiratory signals as input, its accuracy is not hampered by the presence of respiratory diseases. Second, the largest performance gap is associated with neurological conditions, which could be attributed to the interaction between neurological disorders and EEG, which is used to label sleep stages [40].

For additional insight, and to emphasize the ability of the model to work purely based on radio signals, we zoom in on different diseases in each of the above disease categories for the MGH wireless dataset. Figure 4B shows that the model demonstrates consistently high performance across all fine-grained health conditions in the disease categories.

Changes in Sleep Architecture with Diseases and Drugs

Sleep architecture is affected by diseases and drugs. In particular, REM sleep is affected by cardiovascular [48, 49], autoimmune [50, 51], and mental health [52]. In this section, we validate that the sleep metrics derived from our model can faithfully reveal changes in sleep architecture associated with various conditions. We focus on REM architecture and consider four common diseases: Depression, Hypertension, Rheumatoid Arthritis, and Diabetes. For each disease, we study multiple datasets to confirm the association between the observed change in REM and the underlying condition. We also leverage that some of our datasets contain the subject's medication record to investigate whether the impact on REM sleep is due to the disease itself or the pharmaceutical intervention.




Figure 4. Analysis of the sleep staging model in diseased population. (A) Accuracy of the model across different disease categories (Respiratory, Cardiovascular, Immune, and Neurological diseases) in different datasets. ‘All’ refers to all participants in the dataset. Note that the disease categories are not mutually exclusive. **(B)** Accuracy of sleep stage prediction across fine-grained conditions in the MGH dataset using wireless signals as input. **(C)** Disease and drugs’ effects on predicted sleep metrics. We consider Depression, Hypertension, Rheumatoid Arthritis (RA), and Diabetes. For each disease, we analyze the sleep metrics that have differences between the control and the disease groups. For datasets with medication information, we further split the disease group into patients on medications (Meds) and not on medications (No Meds) to study the effect of drugs. The statistical significances are indicated by asterisks. (*: $p < 0.05$, **: $p < 0.01$, ***: $p < 0.001$). Note that, compared to other datasets, patients in the MGH dataset have higher REM latency and less REM duration, which is due to patients visiting the MGH sleep lab are more likely to have sleep problems.

We report the results in Figure 4C. The REM metrics in Figure 4C are based on the output of our model. Figure S5 in the Supplementary Material provides similar graphs for the case where the REM metrics are computed from the ground truth labels, demonstrating the same findings. The statistical significance is computed using multiple linear regression to adjust for age and sex.

Figure 4C shows that individuals suffering from depression tend to experience higher REM latency than those who do not suffer from such a condition. However, the reason for increased REM in these cases is likely due to medications (i.e., antidepressants) as opposed to the condition itself. Similarly, higher REM latency is observed in individuals with hypertension, and the empirical evidence reveals that this latency is due to antihypertension drugs. These results are compatible with the literature which reports that antidepressants that increase serotonin function have a significant effect on REM sleep [53], and anti-hypertension drugs such as β -adrenoceptor and α 2-adrenoceptor agonists reduce the duration of REM sleep [54].

The figure also shows that Rheumatoid Arthritis (RA) is associated with increased REM latency. Yet, unlike hypertension and depression, REM delay seems to stem from the disease as opposed to RA medications. Furthermore, the figure shows that diabetes is associated with less REM sleep. Here too, the data indicates that reduction in REM sleep is due to the underlying disease as opposed to its medications. Our results on the impact of RA and diabetes on REM sleep are consistent with medical literature [37-39].

Discussion

We have demonstrated the ability to extract sleep stages and respiratory events from nocturnal breathing, measured using radio waves or a breathing belt. Our findings have been rigorously validated against the PSG gold standard (Figure 2), revealing a highly accurate sleep staging model that is equitable across various patient demographics (Figure 3) and robust against different diseases and health conditions (Figure 4). Given that our system enables the noninvasive, contactless, and passive measurement of breathing signals through radio signals, it offers a convenient means to continuously monitor sleep stages and respiratory events from the comfort of one's own home, every night.

This capability (i.e., passive and daily assessment of sleep and nocturnal respiration at home) brings several benefits to both clinical care and clinical trials. There is a strong interest in conducting home-based sleep studies for numerous reasons. Traditional sleep lab tests often burden patients with an excessive number of sensors and force them to sleep in an unfamiliar environment, both of which can disrupt sleep and bias the exact phenomenon that the test measures. Moreover, in-lab tests are costly, averaging around \$3,000 per night at the time of writing. Additionally, accessibility remains an issue, particularly in rural areas and on a global scale, where hundreds of millions suffer from sleep disorders, yet PSG labs and qualified sleep technicians are scarce in many regions [55]. Perhaps most importantly, home sleep tests alleviate the burden on patients.

All these factors have spurred interest in at-home sleep tests. However, existing at-home testing kits [56, 57] are typically limited to assessing apnea and respiratory events, as they do not generate a hypnogram [56]. Furthermore, they can be cumbersome for patients, as they require the use of respiratory sensors during sleep. As a result, at-home sleep testing falls short when compared to in-lab testing.

Our results have the potential to significantly expand the scope of at-home sleep testing. On one hand, the ability to accurately extract a sleep hypnogram from respiratory signals means that typical home sleep testing devices can offer more than just apnea testing by including sleep staging. On the other hand, the fact that an equally accurate assessment can be obtained passively through radio signals without the need for wearable sensors allows for continuous monitoring of sleep and respiratory events without any inconvenience for months or even years. This greatly facilitates the conduct of longitudinal at-home sleep studies in clinical trials and the management of mood disorders and other neurological conditions where sleep problems are major contributing factors to the disease burden.

Additionally, our results have the potential to enhance access and equity in healthcare. As mentioned earlier, many communities lack access to nearby PSG labs, and some countries lack sleep labs altogether. The ability to remotely and automatically assess sleep and respiratory events empowers these communities, ensuring more equitable access to healthcare services. Access also improves for children and individuals with dementia or cognitive impairment, for whom wearing sensors can be challenging even when conducting sleep studies at home.

Furthermore, our study underscores the relationship between breathing, sleep, and EEG. We used breathing as an intermediate signal to extract a comprehensive sleep hypnogram with corresponding sleep stages. Typically, sleep staging relies on analyzing the EEG signal. Previous research has reported an association between specific respiratory patterns and sleep stages [58, 59], and the feasibility of sleep staging based on breathing [27, 60]. Our results reinforce and expand upon this body of literature by demonstrating that the relationship between breathing and sleep staging remains consistent across demographics and various health conditions, including those that directly affect respiration, such as COPD and bronchitis.

However, it is essential to acknowledge the limitations of our study. First, a PSG test provides additional data beyond the hypnogram and respiratory events (e.g., ECG, EEG, EMG). While these additional signals are helpful in certain diagnoses, sleep tests typically focus on the hypnogram and respiratory events. Second, identifying the N1 sleep stage poses a challenge for the machine learning model. N1 is characterized by its brief and transitional nature, making accurate labeling difficult even for experienced sleep technicians [30]. Some literature suggests that the labels associated with the N1 stage are inherently unreliable and offer limited clinical utility in practice [61]. Third, our evaluation did not include predicting distinct categories of respiratory events, such as hypopnea and apnea. Inter-rater agreement for these categories is notably weak compared to overall respiratory events [56], primarily due to the subjective nature of distinguishing hypopnea and apnea, introducing greater variability. Finally, our validation is in comparison to PSG lab tests, where participants naturally sleep alone. While this paper's exact evaluation does not address cases where people sleep with a bed partner, prior work has demonstrated that the passive collection of nocturnal breathing signals from radio waves remains

accurate even in the presence of bed partners [22]. As our models solely rely on breathing signals as input, their accuracy depends solely on the quality of breathing data.

Our research aligns with the growing interest in harnessing artificial intelligence to gain a deeper understanding of physiological signals and provide clinicians with valuable clinical insights beyond traditional lab tests. This approach holds enormous potential for offering clinicians comprehensive, longitudinal, ambulatory sleep assessments, paving the way for personalized and data-driven healthcare.

Data Availability

The SHHS, WSC, MROS and MESA datasets are publicly available from the National Sleep Research Resource (<https://sleepdata.org/datasets>). Additional data that support the findings of this study are available from the corresponding author, upon reasonable request.

References

1. Irwin M. R. (2015). Why sleep is important for health: a psychoneuroimmunology perspective. *Annual review of psychology*, 66, 143–172. <https://doi.org/10.1146/annurev-psych-010213-115205>
2. Bishir, M., Bhat, A., Essa, M. M., Ekpo, O., Ihunwo, A. O., Veeraraghavan, V. P., Mohan, S. K., Mahalakshmi, A. M., Ray, B., Tuladhar, S., Chang, S., Chidambaram, S. B., Sakharkar, M. K., Guillemin, G. J., Qoronfleh, M. W., & Ojcius, D. M. (2020). Sleep Deprivation and Neurological Disorders. *BioMed research international*, 2020, 5764017. <https://doi.org/10.1155/2020/5764017>
3. Riemann, D., Berger, M., & Voderholzer, U. (2001). Sleep and depression--results from psychobiological studies: an overview. *Biological psychology*, 57(1-3), 67–103. [https://doi.org/10.1016/s0301-0511\(01\)00090-4](https://doi.org/10.1016/s0301-0511(01)00090-4)
4. Palagini, L., Baglioni, C., Ciapparelli, A., Gemignani, A., & Riemann, D. (2013). REM sleep dysregulation in depression: state of the art. *Sleep medicine reviews*, 17(5), 377–390. <https://doi.org/10.1016/j.smrv.2012.11.001>
5. Wichniak, A., Wierzbicka, A., Wałęcka, M., & Jernajczyk, W. (2017). Effects of Antidepressants on Sleep. *Current psychiatry reports*, 19(9), 63. <https://doi.org/10.1007/s11920-017-0816-4>
6. Mander, B. A., Marks, S. M., Vogel, J. W., Rao, V., Lu, B., Saletin, J. M., Ancoli-Israel, S., Jagust, W. J., & Walker, M. P. (2015). β -amyloid disrupts human NREM slow waves and related hippocampus-dependent memory consolidation. *Nature Neuroscience*, 18(7), 1051–1057. <https://doi.org/10.1038/nn.4035>
7. Kinnucan, J. A., Rubin, D. T., & Ali, T. (2013). Sleep and inflammatory bowel disease: exploring the relationship between sleep disturbances and inflammation. *Gastroenterology & hepatology*, 9(11), 718–727.

8. Marino, M., Li, Y., Rueschman, M. N., Winkelman, J. W., Ellenbogen, J. M., Solet, J. M., Dulin, H., Berkman, L. F., & Buxton, O. M. (2013). Measuring sleep: accuracy, sensitivity, and specificity of wrist actigraphy compared to polysomnography. *Sleep*, 36(11), 1747–1755. <https://doi.org/10.5665/sleep.3142>
9. Grandner, M. A., Williams, N. J., Knutson, K. L., Roberts, D., & Jean-Louis, G. (2016). Sleep disparity, race/ethnicity, and socioeconomic position. *Sleep medicine*, 18, 7–18. <https://doi.org/10.1016/j.sleep.2015.01.020>
10. Kaditis, A., Kheirandish-Gozal, L., & Gozal, D. (2016). Pediatric OSAS: Oximetry can provide answers when polysomnography is not available. *Sleep medicine reviews*, 27, 96–105. <https://doi.org/10.1016/j.smrv.2015.05.008>
11. Mosquera-Lopez, C., Leitschuh, J., Condon, J., Hagen, C. C., Rajhbeharrysingh, U., Hanks, C., & Jacobs, P. G. (2019). Design and Evaluation of a Non-ContactBed-Mounted Sensing Device for Automated In-Home Detection of Obstructive Sleep Apnea: A Pilot Study. *Biosensors*, 9(3), 90. <https://doi.org/10.3390/bios9030090>
12. Harrison, E. I., Roth, R. H., Lobo, J. M., Kang, H., Logan, J., Patel, S. R., Kapur, V. K., & Kwon, Y. (2021). Sleep time and efficiency in patients undergoing laboratory-based polysomnography. *Journal of Clinical Sleep Medicine: JCSM : official publication of the American Academy of Sleep Medicine*, 17(8), 1591–1598. <https://doi.org/10.5664/jcsm.9252>
13. Narla, S., & Silverberg, J. I. (2024). Which Clinical Measurement Tools for Atopic Dermatitis Severity Make the Most Sense in Clinical Practice?. *Dermatitis : contact, atopic, occupational, drug*, 35(S1), S13–S23. <https://doi.org/10.1089/derm.2022.0087>
14. Frase, L., Nissen, C., Spiegelhalter, K., & Feige, B. (2023). The importance and limitations of polysomnography in insomnia disorder—a critical appraisal. *Journal of sleep research*, 32(6), e14036. <https://doi.org/10.1111/jsr.14036>
15. de Zambotti, M., Cellini, N., Goldstone, A., Colrain, I. M., & Baker, F. C. (2019). Wearable Sleep Technology in Clinical and Research Settings. *Medicine and science in sports and exercise*, 51(7), 1538–1557. <https://doi.org/10.1249/MSS.0000000000001947>
16. Zavanelli N, Kim H, Kim J, et al. At-home wireless monitoring of acute hemodynamic disturbances to detect sleep apnea and sleep stages via a soft sternal patch. *Sci Adv*. 2021;7(52):eabl4146. doi:10.1126/sciadv.abl4146
17. Walch O, Huang Y, Forger D, Goldstein C. Sleep stage prediction with raw acceleration and photoplethysmography heart rate data derived from a consumer wearable device. *Sleep*. 2019;42(12):zsz180. doi:10.1093/sleep/zsz180
18. Boe AJ, McGee Koch LL, O'Brien MK, et al. Automating sleep stage classification using wireless, wearable sensors. *NPJ Digit Med*. 2019;2:131. Published 2019 Dec 20. doi:10.1038/s41746-019-0210-1
19. Mitsukura Y, Sumali B, Nagura M, Fukunaga K, Yasui M. Sleep Stage Estimation from Bed Leg Ballistocardiogram Sensors. *Sensors (Basel)*. 2020;20(19):5688. Published 2020 Oct 5. doi:10.3390/s20195688

20. Ding F, Cotton-Clay A, Fava L, et al. Polysomnographic validation of an under-mattress monitoring device in estimating sleep architecture and obstructive sleep apnea in adults. *Sleep Med.* 2022;96:20-27. doi:10.1016/j.sleep.2022.04.010
21. Siyahjani F, Garcia Molina G, Barr S, Mushtaq F. Performance Evaluation of a Smart Bed Technology against Polysomnography. *Sensors (Basel)*. 2022;22(7):2605. Published 2022 Mar 29. doi:10.3390/s22072605
22. Yue, S., He, H., Wang, H., Rahul, H., & Katabi, D. (2018). Extracting multi-person respiration from entangled RF signals. *Proceedings of the ACM on Interactive, Mobile, Wearable and Ubiquitous Technologies*, 2(2), 1-22.
23. Adib, F.M., Mao, H., Kabelac, Z., Katabi, D., & Miller, R. (2015). Smart Homes that Monitor Breathing and Heart Rate. *Proceedings of the 33rd Annual ACM Conference on Human Factors in Computing Systems*.
24. Yu, B., Wang, Y., Niu, K., Zeng, Y., Gu, T., Wang, L., Guan, C., & Zhang, D. (2021). WiFi-Sleep: sleep stage monitoring using commodity Wi-Fi devices. *IEEE Internet of Things Journal*, 8(18), 13900-13913. <https://doi.org/10.1109/JIOT.2021.3068798>
25. Lee JH, Nam H, Kim DH, et al. Developing a deep learning model for sleep stage prediction in obstructive sleep apnea cohort using 60 GHz frequency-modulated continuous-wave radar. *J Sleep Res.* 2024;33(1):e14050. doi:10.1111/jsr.14050
26. Yoo Y-K, Jung C-W, Shin H-C. Unsupervised Detection of Multiple Sleep Stages Using a Single FMCW Radar. *Applied Sciences*. 2023; 13(7):4468. <https://doi.org/10.3390/app13074468>
27. Zhao, M., Yue, S., Katabi, D., Jaakkola, T.S. & Bianchi, M.T.. (2017). Learning Sleep Stages from Radio Signals: A Conditional Adversarial Architecture. *Proceedings of the 34th International Conference on Machine Learning*, in *Proceedings of Machine Learning Research* 70:4100-4109 Available from <https://proceedings.mlr.press/v70/zhao17d.html>.
28. Xue, Z., Gao, Z., Ren, S., & Zhao, H. (2022). The modality focusing hypothesis: Towards understanding crossmodal knowledge distillation. *arXiv preprint arXiv:2206.06487*.
29. Hinton, G., Vinyals, O., & Dean, J. (2015). Distilling the knowledge in a neural network. *arXiv preprint arXiv:1503.02531*.
30. Kang, B., Xie, S., Rohrbach, M., Yan, Z., Gordo, A., Feng, J., & Kalantidis, Y. (2019, September). Decoupling Representation and Classifier for Long-Tailed Recognition. In *International Conference on Learning Representations*.
31. Cohen, J. (1960). A coefficient of agreement for nominal scales. *Educational and psychological measurement*, 20(1), 37-46.
32. Freedman, D., Pisani, R., Purves, R. (2007). *Statistics (international student edition)*. Pisani, R. Purves, 4th edn. WW Norton & Company, New York
33. Koo, T. K., & Li, M. Y. (2016). A guideline of selecting and reporting intraclass correlation coefficients for reliability research. *Journal of chiropractic medicine*, 15(2), 155-163.

34. McKnight, P.E. and Najab, J. (2010). Mann-Whitney U Test. In *The Corsini Encyclopedia of Psychology* (eds I.B. Weiner and W.E. Craighead). <https://doi.org/10.1002/9780470479216.corpsy0524>
35. Jobson, J.D. (1991) Multiple Linear Regression. In: *Applied Multivariate Data Analysis*, Springer, New York, 219-398. <https://doi.org/10.1007/978-1-4612-0955-3>
36. Liu, G. R., Lin, T. Y., Wu, H. T., Sheu, Y. C., Liu, C. L., Liu, W. T., ... & Lo, Y. L. (2021). Large-scale assessment of consistency in sleep stage scoring rules among multiple sleep centers using an interpretable machine learning algorithm. *Journal of Clinical Sleep Medicine*, 17(2), 159-166.
37. Ruehland, W. R., Rochford, P. D., O'Donoghue, F. J., Pierce, R. J., Singh, P., & Thornton, A. T. (2009). The new AASM criteria for scoring hypopneas: impact on the apnea hypopnea index. *sleep*, 32(2), 150-157.
38. Berry, R. B., Brooks, R., Gamaldo, C., Harding, S. M., Lloyd, R. M., Quan, S. F., ... & Vaughn, B. V. (2017). AASM scoring manual updates for 2017 (version 2.4). *Journal of Clinical Sleep Medicine*, 13(5), 665-666.
39. Sibilano, E., Brunetti, A., Buongiorno, D., Lassi, M., Grippo, A., Bessi, V., ... & Bevilacqua, V. (2023). An attention-based deep learning approach for the classification of subjective cognitive decline and mild cognitive impairment using resting-state EEG. *Journal of Neural Engineering*, 20(1), 016048.
40. Supratak, A., Dong, H., Wu, C., & Guo, Y. (2017). DeepSleepNet: A model for automatic sleep stage scoring based on raw single-channel EEG. *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, 25(11), 1998-2008.
41. Sors, A., Bonnet, S., Mirek, S., Vercueil, L., & Payen, J. F. (2018). A convolutional neural network for sleep stage scoring from raw single-channel EEG. *Biomedical Signal Processing and Control*, 42, 107-114.
42. Horie, K., Ota, L., Miyamoto, R., Abe, T., Suzuki, Y., Kawana, F., ... & Kitagawa, H. (2022). Automated sleep stage scoring employing a reasoning mechanism and evaluation of its explainability. *Scientific reports*, 12(1), 12799.
43. Lee, Y. J., Lee, J. Y., Cho, J. H., & Choi, J. H. (2022). Interrater reliability of sleep stage scoring: a meta-analysis. *Journal of Clinical Sleep Medicine*, 18(1), 193-202.
44. Wang, Y. X., Ramanan, D., & Hebert, M. (2017). Learning to model the tail. *Advances in neural information processing systems*, 30.
45. Danker-Hopfe, H., Kunz, D., Gruber, G., Klösch, G., Lorenzo, J. L., Himanen, S. L., ... & Dorffner, G. (2004). Interrater reliability between scorers from eight European sleep laboratories in subjects with different sleep disorders. *Journal of Sleep Research*, 13(1), 63-69.
46. Magalang, U. J., Chen, N. H., Cistulli, P. A., Fedson, A. C., Gislason, T., Hillman, D., ... & SAGIC Investigators. (2013). Agreement in the scoring of respiratory events and sleep among international sleep centers. *Sleep*, 36(4), 591-596.

47. Mandereau-Bruno, L., Leger, D., & Delmas, M. C. (2021). Obstructive sleep apnea: a sharp increase in the prevalence of patients treated with nasal CPAP over the last decade in France. *PLoS one*, 16(1), e0245392.
48. Patel AK, Reddy V, Shumway KR, et al. Physiology, Sleep Stages. [Updated 2024 Jan 26]. In: StatPearls [Internet]. Treasure Island (FL): StatPearls Publishing; 2024 Jan-. Available from: <https://www.ncbi.nlm.nih.gov/books/NBK526132/>
49. Somers VK, Dyken ME, Clary MP, Abboud FM. Sympathetic neural mechanisms in obstructive sleep apnea. *J Clin Invest*. 1995;96(4):1897-1904. doi:10.1172/JCI118235
50. Zhang, Y., Pi, B., Xu, X., Li, Y., Chen, X., Yang, N., & Li, X. (2020). Sleep Characteristics and Influencing Factors of Sleep Quality in Patients With Inflammatory Bowel Disease-Peripheral Arthritis. *Frontiers in medicine*, 6, 190. <https://doi.org/10.3389/fmed.2019.00190>
51. Ju YE, Larson-Prior L, Duntley S. Changing demographics in REM sleep behavior disorder: possible effect of autoimmunity and antidepressants. *Sleep Med*. 2011;12(3):278-283. doi:10.1016/j.sleep.2010.07.022
52. Anderson KN, Bradley AJ. Sleep disturbance in mental health problems and neurodegenerative disease. *Nat Sci Sleep*. 2013;5:61-75. Published 2013 May 31. doi:10.2147/NSS.S34842
53. Wilson S, Argyropoulos S. Antidepressants and sleep: a qualitative review of the literature. *Drugs*. 2005;65(7):927-947. doi:10.2165/00003495-200565070-00003
54. Monti, J. M. (1987). Disturbances of sleep and wakefulness associated with the use of antihypertensive agents. *Life sciences*, 41(17), 1979-1988.
55. Singh R, Juarez PD, Redline S, Jackson CL. Shortage of sleep medicine specialists in federally qualified health centers: an illustrative example of differential access to care. *J Clin Sleep Med*. 2023;19(10):1849-1850. doi:10.5664/jcsm.10688
56. What to know about an at-home sleep test. Johns Hopkins Medicine. (2022, February 8). <https://www.hopkinsmedicine.org/health/wellness-and-prevention/what-to-know-about-an-at-home-sleep-test>
57. At-home sleep apnea tests. Sleep Foundation. (2024, January 2). <https://www.sleepfoundation.org/sleep-studies/at-home-sleep-apnea-test>
58. Douglas NJ, White DP, Pickett CK, Weil JV, Zwillich CW. Respiration during sleep in normal man. *Thorax*. 1982;37(11):840-844. doi:10.1136/thx.37.11.840
59. BULOW K. Respiration and wakefulness in man. *Acta Physiol Scand Suppl*. 1963;209:1-110.
60. Sun H, Ganglberger W, Panneerselvam E, et al. Sleep staging from electrocardiography and respiration with deep learning. *Sleep*. 2020;43(7):zsz306. doi:10.1093/sleep/zsz306
61. Younes, M., Kuna, S. T., Pack, A. I., Walsh, J. K., Kushida, C. A., Staley, B., & Pien, G. W. (2018). Reliability of the American Academy of Sleep Medicine rules for assessing sleep depth in clinical practice. *Journal of Clinical Sleep Medicine*, 14(2), 205-213.

List of Figures

Figure 1. Illustration of traditional and RF-based sleep assessment. (A) PSG sensors during a sleep study. (B) Illustration of the output of sleep studies showing an example sleep hypnogram with 5 distinct sleep stages, and examples of respiratory signals with hypopnea events marked in cyan, and apnea events marked in yellow. (D) Illustration of our RF-based sleep and apnea monitoring system. Our sensor operates like a low power radar. It transmits wireless signals that are 1/1000 the power of a home Wi-Fi router and collects their reflections from nearby people. The breathing signal is extracted from such reflections, and the sleep hypnogram and respiratory events are then predicted from each night of the nocturnal breathing signal.

Figure 2. Performance of our sleep staging model. (A) Accuracy of sleep staging model across breathing channels and datasets. (B) Agreement of model's prediction from different breathing channels, evaluated by Cohen's Kappa. Each entry shows the mean and 95% confidence interval. (C) Confusion matrices of predicted sleep stages from radio waves in the MGH dataset. Left is 4 class staging; Right is 5 class staging. (D) Example of a ground truth hypnogram from PSG data and the corresponding hypnogram output by our model based on radio signals. The 4 sleep cycles during the night are separated by the dashed lines. (E) Scatter plots of predicted sleep metrics from radio signals versus the ground truth from PSG data. Metrics are Total Sleep Time (TST), Sleep Efficiency (SE), Sleep Onset Latency (SOL), and Wake After Sleep Onset (WASO). Pearson correlations and p-values from the two-tailed Wald Test with t-distribution are reported.

Figure 3. Performance of the apnea model. (A) Performance across different channels and datasets. (B) Comparison of predicted AHI from wireless signals (in MGH wireless dataset) across the 4 severity categories of sleep apnea. (C-D) Examples of the PSG labeled respiratory events with the predicted respiratory events for two participants associated with different sleep apnea severity ("Normal" (C) and "Severe" (D))

Figure 4. Analysis of the sleep staging model in diseased population. (A) Accuracy of the model across different disease categories (Respiratory, Cardiovascular, Immune, and Neurological diseases) in different datasets. 'All' refers to all participants in the dataset. Note that the disease categories are not mutually exclusive. (B) Accuracy of sleep stage prediction across fine-grained conditions in the MGH dataset using wireless signals as input. (C) Disease and drugs' effects on predicted sleep metrics. We consider Depression, Hypertension, Rheumatoid Arthritis (RA), and Diabetes. For each disease, we analyze the sleep metrics that have differences between the control and the disease groups. For datasets with medication information, we further split the disease group into patients on medications (Meds) and not on medications (No Meds) to study the effect of drugs. The statistical significances are indicated by asterisks. (*: $p < 0.05$, **: $p < 0.01$, ***: $p < 0.001$). Note that, compared to other datasets, patients in the MGH dataset have higher REM latency and less REM duration, which is due to patients visiting the MGH sleep lab are more likely to have sleep problems.

Acknowledgment

We would like to express our gratitude to the members of the Katabi's lab at MIT for their comments on our manuscript. We would also like to thank all the research participants for their time and contributions. The content is solely the responsibility of the authors and does not necessarily represent the official views of NSF or other sponsors.

Funding

This work is funded by the National Science Foundation Award No. IIS-2014391.

Financial Disclosure

D.K. receives research funding from the NIH, NSF, Sanofi, Takada, IBM, Gwangju Institute of Science and Technology, Michael J Fox Foundation, Helmsley Charitable Trust, and the Rett Syndrome Research Trust. She is a co-founder of Emerald Innovations and has a personal equity interest in the company. She also serves on the board of directors of Cycleron, the scientific advisory board of Janssen, and the data science advisory board of Amgen. B.W. receives funding from the NIH and NSF. He is a co-founder, scientific advisor, and consultant to Beacon Biosignals and has a personal equity interest in the company. J.S. receives research funding from the NSF. These relationships played no role in the present study. The rest of the authors have no financial disclosures. R.H. is a co-founder of Emerald Innovations and has a personal equity interest in the company.

Non-financial Disclosure

None

Author Contributions

D.K. conceived the idea of contactless PSG from radio waves. H.H., C.L., and D.K. developed the machine learning models and algorithms. D.K., J.S. and B.W. conceived the collaboration and secured the funding. B.W. provided clinical insights. R.H. deployed the radio devices in the sleep lab. W.G., K. G., H.S., and B.W. provided the anonymized MGH Sleep Lab dataset and the corresponding demographic and comorbidities information. H.H. and C.L. processed the data and performed experimental validation. H.H., C.L., M.O., and D.K. conducted the analysis. H.H. and C.L. generated the figures. H.H., C.L., and D.K. wrote the original manuscript. D.K. supervised the work. All authors reviewed and approved the manuscript.