

TeleOR: Real-time Telemedicine System for Full-Scene Operating Room

Yixuan Wu^{1,2,3*}, Kaiyuan Hu^{4*}, Qian Shao⁵, Jintai Chen⁶(✉),
Danny Z. Chen⁷, and Jian Wu^{1,2}(✉)

¹ School of Public Health, Zhejiang University, China

² State Key Laboratory of Transvascular Implantation Devices of The Second
Affiliated Hospital, Zhejiang University School of Medicine, China

³ Institute of Wenzhou, Zhejiang University, China

⁴ School of Science and Engineering, The Chinese University of Hong Kong,
Shenzhen, China

⁵ College of Computer Science and Technology, Zhejiang University, China

⁶ Computer Science Department, University of Illinois at Urbana-Champaign, USA

⁷ Department of Computer Science and Engineering, University of Notre Dame, USA
wyx_chloe@zju.edu.cn

Abstract. The advent of telemedicine represents a transformative development in leveraging technology to extend the reach of specialized medical expertise to remote surgeries, a field where the immediacy of expert guidance is paramount. However, the intricate dynamics of Operating Room (OR) scene pose unique challenges for telemedicine, particularly in achieving high-fidelity, real-time scene reconstruction and transmission amidst obstructions and bandwidth limitations. This paper introduces TeleOR, a pioneering system designed to address these challenges through real-time OR scene reconstruction for Tele-intervention. TeleOR distinguishes itself with three innovative approaches: dynamic self-calibration, which leverages inherent scene features for calibration without the need for preset markers, allowing for obstacle avoidance and real-time camera adjustment; selective OR reconstruction, focusing on dynamically changing scene segments to reduce reconstruction complexity; and viewport-adaptive transmission, optimizing data transmission based on real-time client feedback to efficiently deliver high-quality 3D reconstructions within bandwidth constraints. Comprehensive experiments on the 4D-OR surgical scene dataset demonstrate the superiority and applicability of TeleOR, illuminating the potential to revolutionize tele-interventions by overcoming the spatial and technical barriers inherent in remote surgical guidance.

Keywords: Telemedicine · OR reconstruction · Surgical Intervention

1 Introduction

The emergence of telemedicine marks a pivotal advancement in utilizing technology to broaden the accessibility of specialized medical expertise [31,29,8]. This

* Equal Contributions.

shift is crucial for providing timely expert guidance for remote surgery, especially in scenarios where the distance from specialized professionals could delay vital treatment. Telemedicine encompasses a broad range of disciplines, from social and perceptual to behavioral and technical aspects of remote healthcare delivery. In surgical intervention scenarios, the deployment of telemedicine systems necessitates a series of critical steps, including onsite scene reconstruction, transmission, remote rendering, and intervention. A significant technical challenge in telemedicine lies in achieving real-time reconstruction and transmission. This necessity is more pronounced in the context of surgery, as it demands a greater level of real-time responsiveness compared to other telemedicine applications, i.e., ward monitoring or remote consultations.

Previous research has explored various methods by employing multiple static RGB-D cameras to capture and enhance the reconstruction of 3D scenes. The scenarios are often first modeled by separate point clouds from per-camera viewpoints, which are then fused into a common reference frame using intrinsic and extrinsic parameters [1,14]. To achieve higher fidelity reconstructions, some researchers proposed and refined the use of neural implicit functions [9,17,26,28] and generative models [10,18]. However, these methods traditionally relied on designing sophisticated algorithms for robust visual feature extraction [10,30,32] and multiview calibration [9,12], typically conducted in offline settings. On the other hand, certain strategies have aimed to simplify these methods to enhance reconstruction efficiency. However, these simplifications tend to compromise the accuracy of the reconstructions, making them only suitable for static scenes [24,27] rather than dynamic surgical scene. Furthermore, these approaches [5] may restrict the number of remote clients who can access the system, thereby diminishing their applicability for tele-intervention in surgical settings.

The Operating Room (OR) presents a dynamic and complex scenario, teeming with medical staff, patients, and an array of medical equipment, which poses more challenges in scene understanding scenarios [22]. Unlike traditional static settings where full-scene capturing is achieved by deploying camera arrays to surround clear capture regions, OR often involves obstructions from large devices, hanging monitors, and moving staff, leading to frequent occlusions of scene capture cameras. Furthermore, the balance between high-quality reconstruction and efficient real-time transmission underscores the significant challenges in surgical tele-intervention. To address these challenges, in this work, we propose TeleOR, a real-time full-scene **O**perating **R**oom scene reconstruction system for **T**ele-intervention. TeleOR introduces innovative solutions critical for overcoming the complex challenges of telemedicine in surgical scene. Its critical designs and insights involve:

Dynamic Self-Calibration: We introduce an innovative dynamic self-calibration approach for multiview cameras in OR scene. This technique leverages the OR scene’s inherent features as reference points for calibration, eliminating the reliance on pre-set markers or physical calibration tools. Consequently, it enables cameras to move and self-calibrate in real-time, effectively avoiding obstructions.

Selective OR Reconstruction: Recognizing that specific sections of the OR scene remain unchanged for extended durations, our TeleOR system concentrates on identifying and reconstructing only the crucial, dynamic areas. This strategy avoids the unnecessary reconstruction of static parts, thereby simplifying the overall complexity of the reconstruction process.

Viewport-Adaptive Transmission: To address the limitations of network bandwidth, particularly in underdeveloped regions, TeleOR incorporates a novel viewport-adaptive transmission strategy. This technique capitalizes on real-time feedback from the client, i.e., viewport information, to selectively transmit and render only those segments of the scene that the user is likely to view, thus ensuring the transmission of high-quality 3D reconstructions in real-time.

Our TeleOR undergoes comprehensive evaluation on 4D-OR dataset, focusing on reconstruction quality and transmission efficiency under various network constraints. The findings reveal that TeleOR achieves high-quality reconstructions and maintains real-time performance even in bandwidth-limited scenarios, highlighting its effectiveness and reliability for tele-intervention applications.

2 Related Work

Operating Room (OR) Reconstruction. Holistic reconstruction and comprehension of OR marks a pivotal advancement towards the evolution of computer-assisted surgical interventions. The integration of sensing technologies within the OR facilitates the identification of individuals, objects, and their interrelations, serving as a foundation for the development of advanced surgical intervention [9,26,7]. In this regard, Özsoy et al.[23] introduced the first public 4D surgical scene dataset, 4D-OR. Gerats et al.[9] investigated the application of Neural Radiance Fields (NeRF) [17] for dynamic scene reconstruction within the OR, demonstrating depth-supervised regularization notably enhances image fidelity. Additionally, Eck et al. [7] employed a marching cubes variant on point clouds for OR surface reconstruction. Nevertheless, current approaches have yet to achieve real-time remote surgical intervention, a critical capability for enabling professionals to conduct remote surgical interventions.

Virtual Reality (VR) assisted Intervention. Current surgical interventions [3,2,15,20,34,4,6] typically involve the reconstruction of lesions and organs using VR systems and head-mounted displays, aiding physicians in executing on-site procedures. Chong et al. [4] introduced a cutting-edge VR system capable of tracking the laparoscope’s pose and reconstructing the 3D surface structure of organs, thereby facilitating laparoscopic surgery. Similarly, Doughty et al. [6] developed SurgeonAssist-Net, a framework designed for action-and-workflow-driven virtual surgical assistance. However, a gap remains in real-time remote interventions based on reconstructed OR scene, especially for underdeveloped regions with limited network resources.

3 Methodology

In Fig. 1, our proposed TeleOR systematically integrates multiview self-calibration, selective OR reconstruction, and viewport-adaptive transmission to provide remote professionals with real-time fully immersive tele-consulting capabilities.




Fig. 1: TeleOR facilitates real-time observation of surgeries through: (1) self-calibration of multiview cameras, (2) selective reconstruction of changing scene segments, and (3) partial remote transmission tailored to the users’ predicted viewport for the next time step.

3.1 Multiview Self-calibration

Different from traditional indoor reconstruction scenarios, real-time OR reconstruction poses multiple challenges: **(1)** Obstructions of capture devices: Due to frequent movements of medical staff and equipment during surgeries, capture devices are often obstructed, impacting the calibration and synchronization of the capture device array. **(2)** Restriction of pre-recording setups: OR is often equipped with sophisticated instruments and monitors, which prevents the deployment of physical calibration objects exploited by traditional methods. To tackle the above challenges, different from traditional calibration methods [19] that require pre-defined makers priors, we propose a dynamic self-calibration strategy that directly exploits the features extracted in the OR scene as the calibration reference.

In detail, RGB-D cameras are strategically positioned to ensure overlapping fields of view (FoV), with one camera designated as the anchor. The coordinate system of the anchor camera is set as the origin for the entire coordinate system. Initially, the first frame from each camera is processed to extract ORB features [16] for calibration. These features are then matched across adjacent FoV to establish camera positions relative to each other. Subsequently, a transformation matrix is derived to align the reconstructed scene with the reference coordinate system through translational and rotational adjustments.

After the initial calibration, to prevent capture cameras from being obstructed, we mount the cameras on mobile platforms to enable their movement. In this context, preserving calibration accuracy is crucial throughout the surgical procedure due to shifts in the camera’s FoV. Therefore, we perform real-time

updates of the camera positions. The optical flow method [21] is utilized to track changes in camera pose by estimating the motion of feature points across frames, allowing for effective camera movement tracking. Through continuous analysis of optical flow vectors, we dynamically adjust the camera’s calibration parameters, ensuring accurate and synchronized scene capture within the OR system.

3.2 Selective OR Reconstruction

A distinctive characteristic of the OR scene is that many objects remain static for the majority of the surgical process. Leveraging this observation, we propose to reuse already reconstructed content, focusing updates only on the changing parts of the scene. To this end, we use an optical-flow-based motion detection method to guide this selective reconstruction procedure.




Fig. 2: Selective OR Reconstruction Pipeline.

Optical Flow Calculation. As illustrated in Fig. 2, we divide the RGB-D stream from each FoV into $N \times N$ channels, which produces $N \times N$ tiles per frame, named $T_t = \{T_t^1, T_t^2, \dots, T_t^{N^2}\}$ for tiles at time t . During the reconstruction process of each frame, the inter-frame disparity of each tile is examined, by calculating the cumulative magnitude of the optical flow field difference D_c with the previous tile at the same position. To reduce the computational cost of pixel-wise optical flow difference, we exploit sparse optical flow calculation of **Lucas-Kanade algorithm** [33]. Specifically, we compute the frame-wise optical flow difference based on two assumptions:

- **Brightness Constancy Assumption:** The brightness of a pixel does not change as it moves from one frame to the next.
- **Small Motion Assumption:** The motion between frames is assumed to be small, and the displacement vector is constant within the local neighborhood.

The calculation of optical flow vector (U_{xy}, V_{xy}) for point (x,y) between two consecutive tiles T_t^n and T_{t-1}^n , can be attained by solving the following equation:

$$\begin{bmatrix} U_{xy} \\ V_{xy} \end{bmatrix} = \begin{bmatrix} E_x & E_{xy} \\ E_{xy} & E_y \end{bmatrix}^{-1} \begin{bmatrix} -E_{xt} \\ -E_{yt} \end{bmatrix}, \quad (1)$$

where

$$E_x = T_x^2, \quad E_y = T_y^2, \quad E_{xy} = T_x \cdot T_y, \quad E_{xt} = T_x \cdot T_t, \quad E_{yt} = T_y \cdot T_t, \quad (2)$$

$$T_x(x, y) = \left. \frac{\partial T}{\partial x} \right|_{(x,y)}, \quad T_y(x, y) = \left. \frac{\partial T}{\partial y} \right|_{(x,y)}, \quad (3)$$

$$T_t(x, y) = T_t(x, y) - T_{t-1}(x, y). \quad (4)$$

In this way, we can determine which tile has perceptible visual change given an adjustable threshold according to available computational and transmission bandwidth resources. The calculation of the cumulative magnitude of optical flow difference can be represented as:

$$D_c = \sum_{x=1}^{N_{width}} \sum_{y=1}^{N_{height}} M_t(x, y), \quad (5)$$

$$M_t(x, y) = \sqrt{U_{xy}^2 + V_{xy}^2}, \quad (6)$$

where D_c is the cumulative magnitude of optical flow difference between tile T_t^n and T_{t+1}^n , and $M_t(x, y)$ is magnitude of the optical flow difference for point (x,y). **Tile-based Selective Reconstruction.** For each tile, if D_c exceeds the set threshold θ , indicating a perceptible visual difference, the corresponding position in the RGB-D frame will be masked. The value of threshold θ is dynamically determined by the real-time bandwidth. Then, these partially masked frames are forwarded to the synthesis server for conversion into point clouds. Throughout the point cloud generation process, for the masked tiles within each frame, the corresponding sections of the scene are not regenerated. Instead, the previously generated segments in the preceding frame are retained and reused.

3.3 Viewport-Adaptive Transmission

To transmit high-quality 3D reconstructions while overcoming limited bandwidth networks, TeleOR capitalizes on real-time feedback from the client, i.e., viewport information, to adaptively load data for transmission. Initially, we gather data on the user’s past FoV movements to construct a historical trajectory. Utilizing this data, following [11], an efficient LSTM network is employed to predict the user’s FoV for the next time step.

This prediction enables the strategic synthesis of point clouds solely within the predicted FoV, ensuring efficient utilization of bandwidth by transmitting and rendering only the scene segments likely to be in the user’s view. In practice, to ensure accurate predictions and prevent prediction failures, TeleOR employs a prediction range of 120 degrees, expanding beyond the typical user’s viewing frustum [13,25]. This approach not only covers a wider area for improved situational awareness but also strategically reduces the point cloud density from the center of the predicted FoV to minimize data usage.




Fig. 3: Visualization of reconstruction results of different network conditions.

Net Condition	R_{reuse}	$MSSIM$
Limited (20Mbps)	85.8%	0.9012
Medium (50Mbps)	66.9%	0.9615
Adequate (100Mbps)	27.3%	0.9827

Table 1: Reconstruction quality assessment of different network conditions.

In parallel, TeleOR considers the current network bandwidth to adjust the dynamic threshold θ for selective OR scene reconstruction (Sec. 3.2) and the reconstructed point cloud density for the next transmission. This approach allows for the fine-tuning of data transmission, balancing the quality of the point cloud stream against the imperative of maintaining real-time streaming capabilities of the OR scene. Through this balance of predictive modeling and bandwidth management, TeleOR ensures the delivery of essential, high-quality 3D content under the constraints of varying network conditions.

4 Experiment

4.1 Experimental Setup

Datasets and Evaluation Metrics. To evaluate the reconstruction quality and real-time performance of our proposed TeleOR system, we conduct the evaluation on the public operating room dataset 4D-OR [23], which is composed of 6734 scenes captured by six RGB-D Kinect sensors. The evaluation focuses on two key aspects that have the most significant impact on the tele-intervention performance: reconstruction quality (Sec. 4.2) and transmission efficiency (Sec. 4.3).

Implementations. We implement the TeleOR on top of Open3D. We deploy the RGB-D sequences of each FoV on six individual edge computing devices (Nvidia Jetson Nano), each connecting to the synthesis server (equipped with Intel i7-12700 CPU, NVIDIA GeForce RTX 3090 GPU, 64G RAM) via a WiFi connection. For the client-side renderer, we built a real-time player on top of Unity and deployed it on a Meta Quest Pro head-mounted display. The supplementary materials provide a detailed description of the rendering data format.

4.2 Reconstruction Quality

To accurately assess the visual quality of the reconstructed scene, we use the Structural Similarity Index Metric (SSIM) for analyzing 2D renderings derived

Table 2: Transmission performance under different network conditions.

Methods & Net Condition	Frames per Second			Latency (ms)		
	Min.	Avg.	Max.	Min.	Avg.	Max.
Limited (20Mbps)	15.6	22.7	24.3	297	302	357
Medium (50Mbps)	22.5	25.2	29.9	160	192	269
Adequate (100Mbps)	25.9	27.7	29.9	95	105	136
w/o. selective recon. (100Mbps)	3.2	5.7	6.7	132	153	275
w/o. viewport adapt. (100Mbps)	7.5	9.2	12.7	172	197	267

from the 3D scene. This analysis is carried out from a series of predefined viewpoints, termed Multiview SSIM (MSSIM). MSSIM represents the average SSIM value obtained across these multiple predefined viewpoints. Also, Scene Reuse Ratio (R_{reuse}) is adopted to evaluate the effectiveness of utilizing previously reconstructed scene elements in new reconstructions. We evaluate the reconstruction performance of our selective reconstruction approach under various network constraints, with bandwidth limits set to 20, 50, and 100 Mbps. Therefore, our established bandwidth testing range represents a stringent set of extreme conditions. For one thing, in Tab. 1, it is observed that at a bandwidth of 20 Mbps, our method achieves a R_{reuse} of 85.8%, indicating a substantial reuse of segments reconstructed from previous frames. This effectively reduces the complexity of the reconstruction process, ensuring the **real-time performance** of our approach. For another, Fig. 3 demonstrates that our TeleOR achieves high-quality reconstructions at 100 Mbps, and maintains acceptable performance even under severe bandwidth restrictions. This significantly ensures the **visual quality** of real-time reconstruction in bandwidth-constrained environments.

4.3 Transmission Efficiency

The transmission efficiency is assessed through two metrics: Frames per Second (FPS) and Latency, with FPS reflecting the smoothness of video playback and Latency measuring the delay between an action and its visual feedback on the screen. Evaluation is conducted on all ten groups of 4D-OR, with each containing 3000 sample frames for simulation, using the same network constraints as Sec.4.2. As shown in Tab. 2, despite operating under severely constrained bandwidth conditions, TeleOR is still capable of maintaining a frame rate of 22.7 FPS.

4.4 Ablation Study

To explore the effectiveness of TeleOR’s key designs, the selective reconstruction (Sec. 3.2) and viewport adaptive (Sec. 3.3) strategy is removed, respectively. In Tab. 2, without selective reconstruction or viewport adaptation, a noticeable drop in FPS can be observed even under adequate network resources. This observation emphasizes the critical role of TeleOR’s core components, especially under poor network conditions.

5 Conclusion

In this paper, we introduced TeleOR, an advanced system designed for enhancing tele-intervention in OR through real-time, high-fidelity scene reconstructions. By integrating dynamic self-calibration, selective reconstruction, and viewport-adaptive transmission, it addresses issues like camera occlusions, scene complexity, and bandwidth constraints, enhancing remote surgical guidance. Experiments on the 4D-OR dataset indicate that our TeleOR not only achieves exceptional reconstruction accuracy but also superior transmission efficiency under strict bandwidth limitations, enabling seamless real-time interventions.

References

1. Beck, S., Froehlich, B.: Volumetric calibration and registration of multiple rgb-d sensors into a joint coordinate system. In: 2015 IEEE Symposium on 3D User Interfaces (2015)
2. Cannizzaro, D., Zaed, I., Safa, A., Jelmoni, A.J., Composto, A., Bisoglio, A., Schmeizer, K., Becker, A.C., Pizzi, A., Cardia, A., et al.: Augmented reality in neurosurgery, state of art and future projections. a systematic review. *Frontiers in Surgery* (2022)
3. Chen, J., Zheng, X., Yu, H., Chen, D.Z., Wu, J.: Electrocardio Panorama: synthesizing new ECG views with self-supervision (2021)
4. Chong, N., Si, Y., Zhao, W., Zhang, Q., Yin, B., Zhao, Y.: Virtual reality application for laparoscope in clinical surgery based on siamese network and census transformation. In: *Medical Imaging and Computer-Aided Diagnosis* (2022)
5. Dou, M., Khamis, S., Degtyarev, Y., Davidson, P., Fanello, S.R., Kowdle, A., Escolano, S.O., Rhemann, C., Kim, D., Taylor, J., et al.: Fusion4d: Real-time performance capture of challenging scenes. *ACM Transactions on Graphics* (2016)
6. Doughty, M., Singh, K., Ghugre, N.R.: Surgeonassist-net: Towards context-aware head-mounted display-based augmented reality for surgical guidance. In: *MICCAI* (2021)
7. Eck, U., Wechner, M., Pankratz, F., Yu, K., Lazarovici, M., Navab, N.: Real-time 3d reconstruction pipeline for room-scale, immersive, medical teleconsultation. *Applied Sciences* (2023)
8. Ekeland, A.G., Bowes, A., Flottorp, S.: Effectiveness of telemedicine: a systematic review of reviews. *International journal of medical informatics* (2010)
9. Gerats, B.G., Wolterink, J.M., Broeders, I.A.: Dynamic depth-supervised nerf for multi-view rgb-d operating room videos. In: *International Workshop on Predictive Intelligence In Medicine* (2023)
10. He, X., Tan, C., Han, L., Liu, B., Axel, L., Li, K., Metaxas, D.N.: Dmvr: Morphology-guided diffusion model for 3d cardiac volume reconstruction. In: *MICCAI* (2023)
11. Hou, X., Zhang, J., Budagavi, M., Dey, S.: Head and body motion prediction to enable mobile vr experiences with low latency. In: 2019 IEEE Global Communications Conference (2019)
12. Hu, K., Jin, Y., Yang, H., Liu, J., Wang, F.: Fsvvd: A dataset of full scene volumetric video. In: *Proceedings of the 14th Conference on ACM Multimedia Systems* (2023)

13. Hu, K., Yang, H., Jin, Y., Liu, J., Chen, Y., Zhang, M., Wang, F.: Understanding user behavior in volumetric video watching: Dataset, analysis and prediction. In: Proceedings of the 31st ACM International Conference on Multimedia. pp. 1108–1116 (2023)
14. Jin, Y., Hu, K., Liu, J., Wang, F., Liu, X.: From capture to display: A survey on volumetric video (2023)
15. Jud, L., Fotouhi, J., Andronic, O., Aichmair, A., Osgood, G., Navab, N., Farshad, M.: Applicability of augmented reality in orthopedic surgery—a systematic review. *BMC musculoskeletal disorders* (2020)
16. Karami, E., Prasad, S., Shehata, M.: Image matching using sift, surf, brief and orb: performance comparison for distorted images. arXiv preprint arXiv:1710.02726 (2017)
17. Kato, H., Ushiku, Y., Harada, T.: Neural 3d mesh renderer. In: CVPR (2018)
18. Kim, B., Ye, J.C.: Diffusion deformable model for 4d temporal medical image generation. In: MICCAI (2022)
19. Kowalski, M., Naruniec, J., Daniluk, M.: Livescan3d: A fast and inexpensive 3d data acquisition system for multiple kinect v2 sensors. In: 3DV (2015)
20. Linxweiler, M., Pillong, L., Kopanja, D., Kühn, J.P., Wagenpfeil, S., Radosa, J.C., Wang, J., Morris, L.G., Al Kadah, B., Bochen, F., et al.: Augmented reality-enhanced navigation in endoscopic sinus surgery: A prospective, randomized, controlled clinical trial. *Laryngoscope Investigative Otolaryngology* (2020)
21. Liu, J., Subramanian, K.R., Yoo, T.S.: An optical flow approach to tracking colonoscopy video. *Computerized Medical Imaging and Graphics* (2013)
22. Özsoy, E., Czempiel, T., Örnek, E.P., Eck, U., Tombari, F., Navab, N.: Holistic or domain modeling: a semantic scene graph approach. *International Journal of Computer Assisted Radiology and Surgery* (2023)
23. Özsoy, E., Örnek, E.P., Eck, U., et al.: 4D-OR: Semantic scene graphs for or domain modeling. In: MICCAI (2022)
24. Stotko, P., Krumpfen, S., Hullin, M.B., Weinmann, M., Klein, R.: Slamcast: Large-scale, real-time 3d reconstruction and streaming for immersive multi-client live telepresence. *IEEE transactions on visualization and computer graphics* (2019)
25. Sung, K., Shirley, P., Baer, S.: *Essentials of interactive computer graphics: concepts and implementation*. CRC Press (2008)
26. Wang, Y., Long, Y., Fan, S.H., Dou, Q.: Neural rendering for stereo 3d reconstruction of deformable tissues in robotic surgery. In: MICCAI (2022)
27. Weibel, N., Gasques, D., Johnson, J., Sharkey, T., Xu, Z.R., Zhang, X., Zavala, E., Yip, M., Davis, K.: Artemis: Mixed-reality environment for immersive surgical telementoring. In: *Extended Abstracts of the 2020 CHI Conference on Human Factors in Computing Systems* (2020)
28. Wolterink, J.M., Zwienenberg, J.C., Brune, C.: Implicit neural representations for deformable image registration. In: MIDL (2022)
29. Wootton, R.: *Telemedicine*. Bmj (2001)
30. Wu, Y., Chen, J., Yan, J., Zhu, Y., Chen, D.Z., Wu, J.: GCL: Gradient-guided contrastive learning for medical image segmentation with multi-perspective meta labels. In: Proceedings of the 31st ACM International Conference on Multimedia (2023)
31. Wu, Y., Hu, K., Chen, D.Z., Wu, J.: AI-enhanced virtual reality in medicine: A comprehensive survey (2024)
32. Wu, Y., Zheng, B., Chen, J., Chen, D.Z., Wu, J.: Self-learning and one-shot learning based single-slice annotation for 3d medical image segmentation. In: MICCAI (2022)

33. Yedjour, H.: Optical flow based on lucas-kanade method for motion estimation. In: Artificial Intelligence and Renewables Towards an Energy Transition 4 (2021)
34. Zorzal, E.R., Gomes, J.M.C., Sousa, M., Belchior, P., da Silva, P.G., Figueiredo, N., Lopes, D.S., Jorge, J.: Laparoscopy with augmented reality adaptations. Journal of biomedical informatics (2020)