Transformation trees - documentation of multimodal image registration

Agnieszka Anna Tomaka^{a,*}, Dariusz Pojda^a, Michał Tarnawski^a, Leszek Luchowski^a

^aInstitute of Theoretical and Applied Informatics, Polish Academy of Sciences, Baltycka 5, Gliwice, 44-100, Poland

Abstract

The paper presents proposals for the application of a tree structure to the documentation of a set of transformations obtained as a result of various registrations of multimodal images obtained in coordinate systems associated with acquisition devices and being registered in one patient-specific coordinate system. A special file format .dpw (digital patient workspace) is introduced. Examples of different registrations yielded from orthodontic analysis and showing main aspects of the usage of tree structure are illustrated in dpVision software.

Keywords: transformation tree, multimodal image registration, dpVision

1. Introduction

The advancement of computer systems and tools for the acquisition of multidimensional, multimodal images has led to the growing presence of the concepts of the virtual patient and virtual twin in the literature [Joda and Gallucci (2015); Marradi et al. (2020); Seth et al. (2024)]. Both of these concepts are related to the acquisition of digital patient data and matching them in such a way as to obtain the most complete, accurate image of the patient both in terms of his current state and changes over time. The concepts are particularly applicable in in orthodontics and in assisting in the planning of surgical procedures. These concepts consist of the following steps: image

^{*}atomaka@iitis.pl

data collection, segmentation and 3D reconstruction, multimodal data integration, physical and biomechanical simulations, however this workflow is also used in other work [Lo and Lin (2023)].

Integration of multimodal imaging data is associated with such transformations of individual images that they can be visualized in a single common system of patient-related coordinates, enabling precise analysis and comparison. A necessary assumption for data integration is matching, so that the data sets correspond to each other, in the sense of representing the same location in space and in being acquired at the same time. This matching process is called registration and, from a mathematical point of view, it involves finding a transformation which minimizes the distances between the corresponding points.

Image registration employs various techniques, requiring imaging system calibration for proper scaling and metric data acquisition. Registration relies on known geometric relationships between coordinate systems or determining them from images.

When correspondences between multimodal datasets are known, alignment is achieved using the Least Squares method. If correspondences are unknown, the ICP algorithm estimates them iteratively, simultaneously finding the registering transformation. When images lack common parts, additional reference objects ensure consistent spatial relationships.

Motion registration involves distinguishing stable and moving parts, aligning the stable regions to analyze motion. Transformations vary by image type: rigid for fixed structures, affine for CT with gantry tilt, deformable for flexible structures, and projection for 3D-to-2D alignment.

Transformations can be single or a sequence from intermediate registrations, but multiple steps may introduce numerical inaccuracies. Accuracy and data integrity are critical in medical imaging, requiring justification and documentation of all operations.

This article introduces transformation trees for managing multi-modal image registration, with sample orthodontic visualizations using dpVision software [Pojda (2024)], program developed at IITiS PAS, which is used for processing multimodal images.

2. The idea of transformation tree

2.1. Existing Formats of transformation documentation

There are several programs and graphics libraries on the market for registering different types of images. Functions for registering triangular meshes or point clouds, finding rigid-body transformations using ICP can be found in RapidForm, MeshLab, and CloudCompare programs. There are also programs that support deformable registration such as ITK, ANT, Elastix, 3DSlicer. It is more difficult to describe deformation using a single function, and so hierarchical approaches are used that take into account more complicated functions or a local range of influence of a given transformation.

Most 3D graphics processing program propose their own data storage formats. Such formats are usually optimised for the specific applications of the software in question or to optimise I/O operations. With the development of 3D graphics technology, however, interoperability between different formats and software is becoming increasingly important, allowing for seamless data exchange and tool synergy in complex projects.

In addition to dedicated 3D formats, general-purpose formats such as .xml or .json are also used, which allow any information to be stored in a text file. Although these are not strictly 3D data storage formats, they provide a flexible basis for more specialised standards. For example, the glTF (GL Transmission Format) standard was built on top of the .json format. glTF has become very popular due to its lightness and optimisation for real-time data transmission, making it a good choice for web, augmented reality (AR) and virtual reality (VR) applications.

Some formats have become universal for working with 3D data due to their simplicity and wide support in various applications. Among the most popular are the .obj, .stl, and VRML (or its modifications, such as X3D). They are widely used in both modeling and 3D printing.

Binary formats, such as .fbx or .3ds, offer more advanced capabilities, allowing large amounts of data to be stored, including textures, materials and animations. Their complexity ensures efficient processing, but at the same time makes them dependent on specific software, which may limit their use in other environments. Therefore, the choice of the appropriate 3D data format should take into account not only the purpose of the project, but also the requirements for compatibility, performance and the ability to integrate into complex production processes.

2.2. Transformation tree - data structure of dpVision

The proposed scene storage format corresponds to a hierarchical, treebased data structure in the dpVision program [Pojda et al. (2025)]. The root object is the *workspace* of the program. The workspace is the main coordinate system for all objects displayed in the program.

Other objects representing 3D images or other graphic elements (annotations) can be placed in the workspace, as well as special objects such as transformations or motions.

Each graphic object in the dpVision workspace can have at most one parent object and any number of child objects. The coordinates of such an object are specified in the coordinate system of its parent. If it has no parent object, its coordinates are relative to the global coordinate system of the scene.

Graphical objects used to display the data processed in a program do not usually have their own transformation matrices. They are always drawn in the coordinate system inherited from their parent and propagate this coordinate system to their descendants.

2.3. Transformation objects

A transformation is a special object in the dpVision workspace. It has no visible features of its own, but defines a **group** of its child objects. All children are located in a common *local coordinate system*, defined by a transformation matrix provided by their parent.

If successive transformations are nested, the coordinate system applied to their descendant objects will be defined by a matrix that is the result of multiplying all transformation matrices in the sequence.

2.4. Movement objects

A movement is an object in dpVision that can be understood as a timevarying transformation. Instead of a single transformation matrix, it has a sequence of such matrices defined. Each corresponds to one frame of animation. The display time of each frame may also be defined. If an animation is active, the *local coordinate system* for a of its child objects changes over time.

2.5. Structure of .dpw files

A text format .dpw is human-readable and can be easily modified using any text editor. It is also relatively easy to interpret by any application that allows the creation of user scripts or plug-ins.

The basis of the .dpw storage format is a simple scheme that allows a single object to be stored:

```
object_type { <properties_list> | <child_objects_list> }
```

- **object** type is a keyword assigned to a specific type of object represented in the program, e.g. mesh, volumetric, etc.
- **properties** list object properties. Some properties are the same for all objects, such as label, description or keywords, while others are defined only for a specific type of object
- child objects list nested definitions of descendant objects in a format analogous to the one currently described.

A file saved in .dpw format describes relationships between objects, but does not store specific information about their shape. Instead, the object definition has a **file** property that contains a reference to the specific file containing the object data. Data files can be in any format supported by dpVision (include .dpw files).Each property of an object can be a number, an array of numbers or a string.

Template for the definition of an object representing a 3D image:

```
shell {
    label "object label"
    file "absolute/or/relative/file.path"
    <other_optional_properties>
    <optional_child_object>
    . . .
    <optional_child_object>
}
```

Similarly, the definition of a transformation object in .dpw format is:

Optionally, instead of a transformation matrix, partial operations such as rotation about a given axis, translation and scaling can be defined.

The definitions of descendant objects are optional. A transformation defines the coordinate system for its child objects. So if it has no descendant, it will have no effect on the scene layout.

3. Applying transformation tree to document multimodal image registration in orthodontics

Traditional orthodontic records typically consist of 3D dental models, 2D or 3D X-ray images, and either photographs or 3D scans of the face, smile, and intraoral tissues.

This collection of data, diverse in both format and content, already constitutes a multimodal dataset (Figure 1). While each imaging modality can independently contribute to diagnostics, combining them creates a synergistic effect, resulting in a more comprehensive representation, visualization, and understanding of the relevant structures [Tomaka et al. (2019)].

3.1. Patient-specific coordinate system

Even a single image is usually transformed to the patient's coordinate system for visualization and processing (Figure 2). This allows time-based comparisons for the same patient and statistical analysis across patients. The transformation can be defined as follows:



Figure 1: Multimodal Patient Data scheme



Figure 2: Single transformation scheme (note: in most of figures the convention R to L is used)

$$\mathbf{M}' = \begin{bmatrix} x'_1 & \dots & x'_N \\ y'_1 & \dots & y'_N \\ z'_1 & \dots & z'_N \\ 1 & \dots & 1 \end{bmatrix} = \mathbf{P} \cdot \begin{bmatrix} x_1 & \dots & x_N \\ y_1 & \dots & y_N \\ z_1 & \dots & z_N \\ 1 & \dots & 1 \end{bmatrix} = \mathbf{P} \cdot \mathbf{M}$$
(1)

where \mathbf{M} , \mathbf{M}' are image representations (points, vertices of a mesh, vertices of voxels) accordingly in the device and patient-specific coordinate systems; \mathbf{P} is the transformation matrix.

In the case of a rigid body transformation the matrix \mathbf{P} has the form:

$$\begin{bmatrix} x' & y' & z' & 1 \end{bmatrix}^T = \mathbf{P} \cdot \begin{bmatrix} x & y & z & 1 \end{bmatrix}^T = \mathbf{T} \cdot \mathbf{R} \cdot \mathbf{S} \cdot \begin{bmatrix} x & y & z & 1 \end{bmatrix}^T$$
(2)

where \mathbf{S} is the scale, \mathbf{R} is the rotation matrix, \mathbf{T} is the translation matrix.

3.2. List of transformations

Each image has to be transformed from the local device dependant coordinates to patient coordinates. This transformation may be represented by single transformation matrix or by sequence of matrices representing partial registrations. At each level the hierachical structure is preserved, previous transformations become children of the next one and resulting transformation is the multiplication of the matrices from the list, up to the root of the tree - workspace coordinates and down to the the device coordinate system.

$$\mathbf{M}' = (\mathbf{P_m}) \cdot \mathbf{M}$$

where $(\mathbf{P_m}) = \mathbf{P_{m_n}} \cdot \mathbf{P_{m_{n-1}}} \cdot \dots \cdot \mathbf{P_{m_1}} \cdot \mathbf{P_{m_0}}$ is the transformation matrix representing a single transformation or subsequent transformations of the model from device to patient coordinate systems.

The simplest example of the use of the list is the situation of registration using coarse registration, being the first step before the overall registration.

3.3. Embeding intraoral scans in 3D face scans - list of transformations

Another example that demonstrates the linking of a sequence of transformations is the embedding of intraoral scans inside a 3d face scan. The procedure was presented in [Tomaka et al. (2007); Solaberrieta et al. (2013, 2015); Lam et al. (2018)]. The idea is to use a facial arch with intraoral and external surfaces for registration. These additional surfaces are scanned with both objects. The auxiliary scans make it possible to create a list of transformations required to register 3d face and intraoral scans. Adding additional intermediate registrations thus corresponds to nesting the transformation in the list (Figure 3). The overall transformation corresponds to the multiplication of the matrix of component transformations of the list.

3.4. A group - common coordinate system

The tree of transformations gives the possibility to create groups of objects on which a similar transformation acts. This way clearly illustrates the order of the transformations performed. A group is a set of images or objects that can be visualized in a common coordinate system.

Such a group can be formed by segmenting images - the extracted parts of the image are still in coordinate system of the image before segmentation. In this case the volumetric image becomes the parent, and segmented images its children.



Figure 3: Transformation tree

$$\begin{bmatrix} \mathbf{M}_{\mathbf{v}}' & \mathbf{M}_{\mathbf{1}}' & \dots & \mathbf{M}_{\mathbf{i}}' \end{bmatrix} = (\mathbf{P}_{\mathbf{m}}) \cdot \begin{bmatrix} \mathbf{M}_{\mathbf{v}} & \mathbf{M}_{\mathbf{1}} & \dots & \mathbf{M}_{\mathbf{i}} \end{bmatrix}$$

where $\begin{bmatrix} \mathbf{M_v} & \mathbf{M_1} & \dots & \mathbf{M_i} \end{bmatrix} \in \mathbb{R}^{4 \times (n_v + n_1 + \dots + n_i)}$ is a concatenation of matrices $\mathbf{M_v} \in \mathbb{R}^{4 \times n_v}$ and $\mathbf{M_k} \in \mathbb{R}^{4 \times n_k}$ for $k \in 1 \dots i$. The concatenation represents volumetric image and also different segmentations here of teeth, bones, and skin tissues of the same volumetric image (Figure 4). In addition, it is possible to separate the boness and the teeth into maxillary and mandibular parts.

A group can also be formed by elements that have been brought into the same coordinate system as a result of registration. The Figure 5 shows 3D scans of a face obtained from different viewpoints by a scanner moving around, while the scanner's displacements are known only approximately, but these rough transformations are sufficient to initially position the 3D scans so that the ICP algorithm can find the exact transformations.

$$\begin{bmatrix} \mathbf{M}_{1}^{\prime} & \dots & \mathbf{M}_{i}^{\prime} \end{bmatrix} = (\mathbf{P}_{\mathbf{m}}) \cdot \begin{bmatrix} (\mathbf{P}_{1}) \cdot \mathbf{M}_{1} & \dots & (\mathbf{P}_{i}) \cdot \mathbf{M}_{i} \end{bmatrix} =$$
$$= (\mathbf{P}_{\mathbf{m}}) \cdot \begin{bmatrix} \mathbf{P}_{\mathbf{F}_{1}} \cdot \mathbf{P}_{\mathbf{R}_{1}} \cdot \mathbf{M}_{1} & \dots & \mathbf{P}_{\mathbf{F}_{i}} \cdot \mathbf{P}_{\mathbf{R}_{i}} \cdot \mathbf{M}_{i} \end{bmatrix}$$
(3)

where $(\mathbf{P_i}) = \mathbf{P_{F_i}} \cdot \mathbf{P_{R_i}}$ are matrices of transformations for each partial 3D



Figure 4: A group formed by CBCT volume data(top-left) and CBCT segmentation - 3D reconstructions accordingly of: soft tissues (bottom-middle), bones (top-middle), teeth (bottom-left) and the mutual relations in a group (right)



Figure 5: Documentation of registering 3D face scans from different view points. Initial rough registration and final registration



Figure 6: A group formed by the registered partial scans of the head and a resulting model

scan of a head, composed of the initial and the final transformations.

3.5. Groups of objects - transformation and nesting

All elements of the group are subject to the same transformations.

Registration of partial 3D face scans from different viewpoints can be also preformed using pairwise registration between neighbouring scans.

Each newly registered scan is aligned with the previous one, forming a group. This group is then transformed using the preceding registration, progressively expanding until all scans are iteratively registered back to the first scan. It is explained in Figure 6.

$$\begin{bmatrix} \mathbf{M}_{1}' & \dots & \mathbf{M}_{i}' \end{bmatrix} = \\ = (\mathbf{P}_{\mathbf{m}}) \cdot \begin{bmatrix} \mathbf{M}_{1} & \mathbf{P}_{2} \cdot \begin{bmatrix} \mathbf{M}_{2} & \mathbf{P}_{3} \cdot \begin{bmatrix} \mathbf{M}_{3} & \mathbf{P}_{4} \cdot \begin{bmatrix} \dots & \mathbf{P}_{i-1} \cdot \begin{bmatrix} \mathbf{M}_{i-1} & \mathbf{P}_{i} \cdot \mathbf{M}_{i} \end{bmatrix} \end{bmatrix} \end{bmatrix}$$

The transformation tree can be used to document the registration of partial scans of gypsum dental models obtained from 3D scanning with a rotary table. The plaster dental models are models of the patient's teeth. They are created by taking impressions from both dental arches. Their mutual position is fixed by a bite bar. Scanning of each of the models on the rotary table is carried out by performing scans correlated with the rotary table's movement. To obtain a watertight model it is necessary to perform two or more series of scans for different positions of the plaster model on the rotary table.



Figure 7: A diagram of successive registrations of partial 3d scans of plaster dental models.

Each part of plaster model is scanned multiple times at different positions on the rotary table, producing several scan series. Registration occurs within each series, then between series, separately for the maxilla, mandible, and their occlusion. The transformation tree organizes registrations this into three nested groups: registration within groups and then registration of groups between each other to get models of the jaw, mandible and occlusion, and then registration of models of the maxilla and mandible to occlusion.

Registration of scans obtained from scanning with a rotary table at a given angle of rotation comes down to finding the axis of rotation around which the table rotates. Although the article skips over aspects of registration algorithms, it should be mentioned that these specific assumptions impose limitations on the registration process. Once the intra-group scans have been matched, the groups should be registered to each other. Since the transformation affecting a group affects the individual elements in the same way, as long as the single elements of both groups can be matched with sufficient accuracy, the registration of these groups is automatically found. A diagram of successive registrations of 3d scans of dental gypsum models scanned with a rotary table is shown in Figure 7.

3.6. Documentation of motion

In addition to the analysis of the morphology of the stomatognathic system, the temporomandibular joint its structure and its work remains in the



Figure 8: Movement acquisition scheme. All images in the captured sequence are first segmented to static ($\mathbf{M}_{\mathbf{S}}$) and moving part ($\mathbf{M}_{\mathbf{M}}$). The static part is used to find the correspondence $\mathbf{P}_{\mathbf{S}}(t_i)$ between the images. Then, the correspondence of the moving parts is found in order to determine the transformation matrices $\mathbf{P}_{\mathbf{M}}(t_i)$.



Figure 9: Scheme of visualisation of movement. M_S is a static part of 3D image and M_M is a moving part. The matrices determined in the motion capture process are sequentially used to position the moving image fragment.

field of interest of orthodontics. Not all the components that constitute the temporomandibular joint can be imaged with the imaging described in the article. (i.e. the articular disc). However, an indirect examination is used, which allows to assess the function of this joint on the basis of movement. Movement in the stomatognathic system can be described as a change in the position of the mandible relative to the mandible. Defining motion requires additional segmentation - distinguishing stable part from those subject to motion.

The movement can be treated as an object, such as a transformation, can be a group with other objects, and be subject to transformations that register a coordinate system in which the movement occurs to the coordinate system of other objects of the tree. The movement itself only affects the descendants of this object.

$$\begin{bmatrix} \mathbf{M}_{\mathbf{S}}' & \mathbf{M}_{\mathbf{M}}'(t) \end{bmatrix} = (\mathbf{P}_{\mathbf{m}}) \cdot \begin{bmatrix} \mathbf{M}_{\mathbf{S}} & \mathbf{P}_{\mathbf{M}}(t) \cdot \mathbf{M}_{\mathbf{M}} \end{bmatrix}$$

where $\mathbf{P}_{\mathbf{M}}(t)$ is the movement represented as transformations in t moment of time, $\mathbf{M}_{\mathbf{S}}$, $\mathbf{M}_{\mathbf{M}}$ - are representations of stable and moving parts.

The function that defines the motion can be obtained from different systems using different measurement methods, like Zebris JMA–Optic, axiographs, and 3D dynamic scanner or othe vision systems [Zebris Medical (2025); Jakubowska et al. (2023); Tomaka et al. (2016, 2019)].

4. Discussion

The main advantage of using transformation trees to describe patient data is that they introduce order by binding together groups of objects subjected to the same transformations. Storing the transformation coefficients also eliminates the need to save the images in successive coordinate systems, which reduces the memory needed to save the entire set.

While the visualization of the final model is related to the conversion of the model into the coordinate system associated with the user's viewpoint the application of the transformation tree the appropriate calculations are connected with only one multiplication with the latest overall transform.

The tree of the transformation ties together all the objects that constitute the patient's image data set.

This means that by recalculating the transformations from a given image to the root, and then from the root to the second image, the first can be



Figure 10: Simulation of mandible movement



Figure 11: Indirect registration of volumetric images using the alignment of 3D bone reconstructions. The resulting transformation applied to whole groups of images.

represented in the coordinate system of the second, forming a group with it. This method can be used to move the motion to the group of images of the mandible. It can also be integrated with CBCT data and analysed. This integration allows mandibular movement to be simulated (Figure 10) and the change of mandibular condyles in TMJ and occlusion analyzed dynamically.

This also means that registering the new image, when it is possible to find a transformation to any element of the tree it is possible to find registration of this image to overall patient model.

In particular, for example, if a transformation can be found between two bone tissue reconstructions this registration can be used to register the entire group of volumetric images and segmentations derived from it (Figure 11).

The use of a tree structure to record transformations leading to the registration of source image data to create a patient model is also subject to certain drawbacks. From the theory of tree structures it follows that if there are even for one node two paths to the other means that the structure, is not longer a tree, and becomes a graph. And such a situation may occur in the case of registration of two types of images existing in two groups (e.g. separate registration for bone and soft tissues), which registrations may give slightly different transformation, due to possible deformations within the soft tissues, or different segmentation thresholds. Such a situation is an indication for further research on the selection of a structure describing a set of these transformations, with the selection of weights taking into account the quality of the obtained registrations.

5. Conclusions

The most important conclusion is that the transform tree provides interrelationships of all sets of images with each other by organizing them and documenting a list of transformations of each image from the source image to the final patient-defined coordinate system. In this way, it documents all operations performed on the image and provides a path back to the source image. As such, it is essential for creating medical records.

This approach also provides an opportunity to indirectly align images that were impossible to register due to the lack of a common part.

The proposed approach is versatile enough to be used in other fields using computer graphics and multimodal imaging.

Acknowledgements

I would like to thank Krzysztof Domino,PhD for his valuable advice and motivation, which supported the creation of this article.

Declaration of competing interest

The authors have no competing interests to declare that are relevant to the content of this article.

References

- Jakubowska, S., Szerszeń, M.P., Kostrzewa-Janicka, J., 2023. Jaw motion tracking systems – literature review. Prosthodontics 73, 18–28. doi:10. 5114/ps/162663.
- Joda, T., Gallucci, G.O., 2015. The virtual patient in dental medicine. Clinical Oral Implants Research 26, 725–726. doi:10.1111/clr.12379.
- Lam, W.Y., Hsung, R.T., Choi, W.W., Luk, H.W., Cheng, L.Y., Pow, E.H., 2018. A clinical technique for virtual articulator mounting with natural head position by using calibrated stereophotogrammetry. The Journal of Prosthetic Dentistry 119, 902–908. doi:10.1016/j.prosdent.2017.07. 026.

- Lo, L.J., Lin, H.H., 2023. Applications of three-dimensional imaging techniques in craniomaxillofacial surgery: A literature review. Biomedical Journal 46, 100615. doi:10.1016/j.bj.2023.100615.
- Marradi, F., Staderini, E., Zimbalatti, M.A., Rossi, A., Grippaudo, C., Gallenzi, P., 2020. How to obtain an orthodontic virtual patient through superimposition of three-dimensional data: A systematic review. Applied Sciences 10. doi:10.3390/app10155354.
- Pojda, D., 2024. dpVision (data processing for vision). doi:10.5281/zenodo. 13944334. opensource software repository, accessed: 2025-01-10.
- Pojda, D., Zarski, M., Tomaka, A.A., Luchowski, L., 2025. dpVision: Environment for multimodal images. Article sent to SoftwareX Preprint available at SSRN: https://ssrn.com/abstract=5010745.
- Seth, I., Lim, B., Lu, P.Y.J., Xie, Y., Cuomo, R., Ng, S.K.H., Rozen, W.M., Sofiadellis, F., 2024. Digital twins use in plastic surgery: A systematic review. Journal of Clinical Medicine 13. doi:10.3390/jcm13247861.
- Solaberrieta, E., Garmendia, A., Minguez, R., Brizuela, A., Pradies, G., 2015. Virtual facebow technique. The Journal of Prosthetic Dentistry 114, 751-755. doi:10.1016/j.prosdent.2015.06.012.
- Solaberrieta, E., Mínguez, R., Barrenetxea, L., Etxaniz, O., 2013. Direct transfer of the position of digitized casts to a virtual articulator. The Journal of Prosthetic Dentistry 109, 411–414. doi:10.1016/S0022-3913(13) 60330-3.
- Tomaka, A., Tarnawski, M., Luchowski, L., Lisniewska-Machorowska, B., 2007. Digital dental models and 3d patient photographs registration for orthodontic documentation and diagnostic purposes, in: Kurzynski, M., Puchala, E., Wozniak, M., Zolnierek, A. (Eds.), Computer Recognition Systems 2, Springer Berlin Heidelberg, Berlin, Heidelberg. pp. 645–652.
- Tomaka, A.A., Luchowski, L., Pojda, D., Tarnawski, M., Domino, K., 2019. The dynamics of the stomatognathic system from 4d multimodal data, in: Gadomski, A. (ed.) Multiscale Locomotion: Its Active-Matter Addressing Physical Principles. UTP University of Science & Technology, Bydgoszcz, pp. 37–53. URL: http://arxiv.org/abs/1911.08854.

- Tomaka, A.A., Tarnawski, M., Pojda, D., 2016. Multimodal image registration for mandible motion tracking, in: Piętka E., Badura P., Kawa J., Więcławek W. (eds) Information Technologies in Medicine. ITiB 2016. Advances in Intelligent Systems and Computing, vol 471. Springer, Cham, pp. 179–191. doi:10.1007/978-3-319-39796-2_15.
- Zebris Medical, 2025. Zebris JMA-Optic System. List of Dental Publications. Accessed: 2025-01-10. URL: https://www.zebris.de/fileadmin/Editoren/zebris-PDF/zebris-Literatur-PDF/Publications-Dental.pdf.